

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2825

Client ID: SSTDCCC040

Date Analyzed: 08/15/2025

Lab File ID: BP025422.D

Time Analyzed: 10:36

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	913689	17.201	958309	21.648	1096860	25.012
UPPER LIMIT	1827380	17.701	1916620	22.148	2193720	25.512
LOWER LIMIT	456845	16.701	479155	21.148	548430	24.512
EPA SAMPLE NO.						
01 PB169230BL	774408	17.20	884552	21.65	1011780	25.02
02 PB169230BS	812951	17.19	911094	21.64	1094190	25.02
03 PB169230BSD	837407	17.21	883409	21.65	1035260	25.01

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.