

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG No.: Q2825
 Instrument ID: MSVOA_N Calibration Date/Time: 08/13/2025 10:57
 Lab File ID: VN087525.D Init. Calib. Date(s): 07/16/2025 07/16/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 17:05 18:54
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	51.84	-3.678001	20