

Surrogate Summary

SDG No.: **Q2825**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2825-01	TW1	1,2-Dichloroethane-d4	50	60.7	121		70 (74)	130 (125)
		Dibromofluoromethane	50	49.9	100		70 (75)	130 (124)
		Toluene-d8	50	52.3	105		70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.1	104		70 (77)	130 (121)
Q2825-02	FB	1,2-Dichloroethane-d4	50	60.9	122		70 (74)	130 (125)
		Dibromofluoromethane	50	49.5	99		70 (75)	130 (124)
		Toluene-d8	50	51.8	104		70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.5	103		70 (77)	130 (121)
VN0812WBL01	VN0812WBL01	1,2-Dichloroethane-d4	50	58.4	117		70 (74)	130 (125)
		Dibromofluoromethane	50	50.3	101		70 (75)	130 (124)
		Toluene-d8	50	51.3	103		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.1	100		70 (77)	130 (121)
VN0812WBS01	VN0812WBS01	1,2-Dichloroethane-d4	50	53.3	107		70 (74)	130 (125)
		Dibromofluoromethane	50	44.8	90		70 (75)	130 (124)
		Toluene-d8	50	46.2	92		70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.7	97		70 (77)	130 (121)
VN0812WBSD01	VN0812WBSD01	1,2-Dichloroethane-d4	50	54.6	109		70 (74)	130 (125)
		Dibromofluoromethane	50	47.9	96		70 (75)	130 (124)
		Toluene-d8	50	49.1	98		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.4	101		70 (77)	130 (121)
VN0813WBL01	VN0813WBL01	1,2-Dichloroethane-d4	50	59.6	119		70 (74)	130 (125)
		Dibromofluoromethane	50	49.7	99		70 (75)	130 (124)
		Toluene-d8	50	52.0	104		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.2	100		70 (77)	130 (121)
VN0813WBS01	VN0813WBS01	1,2-Dichloroethane-d4	50	58.5	117		70 (74)	130 (125)
		Dibromofluoromethane	50	50.0	100		70 (75)	130 (124)
		Toluene-d8	50	51.1	102		70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.4	105		70 (77)	130 (121)

