

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2825
 Lab File ID: VN087502.D Date Analyzed: 08/12/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:24
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	330662	8.21	639180	9.08	587049	11.85
UPPER LIMIT	661324	8.712	1278360	9.582	1174100	12.347
LOWER LIMIT	165331	7.712	319590	8.582	293525	11.347
EPA SAMPLE NO.						
FB	238840	8.21	519697	9.09	475077	11.85
VN0812WBL01	271221	8.21	578160	9.08	521832	11.85
VN0812WBS01	292778	8.21	576676	9.08	523240	11.85
VN0812WBSD01	316315	8.21	602749	9.08	538329	11.85

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.