

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
Lab Code: ACE SDG NO.: Q2825
Lab File ID: VN087525.D Date Analyzed: 08/13/2025
Instrument ID: MSVOA_N Time Analyzed: 10:57
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	297883	8.21	560249	9.09	506486	11.85
UPPER LIMIT	595766	8.706	1120500	9.588	1012970	12.347
LOWER LIMIT	148942	7.706	280125	8.588	253243	11.347
EPA SAMPLE NO.						
TW1	225954	8.21	487904	9.08	444655	11.85
VN0813WBL01	240605	8.21	517097	9.09	472576	11.85
VN0813WBS01	271136	8.21	532074	9.09	483502	11.85

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.