



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: G Environmental

SDG No.: Q2826

Project: Ave B

RunNo.: LB136800

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	0.0014	0.0025	J	0.00096	0.005	08/13/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	0.0015	0.0025	J	0.00096	0.005	08/13/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	0.0014	0.0025	J	0.00096	0.005	08/13/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	0.0015	0.0025	J	0.00096	0.005	08/13/2025