



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q2826

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	101	70 - 130	CV	08/13/2025	11:26	LB136797
CRI01	Arsenic	20.6	20.0	103	65 - 135	P	08/13/2025	12:37	LB136801
	Barium	93.9	100	94	65 - 135	P	08/13/2025	12:37	LB136801
	Cadmium	5.49	6.0	92	65 - 135	P	08/13/2025	12:37	LB136801
	Chromium	10.00	10.0	100	65 - 135	P	08/13/2025	12:37	LB136801
	Lead	10.0	12.0	83	65 - 135	P	08/13/2025	12:37	LB136801
	Selenium	21.1	20.0	106	65 - 135	P	08/13/2025	12:37	LB136801
	Silver	10.3	10.0	103	65 - 135	P	08/13/2025	12:37	LB136801