



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q2826

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB12	Mercury	0.076	+/-0.2	U	0.20	CV	08/13/2025	11:19	LB136797

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q2826

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB35	Mercury	0.076	+/-0.2	U	0.20	CV	08/13/2025	11:24	LB136797
CCB36	Mercury	0.076	+/-0.2	U	0.20	CV	08/13/2025	11:51	LB136797
CCB37	Mercury	0.076	+/-0.2	U	0.20	CV	08/13/2025	12:10	LB136797

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q2826

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	08/13/2025	12:33	LB136801
	Barium	14.6	+/-50	U	100	P	08/13/2025	12:33	LB136801
	Cadmium	0.50	+/-3	U	6.00	P	08/13/2025	12:33	LB136801
	Chromium	2.12	+/-5	U	10.0	P	08/13/2025	12:33	LB136801
	Lead	2.30	+/-6	U	12.0	P	08/13/2025	12:33	LB136801
	Selenium	9.64	+/-10	U	20.0	P	08/13/2025	12:33	LB136801
	Silver	1.62	+/-5	U	10.0	P	08/13/2025	12:33	LB136801

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q2826

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	08/13/2025	13:03	LB136801
	Barium	14.6	+/-50	U	100	P	08/13/2025	13:03	LB136801
	Cadmium	0.50	+/-3	U	6.00	P	08/13/2025	13:03	LB136801
	Chromium	2.12	+/-5	U	10.0	P	08/13/2025	13:03	LB136801
	Lead	2.30	+/-6	U	12.0	P	08/13/2025	13:03	LB136801
	Selenium	9.64	+/-10	U	20.0	P	08/13/2025	13:03	LB136801
	Silver	1.62	+/-5	U	10.0	P	08/13/2025	13:03	LB136801
CCB02	Arsenic	5.12	+/-10	U	20.0	P	08/13/2025	13:46	LB136801
	Barium	14.6	+/-50	U	100	P	08/13/2025	13:46	LB136801
	Cadmium	0.50	+/-3	U	6.00	P	08/13/2025	13:46	LB136801
	Chromium	2.12	+/-5	U	10.0	P	08/13/2025	13:46	LB136801
	Lead	2.30	+/-6	U	12.0	P	08/13/2025	13:46	LB136801
	Selenium	9.64	+/-10	U	20.0	P	08/13/2025	13:46	LB136801
	Silver	1.62	+/-5	U	10.0	P	08/13/2025	13:46	LB136801
CCB03	Arsenic	5.12	+/-10	U	20.0	P	08/13/2025	14:37	LB136801
	Barium	14.6	+/-50	U	100	P	08/13/2025	14:37	LB136801
	Cadmium	0.50	+/-3	U	6.00	P	08/13/2025	14:37	LB136801
	Chromium	2.12	+/-5	U	10.0	P	08/13/2025	14:37	LB136801
	Lead	2.30	+/-6	U	12.0	P	08/13/2025	14:37	LB136801
	Selenium	9.64	+/-10	U	20.0	P	08/13/2025	14:37	LB136801
	Silver	1.62	+/-5	U	10.0	P	08/13/2025	14:37	LB136801
CCB04	Arsenic	5.12	+/-10	U	20.0	P	08/13/2025	15:27	LB136801
	Barium	14.6	+/-50	U	100	P	08/13/2025	15:27	LB136801
	Cadmium	0.50	+/-3	U	6.00	P	08/13/2025	15:27	LB136801
	Chromium	2.12	+/-5	U	10.0	P	08/13/2025	15:27	LB136801
	Lead	2.30	+/-6	U	12.0	P	08/13/2025	15:27	LB136801
	Selenium	9.64	+/-10	U	20.0	P	08/13/2025	15:27	LB136801
	Silver	1.62	+/-5	U	10.0	P	08/13/2025	15:27	LB136801
CCB05	Arsenic	5.12	+/-10	U	20.0	P	08/13/2025	16:19	LB136801
	Barium	14.6	+/-50	U	100	P	08/13/2025	16:19	LB136801
	Cadmium	0.50	+/-3	U	6.00	P	08/13/2025	16:19	LB136801
	Chromium	2.12	+/-5	U	10.0	P	08/13/2025	16:19	LB136801
	Lead	2.30	+/-6	U	12.0	P	08/13/2025	16:19	LB136801
	Selenium	9.64	+/-10	U	20.0	P	08/13/2025	16:19	LB136801
	Silver	1.62	+/-5	U	10.0	P	08/13/2025	16:19	LB136801
CCB06	Arsenic	5.12	+/-10	U	20.0	P	08/13/2025	17:09	LB136801
	Barium	14.6	+/-50	U	100	P	08/13/2025	17:09	LB136801
	Cadmium	0.50	+/-3	U	6.00	P	08/13/2025	17:09	LB136801
	Chromium	2.12	+/-5	U	10.0	P	08/13/2025	17:09	LB136801

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q2826

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	08/13/2025	17:09	LB136801
	Selenium	9.64	+/-10	U	20.0	P	08/13/2025	17:09	LB136801
	Silver	1.62	+/-5	U	10.0	P	08/13/2025	17:09	LB136801
CCB07	Arsenic	5.12	+/-10	U	20.0	P	08/13/2025	17:50	LB136801
	Barium	14.6	+/-50	U	100	P	08/13/2025	17:50	LB136801
	Cadmium	0.50	+/-3	U	6.00	P	08/13/2025	17:50	LB136801
	Chromium	2.12	+/-5	U	10.0	P	08/13/2025	17:50	LB136801
	Lead	2.30	+/-6	U	12.0	P	08/13/2025	17:50	LB136801
	Selenium	9.64	+/-10	U	20.0	P	08/13/2025	17:50	LB136801
	Silver	1.62	+/-5	U	10.0	P	08/13/2025	17:50	LB136801