

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>G Environmental</u>	SDG No.: <u>Q2826</u>
Contract: <u>GENV01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	13.5			-20	20	08/13/2025	12:41	LB136801
	Barium	-0.47	6.0	8	-94	106	08/13/2025	12:41	LB136801
	Cadmium	2.33	1.0	233	-5	7	08/13/2025	12:41	LB136801
	Chromium	58.2	52.0	112	42	62	08/13/2025	12:41	LB136801
	Lead	-6.38			-12	12	08/13/2025	12:41	LB136801
	Selenium	2.91			-20	20	08/13/2025	12:41	LB136801
	Silver	2.33			-10	10	08/13/2025	12:41	LB136801
ICSAB01	Arsenic	111	104	107	88.4	120	08/13/2025	12:45	LB136801
	Barium	492	537	92	437	637	08/13/2025	12:45	LB136801
	Cadmium	994	972	102	826	1120	08/13/2025	12:45	LB136801
	Chromium	558	542	103	460	624	08/13/2025	12:45	LB136801
	Lead	42.2	49.0	86	37	61	08/13/2025	12:45	LB136801
	Selenium	50.5	46.0	110	26	66	08/13/2025	12:45	LB136801
	Silver	216	201	108	170	232	08/13/2025	12:45	LB136801