

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q2826

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169214BS							
Arsenic	ug/L	4000	3970		99	80 - 120	P
Barium	ug/L	1000	945		94	80 - 120	P
Cadmium	ug/L	1000	950		95	80 - 120	P
Chromium	ug/L	2000	2050		102	80 - 120	P
Lead	ug/L	5000	4780		96	80 - 120	P
Selenium	ug/L	10000	10100		101	80 - 120	P
Silver	ug/L	380	383		101	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q2826

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169233BS Mercury	ug/L	4.0	4.02		100	80 - 120	CV