

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q2826

Continuing Calib Date: 08/12/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 08:54

Initial Calibration Time(s): 11:32

19:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.75	4.76	4.66	4.86	0.01
Aroclor-1016-2	(2)	4.77	4.77	4.67	4.87	0.00
Aroclor-1016-3	(3)	4.83	4.83	4.73	4.93	0.00
Aroclor-1016-4	(4)	4.95	4.95	4.85	5.05	0.00
Aroclor-1016-5	(5)	5.20	5.21	5.11	5.31	0.01
Aroclor-1260-1	(1)	6.24	6.25	6.15	6.35	0.01
Aroclor-1260-2	(2)	6.43	6.43	6.33	6.53	0.00
Aroclor-1260-3	(3)	6.79	6.80	6.70	6.90	0.01
Aroclor-1260-4	(4)	7.05	7.06	6.96	7.16	0.01
Aroclor-1260-5	(5)	7.30	7.30	7.20	7.40	0.00
Tetrachloro-m-xylene		3.67	3.67	3.57	3.77	0.01
Decachlorobiphenyl		8.69	8.69	8.59	8.79	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q2826

Continuing Calib Date: 08/12/2025

Initial Calibration Date(s): 07/23/2025

07/23/2025

Continuing Calib Time: 08:54

Initial Calibration Time(s): 11:32

19:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.73	4.74	4.64	4.84	0.01
Aroclor-1016-2	(2)	4.75	4.76	4.66	4.86	0.01
Aroclor-1016-3	(3)	4.92	4.93	4.83	5.03	0.01
Aroclor-1016-4	(4)	4.97	4.97	4.87	5.07	0.00
Aroclor-1016-5	(5)	5.18	5.19	5.09	5.29	0.01
Aroclor-1260-1	(1)	6.21	6.21	6.11	6.31	0.00
Aroclor-1260-2	(2)	6.39	6.40	6.30	6.50	0.01
Aroclor-1260-3	(3)	6.55	6.55	6.45	6.65	0.00
Aroclor-1260-4	(4)	7.02	7.02	6.92	7.12	0.01
Aroclor-1260-5	(5)	7.26	7.26	7.16	7.36	0.00
Tetrachloro-m-xylene		3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl		8.63	8.64	8.54	8.74	0.01