

Surrogate Summary

SDG No.: **Q2826**

Client: **G Environmental**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO112413.D	PIBLK-PO112413.D	Tetrachloro-m-xyl	1	20	19.1	95		70 (60)	130 (140)
		Decachlorobiphen	1	20	20.0	100		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.5	98		70 (60)	130 (140)
		Decachlorobiphen	2	20	20.4	102		70 (60)	130 (140)
I.BLK-PO112869.D	PIBLK-PO112869.D	Tetrachloro-m-xyl	1	20	20.7	103		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.2	96		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	20.3	102		70 (60)	130 (140)
		Decachlorobiphen	2	20	20.5	103		70 (60)	130 (140)
Q2826-01	WC1	Tetrachloro-m-xyl	1	20	21.0	105		30 (32)	150 (144)
		Decachlorobiphen	1	20	16.3	81		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	20.3	101		30 (32)	150 (144)
		Decachlorobiphen	2	20	18.6	93		30 (32)	150 (175)
I.BLK-PO112884.D	PIBLK-PO112884.D	Tetrachloro-m-xyl	1	20	20.8	104		70 (60)	130 (140)
		Decachlorobiphen	1	20	17.8	89		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	20.3	102		70 (60)	130 (140)
		Decachlorobiphen	2	20	20.8	104		70 (60)	130 (140)
I.BLK-PP074167.D	PIBLK-PP074167.D	Tetrachloro-m-xyl	1	20	19.1	95		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.7	98		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	17.3	86		70 (60)	130 (140)
		Decachlorobiphen	2	20	18.7	93		70 (60)	130 (140)
I.BLK-PP074305.D	PIBLK-PP074305.D	Tetrachloro-m-xyl	1	20	22.0	110		70 (60)	130 (140)
		Decachlorobiphen	1	20	21.1	105		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.7	98		70 (60)	130 (140)
		Decachlorobiphen	2	20	18.5	93		70 (60)	130 (140)
PB169205BL	PB169205BL	Tetrachloro-m-xyl	1	20	23.1	116		30 (32)	150 (144)
		Decachlorobiphen	1	20	22.2	111		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	20.1	101		30 (32)	150 (144)
		Decachlorobiphen	2	20	20.1	100		30 (32)	150 (175)
PB169205BS	PB169205BS	Tetrachloro-m-xyl	1	20	24.3	121		30 (32)	150 (144)
		Decachlorobiphen	1	20	22.5	112		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	20.8	104		30 (32)	150 (144)
		Decachlorobiphen	2	20	20.0	100		30 (32)	150 (175)
Q2830-01MS	BIN0009-DRIVEWAY-TP-SOUT	Tetrachloro-m-xyl	1	20	23.5	117		30 (32)	150 (144)
		Decachlorobiphen	1	20	19.9	100		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	22.0	110		30 (32)	150 (144)
		Decachlorobiphen	2	20	17.5	87		30 (32)	150 (175)
Q2830-01MSD	BIN0009-DRIVEWAY-TP-SOUT	Tetrachloro-m-xyl	1	20	23.3	116		30 (32)	150 (144)
		Decachlorobiphen	1	20	19.9	99		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	22.1	111		30 (32)	150 (144)
		Decachlorobiphen	2	20	17.9	89		30 (32)	150 (175)
I.BLK-PP074319.D	PIBLK-PP074319.D	Tetrachloro-m-xyl	1	20	22.9	115		70 (60)	130 (140)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q2826

Client: G Environmental

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP074319.D	PIBLK-PP074319.D	Decachlorobiphen	1	20	21.7	109		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.8	99		70 (60)	130 (140)
		Decachlorobiphen	2	20	19.8	99		70 (60)	130 (140)