

Surrogate Summary

SW-846

SDG No.: **Q2826**

Client: **G Environmental**

Analytical Method: **8270E**

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169210BL	PB169210BL	2-Fluorophenol	150	123	82		30 (18)	130 (112)
		Phenol-d6	150	120	80		30 (15)	130 (107)
		Nitrobenzene-d5	100	97.5	98		30 (18)	130 (107)
		2-Fluorobiphenyl	100	90.2	90		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	166	110		30 (10)	130 (116)
		Terphenyl-d14	100	101	101		30 (10)	130 (105)
PB169210BS	PB169210BS	2-Fluorophenol	150	136	91		30 (18)	130 (112)
		Phenol-d6	150	138	92		30 (15)	130 (107)
		Nitrobenzene-d5	100	79.6	80		30 (18)	130 (107)
		2-Fluorobiphenyl	100	79.7	80		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	138	92		30 (10)	130 (116)
		Terphenyl-d14	100	83.1	83		30 (10)	130 (105)
Q2826-01	WC1	2-Fluorophenol	150	38.1	25	*	30 (18)	130 (112)
		Phenol-d6	150	41.7	28	*	30 (15)	130 (107)
		Nitrobenzene-d5	100	26.3	26	*	30 (18)	130 (107)
		2-Fluorobiphenyl	100	27.5	27	*	30 (20)	130 (109)
		2,4,6-Tribromophenol	150	23.5	16	*	30 (10)	130 (116)
		Terphenyl-d14	100	28.6	29	*	30 (10)	130 (105)
Q2830-05MS	BIN0009-DRIVEWAY-TP-WESTSID	2-Fluorophenol	150	95.7	64		30 (18)	130 (112)
		Phenol-d6	150	95.4	64		30 (15)	130 (107)
		Nitrobenzene-d5	100	75.4	75		30 (18)	130 (107)
		2-Fluorobiphenyl	100	68.9	69		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	121	81		30 (10)	130 (116)
		Terphenyl-d14	100	58.4	58		30 (10)	130 (105)
Q2830-05MSD	BIN0009-DRIVEWAY-TP-WESTSID	2-Fluorophenol	150	97.8	65		30 (18)	130 (112)
		Phenol-d6	150	101	67		30 (15)	130 (107)
		Nitrobenzene-d5	100	77.3	77		30 (18)	130 (107)
		2-Fluorobiphenyl	100	65.9	66		30 (20)	130 (109)
		2,4,6-Tribromophenol	150	142	95		30 (10)	130 (116)
		Terphenyl-d14	100	77.5	78		30 (10)	130 (105)