

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2826

Client ID : SSTDCCC040

Date Analyzed: 08/13/2025

Lab File ID: BF143365.D

Time Analyzed: 10:00

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	155874	6.934	589243	8.21	314886	9.96
UPPER LIMIT	311748	7.434	1178490	8.71	629772	10.463
LOWER LIMIT	77937	6.434	294622	7.71	157443	9.463
EPA SAMPLE NO.						
01 PB169210BL	176496	6.93	674917	8.21	354793	9.96
02 BIN0009-DRIVEWAY-TP-WESTSIDEMS	137804	6.93	517024	8.21	272568	9.96
03 BIN0009-DRIVEWAY-TP-WESTSIDEMS	182122	6.93	707644	8.21	400799	9.97

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.