

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2826

Client ID : SSTDICCC040

Date Analyzed: 08/14/2025

Lab File ID: BP025396.D

Time Analyzed: 15:19

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	157704	7.796	663553	10.56	441110	14.41
UPPER LIMIT	315408	8.296	1327110	11.06	882220	14.907
LOWER LIMIT	78852	7.296	331777	10.06	220555	13.907
EPA SAMPLE NO.						
01 PB169210BS	156059	7.79	650560	10.56	435539	14.41

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.