

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2826

Client ID : SSTDCCC040

Date Analyzed: 08/18/2025

Lab File ID: BP025455.D

Time Analyzed: 10:39

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	129439	7.778	550485	10.55	359982	14.40
UPPER LIMIT	258878	8.278	1100970	11.048	719964	14.901
LOWER LIMIT	64719.5	7.278	275243	10.048	179991	13.901
EPA SAMPLE NO.						
01 WC1	171293	7.79	683551	10.55	427689	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.