

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2826

Client ID: SSTDCCC040

Date Analyzed: 08/13/2025

Lab File ID: BF143365.D

Time Analyzed: 10:00

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	455320	11.451	268507	14.092	321267	15.58
UPPER LIMIT	910640	11.951	537014	14.592	642534	16.08
LOWER LIMIT	227660	10.951	134254	13.592	160634	15.08
EPA SAMPLE NO.						
01 PB169210BL	604678	11.45	338429	14.09	291238	15.58
02 BIN0009-DRIVEWAY-TP-WESTSI	391965	11.45	309493	14.09	383464	15.58
03 BIN0009-DRIVEWAY-TP-WESTSI	660496	11.45	401162	14.09	380758	15.58

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.