

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2826

Client ID: SSTDICCC040

Date Analyzed: 08/14/2025

Lab File ID: BP025396.D

Time Analyzed: 15:19

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	884529	17.201	899207	21.648	1022140	25.03
UPPER LIMIT	1769060	17.701	1798410	22.148	2044280	25.53
LOWER LIMIT	442265	16.701	449604	21.148	511070	24.53
EPA SAMPLE NO.						
PB169210BS	895607	17.20	937531	21.65	1016820	25.02

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.