

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2826

Client ID: SSTDCCC040

Date Analyzed: 08/18/2025

Lab File ID: BP025455.D

Time Analyzed: 10:39

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	707933	17.189	748424	21.63	846697	25
UPPER LIMIT	1415870	17.689	1496850	22.13	1693390	25.5
LOWER LIMIT	353967	16.689	374212	21.13	423349	24.5
EPA SAMPLE NO.						
WC1	882062	17.20	894833	21.64	1035070	25.01

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.