

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143364.D
Instrument ID: BNA_F

Contract: GENV01
SDG NO.: Q2826
DFTPP Injection Date: 08/13/2025
DFTPP Injection Time: 08:59

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (1.9) 1
69	Mass 69 relative abundance	29.2
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.5
365	Greater than 1% of mass 198	2.8
441	Present, but less than mass 443	15.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.8 (19.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF143365.D	08/13/2025	10:00
PB169210BL	PB169210BL	BF143366.D	08/13/2025	10:30
BIN0009-DRIVEWAY-TP-WESTSIDEMS	Q2830-05MS	BF143367.D	08/13/2025	11:12
BIN0009-DRIVEWAY-TP-WESTSIDEMSD	Q2830-05MSD	BF143368.D	08/13/2025	11:42