

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081225\
 Data File : VY023053.D
 Acq On : 12 Aug 2025 16:02
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/14/2025
 Supervised By :Semsettin Yesilyurt 08/14/2025

Quant Time: Aug 13 01:53:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 13 01:46:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	539282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	856556	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.408	117	751969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	377839	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	249044	48.455	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	96.900%		
35) Dibromofluoromethane	7.628	113	251193	49.010	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	98.020%		
50) Toluene-d8	10.103	98	979082	48.899	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	97.800%		
62) 4-Bromofluorobenzene	12.401	95	316770	49.194	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	98.380%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.861	85	187863	42.677	ug/l	100
3) Chloromethane	2.068	50	327481	45.932	ug/l	100
4) Vinyl Chloride	2.196	62	414609	47.044	ug/l	100
5) Bromomethane	2.592	94	291336	44.263	ug/l	100
6) Chloroethane	2.726	64	277698	47.394	ug/l	100
7) Trichlorofluoromethane	3.049	101	505137	47.136	ug/l	100
8) Diethyl Ether	3.446	74	135373	47.747	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.805	101	249063	47.109	ug/l	100
10) Methyl Iodide	3.994	142	296087	51.899	ug/l	100
11) Tert butyl alcohol	4.860	59	85712	243.383	ug/l	100
12) 1,1-Dichloroethene	3.781	96	249492	48.009	ug/l	100
13) Acrolein	3.647	56	121185	234.966	ug/l	100
14) Allyl chloride	4.378	41	369720	48.350	ug/l	100
15) Acrylonitrile	5.049	53	285675	247.839	ug/l	100
16) Acetone	3.860	43	278510	250.122	ug/l	100
17) Carbon Disulfide	4.098	76	804119	47.456	ug/l	100
18) Methyl Acetate	4.378	43	173642	52.403	ug/l	100
19) Methyl tert-butyl Ether	5.110	73	666514	49.495	ug/l	100
20) Methylene Chloride	4.604	84	275004	47.415	ug/l	100
21) trans-1,2-Dichloroethene	5.104	96	282272	48.334	ug/l	100
22) Diisopropyl ether	6.018	45	815359	49.070	ug/l	100
23) Vinyl Acetate	5.951	43	2268046	246.849	ug/l	100
24) 1,1-Dichloroethane	5.909	63	495073	48.732	ug/l	100
25) 2-Butanone	6.884	43	382711	252.715	ug/l	100
26) 2,2-Dichloropropane	6.884	77	425773	48.571	ug/l	100
27) cis-1,2-Dichloroethene	6.884	96	327272	48.474	ug/l	100
28) Bromochloromethane	7.244	49	196667	48.795	ug/l	100
29) Tetrahydrofuran	7.256	42	235169	250.523	ug/l	100
30) Chloroform	7.415	83	508641	48.547	ug/l	100
31) Cyclohexane	7.695	56	440266	45.989	ug/l	100
32) 1,1,1-Trichloroethane	7.616	97	451521	48.569	ug/l	100
36) 1,1-Dichloropropene	7.829	75	371204	48.734	ug/l	100
37) Ethyl Acetate	6.982	43	160980	50.311	ug/l	100
38) Carbon Tetrachloride	7.817	117	399463	48.285	ug/l	100
39) Methylcyclohexane	9.103	83	484691	48.321	ug/l	100
40) Benzene	8.073	78	1142910	48.683	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081225\
 Data File : VY023053.D
 Acq On : 12 Aug 2025 16:02
 Operator : SY/MD
 Sample : VSTDICCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/14/2025
 Supervised By :Semsettin Yesilyurt 08/14/2025

Quant Time: Aug 13 01:53:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 13 01:46:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.219	41	94123	50.916	ug/l	100
42) 1,2-Dichloroethane	8.152	62	296907	48.627	ug/l	100
43) Isopropyl Acetate	8.195	43	319600	49.419	ug/l	100
44) Trichloroethene	8.859	130	292421	48.199	ug/l	100
45) 1,2-Dichloropropane	9.140	63	264354	48.938	ug/l	100
46) Dibromomethane	9.225	93	152822	49.291	ug/l	100
47) Bromodichloromethane	9.420	83	393523	49.327	ug/l	100
48) Methyl methacrylate	9.213	41	160792	51.958	ug/l	100
49) 1,4-Dioxane	9.225	88	36249	997.124	ug/l	100
51) 4-Methyl-2-Pentanone	9.993	43	853025	255.710	ug/l	100
52) Toluene	10.164	92	735326	49.315	ug/l	100
53) t-1,3-Dichloropropene	10.390	75	359698	49.790	ug/l	100
54) cis-1,3-Dichloropropene	9.853	75	424834	49.731	ug/l	100
55) 1,1,2-Trichloroethane	10.566	97	204368	49.261	ug/l	100
56) Ethyl methacrylate	10.432	69	273801	51.576	ug/l	100
57) 1,3-Dichloropropane	10.713	76	346510	49.249	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.707	63	663351	264.906	ug/l	100
59) 2-Hexanone	10.755	43	584223	257.194	ug/l	100
60) Dibromochloromethane	10.908	129	272171	49.719	ug/l	100
61) 1,2-Dibromoethane	11.005	107	193398	49.598	ug/l	100
64) Tetrachloroethene	10.640	164	327271	48.740	ug/l	100
65) Chlorobenzene	11.438	112	793490	48.920	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.511	131	271839	49.381	ug/l	100
67) Ethyl Benzene	11.511	91	1397704	50.073	ug/l	100
68) m/p-Xylenes	11.621	106	1080401	99.103	ug/l	100
69) o-Xylene	11.950	106	516002	50.536	ug/l	100
70) Styrene	11.962	104	844409	50.569	ug/l	100
71) Bromoform	12.127	173	160769	49.969	ug/l #	100
73) Isopropylbenzene	12.249	105	1324313	49.660	ug/l	100
74) N-amyl acetate	12.066	43	289914	51.264	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.499	83	220853	48.503	ug/l	100
76) 1,2,3-Trichloropropane	12.548	75	202260m	55.430	ug/l	
77) Bromobenzene	12.523	156	312706	49.152	ug/l	100
78) n-propylbenzene	12.590	91	1593368	50.186	ug/l	100
79) 2-Chlorotoluene	12.670	91	899967	49.473	ug/l	100
80) 1,3,5-Trimethylbenzene	12.731	105	1080333	50.126	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.298	75	74567	49.500	ug/l	100
82) 4-Chlorotoluene	12.767	91	930329	49.299	ug/l	100
83) tert-Butylbenzene	12.993	119	978031	50.244	ug/l	100
84) 1,2,4-Trimethylbenzene	13.035	105	1071926	49.963	ug/l	100
85) sec-Butylbenzene	13.170	105	1425553	49.854	ug/l	100
86) p-Isopropyltoluene	13.285	119	1196645	50.417	ug/l	100
87) 1,3-Dichlorobenzene	13.279	146	608249	48.481	ug/l	100
88) 1,4-Dichlorobenzene	13.359	146	605079	48.618	ug/l	100
89) n-Butylbenzene	13.609	91	1099864	50.454	ug/l	100
90) Hexachloroethane	13.871	117	238005	48.984	ug/l	100
91) 1,2-Dichlorobenzene	13.651	146	536637	48.772	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.267	75	35867	50.592	ug/l	100
93) 1,2,4-Trichlorobenzene	14.913	180	320080	49.192	ug/l	100
94) Hexachlorobutadiene	15.017	225	185604	48.543	ug/l	100
95) Naphthalene	15.139	128	562363	50.685	ug/l	100
96) 1,2,3-Trichlorobenzene	15.322	180	272154	48.540	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081225\
Data File : VY023053.D
Acq On : 12 Aug 2025 16:02
Operator : SY/MD
Sample : VSTDICCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/14/2025
Supervised By :Semsettin Yesilyurt 08/14/2025

Quant Time: Aug 13 01:53:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 13 01:46:43 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

