



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081425\
Data File : VY023083.D
Acq On : 14 Aug 2025 09:48
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081225S.M
Title : SW846 8260
Last Update : Wed Aug 13 02:10:11 2025



AutoFind: Scans 1751, 1752, 1753; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	30528	PASS
75	95	30	60	52.9	78941	PASS
95	95	100	100	100.0	149157	PASS
96	95	5	9	6.5	9739	PASS
173	174	0.00	2	0.7	935	PASS
174	95	50	100	84.0	125272	PASS
175	174	5	9	7.6	9545	PASS
176	174	95	101	96.6	121075	PASS
177	176	5	9	6.5	7816	PASS