

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY081225

Review By	Maresh Dadoda	Review On	8/14/2025 7:34:22 PM
Supervise By	Semsettin Yesilyurt	Supervise On	8/14/2025 7:44:13 PM
SubDirectory	VY081225	HP Acquire Method	MSVOA_Y HP Processing Method 82y081225s.m

STD. NAME	STD REF.#
Tune/Reschk	VP135083
Initial Calibration Stds	VP135084,VP135085,VP135086,VP135087,VP135088,VP135089
CCC	
Internal Standard/PEM	VP133934
ICV/I.BLK	VP135090
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023048.D	12 Aug 2025 08:26		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY023049.D	12 Aug 2025 14:32	Not used	SY/MD	Not Ok
3	VSTDICC005	VSTDICC005	VY023050.D	12 Aug 2025 14:54		SY/MD	Ok,M
4	VSTDICC010	VSTDICC010	VY023051.D	12 Aug 2025 15:17		SY/MD	Ok,M
5	VSTDICC020	VSTDICC020	VY023052.D	12 Aug 2025 15:40	LR-20	SY/MD	Ok,M
6	VSTDICCC050	VSTDICCC050	VY023053.D	12 Aug 2025 16:02		SY/MD	Ok,M
7	VSTDICC100	VSTDICC100	VY023054.D	12 Aug 2025 16:25		SY/MD	Ok,M
8	VSTDICC150	VSTDICC150	VY023055.D	12 Aug 2025 16:47		SY/MD	Ok,M
9	VIBLK	VIBLK	VY023056.D	12 Aug 2025 17:31		SY/MD	Ok
10	VSTDICV050	ICVVY081225	VY023057.D	12 Aug 2025 17:53		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY081425

Review By	Maresh Dadoda	Review On	8/19/2025 4:23:08 AM
Supervise By	Semsettin Yesilyurt	Supervise On	8/19/2025 6:52:38 AM
SubDirectory	VY081425	HP Acquire Method	MSVOA_Y HP Processing Method 82y081225s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135134
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135135,VP135136 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023083.D	14 Aug 2025 09:48		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023084.D	14 Aug 2025 10:17		SY/MD	Ok,M
3	VY0814SBL01	VY0814SBL01	VY023085.D	14 Aug 2025 10:50		SY/MD	Ok
4	VY0814SBS01	VY0814SBS01	VY023086.D	14 Aug 2025 11:22		SY/MD	Ok,M
5	VY0814SBSD01	VY0814SBSD01	VY023087.D	14 Aug 2025 11:45		SY/MD	Ok,M
6	Q2830-05	BIN0009-DRIVEWAY-T	VY023088.D	14 Aug 2025 12:26		SY/MD	Ok
7	Q2819-10RE	84SB-SRE	VY023089.D	14 Aug 2025 12:50	Surrogate Fail	SY/MD	Confirms
8	Q2819-11RE	84SB-WRE	VY023090.D	14 Aug 2025 13:13	Internal Standard fail;Surrogate Fail	SY/MD	Confirms
9	Q2819-15	17M-N	VY023091.D	14 Aug 2025 13:37		SY/MD	Ok
10	Q2819-16	38M-S	VY023092.D	14 Aug 2025 14:00		SY/MD	Ok
11	Q2819-20	82H-E	VY023093.D	14 Aug 2025 14:24	Internal Standard fail;Surrogate Fail	SY/MD	ReRun
12	Q2838-03	TP-11-VOC	VY023094.D	14 Aug 2025 14:47		SY/MD	Ok
13	Q2838-07	TP-10-VOC	VY023095.D	14 Aug 2025 15:11		SY/MD	Ok
14	Q2840-01	0804-SOIL	VY023096.D	14 Aug 2025 15:34		SY/MD	Ok
15	Q2858-01	VNJ 216	VY023097.D	14 Aug 2025 15:57		SY/MD	Ok
16	Q2857-03	72-12002	VY023098.D	14 Aug 2025 16:21		SY/MD	Ok
17	Q2826-01	WC1	VY023099.D	14 Aug 2025 16:44		SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY081425

Review By	Maresh Dadoda	Review On	8/19/2025 4:23:08 AM
Supervise By	Semsettin Yesilyurt	Supervise On	8/19/2025 6:52:38 AM
SubDirectory	VY081425	HP Acquire Method	MSVOA_Y HP Processing Method 82y081225s.m
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135134		
CCC	VP135135,VP135136		
Internal Standard/PEM	VP133934		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q2873-01	DRILL CUTTING 1-3 C	VY023100.D	14 Aug 2025 17:08		SY/MD	Ok
19	Q2873-04	DRILL CUTTING 4-6 C	VY023101.D	14 Aug 2025 17:31		SY/MD	Ok
20	Q2871-01	DRILL CUTTING 1-3 C	VY023102.D	14 Aug 2025 17:55		SY/MD	Ok
21	Q2871-04	DRILL CUTTING 4-6 C	VY023103.D	14 Aug 2025 18:18	Internal Standard Fail	SY/MD	ReRun
22	Q2871-07	DRILL CUTTING 1-5 C	VY023104.D	14 Aug 2025 18:42		SY/MD	Ok
23	Q2872-01	DRILL CUTTING 1-2 C	VY023105.D	14 Aug 2025 19:05	Internal Standard Fail	SY/MD	ReRun
24	Q2870-01	DRILL CUTTING 1-5 C	VY023106.D	14 Aug 2025 19:29	Internal Standard Fail	SY/MD	ReRun
25	Q2870-04	DRILL CUTTING 6-9 C	VY023107.D	14 Aug 2025 19:52		SY/MD	Ok
26	VIBLK	VIBLK	VY023108.D	14 Aug 2025 20:15		SY/MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VY023109.D	14 Aug 2025 20:38		SY/MD	Ok,M

M : Manual Integration