

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2826
 Lab File ID: VY023084.D Date Analyzed: 08/14/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:17
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	410930	7.71	647551	8.62	582049	11.41
UPPER LIMIT	821860	8.207	1295100	9.116	1164100	11.914
LOWER LIMIT	205465	7.207	323776	8.116	291025	10.914
EPA SAMPLE NO.						
WC1	266207	7.71	460667	8.61	381025	11.41
VY0814SBL01	514218	7.71	1013413	8.62	885359	11.41
VY0814SBS01	455827	7.71	734618	8.61	641778	11.41
VY0814SBSD01	440114	7.71	717756	8.61	626725	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.