

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q2826
 Lab File ID: VY023084.D Date Analyzed: 08/14/2025
 Instrument ID: MSVOA_Y Time Analyzed: 10:17
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	290049	13.346				
UPPER LIMIT	580098	13.846				
LOWER LIMIT	145025	12.846				
EPA SAMPLE NO.						
WC1	151348	13.34				
VY0814SBL01	340788	13.35				
VY0814SBS01	309825	13.34				
VY0814SBSD01	312365	13.34				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.