

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Viva Nova Inc</u>	SDG No.: <u>Q2828</u>
Contract: <u>VIVA01</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	240000	255000	94	216000	294000	08/13/2025	12:41	LB136801
	Antimony	-4.82			-50	50	08/13/2025	12:41	LB136801
	Arsenic	13.5			-20	20	08/13/2025	12:41	LB136801
	Barium	-0.47	6.0	8	-94	106	08/13/2025	12:41	LB136801
	Beryllium	1.31			-6	6	08/13/2025	12:41	LB136801
	Cadmium	2.33	1.0	233	-5	7	08/13/2025	12:41	LB136801
	Calcium	229000	245000	94	208000	282000	08/13/2025	12:41	LB136801
	Chromium	58.2	52.0	112	42	62	08/13/2025	12:41	LB136801
	Cobalt	1.85			-30	30	08/13/2025	12:41	LB136801
	Copper	4.82	2.0	241	-18	22	08/13/2025	12:41	LB136801
	Iron	99200	101000	98	85600	116500	08/13/2025	12:41	LB136801
	Lead	-6.38			-12	12	08/13/2025	12:41	LB136801
	Magnesium	243000	255000	95	216000	294000	08/13/2025	12:41	LB136801
	Manganese	8.17	7.0	117	-13	27	08/13/2025	12:41	LB136801
	Nickel	5.22	2.0	261	-38	42	08/13/2025	12:41	LB136801
	Potassium	-4.10			0	0	08/13/2025	12:41	LB136801
	Selenium	2.91			-20	20	08/13/2025	12:41	LB136801
	Silver	2.33			-10	10	08/13/2025	12:41	LB136801
	Sodium	3.06			0	0	08/13/2025	12:41	LB136801
	Thallium	2.91			-40	40	08/13/2025	12:41	LB136801
	Vanadium	3.65			-40	40	08/13/2025	12:41	LB136801
	Zinc	2.20			-40	40	08/13/2025	12:41	LB136801
ICSAB01	Aluminum	247000	247000	100	209000	285000	08/13/2025	12:45	LB136801
	Antimony	588	618	95	525	711	08/13/2025	12:45	LB136801
	Arsenic	111	104	107	88.4	120	08/13/2025	12:45	LB136801
	Barium	492	537	92	437	637	08/13/2025	12:45	LB136801
	Beryllium	487	495	98	420	570	08/13/2025	12:45	LB136801
	Cadmium	994	972	102	826	1120	08/13/2025	12:45	LB136801
	Calcium	236000	235000	100	199000	271000	08/13/2025	12:45	LB136801
	Chromium	558	542	103	460	624	08/13/2025	12:45	LB136801
	Cobalt	498	476	105	404	548	08/13/2025	12:45	LB136801
	Copper	486	511	95	434	588	08/13/2025	12:45	LB136801
	Iron	102000	99300	103	84400	114500	08/13/2025	12:45	LB136801
	Lead	42.2	49.0	86	37	61	08/13/2025	12:45	LB136801
	Magnesium	250000	248000	101	210000	286000	08/13/2025	12:45	LB136801
	Manganese	493	507	97	430	584	08/13/2025	12:45	LB136801
	Nickel	992	954	104	810	1100	08/13/2025	12:45	LB136801
	Potassium	42.9			0	0	08/13/2025	12:45	LB136801
	Selenium	50.5	46.0	110	26	66	08/13/2025	12:45	LB136801
	Silver	216	201	108	170	232	08/13/2025	12:45	LB136801
	Sodium	-3.58			0	0	08/13/2025	12:45	LB136801
	Thallium	113	108	105	68	148	08/13/2025	12:45	LB136801

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client:	<u>Viva Nova Inc</u>	SDG No.:	<u>Q2828</u>
Contract:	<u>VIVA01</u>	Lab Code:	<u>ACE</u>
ICS Source:	<u>EPA</u>	Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	479	491	98	417	565	08/13/2025	12:45	LB136801
	Zinc	1040	952	109	809	1095	08/13/2025	12:45	LB136801