

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Viva Nova Inc

SDG No.: Q2828

Contract: VIVA01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169204BS Mercury	mg/Kg	0.27	0.23		85	80 - 124	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Viva Nova Inc
Contract: VIVA01

SDG No.: Q2828
Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169211BS							
Aluminum	mg/Kg	100	98.1		98	74 - 119	P
Antimony	mg/Kg	40.0	40.7		102	79 - 114	P
Arsenic	mg/Kg	40.0	39.2		98	82 - 111	P
Barium	mg/Kg	10.0	9.27		93	83 - 113	P
Beryllium	mg/Kg	10.0	9.71		97	83 - 113	P
Cadmium	mg/Kg	10.0	9.41		94	82 - 113	P
Calcium	mg/Kg	50.0	49.9	J	100	81 - 116	P
Chromium	mg/Kg	20.0	20.3		102	85 - 113	P
Cobalt	mg/Kg	10.0	9.68		97	85 - 112	P
Copper	mg/Kg	15.0	15.4		103	81 - 117	P
Iron	mg/Kg	150	151		101	81 - 118	P
Lead	mg/Kg	50.0	47.4		95	81 - 112	P
Magnesium	mg/Kg	100	97.2	J	97	78 - 115	P
Manganese	mg/Kg	10.0	9.87		99	84 - 114	P
Nickel	mg/Kg	25.0	24.4		98	83 - 113	P
Potassium	mg/Kg	500	486		97	81 - 116	P
Selenium	mg/Kg	100	99.5		100	78 - 111	P
Silver	mg/Kg	3.8	3.78		100	82 - 112	P
Sodium	mg/Kg	150	138		92	83 - 118	P
Thallium	mg/Kg	100	97.7		98	83 - 111	P
Vanadium	mg/Kg	15.0	15.0		100	82 - 114	P
Zinc	mg/Kg	10.0	10.3		103	82 - 113	P