Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside NJ 07092
Tel. 908-789-8900 Fax: 908-789-8922

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following " Results Qualifiers" are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers **"P"** for ICP instrument **"PM"** for ICP when Microwave Digestion is used **"CV"** for Manual Cold Vapor AA **"AV"** for automated Cold Vapor AA **"CA"** for MIDI-Distillation Spectrophotometric **"AS"** for Semi –Automated Spectrophotometric **"C"** for Manual Spectrophotometric **"T"** for Titrimetric **"NR"** for analyte not required to be analyzed
OR	Indicates the analyte's concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q2829	**OrderDate:**	8/11/2025 3:00:13 PM
Client:	Viva Nova Inc	**Project:**	Metal Testing
Contact:	Sanjeev Goyal	**Location:**	D21

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2829-01	**SILICA**	**SOIL**			**08/11/25**			**08/11/25**
			Metals Group3	6010D		08/12/25	08/13/25	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2829 **Order ID:** Q2829

Client: Viva Nova Inc **Project ID:** Metal Testing

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : SILICA								
Q2829-01	SILICA	SOIL	Silicon	213		3.47	16.9	mg/Kg



SAMPLE DATA

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Viva Nova Inc	Date Collected:	08/11/25
Project:	Metal Testing	Date Received:	08/11/25
Client Sample ID:	SILICA	SDG No.:	Q2829
Lab Sample ID:	Q2829-01	Matrix:	SOIL
Level (low/med):	low	% Solid:	100

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units(Dry Weight)	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-21-3	Silicon	213		1	3.47	16.9	mg/Kg	08/12/25 13:05	08/13/25 16:48	6010D	SW3050

Color Before:	White	Clarity Before:		Texture:	Medium
Color After:	Yellow	Clarity After:		Artifacts:	
Comments:	Metals Group3				

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N =Spiked sample recovery not within control limits

METAL CALIBRATION DATA

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Silicon	3740	4000	93	90 - 110	P	08/13/2025	12:21	LB136801

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Silicon	373	400	93	80 - 120	P	08/13/2025	12:28	LB136801

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Silicon	5030	5000	101	90 - 110	P	08/13/2025	12:58	LB136801
CCV02	Silicon	5110	5000	102	90 - 110	P	08/13/2025	13:42	LB136801
CCV03	Silicon	5100	5000	102	90 - 110	P	08/13/2025	14:33	LB136801
CCV04	Silicon	5100	5000	102	90 - 110	P	08/13/2025	15:23	LB136801
CCV05	Silicon	5150	5000	103	90 - 110	P	08/13/2025	16:14	LB136801
CCV06	Silicon	5150	5000	103	90 - 110	P	08/13/2025	17:05	LB136801
CCV07	Silicon	5230	5000	105	90 - 110	P	08/13/2025	17:46	LB136801

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Silicon	3660	4000	92	90 - 110	P	08/26/2025	11:14	LB136973

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Silicon	369	400	92	80 - 120	P	08/26/2025	11:20	LB136973

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Silicon	5010	5000	100	90 - 110	P	08/26/2025	11:50	LB136973
CCV02	Silicon	5230	5000	104	90 - 110	P	08/26/2025	12:41	LB136973
CCV03	Silicon	5090	5000	102	90 - 110	P	08/26/2025	13:30	LB136973
CCV04	Silicon	5230	5000	105	90 - 110	P	08/26/2025	14:20	LB136973

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -
CRDL STANDARD FOR AA & ICP

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Silicon	357	400	89	65 - 135	P	08/13/2025	12:37	LB136801
CRI01	Silicon	371	400	93	65 - 135	P	08/26/2025	11:29	LB136973

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Silicon	40.2	+/-200	U	400	P	08/13/2025	12:33	LB136801

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Silicon	40.2	+/-200	U	400	P	08/13/2025	13:03	LB136801
CCB02	Silicon	40.2	+/-200	U	400	P	08/13/2025	13:46	LB136801
CCB03	Silicon	40.2	+/-200	U	400	P	08/13/2025	14:37	LB136801
CCB04	Silicon	40.2	+/-200	U	400	P	08/13/2025	15:27	LB136801
CCB05	Silicon	40.2	+/-200	U	400	P	08/13/2025	16:19	LB136801
CCB06	Silicon	40.2	+/-200	U	400	P	08/13/2025	17:09	LB136801
CCB07	Silicon	40.2	+/-200	U	400	P	08/13/2025	17:50	LB136801

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Silicon	40.2	+/-200	U	400	P	08/26/2025	11:24	LB136973

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Silicon	40.2	+/-200	U	400	P	08/26/2025	11:54	LB136973
CCB02	Silicon	40.2	+/-200	U	400	P	08/26/2025	12:45	LB136973
CCB03	Silicon	40.2	+/-200	U	400	P	08/26/2025	13:35	LB136973
CCB04	Silicon	40.2	+/-200	U	400	P	08/26/2025	14:24	LB136973

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Viva Nova Inc **SDG No.:** Q2829

Instrument: P4

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB169215BL		**SOLID**		**Batch Number:**	**PB169215**		**Prep Date:**	**08/12/2025**	
	Silicon	4.09	<10	U	20.0	P	08/13/2025	15:15	LB136801

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Silicon	19.2			0	0	08/13/2025	12:41	LB136801
ICSAB01	Silicon	1030			0	0	08/13/2025	12:45	LB136801
ICSA01	Silicon	19.0			0	0	08/26/2025	11:33	LB136973
ICSAB01	Silicon	1000			0	0	08/26/2025	11:37	LB136973

METAL
QC
DATA

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Viva Nova Inc **level:** low **sdg no.:** Q2829

contract: VIVA01 **lab code:** ACE

matrix: Solid **sample id:** Q2832-09 **client id:** TG-S05MS

Percent Solids for Sample: 90.9 **Spiked ID:** Q2832-09MS **Percent Solids for Spike Sample:** 90.9

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Silicon	mg/Kg	75 - 125	1150	D	1240	D	44.4	-216		P

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Viva Nova Inc **level:** low **sdg no.:** Q2829

contract: VIVA01 **lab code:** ACE

matrix: Solid **sample id:** Q2832-09 **client id:** TG-S05MSD

Percent Solids for Sample: 90.9 **Spiked ID:** Q2832-09MSD **Percent Solids for Spike Sample:** 90.9

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Silicon	mg/Kg	75 - 125	1150	D	1240	D	44.2	-209		P

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals
- 5b -

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Matrix: **Level:** LOW **Client ID:**

Sample ID: **Spiked ID:**

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Viva Nova Inc **Level:** LOW **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Matrix: Solid **Sample ID:** Q2832-09 **Client ID:** TG-S05DUP

Percent Solids for Sample: 90.9 **Duplicate ID** Q2832-09DUP **Percent Solids for Spike Sample:** 90.9

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Silicon	mg/Kg	20	1240	D	1350	D	8		P

"A control limit of ±20% RPD for each matrix applies for sample values greater than 10 times Detection Limit"

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 6 -
DUPLICATE SAMPLE SUMMARY

Client: Viva Nova Inc **Level:** LOW **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Matrix: Solid **Sample ID:** Q2832-09MS **Client ID:** TG-S05MSD

Percent Solids for Sample: 90.9 **Duplicate ID** Q2832-09MSD **Percent Solids for Spike Sample:** 90.9

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Silicon	mg/Kg	20	1150	D	1150	D	0		P

"A control limit of ±20% RPD for each matrix applies for sample values greater than 10 times Detection Limit"

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LABORATORY CONTROL SAMPLE SUMMARY

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169215BS							
Silicon	mg/Kg	46.7	46.7		100	80 - 120	P

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

-9 -

ICP SERIAL DILUTIONS

SAMPLE NO.

TG-S05L

Lab Name: Alliance **Contract:** VIVA01

Lab Code: ACE **Lb No.:** lb136801 **Lab Sample ID :** Q2832-09L **SDG No.:** Q2829

Matrix (soil/water): Solid **Level (low/med):** LOW

Concentration Units: mg/Kg

Analyte	Initial Sample Result (I)	C	Serial Dilution Result (S)	C	% Difference	Q	M
Silicon	1100	OR	1190		8		P

METAL PREPARATION & INSTRUMENT DATA

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Metals
- 11 -
ICP INTERELEMENT CORRECTION FACTORS

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE

Instrument ID: **Date:**

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave-Length (nm)	ICP Interelement Correction Factors For:				

METAL PREPARATION & ANALYICAL SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals
- 13 -
SAMPLE PREPARATION SUMMARY

Client: Viva Nova Inc **SDG No.:** Q2829

Contract: VIVA01 **Lab Code:** ACE **Method:**

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(g)	Final Sample Volume (mL)	Percent Solids
Batch Number:	**PB169215**						
PB169215BL	PB169215BL	MB	SOLID	08/12/2025	2.00	100.0	100.00
PB169215BS	PB169215BS	LCS	SOLID	08/12/2025	2.00	100.0	100.00
Q2829-01	SILICA	SAM	SOLID	08/12/2025	2.36	100.0	100.00
Q2832-09DUP	TG-S05DUP	DUP	SOLID	08/12/2025	2.30	100.0	90.90
Q2832-09MS	TG-S05MS	MS	SOLID	08/12/2025	2.31	100.0	90.90
Q2832-09MSD	TG-S05MSD	MSD	SOLID	08/12/2025	2.32	100.0	90.90

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

metals
- 14 -
ANALYSIS RUN LOG

Client: Viva Nova Inc **Contract:** VIVA01

Lab code: ACE **Sdg no.:** Q2829

Instrument id number: **Method:** **Run number:** LB136801

Start date: 08/13/2025 **End date:** 08/13/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1156	Si
S1	S1	1	1200	Si
S2	S2	1	1204	Si
S3	S3	1	1208	Si
S4	S4	1	1212	Si
S5	S5	1	1217	Si
ICV01	ICV01	1	1221	Si
LLICV01	LLICV01	1	1228	Si
ICB01	ICB01	1	1233	Si
CRI01	CRI01	1	1237	Si
ICSA01	ICSA01	1	1241	Si
ICSAB01	ICSAB01	1	1245	Si
CCV01	CCV01	1	1258	Si
CCB01	CCB01	1	1303	Si
CCV02	CCV02	1	1342	Si
CCB02	CCB02	1	1346	Si
CCV03	CCV03	1	1433	Si
CCB03	CCB03	1	1437	Si
PB169215BL	PB169215BL	1	1515	Si
CCV04	CCV04	1	1523	Si
CCB04	CCB04	1	1527	Si
CCV05	CCV05	1	1614	Si
CCB05	CCB05	1	1619	Si
Q2829-01	SILICA	1	1648	Si
Q2832-09L	TG-S05L	5	1700	Si
CCV06	CCV06	1	1705	Si
CCB06	CCB06	1	1709	Si
PB169215BS	PB169215BS	1	1742	Si
CCV07	CCV07	1	1746	Si
CCB07	CCB07	1	1750	Si

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

metals
- 14 -
ANALYSIS RUN LOG

Client: Viva Nova Inc **Contract:** VIVA01

Lab code: ACE **Sdg no.:** Q2829


Instrument id number: **Method:** **Run number:** LB136973

Start date: 08/26/2025 **End date:** 08/26/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1049	Si
S1	S1	1	1053	Si
S2	S2	1	1058	Si
S3	S3	1	1102	Si
S4	S4	1	1106	Si
S5	S5	1	1110	Si
ICV01	ICV01	1	1114	Si
LLICV01	LLICV01	1	1120	Si
ICB01	ICB01	1	1124	Si
CRI01	CRI01	1	1129	Si
ICSA01	ICSA01	1	1133	Si
ICSAB01	ICSAB01	1	1137	Si
CCV01	CCV01	1	1150	Si
CCB01	CCB01	1	1154	Si
Q2832-09DUP	TG-S05DUP	10	1216	Si
Q2832-09MS	TG-S05MS	10	1225	Si
Q2832-09MSD	TG-S05MSD	10	1229	Si
CCV02	CCV02	1	1241	Si
CCB02	CCB02	1	1245	Si
CCV03	CCV03	1	1330	Si
CCB03	CCB03	1	1335	Si
CCV04	CCV04	1	1420	Si
CCB04	CCB04	1	1424	Si

METAL
RAW DATA

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




As 189.042 {479}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.000117	Re-Slope:	1.000000
A1 (Gain):	0.051802	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999647	Status:	OK.
Std Error of Est:	0.000032		
Predicted MDL:	0.002146		
Predicted MQL:	0.007153		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	-.00012	.000	1
S1	.02000	.02154	.002	7.68	.00099	.000	1
S3	2.5000	2.4240	-.076	-3.04	.12531	.001	1
S4	5.0000	4.8667	-.133	-2.67	.25171	.001	1
S5	10.000	10.239	.239	2.39	.52972	.001	1
S2	.80000	.76889	-.031	-3.89	.03967	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Tl 190.856 {477}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000991	Re-Slope:	1.000000
A1 (Gain):	0.064329	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999950	Status:	OK.
Std Error of Est:	0.000021		
Predicted MDL:	0.001756		
Predicted MQL:	0.005854		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00099	.000	1
S1	.04000	.04239	.002	5.97	.00355	.000	1
S3	2.5000	2.5264	.026	1.06	.16058	.003	1
S4	5.0000	4.9348	-.065	-1.30	.31258	.010	1
S5	10.000	10.017	.017	.173	.63366	.014	1
S2	.80000	.81920	.019	2.40	.05275	.001	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Pb 220.353 {453}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000182	Re-Slope:	1.000000
A1 (Gain):	0.194780	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999619	Status:	OK.
Std Error of Est:	0.000097		
Predicted MDL:	0.001287		
Predicted MQL:	0.004290		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00018	.000	1
S1	.01200	.01099	-.001	-8.45	.00228	.000	1
S3	2.5000	2.4317	-.068	-2.73	.47333	.001	1
S4	5.0000	4.8411	-.159	-3.18	.94213	.001	1
S5	10.000	10.251	.251	2.51	1.9948	.007	1
S2	.80000	.77732	-.023	-2.83	.15143	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Se 196.090 {472}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000420	Re-Slope:	1.000000
A1 (Gain):	0.048357	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999648	Status:	OK.
Std Error of Est:	0.000030		
Predicted MDL:	0.002878		
Predicted MQL:	0.009594		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00042	.000	1
S1	.02000	.01951	-.000	-2.46	.00136	.000	1
S3	2.5000	2.4318	-.068	-2.73	.11798	.001	1
S4	5.0000	4.8553	-.145	-2.89	.23514	.002	1
S5	10.000	10.242	.242	2.42	.49555	.001	1
S2	.80000	.77159	-.028	-3.55	.03772	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Sb 206.833 {463}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000614	Re-Slope:	1.000000
A1 (Gain):	0.095945	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999554	Status:	OK.
Std Error of Est:	0.000107		
Predicted MDL:	0.001837		
Predicted MQL:	0.006125		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00061	.000	1
S1	.05000	.05143	.001	2.86	.00556	.000	1
S3	2.5000	2.4172	-.083	-3.31	.23318	.001	1
S4	5.0000	4.8364	-.164	-3.27	.46595	.002	1
S5	10.000	10.272	.272	2.72	.98878	.001	1
S2	.80000	.77287	-.027	-3.39	.07498	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Al 396.152 { 85}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000958	Re-Slope:	1.000000
A1 (Gain):	0.076561	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999527	Status:	OK.
Std Error of Est:	0.000178		
Predicted MDL:	0.010335		
Predicted MQL:	0.034449		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00001	.000	.000	.00096	.001	1
S1	.10000	.09588	-.004	-4.12	.00889	.001	1
S3	5.0000	4.8745	-.125	-2.51	.38158	.001	1
S4	10.000	9.5917	-.408	-4.08	.75016	.001	1
S5	20.000	20.562	.562	2.81	1.6049	.003	1
S2	1.6000	1.5765	-.023	-1.47	.12403	.001	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Ba 493.409 { 68}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.091138	Re-Slope:	1.000000
A1 (Gain):	2.366509	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999400	Status:	OK.
Std Error of Est:	0.006080		
Predicted MDL:	0.000990		
Predicted MQL:	0.003302		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.09115	.002	1
S1	.10000	.10350	.004	3.50	.33608	.001	1
S3	5.0000	4.8646	-.135	-2.71	11.603	.031	1
S4	10.000	9.5554	-.445	-4.45	22.704	.093	1
S5	20.000	20.621	.621	3.10	48.890	.796	1
S2	1.6000	1.5558	-.044	-2.77	3.7728	.008	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Be 234.861 {144}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.000711	Re-Slope:	1.000000
A1 (Gain):	6.529471	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999868	Status:	OK.
Std Error of Est:	0.000304		
Predicted MDL:	0.000068		
Predicted MQL:	0.000227		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	-.00072	.000	1
S1	.00600	.00652	.001	8.65	.04156	.000	1
S3	.12500	.12568	.001	.545	.81626	.006	1
S4	.25000	.24401	-.006	-2.39	1.5853	.010	1
S5	.50000	.50397	.004	.794	3.2753	.018	1
S2	.04000	.04082	.001	2.05	.26465	.001	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Cd 226.502 {449}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000006	Re-Slope:	1.000000
A1 (Gain):	2.935216	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999545	Status:	OK.
Std Error of Est:	0.000805		
Predicted MDL:	0.000073		
Predicted MQL:	0.000244		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00001	.000	1
S1	.00600	.00575	-.000	-4.14	.01695	.000	1
S3	1.2500	1.2035	-.046	-3.72	3.5350	.002	1
S4	2.5000	2.4315	-.068	-2.74	7.1419	.005	1
S5	5.0000	5.1358	.136	2.72	15.084	.047	1
S2	.40000	.37946	-.021	-5.13	1.1146	.001	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Ca 373.690 { 90}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.002587	Re-Slope:	1.000000
A1 (Gain):	0.056985	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999640	Status:	OK.
Std Error of Est:	0.000809		
Predicted MDL:	0.010881		
Predicted MQL:	0.036272		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00004	-.000	.000	.00259	.000	1
S2	4.0000	3.9770	-.023	-.575	.22922	.001	1
S3	12.500	12.280	-.220	-1.76	.70238	.001	1
S4	25.000	24.074	-.926	-3.71	1.3744	.001	1
S5	50.000	51.052	1.05	2.10	2.9118	.009	1
S1	2.0000	2.1169	.117	5.85	.12322	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Cr 267.716 {126}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000235	Re-Slope:	1.000000
A1 (Gain):	0.194899	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999850	Status:	OK.
Std Error of Est:	0.000025		
Predicted MDL:	0.000328		
Predicted MQL:	0.001093		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00023	.000	1
S1	.01000	.01062	.001	6.22	.00231	.000	1
S3	.50000	.49493	-.005	-1.01	.09673	.000	1
S4	1.0000	.97554	-.024	-2.45	.19044	.001	1
S5	2.0000	2.0289	.029	1.44	.39581	.001	1
S2	.16000	.16005	.000	.029	.03144	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Co 228.616 {448}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000233	Re-Slope:	1.000000
A1 (Gain):	1.022573	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999476	Status:	OK.
Std Error of Est:	0.000674		
Predicted MDL:	0.000186		
Predicted MQL:	0.000621		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00024	.000	1
S1	.03000	.03031	.000	1.02	.03109	.000	1
S3	1.2500	1.2002	-.050	-3.99	1.2297	.001	1
S4	2.5000	2.4207	-.079	-3.17	2.4799	.003	1
S5	5.0000	5.1471	.147	2.94	5.2722	.013	1
S2	.40000	.38174	-.018	-4.56	.39128	.001	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Cu 224.700 {450}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000461	Re-Slope:	1.000000
A1 (Gain):	0.649328	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999855	Status:	OK.
Std Error of Est:	0.000131		
Predicted MDL:	0.000410		
Predicted MQL:	0.001367		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00046	.000	1
S1	.02000	.02184	.002	9.19	.01479	.000	1
S3	.62500	.61951	-.005	-.878	.40611	.001	1
S4	1.2500	1.2202	-.030	-2.38	.79954	.001	1
S5	2.5000	2.5328	.033	1.31	1.6586	.001	1
S2	.20000	.20060	.001	.302	.13180	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Fe 240.488 {140}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000061	Re-Slope:	1.000000
A1 (Gain):	0.003475	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999264	Status:	OK.
Std Error of Est:	0.000007		
Predicted MDL:	0.006094		
Predicted MQL:	0.020313		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00001	.000	.000	.00006	.000	1
S1	.10000	.09774	-.002	-2.26	.00040	.000	1
S3	2.5000	2.3953	-.105	-4.19	.00831	.000	1
S4	5.0000	4.7713	-.229	-4.57	.01649	.000	1
S5	10.000	10.342	.342	3.42	.03570	.000	1
S2	.80000	.79341	-.007	-.823	.00279	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Mn 257.610 {131}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000377	Re-Slope:	1.000000
A1 (Gain):	0.410630	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999681	Status:	OK.
Std Error of Est:	0.000172		
Predicted MDL:	0.000548		
Predicted MQL:	0.001828		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00038	.000	1
S1	.02000	.02244	.002	12.2	.00960	.000	1
S3	1.2500	1.2313	-.019	-1.50	.50607	.001	1
S4	2.5000	2.4121	-.088	-3.52	.99105	.004	1
S5	5.0000	5.1044	.104	2.09	2.0968	.010	1
S2	.40000	.39988	-.000	-.029	.16461	.001	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Mg 279.079 {121}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000111	Re-Slope:	1.000000
A1 (Gain):	0.011396	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999505	Status:	OK.
Std Error of Est:	0.000190		
Predicted MDL:	0.025511		
Predicted MQL:	0.085037		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00013	-.000	.000	.00011	.000	1
S2	4.0000	3.9655	-.034	-.861	.04530	.000	1
S3	12.500	12.161	-.339	-2.71	.13869	.001	1
S4	25.000	24.086	-.914	-3.66	.27459	.001	1
S5	50.000	51.054	1.05	2.11	.58192	.003	1
S1	2.0000	2.2343	.234	11.7	.02557	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Ni 231.604 {446}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.000402	Re-Slope:	1.000000
A1 (Gain):	0.592023	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999554	Status:	OK.
Std Error of Est:	0.000415		
Predicted MDL:	0.000342		
Predicted MQL:	0.001142		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	-.00040	.000	1
S1	.04000	.04057	.001	1.42	.02361	.000	1
S3	1.2500	1.2056	-.044	-3.55	.71333	.001	1
S4	2.5000	2.4237	-.076	-3.05	1.4345	.002	1
S5	5.0000	5.1357	.136	2.71	3.0400	.008	1
S2	.40000	.38441	-.016	-3.90	.22718	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Ag 328.068 {103}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.000676	Re-Slope:	1.000000
A1 (Gain):	0.202270	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999619	Status:	OK.
Std Error of Est:	0.000046		
Predicted MDL:	0.000494		
Predicted MQL:	0.001648		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	-.00068	.000	1
S1	.01000	.01010	.000	.965	.00135	.000	1
S3	.62500	.60872	-.016	-2.61	.12184	.000	1
S4	1.2500	1.2064	-.044	-3.49	.24211	.000	1
S5	2.5000	2.5613	.061	2.45	.51494	.001	1
S2	.20000	.19853	-.001	-.735	.03928	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Na 589.592 { 57}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.002148	Re-Slope:	1.000000
A1 (Gain):	0.008998	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.997709	Status:	OK.
Std Error of Est:	0.000323		
Predicted MDL:	0.016861		
Predicted MQL:	0.056205		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00049	.000	.000	.00215	.000	1
S2	4.0000	3.5761	-.424	-10.6	.03433	.000	1
S3	12.500	11.451	-1.05	-8.40	.10518	.001	1
S4	25.000	23.455	-1.54	-6.18	.21320	.002	1
S5	50.000	53.125	3.12	6.25	.48016	.002	1
S1	2.0000	1.8930	-.107	-5.35	.01918	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




V 292.402 {115}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.000121	Re-Slope:	1.000000
A1 (Gain):	0.129419	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999452	Status:	OK.
Std Error of Est:	0.000099		
Predicted MDL:	0.002642		
Predicted MQL:	0.008805		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	-.00012	.000	1
S1	.04000	.04030	.000	.762	.00490	.000	1
S3	1.2500	1.2083	-.042	-3.34	.15389	.001	1
S4	2.5000	2.4029	-.097	-3.88	.30614	.001	1
S5	5.0000	5.1481	.148	2.96	.65670	.002	1
S2	.40000	.39043	-.010	-2.39	.04965	.001	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Zn 206.200 {464}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.002224	Re-Slope:	1.000000
A1 (Gain):	1.086588	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999920	Status:	OK.
Std Error of Est:	0.000323		
Predicted MDL:	0.000148		
Predicted MQL:	0.000494		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00222	.000	1
S1	.04000	.04263	.003	6.58	.04855	.000	1
S3	1.2500	1.2478	-.002	-.173	1.3581	.001	1
S4	2.5000	2.4535	-.046	-1.86	2.6682	.006	1
S5	5.0000	5.0386	.039	.772	5.4771	.019	1
S2	.40000	.40741	.007	1.85	.44491	.001	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Zn 213.856 {158}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.006101	Re-Slope:	1.000000
A1 (Gain):	2.346035	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999718	Status:	OK.
Std Error of Est:	0.001318		
Predicted MDL:	0.000239		
Predicted MQL:	0.000798		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00610	.000	1
S1	.04000	.04146	.001	3.65	.10403	.001	1
S3	1.2500	1.2206	-.029	-2.35	2.8900	.019	1
S4	2.5000	2.4240	-.076	-3.04	5.7337	.007	1
S5	5.0000	5.1050	.105	2.10	12.064	.027	1
S2	.40000	.39894	-.001	-.266	.94855	.003	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




K 766.490 { 44}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.001111	Re-Slope:	1.000000
A1 (Gain):	0.004888	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.998355	Status:	OK.
Std Error of Est:	0.000149		
Predicted MDL:	0.037068		
Predicted MQL:	0.123562		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00040	.000	.000	-.00111	.000	1
S2	4.0000	3.6503	-.350	-8.74	.01673	.000	1
S3	12.500	11.631	-.869	-6.95	.05575	.000	1
S4	25.000	23.647	-1.35	-5.41	.11449	.001	1
S5	50.000	52.652	2.65	5.30	.25628	.001	1
S1	2.0000	1.9200	-.080	-4.00	.00827	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




B 249.678 {135}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.004322	Re-Slope:	1.000000
A1 (Gain):	0.677407	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999891	Status:	OK.
Std Error of Est:	0.000522		
Predicted MDL:	0.000710		
Predicted MQL:	0.002365		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00432	.001	1
S1	.10000	.10509	.005	5.09	.07511	.001	1
S3	2.5000	2.4955	-.005	-.182	1.6887	.015	1
S4	5.0000	4.8874	-.113	-2.25	3.3029	.018	1
S5	10.000	10.103	.103	1.03	6.8239	.024	1
S2	.80000	.80887	.009	1.11	.55030	.003	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Mo 202.030 {467}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000737	Re-Slope:	1.000000
A1 (Gain):	0.664236	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999761	Status:	OK.
Std Error of Est:	0.001081		
Predicted MDL:	0.000254		
Predicted MQL:	0.000847		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00074	.000	1
S1	.20000	.20886	.009	4.43	.13947	.001	1
S3	2.5000	2.4459	-.054	-2.16	1.6254	.005	1
S4	5.0000	4.8633	-.137	-2.73	3.2311	.009	1
S5	10.000	10.187	.187	1.87	6.7674	.007	1
S2	.80000	.79479	-.005	-.651	.52867	.001	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Sn 189.989 {478}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000375	Re-Slope:	1.000000
A1 (Gain):	0.140910	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999689	Status:	OK.
Std Error of Est:	0.000116		
Predicted MDL:	0.000640		
Predicted MQL:	0.002132		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00038	.000	1
S1	.04000	.04056	.001	1.40	.00608	.000	1
S3	2.5000	2.4362	-.064	-2.55	.34330	.000	1
S4	5.0000	4.8659	-.134	-2.68	.68531	.001	1
S5	10.000	10.226	.226	2.26	1.4399	.006	1
S2	.80000	.77101	-.029	-3.62	.10890	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Ti 336.121 {100}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.001461	Re-Slope:	1.000000
A1 (Gain):	0.289410	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999484	Status:	OK.
Std Error of Est:	0.000308		
Predicted MDL:	0.001532		
Predicted MQL:	0.005108		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00146	.000	1
S1	.04000	.04053	.001	1.33	.01318	.000	1
S3	2.5000	2.4300	-.070	-2.80	.70448	.002	1
S4	5.0000	4.7971	-.203	-4.06	1.3893	.004	1
S5	10.000	10.289	.289	2.89	2.9783	.008	1
S2	.80000	.78286	-.017	-2.14	.22794	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Si 288.158 {117}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000168	Re-Slope:	1.000000
A1 (Gain):	0.003278	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.998379	Status:	OK.
Std Error of Est:	0.000020		
Predicted MDL:	0.011631		
Predicted MQL:	0.038769		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00010	.000	.000	.00017	.000	1
S1	.40000	.36203	-.038	-9.49	.00136	.000	1
S3	2.5000	2.3216	-.178	-7.14	.00790	.000	1
S4	5.0000	4.7525	-.248	-4.95	.01599	.000	1
S5	10.000	10.531	.531	5.31	.03518	.000	1
S2	.80000	.73234	-.068	-8.46	.00261	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




P 177.495 {490}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.000608	Re-Slope:	1.000000
A1 (Gain):	0.025515	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999697	Status:	OK.
Std Error of Est:	0.000015		
Predicted MDL:	0.002628		
Predicted MQL:	0.008761		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	-.00061	.000	1
S1	.02000	.01831	-.002	-8.43	-.00014	.000	1
S3	2.5000	2.4392	-.061	-2.43	.06162	.000	1
S4	5.0000	4.8970	-.103	-2.06	.12432	.000	1
S5	10.000	10.213	.213	2.13	.25994	.003	1
S2	.80000	.75251	-.047	-5.94	.01859	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




S 182.034 {485}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000268	Re-Slope:	1.000000
A1 (Gain):	0.028687	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.998847	Status:	OK.
Std Error of Est:	0.000032		
Predicted MDL:	0.002850		
Predicted MQL:	0.009500		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00001	.000	.000	.00027	.000	1
S1	.02000	.01547	-.005	-22.6	.00069	.000	1
S3	2.5000	2.3514	-.149	-5.95	.06753	.000	1
S4	5.0000	4.8167	-.183	-3.67	.13807	.001	1
S5	10.000	10.417	.417	4.17	.29835	.002	1
S2	.80000	.71942	-.081	-10.1	.02085	.000	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Li 670.784 { 50}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.008509	Re-Slope:	1.000000
A1 (Gain):	0.844447	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999199	Status:	OK.
Std Error of Est:	0.000793		
Predicted MDL:	0.001761		
Predicted MQL:	0.005870		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00851	.002	1
S5	10.000	10.365	.365	3.65	8.7676	.018	1
S4	5.0000	4.7650	-.235	-4.70	4.0357	.015	1
S3	2.5000	2.3997	-.100	-4.01	2.0366	.003	1
S1	.02000	.01873	-.001	-6.35	.02462	.001	1
S2	.80000	.77199	-.028	-3.50	.66096	.001	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Sr 407.771 { 83}

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.003479	Re-Slope:	1.000000
A1 (Gain):	7.696765	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999695	Status:	OK.
Std Error of Est:	0.004457		
Predicted MDL:	0.000099		
Predicted MQL:	0.000331		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00348	.001	1
S1	.02000	.02037	.000	1.86	.16101	.000	1
S3	2.5000	2.4354	-.065	-2.58	18.767	.251	1
S4	5.0000	4.8500	-.150	-3.00	37.370	.137	1
S5	10.000	10.222	.222	2.22	78.755	.586	1
S2	.80000	.79198	-.008	-1.00	6.1051	.006	1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801




Y 224.306 {450}*

Date of Fit:	8/13/2025 16:03:25	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000000	Re-Slope:	1.000000
A1 (Gain):	0.000000	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.000000	Status:	Warning Zero Gain
Std Error of Est:	0.000000		
Predicted MDL:	n/a		
Predicted MQL:	n/a		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis



Y 360.073 { 94}*


Date of Fit:	<not fit>	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000000	Re-Slope:	1.000000
A1 (Gain):	0.000000	Y-int:	0.000000

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

A2 (Curvature):	0.000000				
n (Exponent):	1.000000				
Correlation:	0.000000	Status:	Warning	Zero Gain	
Std Error of Est:	0.000000				
Predicted MDL:	n/a				
Predicted MQL:	n/a				

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis




Y 371.030 { 91}*

Date of Fit:	<not fit>	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000000	Re-Slope:	1.000000
A1 (Gain):	0.000000	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.000000	Status:	Warning Zero Gain
Std Error of Est:	0.000000		
Predicted MDL:	n/a		
Predicted MQL:	n/a		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis


Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801



Y 224.306 {150}*

Date of Fit:	<not fit>	Type of Fit:	Linear	Weighting:	1/Conc
A0 (Offset):	0.000000		Re-Slope:	1.000000	
A1 (Gain):	0.000000		Y-int:	0.000000	
A2 (Curvature):	0.000000				
n (Exponent):	1.000000				
Correlation:	0.000000		Status:	Warning	Zero Gain
Std Error of Est:	0.000000				
Predicted MDL:	n/a				
Predicted MQL:	n/a				

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis



In 230.606 {446}*

Date of Fit:	<not fit>	Type of Fit:	Linear	Weighting:	1/Conc
A0 (Offset):	0.000000		Re-Slope:	1.000000	
A1 (Gain):	0.000000		Y-int:	0.000000	

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

A2 (Curvature):		0.000000					
n (Exponent):		1.000000					
Correlation:		0.000000		Status:	Warning	Zero Gain	
Std Error of Est:		0.000000					
Predicted MDL:		n/a					
Predicted MQL:		n/a					

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: S0 Acquired: 8/13/2025 11:56:09 Type: Cal
Method: 6010-200.7 NEW(v96) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934	Be2348
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00012**	**.00099**	**.00018**	**.00042**	**.00061**	**.00096**	**.09115**	**-.00072**
Stddev	.00001	.00008	.00023	.00004	.00002	.00054	.00169	.00032
%RSD	11.798	8.2372	126.08	9.5520	3.8011	56.597	1.8506	45.393
#1	-.00011	.00090	.00044	.00041	.00064	.00033	.08965	-.00044
#2	-.00011	.00105	.00010	.00039	.00060	.00128	.09082	-.00107
#3	-.00013	.00103	.00001	.00047	.00060	.00126	.09297	-.00063

Elem	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404	Mn2576	Mg2790
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00001**	**.00259**	**.00023**	**.00024**	**.00046**	**.00006**	**.00038**	**.00011**
Stddev	.00012	.00023	.00003	.00019	.00014	.00002	.00016	.00026
%RSD	1445.8	8.8787	10.873	79.569	31.413	26.412	41.366	237.07
#1	.00006	.00236	.00021	.00038	.00040	.00005	.00056	-.00017
#2	.00010	.00282	.00025	.00031	.00036	.00005	.00029	.00017
#3	-.00013	.00258	.00024	.00002	.00062	.00008	.00029	.00034

Elem	Ni2316	Ag3280	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00040**	**-.00068**	**.00215**	**-.00012**	**.00610**	**-.00111**	**.00432**	**.00074**
Stddev	.00015	.00004	.00007	.00017	.00034	.00011	.00059	.00007
%RSD	36.354	5.4678	3.4595	141.44	5.5783	10.063	13.565	10.186
#1	-.00051	-.00071	.00207	-.00032	.00577	-.00104	.00382	.00068
#2	-.00046	-.00064	.00221	-.00005	.00645	-.00104	.00418	.00082
#3	-.00024	-.00068	.00218	-.00000	.00608	-.00124	.00496	.00071

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00038**	**.00146**	**.00017**	**-.00061**	**.00027**	**.00851**	**.00348**
Stddev	.00011	.00026	.00004	.00004	.00002	.00192	.00068
%RSD	28.119	17.986	24.511	6.9450	8.5623	22.614	19.590
#1	.00049	.00139	.00018	-.00064	.00029	.01058	.00315
#2	.00035	.00124	.00020	-.00063	.00025	.00817	.00427
#3	.00029	.00175	.00012	-.00056	.00027	.00678	.00303

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: S0 Acquired: 8/13/2025 11:56:09 Type: Cal
Method: 6010-200.7 NEW(v96) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5846.1	103620.	10717.	4580.8	7518.8
Stddev	9.1	117.	61.	12.7	13.9
%RSD	.15484	.11285	.56536	.27636	.18486
#1	5840.1	103630.	10779.	4578.3	7505.2
#2	5856.5	103740.	10713.	4569.6	7533.0
#3	5841.7	103500.	10658.	4594.6	7518.2

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: S1 Acquired: 8/13/2025 12:00:31 Type: Cal
Method: 6010-200.7 NEW(v96) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934	Be2348
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00099**	**.00355**	**.00228**	**.00136**	**.00556**	**.00889**	**.33608**	**.04156**
Stddev	.00005	.00006	.00017	.00003	.00005	.00066	.00075	.00037
%RSD	4.8850	1.5813	7.4028	2.2296	.85816	7.4060	.22369	.88732
#1	.00105	.00362	.00245	.00133	.00561	.00814	.33565	.04114
#2	.00097	.00352	.00211	.00139	.00553	.00921	.33695	.04177
#3	.00096	.00352	.00227	.00137	.00554	.00933	.33565	.04178

Elem	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404	Mn2576	Mg2790
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.01695**	**.12322**	**.00231**	**.03109**	**.01479**	**.00040**	**.00960**	**.02557**
Stddev	.00025	.00033	.00008	.00014	.00016	.00001	.00019	.00029
%RSD	1.5039	.27128	3.6790	.46259	1.1040	3.2185	1.9325	1.1517
#1	.01723	.12357	.00240	.03125	.01493	.00039	.00972	.02591
#2	.01674	.12319	.00224	.03107	.01481	.00039	.00938	.02536
#3	.01687	.12290	.00228	.03097	.01461	.00041	.00969	.02545

Elem	Ni2316	Ag3280	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.02361**	**.00135**	**.01918**	**.00490**	**.10403**	**.00827**	**.07511**	**.13947**
Stddev	.00031	.00010	.00010	.00027	.00097	.00012	.00058	.00109
%RSD	1.3018	7.3903	.54544	5.5792	.93125	1.4218	.77058	.78236
#1	.02361	.00124	.01927	.00490	.10503	.00832	.07555	.14042
#2	.02392	.00143	.01906	.00518	.10309	.00837	.07532	.13828
#3	.02331	.00137	.01921	.00463	.10396	.00814	.07445	.13971

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00608**	**.01318**	**.00136**	**-.00014**	**.00069**	**.02462**	**.16101**
Stddev	.00005	.00035	.00001	.00005	.00004	.00110	.00037
%RSD	.85809	2.6841	1.0584	33.670	5.3741	4.4543	.22731
#1	.00603	.01359	.00137	-.00013	.00065	.02486	.16105
#2	.00611	.01293	.00136	-.00010	.00069	.02558	.16136
#3	.00612	.01303	.00134	-.00019	.00072	.02342	.16063

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: S1 Acquired: 8/13/2025 12:00:31 Type: Cal
Method: 6010-200.7 NEW(v96) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5596.1	98960.	10361.	4343.9	7086.6
Stddev	14.0	81.	27.	17.9	21.9
%RSD	.25060	.08209	.25637	.41295	.30968
#1	5581.6	99054.	10336.	4325.5	7061.3
#2	5609.6	98918.	10358.	4361.3	7099.9
#3	5597.2	98908.	10389.	4344.8	7098.7

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: S2 Acquired: 8/13/2025 12:04:50 Type: Cal
Method: 6010-200.7 NEW(v96) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934	Be2348
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.03967**	**.05275**	**.15143**	**.03772**	**.07498**	**.12403**	**3.7728**	**.26465**
Stddev	.00017	.00134	.00039	.00033	.00047	.00079	.0077	.00071
%RSD	.42943	2.5471	.25485	.86907	.62659	.64084	.20458	.26811
#1	.03986	.05430	.15122	.03806	.07525	.12320	3.7719	.26401
#2	.03952	.05205	.15119	.03740	.07443	.12411	3.7810	.26452
#3	.03963	.05190	.15187	.03770	.07524	.12479	3.7656	.26541

Elem	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404	Mn2576	Mg2790
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**1.1146**	**.22922**	**.03144**	**.39128**	**.13180**	**.00279**	**.16461**	**.04530**
Stddev	.0010	.00072	.00010	.00061	.00041	.00001	.00103	.00012
%RSD	.08950	.31246	.30475	.15648	.30983	.47946	.62526	.25702
#1	1.1137	.22890	.03148	.39071	.13195	.00280	.16344	.04533
#2	1.1144	.22871	.03151	.39122	.13134	.00281	.16537	.04518
#3	1.1156	.23004	.03133	.39193	.13211	.00278	.16503	.04540

Elem	Ni2316	Ag3280	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.22718**	**.03928**	**.03433**	**.04965**	**.94855**	**.01673**	**.55030**	**.52867**
Stddev	.00038	.00007	.00014	.00057	.00336	.00008	.00269	.00118
%RSD	.16859	.18510	.41651	1.1458	.35431	.46016	.48855	.22400
#1	.22712	.03922	.03446	.04908	.95191	.01673	.54997	.52957
#2	.22683	.03928	.03434	.05022	.94855	.01666	.54780	.52733
#3	.22759	.03936	.03418	.04966	.94519	.01681	.55314	.52911

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.10890**	**.22794**	**.00261**	**.01859**	**.02085**	**.66096**	**6.1051**
Stddev	.00025	.00027	.00011	.00016	.00006	.00124	.0056
%RSD	.22625	.11781	4.2783	.84896	.31127	.18771	.09106
#1	.10884	.22767	.00256	.01871	.02082	.65952	6.1067
#2	.10869	.22821	.00253	.01841	.02080	.66166	6.1096
#3	.10917	.22795	.00274	.01864	.02092	.66169	6.0989

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: S2 Acquired: 8/13/2025 12:04:50 Type: Cal
Method: 6010-200.7 NEW(v96) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5834.2	102520.	10899.	4543.5	7278.7
Stddev	20.7	109.	36.	24.3	11.2
%RSD	.35442	.10614	.33016	.53424	.15406
#1	5819.3	102440.	10928.	4519.2	7269.0
#2	5857.8	102480.	10911.	4543.5	7291.0
#3	5825.5	102650.	10859.	4567.7	7276.1

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: S3 Acquired: 8/13/2025 12:08:54 Type: Cal
Method: 6010-200.7 NEW(v96) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934	Be2348
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.12531	.16058	.47333	.11798	.23318	.38158	11.603	.81626
Stddev	.00058	.00332	.00073	.00073	.00074	.00072	.031	.00597
%RSD	.46412	2.0683	.15377	.61790	.31580	.18754	.26425	.73117
#1	.12467	.15726	.47319	.11744	.23264	.38084	11.638	.80962
#2	.12544	.16058	.47269	.11770	.23289	.38227	11.588	.82118
#3	.12581	.16390	.47412	.11881	.23402	.38164	11.583	.81798

Elem	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404	Mn2576	Mg2790
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	3.5350	.70238	.09673	1.2297	.40611	.00831	.50607	.13869
Stddev	.0024	.00099	.00024	.0012	.00073	.00005	.00116	.00076
%RSD	.06726	.14074	.24898	.09641	.17916	.54279	.22970	.54688
#1	3.5374	.70151	.09697	1.2306	.40542	.00836	.50715	.13790
#2	3.5326	.70346	.09674	1.2283	.40603	.00827	.50484	.13941
#3	3.5349	.70218	.09649	1.2300	.40687	.00830	.50623	.13877

Elem	Ni2316	Ag3280	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.71333	.12184	.10518	.15389	2.8900	.05575	1.6887	1.6254
Stddev	.00078	.00030	.00102	.00062	.0188	.00038	.0146	.0050
%RSD	.10954	.24376	.96671	.40450	.65142	.67359	.86512	.30965
#1	.71412	.12213	.10611	.15461	2.8818	.05614	1.6726	1.6213
#2	.71256	.12184	.10409	.15347	2.9115	.05539	1.7012	1.6238
#3	.71329	.12154	.10534	.15360	2.8767	.05571	1.6921	1.6310

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.34330	.70448	.00790	.06162	.06753	2.0366	18.767
Stddev	.00049	.00197	.00009	.00011	.00005	.0029	.251
%RSD	.14245	.27950	1.1216	.17170	.07487	.14186	1.3398
#1	.34346	.70603	.00800	.06173	.06758	2.0394	18.554
#2	.34275	.70515	.00783	.06151	.06748	2.0336	18.702
#3	.34369	.70226	.00788	.06162	.06754	2.0368	19.044

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: S3 Acquired: 8/13/2025 12:08:54 Type: Cal
Method: 6010-200.7 NEW(v96) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5699.2	100690.	10886.	4460.2	6930.0
Stddev	3.9	199.	97.	19.9	11.7
%RSD	.06886	.19730	.89149	.44656	.16855
#1	5696.3	100900.	10995.	4459.0	6920.0
#2	5697.6	100660.	10808.	4440.9	6927.2
#3	5703.7	100510.	10855.	4480.6	6942.8

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: S4 Acquired: 8/13/2025 12:12:59 Type: Cal
Method: 6010-200.7 NEW(v96) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934	Be2348
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.25171**	**.31258**	**.94213**	**.23514**	**.46595**	**.75016**	**22.704**	**1.5853**
Stddev	.00069	.01034	.00065	.00177	.00239	.00078	.093	.0097
%RSD	.27577	3.3092	.06919	.75396	.51292	.10357	.40974	.60926
#1	.25114	.30068	.94248	.23321	.46327	.75065	22.653	1.5887
#2	.25150	.31762	.94254	.23553	.46673	.75057	22.811	1.5927
#3	.25248	.31943	.94138	.23669	.46786	.74927	22.648	1.5743

Elem	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404	Mn2576	Mg2790
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**7.1419**	**1.3744**	**.19044**	**2.4799**	**.79954**	**.01649**	**.99105**	**.27459**
Stddev	.0046	.0008	.00058	.0031	.00135	.00014	.00440	.00068
%RSD	.06505	.06197	.30335	.12447	.16861	.84838	.44438	.24795
#1	7.1419	1.3754	.19006	2.4776	.79959	.01644	.99522	.27420
#2	7.1372	1.3737	.19016	2.4786	.79817	.01638	.98644	.27538
#3	7.1465	1.3742	.19111	2.4834	.80087	.01665	.99148	.27419

Elem	Ni2316	Ag3280	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**1.4345**	**.24211**	**.21320**	**.30614**	**5.7337**	**.11449**	**3.3029**	**3.2311**
Stddev	.0023	.00003	.00243	.00090	.0074	.00085	.0182	.0094
%RSD	.15854	.01145	1.1377	.29499	.12914	.74073	.55153	.29198
#1	1.4328	.24213	.21323	.30617	5.7299	.11433	3.3136	3.2217
#2	1.4336	.24212	.21075	.30522	5.7422	.11373	3.3132	3.2310
#3	1.4371	.24208	.21561	.30702	5.7290	.11540	3.2818	3.2406

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.68531**	**1.3893**	**.01599**	**.12432**	**.13807**	**4.0357**	**37.370**
Stddev	.00107	.0039	.00017	.00015	.00056	.0152	.137
%RSD	.15632	.27905	1.0925	.11901	.40504	.37562	.36677
#1	.68525	1.3930	.01595	.12431	.13743	4.0455	37.528
#2	.68427	1.3852	.01585	.12447	.13836	4.0182	37.305
#3	.68641	1.3896	.01619	.12417	.13843	4.0433	37.278

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: S4 Acquired: 8/13/2025 12:12:59 Type: Cal
Method: 6010-200.7 NEW(v96) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5645.0	100180.	11009.	4394.4	6724.3
Stddev	9.9	387.	45.	13.7	10.8
%RSD	.17453	.38664	.40904	.31203	.16008
#1	5637.2	100530.	11017.	4410.2	6712.5
#2	5656.1	100250.	10960.	4387.4	6733.6
#3	5641.7	99765.	11049.	4385.6	6726.7

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: S5 Acquired: 8/13/2025 12:17:10 Type: Cal
Method: 6010-200.7 NEW(v96) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934	Be2348
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.52972	.63366	1.9948	.49555	.98878	1.6049	48.890	3.2753
Stddev	.00057	.01358	.0073	.00077	.00085	.0029	.796	.0185
%RSD	.10790	2.1431	.36699	.15481	.08636	.18246	1.6274	.56399
#1	.52944	.63123	2.0027	.49576	.98862	1.6046	48.872	3.2867
#2	.53037	.64829	1.9937	.49619	.98802	1.6021	48.104	3.2540
#3	.52933	.62146	1.9881	.49470	.98970	1.6079	49.695	3.2852

Elem	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404	Mn2576	Mg2790
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	15.084	2.9118	.39581	5.2722	1.6586	.03570	2.0968	.58192
Stddev	.047	.0087	.00140	.0125	.0012	.00010	.0099	.00261
%RSD	.31162	.29870	.35251	.23744	.07043	.27258	.47176	.44855
#1	15.136	2.9107	.39742	5.2866	1.6599	.03562	2.0982	.58223
#2	15.073	2.9037	.39514	5.2641	1.6576	.03581	2.0863	.57917
#3	15.044	2.9210	.39488	5.2659	1.6584	.03567	2.1059	.58436

Elem	Ni2316	Ag3280	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	3.0400	.51494	.48016	.65670	12.064	.25628	6.8239	6.7674
Stddev	.0075	.00113	.00190	.00204	.027	.00097	.0245	.0069
%RSD	.24774	.22012	.39602	.31018	.22014	.38010	.35844	.10149
#1	3.0487	.51523	.48100	.65617	12.080	.25638	6.8324	6.7742
#2	3.0363	.51591	.48149	.65498	12.079	.25719	6.7963	6.7676
#3	3.0351	.51370	.47798	.65895	12.034	.25525	6.8429	6.7605

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.4399	2.9783	.03518	.25994	.29835	8.7676	78.755
Stddev	.0062	.0076	.00014	.00260	.00217	.0184	.586
%RSD	.43098	.25457	.39135	1.0012	.72611	.20932	.74368
#1	1.4454	2.9785	.03504	.26244	.29893	8.7630	78.239
#2	1.4412	2.9706	.03531	.26012	.30016	8.7520	78.635
#3	1.4332	2.9857	.03519	.25724	.29595	8.7879	79.392

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: S5 Acquired: 8/13/2025 12:17:10 Type: Cal
Method: 6010-200.7 NEW(v96) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5431.8	95688.	10758.	4147.7	6335.1
Stddev	6.9	332.	45.	6.1	13.8
%RSD	.12679	.34662	.42271	.14602	.21731
#1	5424.4	95521.	10733.	4142.1	6319.3
#2	5438.0	95473.	10810.	4146.9	6341.8
#3	5433.1	96070.	10730.	4154.1	6344.2

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: ICV01 Acquired: 8/13/2025 12:21:22 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICV01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**3.872529**	**3.747317**	**3.774946**	**3.879669**	**3.998496**	**7.680930**
Stddev	.003902	.037013	.002479	.002217	.003616	.014768
%RSD	.1007619	.9877262	.0656677	.0571342	.0904217	.1922650
#1	3.869111	3.761842	3.775273	3.880245	3.994378	7.664547
#2	3.876781	3.705245	3.772320	3.877221	3.999959	7.685023
#3	3.871695	3.774865	3.777245	3.881540	4.001150	7.693219

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**7.787087**	**.1894277**	**1.917346**	**19.17133**	**.7879995**	**1.951427**
Stddev	.066900	.0012904	.002480	.04239	.0021177	.002684
%RSD	.8591090	.6812079	.1293524	.2210986	.2687385	.1375592
#1	7.825703	.1883907	1.918168	19.13004	.7870095	1.952023
#2	7.825719	.1908728	1.914559	19.16923	.7865583	1.948495
#3	7.709838	.1890195	1.919310	19.21473	.7904309	1.953764

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.9971016**	**3.950722**	**1.921770**	**19.09299**	**1.945451**	**1.032062**
Stddev	.0002298	.013018	.007647	.07624	.001243	.001841
%RSD	.0230492	.3295101	.3979201	.3992970	.0638887	.1783370
#1	.9968415	3.961226	1.913017	19.02672	1.946515	1.029977
#2	.9971860	3.936157	1.925135	19.07593	1.944085	1.033462
#3	.9972774	3.954781	1.927157	19.17631	1.945754	1.032747

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**19.09833**	**1.937518**	**1.949359**	**19.37323**	**3.744833**	**3.976290**
Stddev	.11337	.000955	.006075	.14820	.031631	.002192
%RSD	.5936094	.0492879	.3116617	.7649614	.8446475	.0551334
#1	19.13358	1.938246	1.942348	19.41682	3.721577	3.978804
#2	18.97152	1.936437	1.952632	19.20813	3.780851	3.974774
#3	19.18989	1.937872	1.953095	19.49475	3.732071	3.975293

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: ICV01 Acquired: 8/13/2025 12:21:22 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICV01 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	3.853517	3.852559	3.736703	F .0077036	F .0271601	F -.002909
Stddev	.004442	.006257	.022650	.0014054	.0016283	.000859
%RSD	.1152699	.1624205	.6061370	18.24403	5.995220	29.52355
#1	3.852701	3.856439	3.762802	.0088118	.0256722	-.003731
#2	3.849539	3.845341	3.722186	.0061228	.0269086	-.002979
#3	3.858310	3.855899	3.725122	.0081763	.0288995	-.002017

Elem	Sr4077
Units	ppm
Avg	F -.003642
Stddev	.000035
%RSD	.9554361
#1	-.003668
#2	-.003602
#3	-.003654

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5737.902	102195.8	11664.72	4400.772	6955.925
Stddev	8.744	439.9	73.91	18.175	12.061
%RSD	.1523977	.4304161	.6336377	.4129998	.1733879
#1	5729.834	102378.4	11747.31	4404.439	6942.067
#2	5736.678	102514.9	11604.79	4416.834	6964.047
#3	5747.194	101694.0	11642.06	4381.043	6961.661

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: LLICV01 Acquired: 8/13/2025 12:28:47 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: LLICV01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0212869	.0429580	.0111791	.0192117	.0500587	.0951201	.0946827
Stddev	.0006105	.0016753	.0008049	.0015265	.0008149	.0080282	.0009918
%RSD	2.867938	3.899746	7.200015	7.945656	1.627822	8.440043	1.047528
#1	.0218636	.0410657	.0117637	.0203123	.0498314	.0892454	.0958260
#2	.0206474	.0435562	.0115126	.0174690	.0493816	.0918472	.0940538
#3	.0213498	.0442520	.0102611	.0198537	.0509631	.1042677	.0941682

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0060843	.0055941	2.040672	.0104963	.0294114	.0213323	.1068463
Stddev	.0000529	.0000093	.006249	.0000555	.0000699	.0001400	.0039101
%RSD	.8691142	.1670398	.3062115	.5291471	.2377351	.6562415	3.659527
#1	.0060266	.0055895	2.035366	.0104408	.0294738	.0214512	.1066157
#2	.0060959	.0056049	2.039089	.0104963	.0294246	.0213678	.1108666
#3	.0061304	.0055880	2.047559	.0105519	.0293359	.0211780	.1030567

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0210293	2.116955	.0395706	.0101235	1.848759	.0389247	.0424371
Stddev	.0002964	.023230	.0001587	.0003544	.012154	.0002070	.0005068
%RSD	1.409592	1.097332	.4009776	3.500519	.6574014	.5317571	1.194122
#1	.0208042	2.109694	.0397342	.0100837	1.852025	.0391251	.0425327
#2	.0213652	2.098222	.0394174	.0104961	1.835306	.0387117	.0418893
#3	.0209187	2.142948	.0395601	.0097907	1.858946	.0389372	.0428892

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.992274	.1113647	.2068390	.0383012	.0383031	.3725319	.0208482
Stddev	.025784	.0007750	.0005443	.0000473	.0010136	.0088729	.0009790
%RSD	1.294212	.6959354	.2631299	.1234586	2.646283	2.381779	4.695647
#1	2.013299	.1115285	.2071337	.0383426	.0386341	.3822381	.0198549
#2	2.000017	.1120447	.2062110	.0383112	.0391099	.3648378	.0208777
#3	1.963505	.1105208	.2071724	.0382497	.0371654	.3705199	.0218121

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: LLICV01 Acquired: 8/13/2025 12:28:47 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: LLICV01 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0182723	.0169958	.0196862
Stddev	.0009029	.0002173	.0000835
%RSD	4.941298	1.278290	.4239678
#1	.0173798	.0167453	.0197119
#2	.0182516	.0171085	.0195930
#3	.0191853	.0171335	.0197539

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5807.924	104228.6	11501.84	4465.294	7444.260
Stddev	15.704	182.0	26.10	16.352	12.443
%RSD	.2703829	.1746627	.2269126	.3662041	.1671490
#1	5790.519	104032.4	11473.85	4450.376	7441.534
#2	5812.221	104392.1	11525.51	4482.777	7433.405
#3	5821.031	104261.4	11506.15	4462.729	7457.840

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: ICB01 Acquired: 8/13/2025 12:33:04 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICB01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0020461	.0006381	-.000063	-.000772	.0010903	-.002729	-.004066
Stddev	.0012696	.0009465	.000677	.001823	.0010147	.003038	.000165
%RSD	62.04908	148.3491	1080.785	236.2314	93.05941	111.3193	4.051191
#1	.0031922	.0009866	-.000840	.000282	.0013354	-.002780	-.004239
#2	.0022647	.0013609	.000260	-.002877	.0019600	.000334	-.003912
#3	.0006814	-.000433	.000393	.000280	-.000024	-.005740	-.004046

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.000014	.0000246	.0017245	.0003299	.0001895	.0003046	.0079953
Stddev	.000020	.0000281	.0036808	.0000707	.0001241	.0004185	.0035244
%RSD	141.3374	114.2945	213.4394	21.41632	65.47087	137.4151	44.08091
#1	-.000033	-.000007	-.002495	.0004113	.0003019	.0007554	.0052188
#2	-.000017	.000035	.004273	.0002839	.0002102	.0002299	.0119603
#3	.000007	.000046	.003396	.0002946	.0000564	-.000072	.0068069

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0001986	.0058094	.0007518	.0000605	-.049108	.0005254	.0024536
Stddev	.0002193	.0140103	.0003315	.0003838	.004769	.0014947	.0002481
%RSD	110.4549	241.1671	44.09304	634.1919	9.710924	284.4610	10.11297
#1	.0004387	-.005213	.0004171	-.000383	-.054107	-.000838	.0021671
#2	.0000090	.021576	.0010800	.000282	-.044609	.002123	.0025914
#3	.0001479	.001065	.0007581	.000282	-.048608	.000291	.0026021

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0223962	.0069915	.0001878	.0007363	.0006101	-.005488	.0031793
Stddev	.0195386	.0003020	.0002695	.0004124	.0014125	.002475	.0022545
%RSD	87.24066	4.319832	143.5270	56.01670	231.5253	45.09840	70.91221
#1	.0139851	.0073368	-.000110	.0002985	.0011362	-.007188	.0057763
#2	.0447318	.0067764	.000257	.0011175	.0016839	-.006628	.0020369
#3	.0084718	.0068614	.000416	.0007929	-.000990	-.002649	.0017246

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: ICB01 Acquired: 8/13/2025 12:33:04 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICB01 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0009618	-.002107	-.000051
Stddev	.0030958	.000913	.000032
%RSD	321.8765	43.32888	63.30726
#1	.0035907	-.001119	-.000018
#2	-.002450	-.002919	-.000082
#3	.001745	-.002284	-.000053

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5920.177	105800.6	11530.31	4505.262	7705.778
Stddev	8.657	219.9	52.01	12.467	10.457
%RSD	.1462266	.2078285	.4510980	.2767269	.1357041
#1	5911.799	105655.1	11495.27	4490.964	7698.761
#2	5919.644	106053.6	11590.08	4513.864	7700.776
#3	5929.088	105693.2	11505.60	4510.958	7717.797

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CRI01 Acquired: 8/13/2025 12:37:23 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CRI01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0205849	.0406733	.0100082	.0210974	.0501601	.0928898	.0938578
Stddev	.0025719	.0016188	.0006348	.0034708	.0007561	.0136325	.0004453
%RSD	12.49421	3.980102	6.342546	16.45142	1.507352	14.67594	.4744453
#1	.0201310	.0415070	.0106766	.0173189	.0497882	.0820636	.0933436
#2	.0233536	.0417054	.0099343	.0241437	.0510301	.0884066	.0941134
#3	.0182702	.0388075	.0094135	.0218297	.0496620	.1081993	.0941165

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0060104	.0054936	2.009743	.0099982	.0290122	.0208848	.0959440
Stddev	.0000683	.0000424	.002405	.0003607	.0001192	.0002381	.0069841
%RSD	1.135550	.7720568	.1196751	3.607919	.4106902	1.139924	7.279389
#1	.0060431	.0055425	2.007013	.0104143	.0290463	.0207539	.0954112
#2	.0060560	.0054701	2.011551	.0098071	.0291106	.0211595	.0892415
#3	.0059319	.0054681	2.010665	.0097732	.0288798	.0207408	.1031793

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0208250	2.117900	.0386926	.0103327	1.812239	.0422618	.0410103
Stddev	.0000943	.029259	.0001655	.0003137	.002549	.0013977	.0006691
%RSD	.4529240	1.381504	.4276398	3.036040	.1406721	3.307217	1.631641
#1	.0209006	2.088228	.0386578	.0100788	1.809427	.0434929	.0417796
#2	.0208551	2.146728	.0388727	.0102360	1.812888	.0425501	.0406878
#3	.0207193	2.118745	.0385473	.0106834	1.814400	.0407424	.0405634

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.934812	.1007527	.2026697	.0373197	.0381128	.3571710	.0209912
Stddev	.054165	.0010837	.0006563	.0008661	.0007433	.0073746	.0023370
%RSD	2.799475	1.075571	.3238405	2.320852	1.950235	2.064722	11.13306
#1	1.991669	.1000124	.2033662	.0381476	.0385008	.3655936	.0212222
#2	1.883817	.1019965	.2025803	.0373917	.0385818	.3540456	.0185474
#3	1.928950	.1002492	.2020627	.0364199	.0372558	.3518738	.0232041

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CRI01 Acquired: 8/13/2025 12:37:23 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CRI01 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0155669	.0169876	.0194383
Stddev	.0011020	.0005103	.0000266
%RSD	7.079181	3.004134	.1366103
#1	.0166446	.0175666	.0194401
#2	.0156141	.0167930	.0194109
#3	.0144421	.0166033	.0194639

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5878.336	104791.1	11474.11	4497.042	7517.225
Stddev	1.598	168.5	52.55	39.435	13.274
%RSD	.0271765	.1608003	.4579527	.8769002	.1765840
#1	5876.496	104708.3	11520.15	4454.636	7504.717
#2	5879.138	104985.0	11416.86	4532.611	7515.805
#3	5879.373	104679.9	11485.30	4503.879	7531.152

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: ICSA01 Acquired: 8/13/2025 12:41:39 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSA01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0134755**	**.0029051**	**-.006379**	**.0029114**	**-.004818**	**240.2699**
Stddev	.0015729	.0020732	.001490	.0036164	.003585	.5040
%RSD	11.67198	71.36207	23.36131	124.2155	74.40217	.2097530
#1	.0138604	.0051580	-.004993	.0061073	-.003015	240.6353
#2	.0117460	.0010778	-.006189	.0036411	-.008947	240.4794
#3	.0148203	.0024795	-.007955	-.001014	-.002493	239.6949

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000470**	**.0013114**	**.0023261**	**229.2883**	**.0581592**	**.0018541**
Stddev	.000774	.0000873	.0001227	.4618	.0006300	.0001184
%RSD	164.6725	6.658843	5.275484	.2013844	1.083154	6.383889
#1	-.001110	.0013001	.0024222	229.7292	.0588785	.0019160
#2	-.000690	.0012303	.0021878	229.3273	.0578935	.0017176
#3	.000390	.0014038	.0023681	228.8082	.0577057	.0019286

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0048240**	**99.17169**	**.0081692**	**243.0898**	**.0052171**	**.0023332**
Stddev	.0002039	.45403	.0001293	.4697	.0001901	.0004527
%RSD	4.227024	.4578214	1.583067	.1932077	3.643411	19.40112
#1	.0049235	99.35140	.0082993	243.6319	.0051780	.0019210
#2	.0045895	99.50836	.0080407	242.8316	.0050496	.0022609
#3	.0049592	98.65531	.0081676	242.8058	.0054237	.0028176

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0030553**	**.0036474**	**.0021994**	**-.004103**	**.0465611**	**.0004396**
Stddev	.0145555	.0010886	.0002810	.002135	.0024231	.0005469
%RSD	476.4051	29.84718	12.77678	52.04560	5.204049	124.4199
#1	.0192408	.0049021	.0024178	-.001849	.0450116	.0010710
#2	-.008960	.0030859	.0018824	-.004363	.0493534	.0001180
#3	-.001115	.0029541	.0022981	-.006096	.0453184	.0001296

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: ICSA01 Acquired: 8/13/2025 12:41:39 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSA01 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.003408**	**-.005384**	**.0191567**	**.0080433**	**-.007991**	F **-.029667**
Stddev	.000954	.000165	.0046077	.0006923	.006152	.002334
%RSD	27.99310	3.060358	24.05245	8.607079	76.97972	7.868185
#1	-.004471	-.005566	.0231435	.0087184	-.006961	-.031462
#2	-.003127	-.005341	.0141122	.0080764	-.014593	-.030511
#3	-.002626	-.005244	.0202145	.0073350	-.002420	-.027028

Elem	Sr4077
Units	ppm
Avg	**-.000415**
Stddev	.000141
%RSD	33.96231
#1	-.000341
#2	-.000577
#3	-.000326

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5204.532**	**91132.91**	**11089.70**	**3980.441**	**5989.069**
Stddev	4.463	395.10	51.37	18.704	3.737
%RSD	.0857542	.4335453	.4631804	.4698862	.0624027
#1	5200.730	90746.55	11031.60	3965.808	5986.924
#2	5203.419	91115.98	11108.42	3974.002	5993.384
#3	5209.446	91536.21	11129.08	4001.514	5986.898

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: ICSAB01 Acquired: 8/13/2025 12:45:51 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSAB01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1106466	.1128171	.0422017	.0505495	.5881123	246.7037
Stddev	.0050037	.0025893	.0034502	.0021315	.0030349	.6852
%RSD	4.522257	2.295135	8.175500	4.216752	.5160492	.2777456
#1	.1124155	.1106635	.0387382	.0493812	.5915942	247.4761
#2	.1049987	.1120977	.0456384	.0530098	.5867150	246.1690
#3	.1145255	.1156900	.0422284	.0492575	.5860276	246.4660

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.4924139	.4866317	.9942478	235.8991	.5580352	.4980184
Stddev	.0033397	.0019292	.0017136	.8031	.0014520	.0009333
%RSD	.6782315	.3964369	.1723476	.3404239	.2601970	.1873945
#1	.4956595	.4888437	.9932979	236.6501	.5584377	.4975656
#2	.4889874	.4852974	.9932195	235.0525	.5564244	.4973979
#3	.4925949	.4857540	.9962259	235.9946	.5592434	.4990917

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.4858476	101.6958	.4933530	249.5102	.9920598	.2159779
Stddev	.0008161	.4135	.0021714	.7409	.0003657	.0004686
%RSD	.1679772	.4066189	.4401321	.2969430	.0368587	.2169859
#1	.4867018	101.6401	.4950965	250.0856	.9922602	.2158847
#2	.4850758	101.3129	.4909208	248.6742	.9916377	.2155629
#3	.4857651	102.1343	.4940419	249.7708	.9922814	.2164862

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.003581	.4786019	1.036676	.0429249	.9922032	1.025640
Stddev	.008404	.0016721	.002684	.0122727	.0026526	.001398
%RSD	234.7125	.3493630	.2588793	28.59119	.2673424	.1363193
#1	-.000562	.4803976	1.039605	.0438434	.9934008	1.027173
#2	-.013077	.4783184	1.036091	.0302187	.9891630	1.024434
#3	.002897	.4770897	1.034334	.0547126	.9940459	1.025315

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: ICSAB01 Acquired: 8/13/2025 12:45:51 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSAB01 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.014192	.9962127	1.025751	F .0085425	F -.003250	F -.032460
Stddev	.002322	.0030495	.003653	.0045384	.001359	.000917
%RSD	.2289273	.3061054	.3560896	53.12684	41.81546	2.824814
#1	1.011822	.9987402	1.028019	.0051222	-.002467	-.032318
#2	1.014294	.9928257	1.021537	.0068141	-.002463	-.031622
#3	1.016462	.9970721	1.027696	.0136911	-.004819	-.033439

Elem	Sr4077
Units	ppm
Avg	F .0006371
Stddev	.0005195
%RSD	81.53226
#1	.0011466
#2	.0006567
#3	.0001082

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5276.807	92053.53	11147.28	4022.523	6063.212
Stddev	4.006	308.37	48.84	17.541	2.865
%RSD	.0759231	.3349901	.4381269	.4360805	.0472593
#1	5272.726	91974.87	11094.33	4006.839	6063.869
#2	5276.961	92393.61	11190.56	4041.465	6060.075
#3	5280.734	91792.10	11156.96	4019.266	6065.691

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: ICSADLX20 Acquired: 8/13/2025 12:49:58 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSA01DLX20 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0016662	.0004461	-.000930	-.000543	-.002258	11.81398	-.003443
Stddev	.0021118	.0015700	.000423	.001440	.000380	.04807	.000460
%RSD	126.7443	351.9226	45.49889	265.3630	16.81115	.4068997	13.36699
#1	.0016100	-.000817	-.000838	.000136	-.002189	11.85082	-.003069
#2	-.000417	-.000048	-.001391	.000433	-.001918	11.83152	-.003957
#3	.003806	.002204	-.000560	-.002196	-.002668	11.75960	-.003303

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0001197	.0000111	11.78764	.0027379	-.000011	.0003281	4.713823
Stddev	.0000386	.0000244	.00732	.0001998	.000070	.0002010	.011185
%RSD	32.21384	220.3649	.0620667	7.297680	638.5062	61.27001	.2372759
#1	.0001472	.0000170	11.77984	.0029549	.000069	.0002794	4.703251
#2	.0001362	-.000016	11.78872	.0026973	-.000042	.0001559	4.712684
#3	.0000756	.000032	11.79435	.0025615	-.000060	.0005490	4.725533

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0001094	12.22265	.0002019	-.000183	-.067916	.0012792	-.000401
Stddev	.0004360	.04381	.0001770	.000358	.025394	.0010677	.000237
%RSD	398.5806	.3584475	87.67820	195.9442	37.39100	83.46339	59.16842
#1	-.000390	12.24687	.0000780	.000229	-.042532	.0003052	-.000454
#2	.000304	12.17208	.0001230	-.000423	-.093321	.0024207	-.000606
#3	.000414	12.24901	.0004046	-.000355	-.067894	.0011116	-.000141

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.039682	.0002339	-.000140	-.001188	-.000089	.0019297	.0018500
Stddev	.005958	.0006861	.000045	.000313	.001067	.0032203	.0011160
%RSD	15.01318	293.3161	32.09339	26.35886	1197.441	166.8847	60.32268
#1	-.046553	.0008541	-.000191	-.001483	-.001279	.0055024	.0026517
#2	-.036549	-.000503	-.000110	-.000859	.000231	-.000749	.0023227
#3	-.035946	.000351	-.000117	-.001223	.000781	.001036	.0005755

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: ICSADLX20 Acquired: 8/13/2025 12:49:58 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSA01DLX20 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**-.002158**	**-.006689**	**-.000026**
Stddev	.002637	.001510	.000012
%RSD	122.1855	22.57041	45.33292
#1	.000680	-.005457	-.000033
#2	-.004532	-.008373	-.000031
#3	-.002623	-.006236	-.000012

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5846.827**	**102377.5**	**11319.59**	**4431.459**	**7393.786**
Stddev	14.168	530.5	58.08	28.994	6.043
%RSD	.2423127	.5181323	.5131334	.6542756	.0817341
#1	5832.313	101789.3	11252.53	4413.340	7393.536
#2	5860.621	102819.6	11353.47	4464.899	7399.950
#3	5847.547	102523.7	11352.79	4416.138	7387.871

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: ICSABDLX20 Acquired: 8/13/2025 12:54:14 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSAB01DLX20 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0051708	.0080030	.0014689	.0049421	.0284657	12.31042	.0200983
Stddev	.0014910	.0012738	.0004836	.0037558	.0005819	.02913	.0013867
%RSD	28.83515	15.91616	32.92407	75.99624	2.044308	.2366435	6.899455
#1	.0039537	.0066374	.0009183	.0081381	.0280884	12.28041	.0185246
#2	.0068339	.0091589	.0018251	.0008054	.0291359	12.31227	.0211409
#3	.0047248	.0082129	.0016633	.0058827	.0281729	12.33859	.0206294

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0253331	.0471222	12.20490	.0284416	.0237561	.0253892	4.821184
Stddev	.0003163	.0000573	.03281	.0003447	.0000645	.0000652	.016281
%RSD	1.248658	.1215111	.2687919	1.211763	.2715122	.2569712	.3376936
#1	.0249733	.0471883	12.17052	.0288051	.0238197	.0254628	4.839821
#2	.0254583	.0470891	12.20832	.0281195	.0237577	.0253384	4.809724
#3	.0255676	.0470890	12.23587	.0284004	.0236907	.0253666	4.814009

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0251611	12.66865	.0478411	.0100466	-.082355	.0245321	.0501608
Stddev	.0005361	.04374	.0001212	.0004238	.003989	.0017694	.0001986
%RSD	2.130691	.3452777	.2534120	4.218723	4.843279	7.212456	.3959678
#1	.0257800	12.62920	.0479292	.0099463	-.082937	.0260297	.0503138
#2	.0248380	12.66106	.0477029	.0096818	-.078108	.0249868	.0499363
#3	.0248654	12.71569	.0478913	.0105115	-.086021	.0225797	.0502324

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.027170	.0494710	.0516894	.0481216	.0498170	.0454517	.0023753
Stddev	.023706	.0003120	.0001934	.0008298	.0014647	.0036934	.0026308
%RSD	87.25090	.6305877	.3742337	1.724347	2.940181	8.125891	110.7588
#1	-.042417	.0491541	.0519127	.0490324	.0515066	.0421454	.0036628
#2	.000142	.0494810	.0515731	.0474087	.0489078	.0494377	.0041143
#3	-.039233	.0497778	.0515823	.0479235	.0490365	.0447721	-.000651

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: ICSABDLX20 Acquired: 8/13/2025 12:54:14 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSAB01DLX20 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**-.003774**	**-.007824**	**-.000041**
Stddev	.000901	.001075	.000051
%RSD	23.88389	13.74498	124.4206
#1	-.002900	-.008639	-.000082
#2	-.004701	-.008228	.000016
#3	-.003722	-.006605	-.000057

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5788.032**	**102529.6**	**11275.12**	**4431.801**	**7304.387**
Stddev	6.942	244.8	64.99	17.825	10.346
%RSD	.1199305	.2387992	.5763729	.4022091	.1416355
#1	5788.647	102437.3	11342.62	4415.268	7297.166
#2	5780.803	102807.2	11269.78	4450.685	7299.757
#3	5794.645	102344.4	11212.97	4429.449	7316.239

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCV01 Acquired: 8/13/2025 12:58:27 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**4.904034**	**4.815737**	**4.855153**	**4.926587**	**4.929568**	**9.630559**	**9.639505**
Stddev	.003827	.093365	.003356	.000953	.020754	.015530	.074813
%RSD	.0780312	1.938746	.0691173	.0193419	.4210013	.1612545	.7761128
#1	4.903444	4.836858	4.852830	4.925749	4.906877	9.642837	9.631930
#2	4.908121	4.896731	4.853628	4.926387	4.934237	9.613101	9.568767
#3	4.900537	4.713620	4.859000	4.927623	4.947589	9.635739	9.717818

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2396649**	**2.431220**	**24.08287**	**1.004169**	**2.424524**	**1.228070**	**5.033339**
Stddev	.0007196	.002959	.09699	.001192	.004031	.002902	.025154
%RSD	.3002633	.1217028	.4027220	.1187049	.1662660	.2363132	.4997529
#1	.2396428	2.429014	24.14632	1.005306	2.421873	1.227030	5.004453
#2	.2389566	2.430064	23.97123	1.004273	2.422536	1.225830	5.045148
#3	.2403954	2.434583	24.13107	1.002928	2.429163	1.231348	5.050415

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.391793**	**24.03336**	**2.428809**	**1.232624**	**24.64873**	**2.424834**	**2.485846**
Stddev	.014767	.10979	.004260	.001100	.04149	.010607	.001488
%RSD	.6173828	.4568282	.1753993	.0892242	.1683048	.4374388	.0598746
#1	2.385681	24.03945	2.425560	1.232369	24.60117	2.423347	2.487404
#2	2.381063	23.92066	2.427235	1.233829	24.67747	2.415049	2.484439
#3	2.408634	24.13998	2.433632	1.231674	24.66755	2.436106	2.485694

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**25.03939**	**4.708310**	**4.951383**	**4.865971**	**4.797024**	**5.028014**	**4.746392**
Stddev	.11178	.012115	.006688	.003206	.016798	.019018	.004821
%RSD	.4464116	.2573088	.1350742	.0658779	.3501760	.3782440	.1015688
#1	24.92825	4.705204	4.947132	4.863532	4.785107	5.007270	4.750071
#2	25.15180	4.698050	4.947924	4.864781	4.789728	5.032148	4.748170
#3	25.03813	4.721676	4.959092	4.869602	4.816236	5.044626	4.740934

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCV01 Acquired: 8/13/2025 12:58:27 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV01 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**4.763198**	**4.755884**	**4.799049**
Stddev	.007889	.015111	.023342
%RSD	.1656218	.3177372	.4863836
#1	4.763274	4.747278	4.817676
#2	4.771049	4.747042	4.772865
#3	4.755272	4.773333	4.806607

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5716.124**	**98246.45**	**11482.45**	**4313.172**	**6870.297**
Stddev	19.794	201.04	26.66	7.498	22.475
%RSD	.3462904	.2046323	.2322168	.1738287	.3271372
#1	5722.865	98226.65	11474.21	4304.542	6876.671
#2	5731.667	98056.03	11512.26	4316.889	6888.896
#3	5693.839	98456.65	11460.87	4318.085	6845.322

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCB01 Acquired: 8/13/2025 13:03:05 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000665**	**.0002982**	**-.000160**	**.0019303**	**-.000085**	**-.014664**	**-.003595**
Stddev	.000786	.0006068	.000133	.0005636	.001328	.005725	.001019
%RSD	118.1845	203.5272	82.78132	29.19613	1554.203	39.03765	28.33739
#1	-.001490	.0005790	-.000264	.0025654	-.001515	-.020811	-.004488
#2	.000074	-.000398	-.000011	.0014897	.000151	-.013698	-.002485
#3	-.000579	.000714	-.000206	.0017359	.001108	-.009484	-.003813

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000038**	**-.000159**	**-.023724**	**.0014164**	**-.000206**	**.0009549**	**-.007527**
Stddev	.000034	.000037	.000970	.0005836	.000143	.0003958	.000646
%RSD	88.68050	22.99370	4.087030	41.20752	69.44955	41.45068	8.585247
#1	-.000047	-.000191	-.022623	.0020379	-.000325	.0014098	-.006947
#2	-.000001	-.000119	-.024447	.0008800	-.000047	.0007663	-.008224
#3	-.000068	-.000168	-.024103	.0013311	-.000245	.0006887	-.007412

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000097**	**-.022407**	**-.000059**	**.0003136**	**-.097923**	**-.001545**	**.0010055**
Stddev	.000325	.006411	.000094	.0005164	.006892	.000712	.0001678
%RSD	333.1988	28.61382	159.2137	164.6458	7.038613	46.12859	16.68869
#1	-.000233	-.020516	-.000124	.0001423	-.096068	-.001737	.0010634
#2	-.000332	-.017153	.000049	-.000095	-.105553	-.002141	.0008164
#3	.000273	-.029551	-.000102	.000894	-.092147	-.000755	.0011367

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.014650**	**.0059156**	**-.000159**	**.0000338**	**-.001152**	**.0038132**	**.0010622**
Stddev	.016514	.0003154	.000202	.0004324	.000530	.0062807	.0014760
%RSD	112.7205	5.331710	127.4271	1278.706	45.99109	164.7087	138.9568
#1	.002381	.0060799	-.000132	.0002199	-.001375	-.003419	.0027407
#2	-.030592	.0061148	-.000373	-.000460	-.000547	.006957	.0004786
#3	-.015740	.0055519	.000029	.000342	-.001533	.007902	-.000033

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCB01 Acquired: 8/13/2025 13:03:05 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB01 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**-.001350**	**-.004000**	**-.000299**
Stddev	.001798	.001573	.000052
%RSD	133.2183	39.32689	17.41513
#1	-.001136	-.003459	-.000260
#2	.000332	-.002769	-.000279
#3	-.003245	-.005772	-.000358

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5851.750**	**104344.0**	**11399.46**	**4448.811**	**7637.193**
Stddev	5.292	478.6	120.53	11.758	4.517
%RSD	.0904386	.4586607	1.057325	.2642944	.0591440
#1	5853.156	104753.0	11439.63	4458.315	7638.654
#2	5856.197	103817.6	11263.97	4452.457	7640.799
#3	5845.897	104461.4	11494.76	4435.662	7632.127

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2828-01 Acquired: 8/13/2025 13:07:24 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1134400	.0091516	1.533352	.0108986	.1062454	166.4215	2.636055
Stddev	.0015758	.0040709	.001904	.0014848	.0027283	1.3196	.011931
%RSD	1.389133	44.48259	.1241626	13.62367	2.567924	.7929354	.4526193
#1	.1134908	.0050198	1.532807	.0121019	.1073272	166.8840	2.626658
#2	.1118394	.0092765	1.531779	.0113547	.1082669	164.9329	2.632029
#3	.1149898	.0131587	1.535469	.0092393	.1031421	167.4476	2.649479

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0202952	.1062010	186.2417	1.308005	.8657477	1.634701	256.9624
Stddev	.0001481	.0004095	2.7180	.004477	.0004183	.003580	4.5732
%RSD	.7299014	.3855493	1.459390	.3423139	.0483169	.2189785	1.779725
#1	.0204270	.1057696	187.4768	1.309300	.8659356	1.631343	260.6307
#2	.0203238	.1065842	188.1228	1.303023	.8660392	1.638467	251.8385
#3	.0201349	.1062493	183.1255	1.311693	.8652684	1.634294	258.4179

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	4.443344	25.86734	3.062504	.0338287	18.71360	.6835053	5.160948
Stddev	.009517	.03745	.001462	.0016321	.22414	.0004427	.011324
%RSD	.2141747	.1447709	.0477248	4.824718	1.197735	.0647699	.2194135
#1	4.437145	25.90887	3.063683	.0353896	18.54055	.6837543	5.173606
#2	4.438587	25.85698	3.060869	.0321336	18.63345	.6829942	5.157459
#3	4.454301	25.83616	3.062961	.0339629	18.96680	.6837675	5.151780

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	6.638579	.9711255	.1761253	.1868855	14.96628	4.791280	3.820196
Stddev	.038190	.0151179	.0001415	.0008033	.02037	.037473	.012040
%RSD	.5752779	1.556740	.0803617	.4298641	.1361015	.7821169	.3151649
#1	6.607898	.9861273	.1759761	.1872383	14.96144	4.786174	3.810288
#2	6.626488	.9558942	.1761422	.1859661	14.94876	4.756621	3.816704
#3	6.681351	.9713551	.1762576	.1874521	14.98863	4.831044	3.833596

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2828-01 Acquired: 8/13/2025 13:07:24 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	7.773174	.3623238	1.257674
Stddev	.036608	.0018432	.014664
%RSD	.4709588	.5087047	1.165947
#1	7.764086	.3632790	1.273092
#2	7.741966	.3601991	1.256028
#3	7.813471	.3634933	1.243903

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	10847.07	190843.4	24787.89	8059.269	5851.491
Stddev	23.51	876.4	139.91	14.056	6.464
%RSD	.2167710	.4592384	.5644095	.1744020	.1104658
#1	10874.18	191183.3	24628.60	8043.386	5858.912
#2	10832.27	191499.0	24844.22	8064.318	5847.087
#3	10834.77	189847.9	24890.85	8070.102	5848.474

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2828-01DUP Acquired: 8/13/2025 13:12:06 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1158155	.0148352	1.532858	.0134744	.1128472	165.6223	2.601885
Stddev	.0015015	.0029402	.002353	.0008166	.0007135	2.9241	.010841
%RSD	1.296437	19.81900	.1535243	6.060074	.6322709	1.765542	.4166676
#1	.1172474	.0115629	1.530339	.0142994	.1134845	162.3270	2.595097
#2	.1142530	.0172546	1.535001	.0134573	.1120763	166.6324	2.596170
#3	.1159463	.0156882	1.533233	.0126666	.1129807	167.9075	2.614388

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0203181	.1055997	186.2800	1.293369	.8525037	1.653211	255.0616
Stddev	.0001062	.0001265	1.5873	.005095	.0013593	.002973	2.7857
%RSD	.5225224	.1198118	.8521250	.3939655	.1594527	.1798595	1.092172
#1	.0202344	.1055010	185.3925	1.288126	.8524774	1.650434	258.2317
#2	.0202824	.1057423	188.1126	1.293677	.8538760	1.656348	253.0042
#3	.0204375	.1055557	185.3349	1.298303	.8511577	1.652852	253.9490

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	4.454185	25.99138	3.023303	.0339602	18.67039	.6861745	5.142791
Stddev	.020187	.05425	.003063	.0012542	.03486	.0049921	.006382
%RSD	.4532054	.2087361	.1013161	3.693023	.1867154	.7275241	.1240934
#1	4.439820	25.92941	3.019793	.0353625	18.63341	.6808485	5.135883
#2	4.445471	26.01442	3.025437	.0329458	18.70265	.6869278	5.144023
#3	4.477265	26.03032	3.024679	.0335723	18.67511	.6907471	5.148467

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	6.603250	.9747618	.1783009	.1853271	14.84422	3.869544	3.781350
Stddev	.030996	.0059859	.0001001	.0016453	.05866	.022660	.011464
%RSD	.4694029	.6140847	.0561218	.8877616	.3951431	.5856026	.3031768
#1	6.637791	.9810424	.1784010	.1870788	14.81215	3.844501	3.782269
#2	6.594100	.9691223	.1783006	.1850881	14.80859	3.888632	3.769455
#3	6.577859	.9741207	.1782009	.1838145	14.91192	3.875500	3.792328

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2828-01DUP Acquired: 8/13/2025 13:12:06 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	7.734245	.3603983	1.261066
Stddev	.040429	.0010436	.015900
%RSD	.5227262	.2895702	1.260865
#1	7.738389	.3607265	1.251076
#2	7.691904	.3592301	1.279401
#3	7.772442	.3612384	1.252720

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	10603.16	189222.5	24470.91	8067.296	5787.316
Stddev	22.72	175.1	131.64	26.251	12.262
%RSD	.2143126	.0925284	.5379531	.3254063	.2118816
#1	10617.09	189391.1	24546.45	8045.391	5799.051
#2	10576.94	189041.6	24547.38	8096.394	5774.587
#3	10615.46	189234.8	24318.91	8060.102	5788.311

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2828-01LX5 Acquired: 8/13/2025 13:16:46 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0374715	.0054799	.3129119	.0034262	.0352743	56.94094	.8923724
Stddev	.0030387	.0019713	.0007200	.0013189	.0009382	.14880	.0014934
%RSD	8.109254	35.97277	.2301101	38.49380	2.659784	.2613265	.1673533
#1	.0384185	.0032081	.3121144	.0021354	.0354689	56.76937	.8911017
#2	.0399239	.0064933	.3135144	.0047714	.0360999	57.03476	.8940173
#3	.0340721	.0067382	.3131068	.0033718	.0342540	57.01870	.8919981

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0072923	.0167536	65.96660	.4662722	.1679308	.5963318	87.31986
Stddev	.0000975	.0000931	.09158	.0050144	.0005588	.0009259	.27581
%RSD	1.337256	.5558538	.1388234	1.075420	.3327414	.1552667	.3158593
#1	.0072447	.0166691	65.86390	.4715687	.1673239	.5952874	87.59274
#2	.0074045	.0167382	66.03976	.4656498	.1680448	.5966561	87.32562
#3	.0072278	.0168534	65.99613	.4615980	.1684239	.5970519	87.04121

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.597058	9.332307	.6110464	.0122545	5.451154	.2416844	1.868615
Stddev	.003012	.015072	.0011982	.0002839	.017208	.0020022	.007942
%RSD	.1885819	.1615043	.1960946	2.316671	.3156702	.8284214	.4250164
#1	1.593591	9.334797	.6099449	.0121666	5.462255	.2423569	1.860546
#2	1.599027	9.316146	.6108720	.0125719	5.459874	.2432637	1.876424
#3	1.598556	9.345980	.6123222	.0120249	5.431331	.2394325	1.868875

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	2.008778	.3605193	.0609411	.0368072	5.260559	1.535192	.7285319
Stddev	.030265	.0011303	.0002233	.0000442	.003747	.026223	.0044502
%RSD	1.506614	.3135261	.3664465	.1200149	.0712353	1.708100	.6108499
#1	2.017505	.3592621	.0611443	.0367570	5.260691	1.561722	.7238226
#2	1.975108	.3614515	.0609771	.0368244	5.264239	1.534565	.7291056
#3	2.033720	.3608444	.0607020	.0368402	5.256747	1.509288	.7326675

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2828-01LX5 Acquired: 8/13/2025 13:16:46 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	1.417286	.1153787	.4399982
Stddev	.005971	.0004917	.0016475
%RSD	.4212716	.4261647	.3744271
#1	1.424138	.1157406	.4381381
#2	1.413197	.1155766	.4412736
#3	1.414524	.1148189	.4405829

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6797.836	120467.2	14251.06	5159.481	6763.759
Stddev	4.545	191.4	72.55	18.358	6.699
%RSD	.0668549	.1589031	.5090718	.3558027	.0990436
#1	6800.417	120479.8	14330.11	5172.609	6771.340
#2	6792.589	120269.8	14187.52	5138.503	6761.305
#3	6800.503	120652.0	14235.56	5167.330	6758.634

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2828-01MS Acquired: 8/13/2025 13:20:51 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.4351764	1.631806	2.687091	.7305922	.2975854	226.5995	2.854597
Stddev	.0027281	.050930	.004847	.0045100	.0030584	2.1090	.021283
%RSD	.6268927	3.121092	.1803775	.6173082	1.027749	.9307268	.7455781
#1	.4355422	1.578458	2.686194	.7279700	.2973353	224.1644	2.830455
#2	.4322839	1.637048	2.682755	.7280067	.2946597	227.7955	2.870649
#3	.4377032	1.679913	2.692323	.7357998	.3007612	227.8387	2.862687

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0971046	.3053206	253.8815	1.519295	1.149584	1.963534	249.5014
Stddev	.0007481	.0006048	2.5695	.010415	.002146	.000298	2.4786
%RSD	.7704324	.1980948	1.012103	.6855172	.1866773	.0151547	.9934152
#1	.0962449	.3052109	250.9508	1.507276	1.148241	1.963518	246.8062
#2	.0976077	.3047781	255.7482	1.525667	1.148453	1.963244	251.6829
#3	.0974611	.3059728	254.9454	1.524943	1.152059	1.963839	250.0152

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	5.333563	35.79715	3.824304	.0783584	24.96618	1.366538	5.945174
Stddev	.014569	.18385	.005270	.0007498	.00881	.006098	.024668
%RSD	.2731514	.5135933	.1378003	.9568491	.0352953	.4462156	.4149186
#1	5.317401	35.61445	3.825148	.0774946	24.95866	1.359577	5.966449
#2	5.345686	35.98213	3.818663	.0787399	24.97587	1.370936	5.918134
#3	5.337603	35.79488	3.829101	.0788406	24.96401	1.369101	5.950938

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	13.41596	1.260318	.3335284	.8504869	15.32760	5.942173	9.990505
Stddev	.01873	.012414	.0010168	.0028623	.04145	.028959	.025481
%RSD	.1395949	.9849807	.3048548	.3365440	.2703969	.4873392	.2550539
#1	13.39991	1.246265	.3325826	.8484021	15.27986	5.910251	10.00195
#2	13.41145	1.269790	.3333989	.8493084	15.34852	5.949513	9.96131
#3	13.43654	1.264899	.3346037	.8537503	15.35441	5.966756	10.00826

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2828-01MS Acquired: 8/13/2025 13:20:51 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	9.405993	.4839844	1.580557
Stddev	.052539	.0012301	.017248
%RSD	.5585735	.2541656	1.091286
#1	9.353141	.4831675	1.576974
#2	9.406625	.4833866	1.565381
#3	9.458214	.4853992	1.599315

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	11150.53	196965.1	26052.92	8416.630	5545.990
Stddev	18.99	1494.4	236.39	16.122	5.603
%RSD	.1703315	.7587226	.9073641	.1915531	.1010329
#1	11166.06	198514.7	26310.19	8428.839	5552.447
#2	11129.35	195532.8	25845.28	8398.355	5542.405
#3	11156.16	196847.8	26003.31	8422.697	5543.117

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2828-01MSD Acquired: 8/13/2025 13:25:37 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.4615840	1.715748	2.650532	.7844096	.2400832	241.6083	2.750904
Stddev	.0014728	.030870	.001999	.0023306	.0007801	2.2615	.042241
%RSD	.3190697	1.799219	.0754252	.2971139	.3249339	.9360124	1.535533
#1	.4603177	1.687476	2.648324	.7818328	.2407470	244.0371	2.751524
#2	.4612341	1.711081	2.651052	.7863701	.2402786	239.5632	2.792832
#3	.4632002	1.748685	2.652220	.7850259	.2392239	241.2247	2.708357

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1061506	.3148512	282.9101	1.746571	1.203677	1.923987	288.2252
Stddev	.0002332	.0006359	1.8151	.007134	.001830	.002565	3.2417
%RSD	.2196529	.2019656	.6415928	.4084683	.1520090	.1333316	1.124696
#1	.1063684	.3141310	284.3286	1.748688	1.201599	1.921458	291.5038
#2	.1059046	.3153350	283.5371	1.738618	1.204387	1.926587	285.0218
#3	.1061787	.3150876	280.8646	1.752407	1.205046	1.923916	288.1502

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	5.736477	39.05291	4.079435	.0788640	24.11455	.9461061	5.724962
Stddev	.029504	.08312	.005347	.0010605	.10649	.0024564	.021914
%RSD	.5143223	.2128394	.1310627	1.344657	.4416175	.2596280	.3827767
#1	5.757364	39.14467	4.073860	.0799670	24.15044	.9489230	5.730063
#2	5.749342	38.98265	4.084519	.0778519	24.19847	.9449851	5.700947
#3	5.702725	39.03143	4.079925	.0787732	23.99475	.9444103	5.743875

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	13.30633	1.279237	.3421945	.8797464	14.67077	1.143119	9.957083
Stddev	.03329	.008530	.0006492	.0033389	.04647	.017539	.045158
%RSD	.2502135	.6668188	.1897172	.3795345	.3167335	1.534312	.4535276
#1	13.32852	1.287682	.3420382	.8771325	14.68701	1.162189	10.00213
#2	13.32244	1.270624	.3416377	.8785990	14.70694	1.139488	9.95730
#3	13.26805	1.279405	.3429076	.8835078	14.61836	1.127679	9.91182

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2828-01MSD Acquired: 8/13/2025 13:25:37 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	9.579984	.4892810	1.560177
Stddev	.022632	.0022723	.017256
%RSD	.2362426	.4644240	1.106057
#1	9.568223	.4909751	1.564974
#2	9.565654	.4901691	1.574528
#3	9.606075	.4866987	1.541030

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	10569.60	185523.5	24488.62	8022.232	5387.879
Stddev	8.81	562.7	26.55	28.638	5.614
%RSD	.0833058	.3033084	.1084259	.3569804	.1042035
#1	10575.21	185376.6	24463.69	8033.160	5388.470
#2	10574.14	186145.1	24516.54	8043.797	5393.175
#3	10559.46	185048.8	24485.63	7989.739	5381.993

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2828-01A Acquired: 8/13/2025 13:30:23 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.4985563	1.727627	2.394146	.9585126	.5136132	174.0700	2.651098
Stddev	.0127946	.072811	.112491	.0253087	.0159573	.9918	.008567
%RSD	2.566324	4.214509	4.698568	2.640417	3.106871	.5697729	.3231585
#1	.5020827	1.759048	2.438899	.9679940	.5192490	175.2100	2.659650
#2	.5092178	1.779450	2.477372	.9777115	.5259879	173.4052	2.642516
#3	.4843683	1.644381	2.266166	.9298321	.4956028	173.5948	2.651128

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1078394	.2920576	192.9332	1.479441	1.018800	1.764613	257.2287
Stddev	.0002290	.0151426	1.1552	.004351	.051056	.045010	1.6002
%RSD	.2123098	5.184791	.5987651	.2940722	5.011344	2.550723	.6220730
#1	.1076176	.2988290	191.6937	1.475171	1.041834	1.784497	257.5228
#2	.1078256	.3026328	193.9798	1.479284	1.054280	1.796258	255.5018
#3	.1080749	.2747109	193.1260	1.483868	.960285	1.713085	258.6613

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	4.586002	27.82926	3.425555	.0742261	20.53416	.8293754	5.280278
Stddev	.011470	.06904	.164230	.0007513	.06541	.0017896	.006902
%RSD	.2501002	.2480940	4.794263	1.012168	.3185292	.2157720	.1307175
#1	4.592972	27.89665	3.498943	.0742723	20.58898	.8306126	5.284897
#2	4.572764	27.75868	3.540296	.0734528	20.46176	.8273234	5.283594
#3	4.592270	27.83246	3.237428	.0749533	20.55175	.8301902	5.272344

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	12.34843	1.100116	.3691177	.8588195	14.91686	3.738774	9.330480
Stddev	.02891	.006739	.0088644	.0359762	.03014	.020807	.451604
%RSD	.2340916	.6125717	2.401496	4.189034	.2020420	.5565135	4.840099
#1	12.34718	1.099619	.3739229	.8760554	14.93945	3.717871	9.532941
#2	12.32017	1.093640	.3745419	.8829352	14.88264	3.738969	9.645426
#3	12.37794	1.107091	.3588883	.8174681	14.92848	3.759484	8.813072

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2828-01A Acquired: 8/13/2025 13:30:23 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**7.567011**	**.4534073**	**1.365456**
Stddev	.382646	.0018455	.006631
%RSD	5.056764	.4070324	.4855914
#1	7.729611	.4554836	1.370162
#2	7.841503	.4519537	1.357873
#3	7.129917	.4527845	1.368333

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**10592.85**	**185875.3**	**24225.45**	**7907.430**	**5852.049**
Stddev	136.59	303.8	18.83	33.349	226.891
%RSD	1.289441	.1634175	.0777375	.4217425	3.877115
#1	10528.16	186222.7	24242.12	7875.736	5746.291
#2	10500.63	185659.7	24205.02	7904.335	5697.344
#3	10749.77	185743.6	24229.21	7942.219	6112.511

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCV02 Acquired: 8/13/2025 13:42:20 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV02 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**4.949760**	**4.959503**	**4.852074**	**4.982082**	**4.970518**	**9.695432**	**9.662181**
Stddev	.008222	.043373	.003983	.004322	.002979	.013974	.133465
%RSD	.1661075	.8745405	.0820986	.0867538	.0599281	.1441246	1.381311
#1	4.945489	4.927510	4.855461	4.981265	4.973161	9.707947	9.655441
#2	4.959239	4.942129	4.847685	4.978226	4.971104	9.680354	9.532215
#3	4.944553	5.008869	4.853075	4.986754	4.967290	9.697994	9.798889

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2404294**	**2.432348**	**24.24217**	**.9874279**	**2.426362**	**1.233085**	**4.979648**
Stddev	.0010260	.003406	.06117	.0023567	.002690	.001601	.035100
%RSD	.4267275	.1400448	.2523355	.2386671	.1108509	.1298123	.7048683
#1	.2401409	2.435142	24.26725	.9899564	2.428882	1.231449	5.014956
#2	.2395785	2.428553	24.17244	.9852925	2.423530	1.233160	4.979228
#3	.2415687	2.433349	24.28681	.9870347	2.426673	1.234647	4.944760

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.415383**	**24.10653**	**2.434774**	**1.235911**	**24.10775**	**2.445297**	**2.459773**
Stddev	.009072	.14085	.003600	.000164	.17284	.004068	.006594
%RSD	.3755825	.5842922	.1478514	.0132602	.7169484	.1663490	.2680792
#1	2.421927	24.10849	2.438770	1.235770	24.27911	2.449432	2.452422
#2	2.405027	23.96471	2.431785	1.236091	24.11068	2.441300	2.461730
#3	2.419194	24.24639	2.433767	1.235874	23.93347	2.445160	2.465167

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**24.58278**	**4.818771**	**4.982915**	**4.884147**	**4.850847**	**5.108879**	**4.740257**
Stddev	.14503	.020018	.005117	.009146	.017529	.024578	.019218
%RSD	.5899829	.4154131	.1026955	.1872541	.3613697	.4810744	.4054274
#1	24.66268	4.807965	4.979921	4.893465	4.869623	5.131006	4.761259
#2	24.67030	4.806478	4.979999	4.883790	4.834910	5.082425	4.735965
#3	24.41537	4.841870	4.988823	4.875184	4.848008	5.113205	4.723548

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCV02 Acquired: 8/13/2025 13:42:20 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV02 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**4.789228**	**4.691877**	**4.874356**
Stddev	.008200	.014598	.024380
%RSD	.1712100	.3111262	.5001740
#1	4.790081	4.708733	4.850932
#2	4.796967	4.683532	4.899591
#3	4.780635	4.683367	4.872544

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5628.741**	**99513.57**	**11398.15**	**4267.446**	**6777.463**
Stddev	7.659	284.50	60.04	14.171	10.063
%RSD	.1360747	.2858916	.5267156	.3320646	.1484718
#1	5629.214	99203.61	11377.40	4278.198	6772.061
#2	5636.153	99762.81	11465.81	4272.752	6789.073
#3	5620.856	99574.29	11351.24	4251.388	6771.255

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCB02 Acquired: 8/13/2025 13:46:31 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB02 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0007395	-.000110	-.000212	.0008023	-.001855	-.017428	-.005219
Stddev	.0009079	.001248	.001121	.0027119	.001276	.002048	.001690
%RSD	122.7650	1135.386	529.6019	338.0241	68.78618	11.75390	32.37903
#1	.0002206	-.000462	.000716	-.001961	-.003304	-.018412	-.003322
#2	.0017879	.001277	.000106	.000907	-.000898	-.015073	-.006563
#3	.0002101	-.001144	-.001457	.003460	-.001364	-.018798	-.005773

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.000071	-.000106	-.014381	.0001949	-.000093	-.000229	-.007042
Stddev	.000017	.000042	.008184	.0003369	.000093	.000264	.000950
%RSD	23.97527	39.50109	56.90658	172.9026	99.95443	115.5440	13.49072
#1	-.000055	-.000104	-.022739	.0005834	-.000196	-.000451	-.008084
#2	-.000070	-.000148	-.014022	-.000016	-.000014	.000064	-.006816
#3	-.000089	-.000065	-.006383	.000018	-.000070	-.000299	-.006225

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.000148	-.012578	-.000139	-.000051	-.007774	-.000516	-.000457
Stddev	.000238	.004012	.000274	.000282	.008801	.000422	.000184
%RSD	160.4225	31.89640	197.0884	549.7045	113.2215	81.75090	40.21013
#1	-.000381	-.010357	-.000275	.000272	-.005774	-.000908	-.000394
#2	-.000158	-.017209	.000177	-.000245	-.017403	-.000070	-.000664
#3	.000094	-.010168	-.000319	-.000180	-.000144	-.000570	-.000313

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1141832	.0063237	.0000460	.0003722	-.000269	.0103934	.0006505
Stddev	.0261732	.0003078	.0001857	.0001704	.001067	.0175891	.0015128
%RSD	22.92217	4.867042	403.9285	45.78533	396.2427	169.2337	232.5626
#1	.0840241	.0066199	.0000858	.0002075	-.001092	.0031930	.0022614
#2	.1275707	.0063457	.0002085	.0003612	-.000652	-.002453	-.000740
#3	.1309546	.0060055	-.000156	.0005478	.000936	.030440	.000430

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCB02 Acquired: 8/13/2025 13:46:31 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB02 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**-.000826**	**-.002222**	**-.000385**
Stddev	.001082	.000297	.000013
%RSD	130.9256	13.36158	3.308400
#1	.000422	-.002151	-.000370
#2	-.001416	-.001967	-.000393
#3	-.001484	-.002548	-.000392

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5806.516**	**104236.5**	**11409.06**	**4447.255**	**7564.010**
Stddev	9.174	248.6	90.83	17.722	3.764
%RSD	.1580008	.2385264	.7961604	.3984855	.0497651
#1	5796.239	103984.8	11416.82	4427.024	7564.422
#2	5809.427	104242.7	11495.76	4454.709	7567.552
#3	5813.882	104482.0	11314.59	4460.032	7560.057

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2830-01 Acquired: 8/13/2025 13:50:51 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1453111	-.056883	.3865274	.0131982	-.022559	152.2161	1.736366
Stddev	.0009073	.001454	.0008237	.0036534	.002791	.1455	.004921
%RSD	.6244071	2.556802	.2131029	27.68147	12.37309	.0956025	.2834265
#1	.1454709	-.055205	.3867949	.0146423	-.019433	152.2886	1.740244
#2	.1461279	-.057657	.3871841	.0159088	-.024800	152.0486	1.730830
#3	.1443344	-.057787	.3856032	.0090434	-.023445	152.3112	1.738026

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0173701	.0147549	24.50382	.2734636	.1620453	.1345569	440.7952
Stddev	.0000865	.0008025	.05598	.0015646	.0002981	.0032392	7.2620
%RSD	.4981133	5.438563	.2284726	.5721374	.1839742	2.407349	1.647467
#1	.0173360	.0138483	24.56844	.2752627	.1618765	.1308169	448.8563
#2	.0174684	.0150422	24.46997	.2727076	.1623895	.1363828	438.7644
#3	.0173057	.0153741	24.47305	.2724207	.1618698	.1364711	434.7647

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	7.053170	43.39191	.2419968	.0125469	1.615406	.5301259	.7093288
Stddev	.008929	.11217	.0005779	.0023185	.023901	.0024664	.0028820
%RSD	.1265909	.2584947	.2387986	18.47885	1.479546	.4652570	.4063013
#1	7.063184	43.45313	.2413544	.0150755	1.641766	.5327165	.7093729
#2	7.046041	43.46015	.2424743	.0120445	1.609304	.5278059	.7121884
#3	7.050282	43.26246	.2421617	.0105208	1.595148	.5298554	.7064249

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	11.93763	.2061289	.0039361	.0175232	2.264388	6.567941	9.327685
Stddev	.06222	.0197373	.0001829	.0016550	.003505	.048097	.006302
%RSD	.5212011	9.575225	4.645983	9.444708	.1547686	.7323005	.0675626
#1	12.00235	.2284110	.0038572	.0161573	2.266175	6.609928	9.323849
#2	11.87826	.1991339	.0038059	.0193637	2.260350	6.578430	9.324248
#3	11.93228	.1908417	.0041452	.0170487	2.266640	6.515465	9.334958

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2830-01 Acquired: 8/13/2025 13:50:51 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**2.041949**	**.6125078**	**-.279202**
Stddev	.016170	.0010017	.006923
%RSD	.7919129	.1635426	2.479715
#1	2.026327	.6136305	-.286871
#2	2.040902	.6117053	-.277326
#3	2.058617	.6121878	-.273411

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**6119.552**	**107055.6**	**13022.57**	**4596.786**	**6449.588**
Stddev	7.994	543.1	11.22	22.320	6.216
%RSD	.1306277	.5073419	.0861711	.4855565	.0963758
#1	6110.346	106533.9	13031.54	4571.547	6453.282
#2	6124.736	107014.8	13009.99	4604.886	6442.412
#3	6123.575	107617.9	13026.18	4613.925	6453.071

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2830-03 Acquired: 8/13/2025 13:55:01 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1069329	-.058267	.3408371	.0184989	-.023934	285.4843	1.256137
Stddev	.0054180	.001836	.0010438	.0029298	.001281	1.7627	.009323
%RSD	5.066772	3.151502	.3062355	15.83794	5.350297	.6174488	.7421650
#1	.1059495	-.060360	.3403176	.0152174	-.025344	287.4928	1.256099
#2	.1020739	-.056925	.3420387	.0194269	-.022843	284.1942	1.265479
#3	.1127753	-.057517	.3401551	.0208524	-.023616	284.7659	1.246834

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0200945	.0183256	13.93393	.4177108	.2867184	.1966260	546.8259
Stddev	.0000651	.0003885	.03499	.0007056	.0009352	.0044343	4.2256
%RSD	.3237473	2.119844	.2511435	.1689157	.3261755	2.255180	.7727529
#1	.0200965	.0184833	13.90667	.4177481	.2865534	.1995388	544.2780
#2	.0201585	.0186104	13.97339	.4183969	.2877250	.1988165	544.4962
#3	.0200284	.0178830	13.92172	.4169873	.2858766	.1915228	551.7036

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	7.443343	93.59478	.5106256	.0164036	2.137585	.5050311	1.294008
Stddev	.039951	.38588	.0018568	.0021047	.020426	.0025238	.003464
%RSD	.5367407	.4122851	.3636318	12.83043	.9555561	.4997373	.2676837
#1	7.436096	93.47368	.5093758	.0148412	2.142030	.5021470	1.292839
#2	7.486422	94.02668	.5127593	.0155727	2.155422	.5068353	1.297905
#3	7.407512	93.28397	.5097418	.0187968	2.115302	.5061110	1.291280

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	31.98573	^ *****	.0084098	.0201118	4.861828	6.028986	3.927140
Stddev	.02819	-----	.0005355	.0002997	.016239	.030391	.010970
%RSD	.0881428	-----	6.367142	1.490399	.3340137	.5040852	.2793332
#1	31.98346	^ -----	.0090075	.0202926	4.852424	6.037934	3.927290
#2	32.01500	^ -----	.0079740	.0197658	4.880580	5.995126	3.938034
#3	31.95875	^ -----	.0082479	.0202769	4.852481	6.053899	3.916096

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2830-03 Acquired: 8/13/2025 13:55:01 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	1.148863	.9060076	-.400646
Stddev	.010284	.0038149	.004533
%RSD	.8951699	.4210639	1.131329
#1	1.141996	.9036094	-.398279
#2	1.160687	.9104067	-.397786
#3	1.143906	.9040068	-.405872

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5872.360	103376.9	12822.03	4485.751	6090.982
Stddev	16.435	298.3	80.93	24.339	17.017
%RSD	.2798652	.2885852	.6311699	.5425853	.2793761
#1	5856.661	103071.6	12809.47	4492.041	6081.982
#2	5870.977	103391.2	12748.11	4458.884	6080.356
#3	5889.443	103667.8	12908.50	4506.327	6110.609

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2830-05 Acquired: 8/13/2025 13:59:20 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0981097	-.049740	.3946600	.0048797	-.022567	210.1859	2.187957
Stddev	.0010152	.003686	.0005067	.0022408	.000572	2.9339	.004397
%RSD	1.034745	7.409790	.1283799	45.92187	2.533900	1.395850	.2009432
#1	.0988646	-.045697	.3949134	.0052919	-.022903	213.3367	2.185531
#2	.0969556	-.050611	.3940766	.0068858	-.021906	207.5326	2.193032
#3	.0985090	-.052913	.3949900	.0024613	-.022891	209.6885	2.185309

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0251375	.0285491	20.79386	.3784151	.3824168	.1868235	469.9531
Stddev	.0002079	.0002725	.02010	.0005088	.0003619	.0012097	1.2231
%RSD	.8268877	.9545306	.0966698	.1344550	.0946473	.6475056	.2602682
#1	.0253632	.0287931	20.80846	.3788631	.3820680	.1859195	468.5525
#2	.0249540	.0282550	20.80218	.3785203	.3823918	.1863532	470.8111
#3	.0250954	.0285993	20.77093	.3778619	.3827906	.1881977	470.4956

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	12.99003	91.00166	.6855060	.0110437	2.210052	.4841577	1.257853
Stddev	.01316	.10038	.0021833	.0003044	.019969	.0007250	.002923
%RSD	.1013234	.1103038	.3184924	2.756213	.9035583	.1497453	.2323714
#1	13.00512	91.06672	.6833150	.0106936	2.188100	.4844402	1.254484
#2	12.98403	90.88605	.6855213	.0111919	2.227138	.4846989	1.259366
#3	12.98094	91.05220	.6876815	.0112456	2.214919	.4833340	1.259709

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	32.36584	^ *****	.0106158	.0188448	5.040346	6.474977	6.195487
Stddev	.15145	-----	.0000928	.0016341	.002647	.029825	.013510
%RSD	.4679389	-----	.8743992	8.671432	.0525142	.4606186	.2180576
#1	32.19118	^ -----	.0105786	.0169585	5.043362	6.445573	6.182473
#2	32.46086	^ -----	.0105473	.0198295	5.038409	6.505206	6.209443
#3	32.44548	^ -----	.0107215	.0197465	5.039267	6.474154	6.194545

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2830-05 Acquired: 8/13/2025 13:59:20 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**2.052204**	**.9037890**	**-.277054**
Stddev	.008756	.0020129	.001073
%RSD	.4266834	.2227147	.3874281
#1	2.062032	.9015527	-.275819
#2	2.045234	.9054556	-.277761
#3	2.049345	.9043587	-.277582

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**6912.601**	**123022.2**	**15099.17**	**5259.239**	**6026.339**
Stddev	4.661	262.8	57.80	16.562	4.823
%RSD	.0674281	.2135984	.3828035	.3149163	.0800352
#1	6907.804	123213.6	15054.11	5277.778	6026.448
#2	6917.113	123130.5	15164.34	5254.035	6031.106
#3	6912.885	122722.6	15079.06	5245.903	6021.462

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2830-07 Acquired: 8/13/2025 14:03:39 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2164020**	**-.086202**	**.3725533**	**.0126132**	**-.034533**	**232.5435**	**2.703325**
Stddev	.0020258	.002440	.0048387	.0019213	.003323	2.7196	.011360
%RSD	.9361312	2.830043	1.298802	15.23228	9.623952	1.169486	.4202122
#1	.2150516	-.084667	.3724754	.0146163	-.038242	230.4673	2.707604
#2	.2154232	-.089015	.3677541	.0107858	-.033529	231.5412	2.711924
#3	.2187314	-.084925	.3774306	.0124376	-.031827	235.6220	2.690447

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0318307**	**.0331737**	**26.50333**	**.4026310**	**.2514180**	**.1070152**	**635.3351**
Stddev	.0001591	.0004114	.02141	.0005523	.0004308	.0017124	2.7732
%RSD	.4999650	1.240123	.0807693	.1371638	.1713455	1.600164	.4364864
#1	.0318912	.0336477	26.52071	.4020195	.2515941	.1073989	632.2128
#2	.0316502	.0329649	26.47941	.4027800	.2509270	.1051435	637.5117
#3	.0319507	.0329086	26.50986	.4030935	.2517328	.1085033	636.2809

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.406815**	**69.61454**	**.4438670**	**.0160380**	**2.111442**	**.7571068**	**.9436670**
Stddev	.027553	.13335	.0004000	.0011218	.009959	.0021030	.0014319
%RSD	.2929047	.1915481	.0901211	6.994476	.4716822	.2777619	.1517369
#1	9.420118	69.54421	.4440964	.0147561	2.122432	.7581159	.9449674
#2	9.425192	69.53109	.4434051	.0168399	2.103015	.7546894	.9439012
#3	9.375135	69.76833	.4440995	.0165181	2.108878	.7585150	.9421325

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**20.91657**	**^ *****	**.0071506**	**.0224507**	**3.350843**	**7.718096**	**16.67951**
Stddev	.05597	-----	.0001171	.0008657	.006427	.022225	.05170
%RSD	.2675800	-----	1.637720	3.856207	.1917944	.2879593	.3099444
#1	20.95895	^ -----	.0071322	.0216211	3.355016	7.694176	16.66228
#2	20.93763	^ -----	.0070437	.0223826	3.354071	7.738108	16.63863
#3	20.85312	^ -----	.0072758	.0233485	3.343442	7.722005	16.73762

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2830-07 Acquired: 8/13/2025 14:03:39 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	1.752929	.8552609	-.447667
Stddev	.024383	.0016938	.002827
%RSD	1.391001	.1980485	.6314977
#1	1.725597	.8565158	-.444412
#2	1.760741	.8559327	-.449082
#3	1.772449	.8533342	-.449507

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6521.479	116295.9	14513.92	4946.778	5954.749
Stddev	13.480	81.2	47.96	18.246	7.674
%RSD	.2066945	.0698483	.3304376	.3688440	.1288658
#1	6505.990	116366.6	14528.89	4965.887	5946.839
#2	6527.894	116207.1	14552.61	4929.539	5962.162
#3	6530.553	116313.8	14460.26	4944.908	5955.247

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2831-01 Acquired: 8/13/2025 14:07:58 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0842644	-.028549	.7568970	.0092254	-.001733	129.3234	.9619753
Stddev	.0018064	.003787	.0025219	.0046695	.001596	.2170	.0032684
%RSD	2.143742	13.26499	.3331951	50.61560	92.09078	.1677751	.3397588
#1	.0857302	-.029119	.7546658	.0103259	-.003434	129.3041	.9582198
#2	.0822463	-.024510	.7596333	.0041040	-.001495	129.1166	.9635293
#3	.0848167	-.032019	.7563920	.0132464	-.000269	129.5493	.9641767

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0110824	.0130389	29.22138	.2229414	.1304059	.4015786	287.3479
Stddev	.0000183	.0001943	.03085	.0004774	.0007261	.0005560	.9555
%RSD	.1649328	1.490445	.1055895	.2141515	.5568365	.1384538	.3325383
#1	.0110677	.0128947	29.24894	.2230085	.1298015	.4016671	286.4096
#2	.0110767	.0129622	29.18805	.2224339	.1302048	.4009837	288.3198
#3	.0111029	.0132599	29.22716	.2233817	.1312114	.4020851	287.3142

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	7.318509	44.53995	.2586293	.0106994	2.636514	.3195009	1.659430
Stddev	.015617	.11522	.0008225	.0000502	.014832	.0014166	.006740
%RSD	.2133840	.2586944	.3180095	.4688305	.5625434	.4433748	.4061625
#1	7.322214	44.64692	.2576882	.0107320	2.620050	.3178988	1.665255
#2	7.301373	44.41795	.2592102	.0106417	2.648828	.3200162	1.652048
#3	7.331940	44.55498	.2589895	.0107246	2.640665	.3205878	1.660988

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	12.90753	.1555091	.0052800	.0739703	3.042918	11.72958	7.082379
Stddev	.03364	.0093015	.0002461	.0008128	.002422	.04835	.013683
%RSD	.2606456	5.981301	4.661024	1.098832	.0795903	.4122122	.1931991
#1	12.87202	.1485422	.0050530	.0748184	3.044507	11.67378	7.087302
#2	12.93892	.1660717	.0055416	.0731981	3.040131	11.75898	7.066916
#3	12.91166	.1519134	.0052453	.0738942	3.044117	11.75599	7.092920

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2831-01 Acquired: 8/13/2025 14:07:58 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	7.757684	.2486897	-.132920
Stddev	.044599	.0011034	.001025
%RSD	.5748951	.4436814	.7711807
#1	7.706414	.2483096	-.131963
#2	7.787511	.2478265	-.134002
#3	7.779126	.2499329	-.132795

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6320.222	111627.8	13570.13	4762.696	6503.062
Stddev	13.159	29.7	59.74	6.222	11.252
%RSD	.2082032	.0266202	.4402398	.1306349	.1730257
#1	6327.456	111641.2	13526.41	4769.809	6511.421
#2	6328.177	111593.7	13638.20	4760.011	6507.496
#3	6305.033	111648.5	13545.78	4758.267	6490.269

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2818-01 Acquired: 8/13/2025 14:11:59 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.2462969	-.069644	.3354100	.0130213	-.018397	112.4632
Stddev	.0012952	.003752	.0013166	.0048755	.000980	.2340
%RSD	.5258676	5.387055	.3925382	37.44268	5.327224	.2080750
#1	.2449346	-.067971	.3359474	.0186501	-.017518	112.5226
#2	.2464436	-.067019	.3339097	.0102981	-.018220	112.6618
#3	.2475125	-.073941	.3363729	.0101158	-.019454	112.2053

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.8276041	.0201980	.0267863	35.14157	1.105311	.0887568
Stddev	.0039778	.0000887	.0003532	.10878	.004022	.0002024
%RSD	.4806467	.4393552	1.318600	.3095544	.3639050	.2279821
#1	.8265402	.0201625	.0271701	35.13530	1.103455	.0889904
#2	.8320057	.0202990	.0264749	35.25336	1.109926	.0886411
#3	.8242664	.0201326	.0267140	35.03606	1.102551	.0886389

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.5169805	540.4245	2.067864	12.72622	.1702107	.0153687
Stddev	.0014479	1.1601	.010081	.07795	.0007809	.0013469
%RSD	.2800703	.2146682	.4875018	.6125339	.4587622	8.763910
#1	.5172343	539.9670	2.066710	12.76450	.1693470	.0158187
#2	.5154225	539.5629	2.078472	12.77763	.1708666	.0138544
#3	.5182848	541.7437	2.058410	12.63653	.1704185	.0164329

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	2.627463	1.127048	1.580430	23.04248	^ *****	.0151367
Stddev	.013806	.004197	.003761	.14557	-----	.0003061
%RSD	.5254600	.3724033	.2379777	.6317488	-----	2.022111
#1	2.615976	1.129014	1.583086	22.95315	^ -----	.0154716
#2	2.642780	1.129901	1.582077	23.21046	^ -----	.0148713
#3	2.623633	1.122229	1.576126	22.96385	^ -----	.0150672

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2818-01 Acquired: 8/13/2025 14:11:59 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0584856	2.938632	3.014497	F 46.56130	16.56324	.0715998
Stddev	.0018447	.010070	.024435	.18311	.10076	.0004075
%RSD	3.154063	.3426813	.8105747	.3932588	.6083331	.5691201
#1	.0566042	2.934504	3.008757	46.46829	16.53407	.0711311
#2	.0585614	2.950111	3.041290	46.44336	16.48029	.0718701
#3	.0602912	2.931282	2.993443	46.77224	16.67537	.0717983

Elem	Sr4077
Units	ppm
Avg	-.203737
Stddev	.002307
%RSD	1.132247
#1	-.203987
#2	-.201316
#3	-.205909

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6505.282	116815.1	14586.43	4966.785	6412.573
Stddev	6.301	286.7	28.92	21.413	2.696
%RSD	.0968644	.2454425	.1982957	.4311184	.0420488
#1	6503.059	116812.9	14567.60	4942.072	6409.605
#2	6512.393	116529.4	14571.96	4979.821	6414.872
#3	6500.393	117102.8	14619.74	4978.461	6413.243

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2818-02 Acquired: 8/13/2025 14:16:10 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.2545211	-.069890	.3866487	.0127639	-.017780	108.6222
Stddev	.0016831	.003461	.0014114	.0036067	.001657	.2133
%RSD	.6612892	4.952637	.3650320	28.25681	9.320586	.1963731
#1	.2553146	-.073679	.3851463	.0140571	-.019657	108.4475
#2	.2556607	-.066894	.3868531	.0086890	-.016516	108.8599
#3	.2525879	-.069096	.3879468	.0155457	-.017168	108.5592

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.7594194	.0201688	.0271667	32.80926	1.098975	.0855620
Stddev	.0033874	.0001196	.0006660	.08813	.001460	.0000602
%RSD	.4460551	.5927880	2.451716	.2686042	.1328341	.0703293
#1	.7557895	.0200884	.0264044	32.71273	1.100395	.0855333
#2	.7624963	.0203062	.0276359	32.88541	1.099053	.0856311
#3	.7599725	.0201118	.0274599	32.82965	1.097478	.0855215

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.4744454	534.4423	1.931949	11.77155	.1668473	.0150368
Stddev	.0022762	3.1452	.009811	.02410	.0001242	.0013096
%RSD	.4797657	.5885101	.5078149	.2047647	.0744245	8.708974
#1	.4722957	536.7234	1.921135	11.74549	.1669135	.0153019
#2	.4768300	530.8543	1.940279	11.77610	.1669242	.0136151
#3	.4742105	535.7491	1.934433	11.79305	.1667040	.0161936

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	2.452424	1.133393	1.477880	21.72002	^ *****	.0141732
Stddev	.004034	.005879	.004034	.03342	-----	.0001423
%RSD	.1645005	.5186950	.2729910	.1538895	-----	1.004391
#1	2.456539	1.129117	1.481070	21.68251	^ -----	.0140227
#2	2.448476	1.140097	1.479226	21.74666	^ -----	.0141912
#3	2.452256	1.130965	1.473345	21.73087	^ -----	.0143057

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2818-02 Acquired: 8/13/2025 14:16:10 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0578646	2.805185	3.128855	F 44.96693	14.47537	.0694440
Stddev	.0006240	.009307	.023205	.06087	.10163	.0007622
%RSD	1.078428	.3317661	.7416472	.1353728	.7020610	1.097547
#1	.0572046	2.797529	3.155627	44.90487	14.36199	.0702671
#2	.0579440	2.815544	3.116429	45.02654	14.50585	.0687626
#3	.0584451	2.802483	3.114510	44.96938	14.55826	.0693022

Elem	Sr4077
Units	ppm
Avg	-.224434
Stddev	.004153
%RSD	1.850307
#1	-.227936
#2	-.219846
#3	-.225521

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6436.522	116239.7	14489.29	4933.282	6323.225
Stddev	9.761	77.4	57.37	12.272	10.112
%RSD	.1516498	.0666187	.3959303	.2487548	.1599176
#1	6443.927	116314.2	14530.81	4919.397	6330.209
#2	6440.177	116245.2	14423.83	4937.773	6327.836
#3	6425.461	116159.7	14513.23	4942.676	6311.629

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2823-01 Acquired: 8/13/2025 14:20:20 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0595789	-.019709	.6776581	.0015832	-.010210	79.00177	.7523120
Stddev	.0034710	.001638	.0010042	.0019171	.000188	.08725	.0019272
%RSD	5.825966	8.312127	.1481942	121.0878	1.845286	.1104444	.2561695
#1	.0632265	-.018900	.6786612	.0033229	-.010356	79.07834	.7541088
#2	.0591935	-.018632	.6766527	-.000472	-.009997	78.90677	.7525504
#3	.0563166	-.021594	.6776602	.001899	-.010276	79.02019	.7502766

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0071076	.0087216	39.35586	.2045438	.0829697	.2532587	208.5730
Stddev	.0001596	.0002049	.07032	.0009577	.0005058	.0008350	.6916
%RSD	2.245017	2.348884	.1786734	.4682113	.6095739	.3297056	.3315707
#1	.0072905	.0088676	39.43584	.2040452	.0824443	.2528842	208.2867
#2	.0070354	.0084874	39.32795	.2056479	.0830114	.2526764	209.3618
#3	.0069969	.0088099	39.30377	.2039382	.0834533	.2542154	208.0706

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	3.078888	31.45919	.1655363	.0060345	4.077726	.2871253	.6014302
Stddev	.007116	.06920	.0005944	.0008514	.040903	.0008044	.0012764
%RSD	.2311380	.2199715	.3590567	14.10940	1.003078	.2801656	.2122278
#1	3.086641	31.53791	.1660194	.0063611	4.057113	.2874946	.6012849
#2	3.077368	31.43171	.1648725	.0066744	4.124834	.2876787	.6002326
#3	3.072654	31.40795	.1657169	.0050682	4.051231	.2862025	.6027730

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	7.947202	.0832210	.0074125	.0262555	2.867014	1.927358	3.548915
Stddev	.038130	.0017401	.0001892	.0001883	.002132	.006536	.003031
%RSD	.4797947	2.090957	2.552212	.7173225	.0743457	.3391339	.0853944
#1	7.929443	.0827836	.0075810	.0264117	2.868529	1.920358	3.549962
#2	7.990973	.0851380	.0074488	.0263084	2.864577	1.933302	3.545500
#3	7.921191	.0817413	.0072079	.0260463	2.867938	1.928414	3.551283

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2823-01 Acquired: 8/13/2025 14:20:20 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**2.176423**	**.1429790**	**-.052955**
Stddev	.005012	.0009951	.000635
%RSD	.2302675	.6959702	1.198276
#1	2.182205	.1440108	-.052615
#2	2.173326	.1420252	-.053687
#3	2.173737	.1429009	-.052562

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**6096.054**	**109663.4**	**13198.58**	**4675.450**	**6669.800**
Stddev	15.783	115.6	45.71	6.149	15.250
%RSD	.2589088	.1053905	.3463608	.1315232	.2286411
#1	6108.999	109642.7	13155.17	4668.473	6681.476
#2	6100.691	109559.7	13246.29	4677.796	6675.378
#3	6078.471	109788.0	13194.28	4680.081	6652.547

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2831-03 Acquired: 8/13/2025 14:24:23 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.001782**	**.0022204**	**.0072347**	**.0057846**	**-.000345**	**.2128086**	**.2527329**
Stddev	.001345	.0018925	.0013501	.0020184	.001414	.0027700	.0016029
%RSD	75.48122	85.23361	18.66155	34.89236	410.2523	1.301650	.6342331
#1	-.000765	.0042956	.0075343	.0059234	.001027	.2155533	.2535139
#2	-.003307	.0017760	.0084098	.0077301	-.001796	.2128587	.2508892
#3	-.001274	.0005897	.0057599	.0037004	-.000265	.2100139	.2537956

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0000979**	**.0009984**	**18.27654**	**.0001951**	**.0007236**	**.0133758**	**.0766454**
Stddev	.0000255	.0000861	.02150	.0002028	.0001225	.0002703	.0055073
%RSD	26.09615	8.620233	.1176456	103.9375	16.92504	2.020487	7.185439
#1	.0000888	.0010729	18.29203	-.000013	.0006898	.0134573	.0800146
#2	.0001267	.0010182	18.25199	.000392	.0006216	.0130742	.0796315
#3	.0000781	.0009042	18.28560	.000206	.0008594	.0135959	.0702899

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.4360914**	**5.240611**	**.0046601**	**.0002476**	**346.6346**	**.0004213**	**.4801889**
Stddev	.0001023	.028827	.0004023	.0001374	4.4560	.0023461	.0003268
%RSD	.0234688	.5500614	8.633362	55.48468	1.285492	556.8789	.0680459
#1	.4359813	5.214469	.0050681	.0003150	347.2159	-.002265	.4800080
#2	.4361837	5.271526	.0046484	.0000895	350.7714	.001462	.4805661
#3	.4361091	5.235838	.0042637	.0003382	341.9165	.002067	.4799926

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.283750**	**.0615317**	**-.000045**	**.0126468**	**-.000006**	**1.012198**	**.0363648**
Stddev	.029012	.0002360	.000100	.0009100	.000794	.011543	.0018684
%RSD	2.259923	.3836309	223.5305	7.195346	12531.79	1.140378	5.137891
#1	1.252012	.0617605	-.000143	.0133302	-.000835	1.000759	.0373706
#2	1.308903	.0612890	.000057	.0116139	.000068	1.023842	.0342090
#3	1.290334	.0615455	-.000048	.0129963	.000748	1.011991	.0375149

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2831-03 Acquired: 8/13/2025 14:24:23 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**2.475763**	**-.011109**	**.0848972**
Stddev	.020844	.000913	.0001657
%RSD	.8419229	8.215181	.1951584
#1	2.499631	-.010942	.0849354
#2	2.461143	-.010291	.0847158
#3	2.466515	-.012094	.0850406

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5385.989**	**94690.45**	**11623.75**	**4075.253**	**6446.534**
Stddev	4.937	253.82	57.76	10.361	4.117
%RSD	.0916672	.2680559	.4969014	.2542410	.0638679
#1	5390.458	94887.34	11557.40	4081.787	6446.719
#2	5386.819	94403.99	11662.82	4063.306	6450.557
#3	5380.689	94780.03	11651.02	4080.665	6442.328

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2831-03DUP Acquired: 8/13/2025 14:28:47 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.003354**	**.0014374**	**.0034986**	**.0042349**	**.0005220**	**.1115003**	**.1365477**
Stddev	.000283	.0009253	.0011448	.0008564	.0013241	.0033446	.0010934
%RSD	8.424692	64.37288	32.72062	20.22219	253.6791	2.999625	.8007627
#1	-.003031	.0020996	.0047531	.0032468	.0011585	.1123039	.1361672
#2	-.003473	.0018325	.0032321	.0047640	.0014076	.1078271	.1377805
#3	-.003557	.0003802	.0025106	.0046938	-.001000	.1143698	.1356954

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000020**	**.0004604**	**10.20916**	**-.000284**	**.0000906**	**.0075155**	**.0398376**
Stddev	.000058	.0000458	.02268	.000087	.0000781	.0003572	.0037292
%RSD	295.5362	9.953500	.2221622	30.70127	86.22854	4.753048	9.361138
#1	.000045	.0004936	10.20332	-.000298	.0000717	.0071280	.0441430
#2	-.000069	.0004081	10.23418	-.000364	.0000236	.0075867	.0377548
#3	-.000035	.0004796	10.18996	-.000191	.0001764	.0078317	.0376150

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2470744**	**2.956949**	**.0029505**	**.0001497**	**189.9947**	**-.000525**	**.2629899**
Stddev	.0009725	.015717	.0001315	.0002272	.5586	.001145	.0009218
%RSD	.3936256	.5315327	4.457111	151.7977	.2939906	218.1441	.3505160
#1	.2460813	2.939145	.0029633	-.000055	190.6256	-.000821	.2620424
#2	.2471167	2.968900	.0028131	.000394	189.5632	-.001494	.2638837
#3	.2480250	2.962801	.0030752	.000110	189.7952	.000739	.2630436

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.6666184**	**.0327151**	**.0000563**	**-.000364**	**-.000543**	**.5428755**	**.0242340**
Stddev	.0221596	.0001096	.0000735	.000304	.001098	.0123307	.0026964
%RSD	3.324182	.3351035	130.6769	83.40287	202.2671	2.271361	11.12637
#1	.6918465	.0325885	.0000677	-.000625	-.001644	.5383948	.0212463
#2	.6577061	.0327748	-.000022	-.000031	.000551	.5568200	.0264865
#3	.6503024	.0327818	.000123	-.000436	-.000535	.5334116	.0249691

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2831-03DUP Acquired: 8/13/2025 14:28:47 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	1.343508	-.009309	.0470925
Stddev	.003919	.001076	.0001571
%RSD	.2916972	11.55872	.3336625
#1	1.347467	-.009336	.0470512
#2	1.339631	-.010371	.0472661
#3	1.343427	-.008220	.0469601

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5510.051	97698.65	11644.48	4184.562	6724.039
Stddev	3.895	466.06	34.27	16.117	10.535
%RSD	.0706947	.4770432	.2942668	.3851555	.1566725
#1	5512.259	97160.55	11642.56	4177.265	6734.819
#2	5512.340	97960.96	11611.21	4173.384	6723.532
#3	5505.553	97974.46	11679.66	4203.037	6713.768

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCV03 Acquired: 8/13/2025 14:33:12 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV03 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**4.986424**	**4.822139**	**4.858771**	**5.024045**	**5.017082**	**9.649308**	**9.671877**
Stddev	.003959	.085199	.007191	.017583	.010405	.020193	.045147
%RSD	.0793983	1.766839	.1480049	.3499746	.2073991	.2092710	.4667852
#1	4.990412	4.914749	4.851550	5.029091	5.009077	9.640154	9.679255
#2	4.982494	4.747085	4.858832	5.004491	5.013324	9.672457	9.712880
#3	4.986365	4.804581	4.865932	5.038553	5.028844	9.635314	9.623495

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2382256**	**2.432166**	**24.13075**	**.9926120**	**2.431683**	**1.239833**	**5.013994**
Stddev	.0018072	.002384	.04250	.0020740	.003451	.001124	.046923
%RSD	.7585863	.0980048	.1761411	.2089462	.1419019	.0906852	.9358501
#1	.2401196	2.431048	24.17977	.9908566	2.428915	1.240269	4.971475
#2	.2380372	2.430546	24.10842	.9920788	2.430586	1.238556	5.006168
#3	.2365201	2.434903	24.10406	.9949005	2.435549	1.240675	5.064338

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.389669**	**23.92152**	**2.442080**	**1.238723**	**23.27223**	**2.445371**	**2.483963**
Stddev	.008551	.01033	.003269	.002431	.13607	.007631	.008131
%RSD	.3578296	.0431736	.1338689	.1962353	.5846781	.3120772	.3273534
#1	2.397991	23.92854	2.441360	1.237511	23.13046	2.448595	2.491076
#2	2.390109	23.92636	2.439231	1.237136	23.28446	2.450861	2.485714
#3	2.380906	23.90966	2.445649	1.241521	23.40177	2.436657	2.475099

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**24.67334**	**4.759229**	**5.020476**	**4.883955**	**4.832826**	**5.100712**	**4.779387**
Stddev	.16418	.037341	.006420	.010498	.003462	.038615	.011436
%RSD	.6654027	.7846068	.1278746	.2149406	.0716289	.7570491	.2392803
#1	24.48716	4.796477	5.021408	4.892453	4.829107	5.059807	4.778083
#2	24.73554	4.759413	5.013642	4.872220	4.833417	5.105796	4.768659
#3	24.79734	4.721796	5.026380	4.887192	4.835954	5.136533	4.791419

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCV03 Acquired: 8/13/2025 14:33:12 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV03 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**4.787702**	**4.582997**	**4.827607**
Stddev	.033485	.005874	.012223
%RSD	.6993966	.1281670	.2531995
#1	4.808043	4.579106	4.829398
#2	4.749054	4.589754	4.838837
#3	4.806008	4.580131	4.814587

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5574.767**	**98989.92**	**11516.30**	**4230.194**	**6730.322**
Stddev	3.921	397.78	60.67	5.893	7.412
%RSD	.0703324	.4018375	.5268301	.1393006	.1101267
#1	5573.760	99432.30	11463.50	4224.832	6729.355
#2	5579.094	98875.74	11502.83	4229.246	6738.170
#3	5571.448	98661.72	11582.57	4236.503	6723.441

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCB03 Acquired: 8/13/2025 14:37:23 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB03 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0013043**	**-.000444**	**-.000035**	**.0014647**	**-.000561**	**-.013096**	**-.007571**
Stddev	.0010760	.000922	.000585	.0012494	.000295	.002866	.000470
%RSD	82.49632	207.4499	1672.267	85.29743	52.54865	21.88644	6.204061
#1	.0000706	.000239	.000442	.0020247	-.000875	-.012492	-.007045
#2	.0020486	-.001493	-.000688	.0023362	-.000291	-.010581	-.007949
#3	.0017939	-.000079	.000140	.0000333	-.000517	-.016217	-.007719

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0000034**	**-.000120**	**-.010327**	**.0001169**	**-.000137**	**.0004249**	**-.007554**
Stddev	.0000183	.000021	.005361	.0002231	.000052	.0002892	.002689
%RSD	541.4829	17.32571	51.91306	190.8846	38.10200	68.06249	35.60151
#1	.0000243	-.000125	-.010753	.0001450	-.000078	.0007587	-.006192
#2	-.000004	-.000138	-.015462	.0003246	-.000175	.0002634	-.005818
#3	-.000010	-.000097	-.004765	-.000119	-.000159	.0002525	-.010651

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000657**	**-.023313**	**-.000155**	**.0007437**	**.1904242**	**.0000452**	**-.000538**
Stddev	.000429	.009521	.000080	.0004234	.0047336	.0020678	.000099
%RSD	65.34766	40.83963	51.30466	56.92453	2.485802	4578.159	18.42797
#1	-.000597	-.016827	-.000124	.0002873	.1873210	.0021762	-.000624
#2	-.001113	-.018867	-.000246	.0011237	.1880790	-.001953	-.000430
#3	-.000261	-.034243	-.000096	.0008202	.1958725	-.000088	-.000561

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2739713**	**.0035535**	**.0000632**	**-.000024**	**-.000684**	**.0102449**	**.0004323**
Stddev	.0411817	.0004324	.0001732	.000330	.000986	.0066173	.0012958
%RSD	15.03137	12.16744	274.1544	1388.755	144.2009	64.59191	299.7441
#1	.2747650	.0032845	.0001078	-.000324	-.001277	.0150189	.0003350
#2	.3147504	.0033237	-.000128	-.000077	.000455	.0130246	.0017740
#3	.2323986	.0040522	.000210	.000329	-.001230	.0026911	-.000812

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCB03 Acquired: 8/13/2025 14:37:23 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB03 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0001712	.0016517	-.000300
Stddev	.0001214	.0005783	.000041
%RSD	70.91110	35.01581	13.72156
#1	.0000660	.0023193	-.000345
#2	.0003040	.0013039	-.000293
#3	.0001435	.0013318	-.000263

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5808.967	103706.5	11557.20	4391.536	7615.299
Stddev	13.586	22.7	49.75	20.429	14.576
%RSD	.2338870	.0219044	.4304791	.4652011	.1914060
#1	5810.487	103729.6	11515.66	4369.999	7611.062
#2	5821.730	103705.7	11543.61	4410.640	7631.524
#3	5794.685	103684.1	11612.34	4393.970	7603.310

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2831-03LX5 Acquired: 8/13/2025 14:41:43 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000224**	**.0008080**	**.0007574**	**.0013239**	**-.000649**	**.0318481**	**.0454297**
Stddev	.000812	.0002989	.0003265	.0019765	.001485	.0010529	.0008129
%RSD	362.5866	36.99664	43.11248	149.2915	228.7796	3.306092	1.789435
#1	-.001148	.0011313	.0009554	-.000886	-.001682	.0322999	.0462405
#2	.000100	.0007510	.0003805	.001933	-.001318	.0325996	.0446146
#3	.000376	.0005416	.0009362	.002924	.001053	.0306446	.0454341

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0000047**	**.0000973**	**3.744385**	**.0000387**	**.0000530**	**.0024554**	**.0085002**
Stddev	.0000088	.0000408	.011643	.0000553	.0000989	.0001339	.0007300
%RSD	189.2416	41.95617	.3109418	143.1035	186.5810	5.452624	8.588246
#1	-.000004	.0000666	3.757727	.0000986	.0000058	.0023697	.0085723
#2	.000013	.0001436	3.739145	.0000279	-.000013	.0023868	.0077368
#3	.000005	.0000817	3.736283	-.000010	.000167	.0026097	.0091915

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0880318**	**1.067370**	**.0005679**	**.0002273**	**65.56280**	**.0001943**	**.0933207**
Stddev	.0002132	.007704	.0000554	.0003818	.53608	.0014219	.0006836
%RSD	.2421539	.7217580	9.747564	167.9915	.8176580	731.7425	.7324795
#1	.0882504	1.066151	.0006241	.0002473	65.17578	.0005096	.0928510
#2	.0878245	1.060348	.0005134	-.000164	66.17469	-.001359	.0941049
#3	.0880206	1.075610	.0005663	.000599	65.33794	.001432	.0930062

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.3965807**	**.0144797**	**-.000184**	**.0026916**	**-.000028**	**.2001550**	**.0066870**
Stddev	.0160923	.0005667	.000141	.0001849	.001149	.0095789	.0022604
%RSD	4.057775	3.913686	76.91514	6.869458	4079.186	4.785760	33.80202
#1	.3871823	.0143500	-.000126	.0027694	.001112	.2107249	.0040900
#2	.4151621	.0139891	-.000080	.0024805	-.001187	.1920481	.0082107
#3	.3873976	.0151000	-.000345	.0028250	-.000009	.1976919	.0077605

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2831-03LX5 Acquired: 8/13/2025 14:41:43 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.4870356	.0004293	.0170986
Stddev	.0016418	.0011333	.0000646
%RSD	.3371069	264.0082	.3778988
#1	.4886342	.0017373	.0170310
#2	.4853537	-.000192	.0171597
#3	.4871187	-.000257	.0171052

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5585.186	100202.2	11652.56	4255.730	6912.357
Stddev	1.550	72.2	89.21	18.708	5.359
%RSD	.0277572	.0721033	.7655904	.4395848	.0775208
#1	5585.117	100261.2	11574.69	4275.163	6906.259
#2	5586.770	100121.6	11749.90	4237.844	6916.315
#3	5583.672	100223.8	11633.10	4254.183	6914.497

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2831-03MS Acquired: 8/13/2025 14:46:01 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.8458270**	**1.737286**	**.9652560**	**2.107903**	**.8443818**	**2.148898**	**.3826328**
Stddev	.0032919	.055051	.0019581	.005155	.0007695	.011355	.0004337
%RSD	.3891977	3.168775	.2028598	.2445769	.0911301	.5284083	.1133377
#1	.8420374	1.792371	.9674898	2.110184	.8451014	2.159785	.3821437
#2	.8479788	1.737216	.9638363	2.111525	.8435706	2.137127	.3829703
#3	.8474648	1.682270	.9644419	2.102001	.8444732	2.149782	.3827843

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1909194**	**.2022705**	**14.48074**	**.4198337**	**.2044384**	**.3160573**	**3.317671**
Stddev	.0006541	.0005574	.03179	.0017132	.0006591	.0005505	.022368
%RSD	.3425861	.2755808	.2195316	.4080718	.3224065	.1741778	.6742153
#1	.1906266	.2027879	14.50251	.4184322	.2051783	.3160583	3.327126
#2	.1904629	.2016802	14.44426	.4193253	.2039140	.3155063	3.333758
#3	.1916687	.2023436	14.49546	.4217436	.2042230	.3166073	3.292127

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.5195057**	**5.857036**	**.5152142**	**.0776577**	**257.9875**	**.3094912**	**.5653778**
Stddev	.0007524	.036244	.0015212	.0004349	2.0905	.0026752	.0021538
%RSD	.1448350	.6188112	.2952615	.5600181	.8103231	.8643875	.3809552
#1	.5199643	5.825510	.5162708	.0781599	259.7102	.3074428	.5641219
#2	.5199155	5.896636	.5134706	.0774043	255.6617	.3125178	.5678648
#3	.5186373	5.848962	.5159011	.0774089	258.5907	.3085129	.5641467

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**11.81714**	**.3078938**	**.4365804**	**.7112207**	**.2010500**	**1.682962**	**5.963440**
Stddev	.05405	.0015414	.0007741	.0016633	.0003920	.007348	.047251
%RSD	.4573821	.5006193	.1773155	.2338691	.1949985	.4366399	.7923422
#1	11.75549	.3075952	.4370464	.7127760	.2013333	1.685239	6.017025
#2	11.85636	.3065235	.4356868	.7094671	.2012142	1.674745	5.945545
#3	11.83957	.3095626	.4370079	.7114188	.2006026	1.688903	5.927751

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2831-03MS Acquired: 8/13/2025 14:46:01 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	1.831955	.1880088	.2552432
Stddev	.004998	.0007590	.0000944
%RSD	.2728126	.4037142	.0369676
#1	1.835144	.1885904	.2553116
#2	1.834526	.1871502	.2551356
#3	1.826195	.1882857	.2552825

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5439.391	97001.11	11758.90	4135.885	6517.757
Stddev	8.286	426.72	49.67	13.807	13.219
%RSD	.1523291	.4399137	.4224153	.3338441	.2028180
#1	5429.842	97463.44	11784.21	4151.551	6505.464
#2	5443.642	96917.51	11790.81	4130.611	6531.740
#3	5444.688	96622.38	11701.67	4125.491	6516.068

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2831-03MSD Acquired: 8/13/2025 14:50:08 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.8248648	1.750507	.9362422	2.071232	.8250552	2.146683	.4082486
Stddev	.0033446	.018036	.0017761	.009606	.0010754	.016818	.0020494
%RSD	.4054669	1.030313	.1897025	.4637894	.1303381	.7834435	.5019883
#1	.8263015	1.768061	.9345192	2.082132	.8241325	2.151418	.4096091
#2	.8210419	1.732026	.9380670	2.064003	.8247969	2.160626	.4058915
#3	.8272509	1.751433	.9361406	2.067560	.8262362	2.128005	.4092451

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1864374	.1978592	16.49128	.4134135	.1997713	.3086908	3.292396
Stddev	.0015550	.0004035	.00567	.0014333	.0001766	.0006836	.027436
%RSD	.8340785	.2039125	.0343917	.3467045	.0884234	.2214542	.8333133
#1	.1881190	.1980268	16.48739	.4137637	.1996369	.3082781	3.262573
#2	.1861416	.1973989	16.48867	.4118375	.1997055	.3094799	3.298051
#3	.1850514	.1981519	16.49779	.4146392	.1999714	.3083144	3.316563

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.5638115	6.376995	.5031403	.0762612	299.6543	.3025380	.6117689
Stddev	.0036059	.039751	.0005890	.0003046	4.3223	.0038459	.0018888
%RSD	.6395530	.6233514	.1170613	.3994313	1.442421	1.271200	.3087506
#1	.5651357	6.342419	.5026829	.0759154	294.6635	.3018415	.6129626
#2	.5597308	6.420427	.5029330	.0764897	302.1703	.3066846	.6127529
#3	.5665681	6.368138	.5038049	.0763785	302.1293	.2990880	.6095912

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	11.87250	.3056645	.4329779	.6947003	.1979171	1.791201	5.774218
Stddev	.13244	.0025377	.0006989	.0005534	.0006869	.020793	.008269
%RSD	1.115480	.8302192	.1614189	.0796592	.3470739	1.160863	.1432019
#1	11.74365	.3085916	.4337263	.6947725	.1977898	1.767312	5.777394
#2	11.86559	.3043194	.4323421	.6941144	.1973027	1.801060	5.764832
#3	12.00825	.3040825	.4328652	.6952141	.1986588	1.805231	5.780427

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2831-03MSD Acquired: 8/13/2025 14:50:08 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	2.043275	.1853131	.2619561
Stddev	.006391	.0008483	.0012361
%RSD	.3127669	.4577749	.4718664
#1	2.047637	.1849999	.2622989
#2	2.035939	.1846659	.2605848
#3	2.046250	.1862735	.2629846

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5412.546	95621.60	11643.53	4093.967	6490.141
Stddev	5.226	213.61	87.78	14.727	6.093
%RSD	.0965558	.2233897	.7539105	.3597244	.0938754
#1	5411.979	95831.57	11557.82	4098.980	6491.923
#2	5407.627	95404.52	11639.54	4077.388	6483.357
#3	5418.032	95628.70	11733.25	4105.532	6495.145

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2831-03A Acquired: 8/13/2025 14:54:16 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.8161852**	**1.719967**	**.9271782**	**2.051999**	**.8218899**	**2.112069**	**.4152201**
Stddev	.0025675	.051805	.0016566	.006699	.0029110	.013334	.0011182
%RSD	.3145669	3.011978	.1786717	.3264794	.3541796	.6313132	.2693077
#1	.8133408	1.661563	.9290787	2.047123	.8245381	2.114963	.4163468
#2	.8183312	1.737965	.9264162	2.049236	.8187730	2.123719	.4141106
#3	.8168838	1.760372	.9260397	2.059638	.8223585	2.097526	.4152031

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1823069**	**.1958270**	**17.37746**	**.4037123**	**.1970383**	**.3050489**	**3.259105**
Stddev	.0014486	.0001843	.03862	.0003254	.0002935	.0003343	.021954
%RSD	.7945675	.0941342	.2222462	.0805935	.1489609	.1096017	.6736165
#1	.1823624	.1956581	17.42196	.4036041	.1968462	.3051724	3.261160
#2	.1808314	.1957992	17.35258	.4040780	.1968926	.3046704	3.279959
#3	.1837269	.1960236	17.35786	.4034549	.1973762	.3053040	3.236196

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.5818434**	**6.601214**	**.4985090**	**.0749881**	**315.4277**	**.2977980**	**.6383998**
Stddev	.0018705	.047845	.0001743	.0004465	2.4947	.0015731	.0014964
%RSD	.3214869	.7247877	.0349711	.5953886	.7908790	.5282555	.2344010
#1	.5838508	6.610990	.4983779	.0748782	313.5838	.2996047	.6395410
#2	.5801493	6.549237	.4987068	.0754792	318.2663	.2970575	.6367055
#3	.5815301	6.643416	.4984423	.0746068	314.4331	.2967318	.6389528

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**11.90156**	**.3035321**	**.4269753**	**.7003077**	**.1937429**	**1.872723**	**5.769628**
Stddev	.08156	.0031720	.0007963	.0021545	.0006656	.012466	.014138
%RSD	.6852792	1.045028	.1865055	.3076533	.3435298	.6656707	.2450370
#1	11.93287	.3038839	.4269369	.6982305	.1936932	1.864609	5.757904
#2	11.96281	.3001988	.4261989	.7025320	.1931036	1.887077	5.785328
#3	11.80898	.3065135	.4277901	.7001606	.1944320	1.866484	5.765653

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2831-03A Acquired: 8/13/2025 14:54:16 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	2.199047	.1836694	.2632172
Stddev	.016043	.0014938	.0005194
%RSD	.7295647	.8133028	.1973294
#1	2.183032	.1846845	.2633951
#2	2.198991	.1843696	.2636242
#3	2.215119	.1819542	.2626322

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5378.167	95313.66	11712.23	4072.355	6425.140
Stddev	7.429	216.53	61.11	22.964	8.912
%RSD	.1381366	.2271780	.5217566	.5639076	.1387091
#1	5375.971	95335.45	11711.70	4076.893	6418.966
#2	5386.447	95087.05	11773.60	4047.461	6435.357
#3	5372.084	95518.47	11651.39	4092.712	6421.097

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: PB169214BL Acquired: 8/13/2025 14:58:22 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0010235	-.001369	-.000749	.0004666	-.001921	-.017821
Stddev	.0016576	.000870	.001029	.0018969	.001751	.008505
%RSD	161.9481	63.52798	137.3356	406.5549	91.13790	47.72294
#1	.0009849	-.000381	-.001178	-.001717	-.001667	-.008046
#2	.0027000	-.001705	-.001494	.001709	-.000311	-.021890
#3	-.000614	-.002020	.000425	.001407	-.003785	-.023527

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.006645	-.000032	-.000155	-.010913	.0002165	-.000203
Stddev	.000943	.000014	.000064	.008788	.0001215	.000101
%RSD	14.19433	42.72825	41.49957	80.52361	56.12477	49.86982
#1	-.006678	-.000028	-.000123	-.009754	.0000770	-.000108
#2	-.005685	-.000021	-.000112	-.002763	.0002994	-.000191
#3	-.007571	-.000048	-.000229	-.020223	.0002732	-.000310

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.000247	-.003087	-.000551	-.016681	-.000100	.0007818
Stddev	.000437	.005731	.000210	.021290	.000125	.0000986
%RSD	177.3000	185.6381	38.16177	127.6314	125.4901	12.60749
#1	.000257	.000946	-.000792	-.030047	-.000101	.0007171
#2	-.000469	-.009647	-.000402	.007871	.000026	.0008952
#3	-.000528	-.000560	-.000459	-.027867	-.000224	.0007330

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.6899983	-.000811	-.000316	.6461554	-.002127	-.000022
Stddev	.0482194	.001431	.000049	.0117819	.000504	.000146
%RSD	6.988336	176.4536	15.60825	1.823382	23.70824	657.4477
#1	.7288618	-.001793	-.000363	.6589839	-.002709	.000128
#2	.7050966	-.001472	-.000265	.6436634	-.001817	-.000032
#3	.6360364	.000831	-.000321	.6358188	-.001854	-.000163

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: PB169214BL Acquired: 8/13/2025 14:58:22 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.000303	-.000362	.0126520	.0001514	-.000187	F .0167851
Stddev	.000239	.001056	.0168624	.0017202	.001669	.0009989
%RSD	78.93664	291.4126	133.2787	1136.296	892.6703	5.951275
#1	-.000292	.000187	.0320242	-.000502	.001722	.0178294
#2	-.000548	.000306	.0046624	-.001146	-.001373	.0166874
#3	-.000070	-.001580	.0012693	.002103	-.000909	.0158387

Elem	Sr4077
Units	ppm
Avg	-.000385
Stddev	.000022
%RSD	5.690411
#1	-.000392
#2	-.000403
#3	-.000361

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5755.226	103286.9	11494.17	4335.281	7536.664
Stddev	12.533	91.4	21.48	17.691	19.896
%RSD	.2177708	.0884516	.1868431	.4080599	.2639897
#1	5743.195	103363.4	11470.19	4353.659	7521.600
#2	5768.207	103185.7	11511.64	4333.814	7559.218
#3	5754.276	103311.5	11500.67	4318.370	7529.176

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: PB169214BS Acquired: 8/13/2025 15:02:41 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.7938186	2.189076	.9560042	2.026724	.8034402	1.975594
Stddev	.0021161	.023292	.0021111	.007024	.0023960	.007247
%RSD	.2665748	1.064003	.2208226	.3465452	.2982235	.3668485
#1	.7938879	2.177530	.9583097	2.022218	.8009094	1.983903
#2	.7958992	2.173812	.9541659	2.023136	.8056737	1.970578
#3	.7916687	2.215885	.9555368	2.034816	.8037376	1.972301

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1889560	.2000878	.1899912	1.011083	.4100973	.1953476
Stddev	.0005085	.0017927	.0002470	.003965	.0009950	.0001890
%RSD	.2691272	.8959469	.1300109	.3921409	.2426386	.0967637
#1	.1887475	.2013838	.1902228	1.007401	.4097026	.1952237
#2	.1885849	.2008377	.1900197	1.015280	.4093602	.1955651
#3	.1895357	.1980420	.1897312	1.010568	.4112291	.1952539

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.3102760	3.023575	.2008767	1.981185	.4909465	.0765581
Stddev	.0002951	.051688	.0005262	.019595	.0004238	.0005259
%RSD	.0950965	1.709502	.2619648	.9890339	.0863225	.6869422
#1	.3104296	2.989808	.2002727	1.969465	.4913109	.0761178
#2	.3104626	2.997838	.2011220	1.970283	.4910471	.0764162
#3	.3099358	3.083079	.2012356	2.003805	.4904814	.0771405

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	2.969037	.3040168	.2082205	9.866256	.2916774	.4093377
Stddev	.021020	.0026288	.0010499	.076192	.0014277	.0004565
%RSD	.7079573	.8647066	.5042331	.7722504	.4894870	.1115208
#1	2.952377	.3013267	.2094113	9.845729	.2913583	.4089553
#2	2.962081	.3065798	.2074280	9.802431	.2932376	.4092147
#3	2.992652	.3041438	.2078222	9.950610	.2904362	.4098431

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: PB169214BS Acquired: 8/13/2025 15:02:41 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.6666399	.1919342	.8958332	5.371450	F -.001421	.2000307
Stddev	.0007908	.0018770	.0019460	.020209	.001956	.0004630
%RSD	.1186221	.9779166	.2172320	.3762335	137.6998	.2314649
#1	.6674579	.1898465	.8941467	5.389964	-.001980	.2005640
#2	.6665823	.1934822	.8953905	5.374494	.000754	.1997321
#3	.6658795	.1924739	.8979625	5.349891	-.003036	.1997958

Elem	Sr4077
Units	ppm
Avg	.1943626
Stddev	.0001925
%RSD	.0990383
#1	.1944057
#2	.1941522
#3	.1945299

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5751.688	102143.9	11497.78	4323.500	7336.115
Stddev	10.907	158.0	94.96	4.580	21.047
%RSD	.1896324	.1547224	.8259352	.1059422	.2868955
#1	5742.439	102201.5	11455.24	4326.232	7317.695
#2	5748.910	102265.1	11431.51	4318.212	7331.595
#3	5763.716	101965.1	11606.57	4326.056	7359.054

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: PB169211BL Acquired: 8/13/2025 15:06:42 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0004111**	**-.000199**	**-.001120**	**-.003356**	**-.000939**	**-.020048**	**-.008017**
Stddev	.0007195	.000661	.000975	.001254	.000843	.002858	.001327
%RSD	175.0380	332.3926	87.00738	37.36476	89.71201	14.25701	16.55008
#1	-.000324	-.000044	-.002174	-.002179	-.001320	-.018584	-.006525
#2	.000443	-.000924	-.000933	-.004675	.000027	-.018219	-.009064
#3	.001114	.000371	-.000252	-.003215	-.001525	-.023342	-.008461

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0000183**	**-.000076**	**-.015517**	**.0000725**	**-.000194**	**-.000221**	**-.007885**
Stddev	.0000233	.000047	.008146	.0001255	.000098	.000184	.002626
%RSD	127.5846	60.99397	52.49584	173.1570	50.42397	83.06363	33.30239
#1	-.000001	-.000106	-.006967	.0001875	-.000271	-.000428	-.006194
#2	.000012	-.000023	-.023187	.0000914	-.000226	-.000156	-.006551
#3	.000044	-.000100	-.016396	-.000061	-.000084	-.000078	-.010911

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000367**	**-.004463**	**-.000069**	**.0002437**	**.1680191**	**-.000545**	**-.000495**
Stddev	.000595	.015803	.000434	.0004300	.0025115	.002112	.000291
%RSD	161.9412	354.1147	626.6471	176.4375	1.494784	387.6116	58.74022
#1	-.000771	.009089	-.000367	-.000130	.1681156	-.002955	-.000160
#2	.000316	-.000656	-.000269	.000147	.1654607	.000984	-.000649
#3	-.000647	-.021821	.000428	.000714	.1704810	.000336	-.000676

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2929585**	**-.002242**	**-.000193**	**-.000169**	**.0003801**	**-.002124**	**.0002534**
Stddev	.0238256	.000233	.000304	.000568	.0008979	.002269	.0022480
%RSD	8.132756	10.39470	157.4197	336.4873	236.2466	106.8311	887.0102
#1	.3100437	-.002024	.000138	.000487	.0013544	-.001325	.0028152
#2	.2657415	-.002214	-.000460	-.000501	.0002000	-.004684	-.000665
#3	.3030903	-.002488	-.000258	-.000492	-.000414	-.000363	-.001390

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: PB169211BL Acquired: 8/13/2025 15:06:42 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**-.002507**	**-.001726**	**-.000380**
Stddev	.002432	.001021	.000048
%RSD	97.01249	59.15866	12.57555
#1	.000001	-.001593	-.000348
#2	-.002666	-.002807	-.000435
#3	-.004856	-.000778	-.000357

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5798.323**	**102055.2**	**11545.26**	**4324.539**	**7614.703**
Stddev	7.539	466.9	61.26	16.988	8.538
%RSD	.1300236	.4575016	.5305655	.3928282	.1121196
#1	5789.658	101640.4	11504.38	4310.325	7608.147
#2	5801.929	102560.9	11515.70	4343.354	7611.604
#3	5803.383	101964.2	11615.69	4319.939	7624.357

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: PB169215BL Acquired: 8/13/2025 15:15:01 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0015062	-.000428	-.000812	-.002134	-.001279	-.023435	-.005510
Stddev	.0001711	.000552	.000620	.000704	.001398	.003566	.000870
%RSD	11.36091	129.1807	76.33808	32.99679	109.2534	15.21467	15.78782
#1	.0013103	-.000504	-.001492	-.001385	-.002614	-.020095	-.005545
#2	.0015813	-.000938	-.000666	-.002235	.000174	-.023020	-.004624
#3	.0016269	.000159	-.000278	-.002783	-.001397	-.027190	-.006363

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.000047	-.000135	-.015382	.0000405	-.000207	-.000469	-.013124
Stddev	.000058	.000064	.006357	.0002126	.000163	.000147	.001165
%RSD	124.0776	47.38313	41.32737	524.7280	78.42537	31.31782	8.875237
#1	.000009	-.000102	-.021738	.0000395	-.000136	-.000455	-.013966
#2	-.000041	-.000209	-.009023	-.000172	-.000394	-.000622	-.011794
#3	-.000107	-.000095	-.015386	.000254	-.000093	-.000329	-.013610

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.000145	.0010432	.0003212	.0004175	.0690850	-.000806	-.000630
Stddev	.000564	.0173742	.0000882	.0000908	.0156023	.001863	.000149
%RSD	390.2070	1665.397	27.45343	21.75538	22.58418	231.1434	23.67576
#1	.000463	.0005179	.0002315	.0005218	.0543786	-.002948	-.000591
#2	-.000652	-.016062	.0003242	.0003559	.0674258	.000088	-.000794
#3	-.000245	.018674	.0004078	.0003747	.0854505	.000442	-.000504

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1915622	-.002501	-.000308	-.000352	-.000664	.0001007	.0002412
Stddev	.0137264	.000266	.000173	.000475	.001153	.0030708	.0008293
%RSD	7.165503	10.63528	56.21316	134.8013	173.6091	3048.390	343.7718
#1	.1860739	-.002255	-.000109	-.000887	-.000429	-.000044	.0006213
#2	.1814292	-.002784	-.000424	-.000187	-.001916	.003242	-.000710
#3	.2071836	-.002465	-.000391	.000018	.000353	-.002895	.000812

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: PB169215BL Acquired: 8/13/2025 15:15:01 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**-.000445**	**-.004324**	**-.000377**
Stddev	.000902	.000723	.000033
%RSD	202.4703	16.71955	8.805651
#1	-.000192	-.004431	-.000339
#2	-.001447	-.003553	-.000390
#3	.000303	-.004987	-.000402

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5744.394**	**102711.1**	**11389.97**	**4334.163**	**7539.356**
Stddev	11.519	111.5	203.83	12.352	15.660
%RSD	.2005314	.1085839	1.789570	.2849853	.2077039
#1	5731.097	102613.7	11497.06	4322.124	7522.271
#2	5750.756	102832.8	11154.91	4333.561	7553.027
#3	5751.331	102686.7	11517.93	4346.805	7542.769

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: PB169211BS Acquired: 8/13/2025 15:19:21 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.7848174**	**1.954682**	**.9478664**	**1.989791**	**.8133987**	**1.961216**
Stddev	.0024503	.035435	.0011982	.001370	.0026707	.004952
%RSD	.3122066	1.812850	.1264075	.0688278	.3283425	.2525182
#1	.7837395	1.959969	.9491421	1.991090	.8104130	1.959730
#2	.7830909	1.916899	.9467649	1.989923	.8155600	1.966742
#3	.7876218	1.987176	.9476922	1.988361	.8142233	1.957177

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1853821**	**.1942511**	**.1881470**	**.9979744**	**.4062653**	**.1935854**
Stddev	.0009630	.0012350	.0003905	.0080823	.0008658	.0003802
%RSD	.5194793	.6357772	.2075344	.8098726	.2131039	.1964173
#1	.1858355	.1929472	.1885978	.9887545	.4068685	.1939572
#2	.1842761	.1944030	.1879177	1.001331	.4066540	.1931972
#3	.1860348	.1954031	.1879253	1.003837	.4052733	.1936016

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.3085259**	**3.018975**	**.1973862**	**1.944879**	**.4873920**	**.0755795**
Stddev	.0002140	.028553	.0008450	.020887	.0003084	.0002915
%RSD	.0693534	.9457873	.4280677	1.073929	.0632701	.3857110
#1	.3086825	3.051796	.1982846	1.941037	.4877408	.0757756
#2	.3082821	3.005280	.1966075	1.967421	.4871558	.0757185
#3	.3086132	2.999849	.1972664	1.926181	.4872792	.0752445

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.751885**	**.2998950**	**.2065639**	**9.713072**	**.2837502**	**.4137481**
Stddev	.013668	.0020729	.0003828	.051070	.0016970	.0007590
%RSD	.4966925	.6911981	.1853020	.5257830	.5980526	.1834400
#1	2.767391	.2975351	.2063815	9.770009	.2828754	.4143840
#2	2.746679	.3014212	.2063065	9.697895	.2826690	.4129078
#3	2.741584	.3007286	.2070038	9.671310	.2857060	.4139524

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: PB169211BS Acquired: 8/13/2025 15:19:21 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.6712240	.1938734	.9436818	5.420376	F .0001134	.1906200
Stddev	.0022518	.0016192	.0098181	.006628	.0005296	.0005905
%RSD	.3354744	.8351992	1.040398	.1222854	467.1833	.3097584
#1	.6734125	.1922009	.9510521	5.426410	.0004135	.1901507
#2	.6689139	.1954336	.9474568	5.413281	-.000498	.1912830
#3	.6713454	.1939856	.9325366	5.421438	.000425	.1904264

Elem	Sr4077
Units	ppm
Avg	.1907465
Stddev	.0004953
%RSD	.2596457
#1	.1912985
#2	.1906001
#3	.1903409

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5760.607	102514.7	11697.02	4335.838	7356.311
Stddev	7.859	331.6	37.82	32.091	11.319
%RSD	.1364212	.3234640	.3233563	.7401422	.1538715
#1	5753.008	102132.5	11736.44	4300.527	7343.296
#2	5760.110	102686.2	11693.59	4363.225	7363.864
#3	5768.702	102725.5	11661.02	4343.761	7361.772

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCV04 Acquired: 8/13/2025 15:23:20 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV04 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**5.006582**	**4.764937**	**4.861840**	**5.066387**	**5.057073**	**9.759803**	**9.663803**
Stddev	.008391	.079538	.005222	.007827	.008178	.026934	.120905
%RSD	.1676081	1.669233	.1074077	.1544924	.1617067	.2759644	1.251114
#1	4.998108	4.796499	4.865970	5.060269	5.051677	9.764857	9.693351
#2	5.006752	4.823850	4.855970	5.063684	5.053061	9.730700	9.530863
#3	5.014888	4.674462	4.863579	5.075207	5.066482	9.783851	9.767196

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2390305**	**2.429669**	**24.30502**	**.9910053**	**2.431002**	**1.242334**	**4.996794**
Stddev	.0013599	.001403	.02608	.0002695	.001309	.001685	.041713
%RSD	.5689187	.0577432	.1073167	.0271950	.0538314	.1356615	.8348001
#1	.2388577	2.430933	24.31981	.9913162	2.431519	1.240533	4.994554
#2	.2377653	2.429915	24.27491	.9908622	2.431973	1.243873	5.039583
#3	.2404685	2.428160	24.32035	.9908375	2.429514	1.242595	4.956246

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.394636**	**24.17056**	**2.443000**	**1.242925**	**23.74288**	**2.466601**	**2.504263**
Stddev	.001663	.01215	.001712	.001103	.12195	.004611	.004576
%RSD	.0694360	.0502537	.0700656	.0887644	.5136294	.1869500	.1827204
#1	2.395846	24.18294	2.443771	1.243739	23.69289	2.471856	2.509547
#2	2.395321	24.15866	2.444191	1.241670	23.88188	2.463230	2.501620
#3	2.392740	24.17007	2.441039	1.243367	23.65387	2.464716	2.501623

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**24.65721**	**4.784846**	**5.044223**	**4.871666**	**4.871899**	**5.099103**	**4.769733**
Stddev	.12776	.025705	.005817	.006420	.005393	.016541	.000592
%RSD	.5181358	.5372075	.1153114	.1317765	.1107047	.3243961	.0124219
#1	24.56000	4.785157	5.038909	4.875508	4.874707	5.080005	4.769991
#2	24.80191	4.758987	5.043322	4.875234	4.865681	5.108392	4.769055
#3	24.60971	4.810393	5.050437	4.864255	4.875310	5.108911	4.770153

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCV04 Acquired: 8/13/2025 15:23:20 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV04 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**4.797300**	**4.605971**	**4.852389**
Stddev	.012976	.011410	.053692
%RSD	.2704760	.2477323	1.106515
#1	4.794412	4.592854	4.899625
#2	4.811476	4.611453	4.793994
#3	4.786011	4.613606	4.863549

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5545.487**	**99200.59**	**11507.41**	**4182.301**	**6713.448**
Stddev	8.693	130.22	68.47	19.057	6.975
%RSD	.1567527	.1312741	.5950074	.4556596	.1038959
#1	5555.224	99051.88	11472.81	4160.894	6716.892
#2	5542.728	99294.28	11586.28	4197.417	6705.420
#3	5538.508	99255.60	11463.14	4188.593	6718.030

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCB04 Acquired: 8/13/2025 15:27:31 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB04 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0006266	-.000228	-.000818	-.001163	-.001750	-.015706	-.008146
Stddev	.0011529	.000821	.001139	.000186	.001919	.007660	.001159
%RSD	183.9847	360.0616	139.3426	15.99170	109.6717	48.77059	14.22457
#1	.0019305	-.000954	-.001659	-.001325	-.003960	-.022755	-.008678
#2	-.000258	-.000392	.000479	-.001204	-.000771	-.016810	-.008944
#3	.000208	.000662	-.001273	-.000960	-.000518	-.007554	-.006817

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.000045	-.000098	-.012792	.0002817	-.000097	-.000294	-.005749
Stddev	.000032	.000042	.010021	.0000558	.000108	.000057	.002936
%RSD	70.34791	42.49146	78.33554	19.80702	111.6106	19.42404	51.06848
#1	-.000015	-.000140	-.024356	.0003404	-.000175	-.000357	-.008842
#2	-.000042	-.000057	-.006663	.0002755	-.000141	-.000246	-.003000
#3	-.000078	-.000096	-.007358	.0002293	.000026	-.000278	-.005405

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.000646	.0021441	.0002269	.0004995	.0345725	-.000698	.0002276
Stddev	.000229	.0139189	.0000453	.0004575	.0111349	.001094	.0000262
%RSD	35.49877	649.1859	19.96302	91.58869	32.20728	156.6740	11.50061
#1	-.000681	-.002513	.0001940	.0008650	.0273777	.000464	.0001993
#2	-.000857	-.008849	.0002786	-.000014	.0473982	-.000850	.0002324
#3	-.000402	.017794	.0002082	.000647	.0289416	-.001708	.0002510

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1786227	.0052852	-.000120	-.000082	.0005844	.0214060	.0023649
Stddev	.0182456	.0003073	.000182	.000304	.0005932	.0071425	.0027215
%RSD	10.21458	5.814260	152.1822	370.0067	101.5089	33.36692	115.0769
#1	.1995735	.0056400	.000091	-.000357	.0012520	.0288243	.0025590
#2	.1662245	.0051011	-.000221	.000245	.0003831	.0145757	-.000448
#3	.1700700	.0051146	-.000228	-.000134	.0001180	.0208179	.004984

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCB04 Acquired: 8/13/2025 15:27:31 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB04 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	-.000558	-.002100	-.000347
Stddev	.001640	.000751	.000021
%RSD	293.8440	35.75941	6.084275
#1	.001310	-.002774	-.000371
#2	-.001761	-.002235	-.000331
#3	-.001223	-.001291	-.000338

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5812.099	104162.6	11752.27	4374.846	7663.584
Stddev	9.963	483.3	12.47	16.746	9.584
%RSD	.1714149	.4639507	.1061224	.3827870	.1250534
#1	5816.245	103903.7	11765.60	4357.058	7670.426
#2	5819.319	103863.8	11740.88	4377.175	7667.695
#3	5800.732	104720.1	11750.34	4390.307	7652.631

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: PB169200TB Acquired: 8/13/2025 15:31:51 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.001733**	**-.001251**	**-.000954**	**.0064789**	**.0001909**	**-.012922**	**-.009606**
Stddev	.000746	.001418	.001236	.0009879	.0013411	.001543	.000140
%RSD	43.05275	113.3527	129.6193	15.24854	702.6429	11.94053	1.461455
#1	-.001163	.000386	-.002222	.0064606	.0003743	-.014377	-.009503
#2	-.001458	-.002054	-.000887	.0055003	-.001232	-.013084	-.009549
#3	-.002577	-.002084	.000248	.0074759	.001431	-.011304	-.009766

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000145**	**-.000210**	**.0105569**	**.0000744**	**-.000443**	**-.000113**	**-.005603**
Stddev	.000013	.000035	.0048680	.0003986	.000130	.000396	.008326
%RSD	8.948788	16.85594	46.11250	535.6730	29.44306	349.7343	148.6179
#1	-.000136	-.000169	.0156342	.0001001	-.000449	.000042	-.003759
#2	-.000159	-.000230	.0059294	.0004595	-.000570	.000181	-.014696
#3	-.000138	-.000230	.0101069	-.000336	-.000309	-.000564	.001648

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0055757**	**-.019547**	**.0009463**	**.0002576**	**441.2733**	**-.000780**	**.0027015**
Stddev	.0000689	.014846	.0002477	.0002575	2.2551	.000664	.0002476
%RSD	1.236446	75.95006	26.17142	99.96742	.5110384	85.09278	9.165545
#1	.0054992	-.008862	.0012137	.0001590	443.4230	-.000017	.0024160
#2	.0056329	-.013281	.0007247	.0005497	438.9259	-.001098	.0028299
#3	.0055950	-.036499	.0009007	.0000639	441.4710	-.001225	.0028585

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1628629**	**.0004407**	**.0002781**	**.0017237**	**-.001204**	**.0075056**	**.0085583**
Stddev	.0196115	.0001943	.0000857	.0007288	.001820	.0104390	.0005957
%RSD	12.04171	44.08205	30.80454	42.27880	151.1821	139.0824	6.960305
#1	.1742980	.0002171	.0002634	.0015876	-.000631	.0191918	.0080713
#2	.1402179	.0005368	.0003702	.0010726	-.003240	.0042211	.0083812
#3	.1740728	.0005682	.0002008	.0025110	.000261	-.000896	.0092225

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: PB169200TB Acquired: 8/13/2025 15:31:51 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.1028890	-.009087	-.000173
Stddev	.0029027	.000353	.000020
%RSD	2.821224	3.887003	11.44720
#1	.1009599	-.009440	-.000151
#2	.1014799	-.008733	-.000179
#3	.1062273	-.009089	-.000189

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5259.474	93740.99	11870.25	3963.272	6334.302
Stddev	10.789	129.24	42.59	19.409	6.300
%RSD	.2051437	.1378663	.3588237	.4897223	.0994563
#1	5267.172	93670.20	11900.24	3984.606	6327.756
#2	5264.108	93662.61	11889.01	3946.660	6340.322
#3	5247.142	93890.15	11821.49	3958.548	6334.827

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2808-04 Acquired: 8/13/2025 15:36:20 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0012084	-.007142	-.001966	.0008957	-.001231	.0353977	.7543543
Stddev	.0011546	.001331	.000905	.0020298	.001383	.0169755	.0023880
%RSD	95.54893	18.63047	46.02128	226.6194	112.3448	47.95647	.3165628
#1	.0011245	-.008650	-.001111	.0005554	-.001121	.0507281	.7568897
#2	.0024027	-.006640	-.002913	.0030741	-.002667	.0171542	.7540254
#3	.0000980	-.006135	-.001873	-.000942	.000094	.0383109	.7521478

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0003242	.0004314	143.6534	-.000585	.0014674	-.001415	44.79447
Stddev	.0000299	.0000682	.2709	.000306	.0000678	.000107	.07574
%RSD	9.231972	15.79592	.1885467	52.29269	4.623183	7.544293	.1690826
#1	.0002944	.0004456	143.7573	-.000726	.0014798	-.001318	44.78370
#2	.0003543	.0004914	143.8570	-.000795	.0015283	-.001530	44.72469
#3	.0003239	.0003573	143.3460	-.000234	.0013943	-.001399	44.87502

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.313864	7.095835	.0203040	.0025133	336.3410	.0000097	.1752275
Stddev	.005340	.024980	.0002070	.0003095	2.3443	.0007534	.0002137
%RSD	.4064632	.3520442	1.019769	12.31277	.6970086	7762.912	.1219832
#1	1.318051	7.085513	.0204942	.0023224	339.0220	-.000633	.1750853
#2	1.315692	7.124322	.0203345	.0023472	335.3247	-.000177	.1751240
#3	1.307850	7.077669	.0200834	.0028704	334.6763	.000839	.1754733

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.853296	.0928011	.0002084	-.002211	-.005004	.9744557	.0088558
Stddev	.026461	.0007300	.0000739	.000550	.000903	.0149631	.0013873
%RSD	1.427756	.7865870	35.43935	24.88461	18.03725	1.535536	15.66535
#1	1.845325	.0920627	.0001431	-.002834	-.004727	.9592809	.0085364
#2	1.831737	.0928183	.0002886	-.001790	-.004273	.9748884	.0076561
#3	1.882825	.0935223	.0001936	-.002010	-.006013	.9891978	.0103750

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2808-04 Acquired: 8/13/2025 15:36:20 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**2.711850**	**-.023660**	**.7856772**
Stddev	.008461	.000470	.0019203
%RSD	.3119986	1.984925	.2444132
#1	2.717472	-.023555	.7878258
#2	2.715960	-.024173	.7850774
#3	2.702120	-.023252	.7841284

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5158.161**	**92861.56**	**11796.74**	**3888.707**	**6164.053**
Stddev	4.954	17.34	35.72	6.322	4.149
%RSD	.0960425	.0186783	.3028281	.1625823	.0673081
#1	5160.742	92862.52	11779.42	3882.122	6160.774
#2	5161.292	92843.75	11772.99	3894.728	6162.668
#3	5152.450	92878.40	11837.83	3889.272	6168.718

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2826-02 Acquired: 8/13/2025 15:40:42 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.003870**	**.0013367**	**.0016224**	**.0009300**	**.0031688**	**.0630191**	**.1632868**
Stddev	.000858	.0014517	.0000766	.0008565	.0014430	.0067992	.0004741
%RSD	22.16069	108.6039	4.720620	92.10016	45.53870	10.78918	.2903508
#1	-.004848	.0001936	.0015644	-.000052	.0019344	.0612981	.1629348
#2	-.003519	.0029701	.0015937	.001318	.0047553	.0705135	.1630997
#3	-.003244	.0008464	.0017093	.001524	.0028168	.0572458	.1638259

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000201**	**.0074130**	**215.8080**	**.0307831**	**.0034087**	**.0692483**	**-.005556**
Stddev	.000062	.0000339	.2169	.0002474	.0002559	.0004980	.005497
%RSD	31.11475	.4579942	.1005035	.8035180	7.508624	.7190735	98.94037
#1	-.000248	.0074473	215.5626	.0304976	.0032664	.0688442	-.010281
#2	-.000224	.0074122	215.8875	.0309326	.0032554	.0690961	-.006864
#3	-.000130	.0073794	215.9740	.0309190	.0037041	.0698045	.000477

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.8521692**	**7.864127**	**.2856859**	**.0001217**	**353.6145**	**.0012170**	**.2019797**
Stddev	.0007872	.071823	.0007400	.0000527	.5166	.0003557	.0006402
%RSD	.0923748	.9133012	.2590378	43.32778	.1461046	29.23057	.3169736
#1	.8521487	7.805185	.2849488	.0001823	354.1647	.0013085	.2021213
#2	.8513925	7.843071	.2856803	.0000960	353.1396	.0015180	.2012806
#3	.8529665	7.944125	.2864288	.0000868	353.5392	.0008244	.2025373

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**6.455178**	**.1522442**	**.0017668**	**-.002519**	**-.009678**	**7.306074**	**.0321874**
Stddev	.046069	.0007788	.0000610	.001176	.000823	.051828	.0025756
%RSD	.7136775	.5115451	3.453233	46.67749	8.501667	.7093882	8.001861
#1	6.469983	.1519378	.0018062	-.003867	-.008792	7.337459	.0341779
#2	6.492024	.1531296	.0016965	-.001982	-.009823	7.334511	.0292785
#3	6.403526	.1516652	.0017977	-.001708	-.010419	7.246252	.0331058

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2826-02 Acquired: 8/13/2025 15:40:42 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**8.226027**	**-.034296**	**.7481124**
Stddev	.050011	.001144	.0016771
%RSD	.6079566	3.335712	.2241805
#1	8.191122	-.033056	.7495049
#2	8.203639	-.035310	.7485817
#3	8.283321	-.034521	.7462506

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5146.808**	**92279.32**	**11784.68**	**3808.550**	**6169.524**
Stddev	1.627	89.97	46.13	4.458	5.498
%RSD	.0316022	.0974960	.3914651	.1170558	.0891226
#1	5148.622	92262.71	11831.06	3809.639	6173.699
#2	5146.320	92198.82	11784.21	3803.648	6171.580
#3	5145.480	92376.44	11738.79	3812.362	6163.294

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2827-04 Acquired: 8/13/2025 15:45:01 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0033375	-.003074	-.001057	.0014729	-.000308	.0380344	.2495645
Stddev	.0016837	.001029	.000641	.0027602	.000919	.0160508	.0014078
%RSD	50.44936	33.47081	60.59839	187.4022	298.6974	42.20084	.5641090
#1	.0052817	-.003587	-.000964	-.000025	.000667	.0565648	.2511639
#2	.0023584	-.001890	-.000468	-.000215	-.000432	.0284558	.2490165
#3	.0023724	-.003746	-.001739	.004658	-.001159	.0290826	.2485131

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0000982	.0001396	138.2531	-.000157	.0059795	-.001708	25.64333
Stddev	.0000389	.0000339	.4155	.000235	.0000543	.000679	.70712
%RSD	39.61762	24.23960	.3005079	149.3777	.9088097	39.74215	2.757535
#1	.0001155	.0001688	138.6192	-.000254	.0059409	-.001211	25.16834
#2	.0001254	.0001025	137.8015	-.000329	.0059559	-.002482	26.45599
#3	.0000536	.0001476	138.3385	.000111	.0060417	-.001432	25.30567

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.504102	8.245154	.0126765	.0019749	369.3551	.0004210	.1281643
Stddev	.003141	.015705	.0001787	.0004440	8.0391	.0016323	.0038640
%RSD	.2088171	.1904801	1.409777	22.48106	2.176532	387.7670	3.014896
#1	1.507721	8.255025	.0127857	.0022206	363.2808	.0022615	.1259379
#2	1.502497	8.227043	.0124703	.0022418	378.4713	-.000851	.1326260
#3	1.502087	8.253393	.0127735	.0014624	366.3131	-.000148	.1259288

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	3.360798	.1614519	.0019197	-.003109	-.005713	.9725995	.2032524
Stddev	.080698	.0013615	.0003764	.000292	.000385	.0164030	.0038331
%RSD	2.401147	.8433026	19.60880	9.388671	6.747460	1.686509	1.885884
#1	3.302088	.1615220	.0016286	-.002775	-.005993	.9561299	.2075059
#2	3.452818	.1627771	.0023448	-.003312	-.005872	.9889350	.2000657
#3	3.327487	.1600567	.0017858	-.003241	-.005273	.9727336	.2021856

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2827-04 Acquired: 8/13/2025 15:45:01 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**2.020646**	**-.025564**	**.5703315**
Stddev	.005679	.000964	.0015278
%RSD	.2810378	3.770428	.2678728
#1	2.014401	-.024725	.5715272
#2	2.025501	-.026617	.5686103
#3	2.022034	-.025351	.5708568

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5140.094**	**91969.88**	**11910.66**	**3834.611**	**6150.932**
Stddev	14.189	2231.06	79.44	103.400	7.563
%RSD	.2760359	2.425855	.6670009	2.696489	.1229531
#1	5145.943	93440.29	11823.35	3896.357	6155.400
#2	5150.423	89402.73	11978.69	3715.239	6155.196
#3	5123.916	93066.63	11929.93	3892.237	6142.200

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2827-08 Acquired: 8/13/2025 15:49:23 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0004648**	**-.001267**	**.0023193**	**.0023266**	**.0015198**	**.0487123**	**.3903590**
Stddev	.0009130	.000683	.0011875	.0004222	.0007329	.0045938	.0011785
%RSD	196.4136	53.92181	51.19976	18.14667	48.22536	9.430496	.3019021
#1	-.000413	-.001527	.0016642	.0027112	.0020389	.0472627	.3909608
#2	.000398	-.000492	.0036900	.0018749	.0018390	.0538561	.3890011
#3	.001410	-.001782	.0016037	.0023938	.0006814	.0450182	.3911151

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000030**	**.0000027**	**161.0306**	**-.000177**	**.0029226**	**.0004742**	**11.32621**
Stddev	.000045	.0000286	.1109	.000292	.0001423	.0002516	.01758
%RSD	151.4507	1055.262	.0688726	165.3530	4.868289	53.06499	.1552358
#1	-.000075	-.000015	161.0566	.000157	.0029378	.0002823	11.33421
#2	.000014	.000036	160.9090	-.000385	.0030567	.0007591	11.30605
#3	-.000027	-.000013	161.1262	-.000302	.0027733	.0003812	11.33838

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.6648473**	**3.602779**	**.0131446**	**.0006610**	**256.3784**	**.0010493**	**.0921859**
Stddev	.0026927	.023224	.0002111	.0005121	1.4984	.0022454	.0004844
%RSD	.4050022	.6446002	1.606094	77.47172	.5844562	213.9792	.5254905
#1	.6662891	3.610170	.0129298	.0011115	258.0277	.0026946	.0922620
#2	.6617407	3.621408	.0131522	.0001041	256.0068	-.001509	.0926278
#3	.6665120	3.576760	.0133518	.0007673	255.1008	.001962	.0916680

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.665077**	**.0347487**	**.0018025**	**-.002590**	**-.006114**	**1.242884**	**.0140477**
Stddev	.034436	.0005577	.0001830	.000708	.000471	.016625	.0014188
%RSD	1.292139	1.605026	10.15119	27.34576	7.704667	1.337593	10.10002
#1	2.700511	.0350012	.0016436	-.003080	-.006611	1.223719	.0139328
#2	2.631733	.0341094	.0020025	-.001778	-.005674	1.251503	.0155204
#3	2.662988	.0351356	.0017612	-.002912	-.006056	1.253428	.0126898

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2827-08 Acquired: 8/13/2025 15:49:23 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	2.260382	-.027707	.6373880
Stddev	.019203	.001200	.0022095
%RSD	.8495358	4.331528	.3466420
#1	2.242359	-.027369	.6368776
#2	2.258206	-.029040	.6354785
#3	2.280580	-.026712	.6398080

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5038.797	90754.12	11402.74	3777.226	6099.754
Stddev	37.033	188.74	12.15	19.547	35.239
%RSD	.7349610	.2079649	.1065776	.5175070	.5777085
#1	5059.530	90665.48	11412.82	3770.968	6120.440
#2	5060.819	90626.02	11389.25	3761.575	6119.756
#3	4996.041	90970.86	11406.17	3799.137	6059.066

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2808-01 Acquired: 8/13/2025 15:53:46 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1800329**	**.0741751**	**2.631210**	**.0111485**	**-.056125**	**138.0349**
Stddev	.0032144	.0075606	.001568	.0039018	.000925	.1482
%RSD	1.785433	10.19287	.0595839	34.99810	1.647939	.1073441
#1	.1826189	.0705985	2.629903	.0156013	-.056725	137.9216
#2	.1810458	.0690665	2.630779	.0095167	-.056590	137.9805
#3	.1764341	.0828604	2.632948	.0083276	-.055060	138.2026

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.094576**	**.0145821**	**.0838653**	**225.7189**	**1.039326**	**.3533821**
Stddev	.002682	.0001221	.0006281	.4492	.004628	.0002190
%RSD	.1280380	.8375648	.7489187	.1990023	.4452901	.0619682
#1	2.093157	.0144809	.0832362	225.2066	1.039877	.3532155
#2	2.092901	.0145476	.0844923	225.9045	1.043655	.3533007
#3	2.097669	.0147177	.0838674	226.0455	1.034448	.3536301

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.759315**	**895.1544**	**6.072521**	**86.12084**	**2.293298**	**.0454620**
Stddev	.003353	6.4741	.010794	.25625	.000710	.0019672
%RSD	.1215271	.7232422	.1777488	.2975493	.0309422	4.327147
#1	2.756124	901.7589	6.062760	85.82496	2.292673	.0472833
#2	2.759012	894.8853	6.084113	86.26598	2.294069	.0457271
#3	2.762810	888.8190	6.070690	86.27157	2.293153	.0433757

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**17.08201**	**10.35382**	**5.291721**	**6.838310**	**^ *****	**.0751818**
Stddev	.14609	.01088	.008489	.027116	-----	.0004492
%RSD	.8552481	.1050630	.1604148	.3965274	-----	.5974492
#1	17.23668	10.34234	5.282000	6.858528	^ -----	.0751910
#2	17.06298	10.35516	5.295499	6.848906	^ -----	.0756264
#3	16.94636	10.36397	5.297665	6.807496	^ -----	.0747282

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2808-01 Acquired: 8/13/2025 15:53:46 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.143629	3.997180	8.950507	9.454075	F 62.82702	.1715536
Stddev	.002432	.007489	.143087	.024550	.40623	.0022124
%RSD	.2126665	.1873633	1.598650	.2596788	.6465844	1.289631
#1	1.141105	3.988535	8.940798	9.460310	62.80098	.1720816
#2	1.145958	4.001304	9.098201	9.474906	63.24565	.1734543
#3	1.143825	4.001701	8.812521	9.427008	62.43444	.1691250

Elem	Sr4077
Units	ppm
Avg	.7413666
Stddev	.0226181
%RSD	3.050870
#1	.7157265
#2	.7498832
#3	.7584901

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5696.821	104518.4	13894.88	4253.124	5664.939
Stddev	8.061	192.8	102.65	3.394	3.323
%RSD	.1414970	.1844623	.7387386	.0798042	.0586514
#1	5700.852	104368.8	14011.48	4251.763	5667.372
#2	5702.072	104450.5	13854.96	4250.623	5666.292
#3	5687.540	104736.0	13818.18	4256.988	5661.154

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2817-01 Acquired: 8/13/2025 15:58:02 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0159929	.0214913	.0569516	-.011547	.0096765	119.7184	7.361722
Stddev	.0043442	.0021804	.0011669	.003261	.0025923	.2278	.017733
%RSD	27.16340	10.14534	2.049020	28.24337	26.78950	.1902590	.2408823
#1	.0126426	.0189748	.0561640	-.009728	.0120109	119.4951	7.351709
#2	.0144349	.0228155	.0563986	-.009600	.0068867	119.7096	7.351259
#3	.0209013	.0226835	.0582923	-.015312	.0101321	119.9504	7.382197

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0019774	.0082569	2187.873	.0593914	.0249562	.0572867	111.2777
Stddev	.0000469	.0001407	15.838	.0002033	.0004285	.0015108	.3812
%RSD	2.370089	1.704371	.7239144	.3422926	1.717001	2.637314	.3425388
#1	.0020309	.0084009	2206.115	.0595530	.0247352	.0588926	110.9173
#2	.0019574	.0082502	2179.880	.0594580	.0246832	.0570739	111.2391
#3	.0019438	.0081197	2177.624	.0591631	.0254500	.0558935	111.6767

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.6303406	294.7570	.1856770	.0184864	52.26184	.0976766	2.813455
Stddev	.0012775	.0737	.0007114	.0004646	.26752	.0008193	.005626
%RSD	.2026767	.0250085	.3831118	2.513492	.5118759	.8388312	.1999734
#1	.6317361	294.7615	.1863830	.0179514	51.99040	.0968175	2.816595
#2	.6292286	294.8284	.1849604	.0187179	52.26986	.0977629	2.816811
#3	.6300571	294.6812	.1856877	.0187898	52.52525	.0984493	2.806960

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	28.83225	.1376957	.0047863	.0777810	9.862906	3.127149	9.469285
Stddev	.22718	.0026629	.0006540	.0017422	.046872	.032871	.002983
%RSD	.7879519	1.933894	13.66310	2.239854	.4752354	1.051153	.0315036
#1	28.60651	.1354698	.0040317	.0794891	9.875588	3.136849	9.470172
#2	28.82938	.1369716	.0051871	.0778472	9.902131	3.154078	9.471724
#3	29.06085	.1406458	.0051401	.0760066	9.810997	3.090519	9.465959

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2817-01 Acquired: 8/13/2025 15:58:02 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**17.77725**	**-.286456**	**1.244542**
Stddev	.05895	.003348	.002127
%RSD	.3316002	1.168719	.1708942
#1	17.83882	-.289100	1.245300
#2	17.72133	-.287577	1.242140
#3	17.77162	-.282692	1.246185

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**4661.789**	**84226.01**	**11981.89**	**3484.731**	**4867.015**
Stddev	6.606	157.36	40.63	7.452	15.862
%RSD	.1417051	.1868335	.3391069	.2138553	.3259127
#1	4669.383	84307.83	11935.51	3489.956	4877.086
#2	4658.606	84325.61	12011.17	3488.041	4875.228
#3	4657.377	84044.59	11999.00	3476.197	4848.730

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2819-06 Acquired: 8/13/2025 16:02:22 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0295681	-.002025	.3391143	.0025716	-.002239	40.53668	.1649028
Stddev	.0009381	.000952	.0003256	.0008557	.001405	.10716	.0003091
%RSD	3.172670	47.03529	.0960099	33.27358	62.75319	.2643643	.1874691
#1	.0305262	-.002576	.3387664	.0016127	-.000651	40.65830	.1651227
#2	.0286514	-.000925	.3394116	.0028447	-.003320	40.45612	.1650363
#3	.0295267	-.002573	.3391649	.0032574	-.002747	40.49561	.1645493

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0027959	.0017242	3.648005	.1100694	.0159814	.0459838	59.94381
Stddev	.0000498	.0000771	.018676	.0002597	.0002884	.0003141	.13663
%RSD	1.780254	4.473007	.5119561	.2359039	1.804572	.6830040	.2279226
#1	.0028531	.0017643	3.643727	.1103691	.0158759	.0463248	59.89063
#2	.0027622	.0016352	3.631839	.1099141	.0157606	.0457064	60.09903
#3	.0027724	.0017729	3.668449	.1099248	.0163077	.0459203	59.84177

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.7734115	6.536183	.0359841	.0027939	.9287686	.0886997	.1848661
Stddev	.0021155	.010341	.0004410	.0004057	.0127441	.0013520	.0006991
%RSD	.2735216	.1582177	1.225477	14.52069	1.372151	1.524231	.3781866
#1	.7758503	6.547506	.0355489	.0027385	.9178605	.0890158	.1846904
#2	.7720725	6.527237	.0359727	.0032245	.9256685	.0898657	.1842716
#3	.7723117	6.533808	.0364306	.0024188	.9427767	.0872177	.1856363

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	2.593731	.0122503	.0041230	.0166402	1.202571	2.058589	1.973729
Stddev	.026574	.0017682	.0000417	.0004858	.002886	.007744	.011679
%RSD	1.024540	14.43363	1.010701	2.919133	.2400049	.3762024	.5917076
#1	2.563046	.0109268	.0041248	.0171662	1.205589	2.049676	1.967975
#2	2.609104	.0142584	.0041637	.0165460	1.199838	2.063671	1.966045
#3	2.609043	.0115657	.0040805	.0162085	1.202286	2.062421	1.987169

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2819-06 Acquired: 8/13/2025 16:02:22 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	1.081090	.0601279	-.033818
Stddev	.008608	.0007097	.000099
%RSD	.7962149	1.180226	.2933515
#1	1.087160	.0604386	-.033728
#2	1.071239	.0593159	-.033924
#3	1.084871	.0606292	-.033802

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5828.270	105927.1	12790.54	4428.850	7233.982
Stddev	17.873	74.6	74.31	8.621	22.005
%RSD	.3066564	.0704444	.5809669	.1946463	.3041915
#1	5847.456	106013.0	12727.90	4437.973	7254.738
#2	5825.261	105878.5	12872.65	4427.736	7236.297
#3	5812.092	105889.7	12771.09	4420.840	7210.911

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2819-11 Acquired: 8/13/2025 16:06:31 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0704648	-.004367	.4667449	.0017330	-.003384	29.56425	.1721573
Stddev	.0034987	.002492	.0009880	.0015615	.002931	.04793	.0004831
%RSD	4.965171	57.05375	.2116759	90.09899	86.61258	.1621139	.2805978
#1	.0728630	-.001640	.4677795	.0002809	-.001866	29.53883	.1718935
#2	.0720812	-.006524	.4666442	.0033846	-.001524	29.61954	.1727149
#3	.0664501	-.004938	.4658112	.0015335	-.006763	29.53439	.1718637

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0056646	.0029592	110.3061	.0841845	.0187954	.1282340	72.07005
Stddev	.0000424	.0000212	.0224	.0003283	.0000981	.0008094	.18015
%RSD	.7489706	.7158564	.0203489	.3899735	.5218511	.6311963	.2499690
#1	.0056856	.0029378	110.3125	.0844252	.0187189	.1283444	72.10278
#2	.0056923	.0029802	110.3247	.0838106	.0189060	.1289825	71.87578
#3	.0056157	.0029596	110.2812	.0843178	.0187613	.1273751	72.23160

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.6079634	53.14769	.0498847	.0030106	1.874281	.0886187	.4421797
Stddev	.0012363	.08071	.0004164	.0001571	.003327	.0023310	.0007953
%RSD	.2033503	.1518658	.8348100	5.217064	.1775170	2.630412	.1798579
#1	.6068666	53.16746	.0499115	.0031041	1.877028	.0885502	.4425090
#2	.6093031	53.21669	.0502871	.0030985	1.870582	.0909833	.4427575
#3	.6077204	53.05893	.0494556	.0028293	1.875234	.0863227	.4412726

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	2.859434	.0272433	.0034364	.0181882	1.074529	2.073624	2.414830
Stddev	.023004	.0008122	.0003026	.0005704	.002707	.007259	.012654
%RSD	.8044803	2.981392	8.805798	3.135907	.2518943	.3500707	.5240094
#1	2.836883	.0267864	.0033852	.0176564	1.071978	2.065749	2.426694
#2	2.858556	.0267624	.0037614	.0187906	1.074240	2.075073	2.416285
#3	2.882865	.0281810	.0031627	.0181176	1.077368	2.080048	2.401512

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2819-11 Acquired: 8/13/2025 16:06:31 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**8.773774**	**.0186115**	**.0773708**
Stddev	.034542	.0003194	.0004114
%RSD	.3937015	1.716425	.5316864
#1	8.778689	.0186403	.0772887
#2	8.805596	.0182786	.0778170
#3	8.737037	.0189155	.0770066

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5512.647**	**100173.0**	**12390.82**	**4162.383**	**6647.578**
Stddev	9.278	170.2	36.82	6.630	8.815
%RSD	.1683061	.1698899	.2971413	.1592923	.1326044
#1	5521.993	99977.9	12373.84	4166.015	6657.754
#2	5503.438	100250.9	12365.56	4154.731	6642.672
#3	5512.510	100290.4	12433.07	4166.404	6642.307

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2819-14 Acquired: 8/13/2025 16:10:37 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0562086	-.001838	.0944071	.0016999	-.002344	18.90034	.0493644
Stddev	.0009465	.000624	.0009691	.0011157	.001170	.02609	.0016606
%RSD	1.683891	33.92103	1.026491	65.63104	49.91350	.1380288	3.364029
#1	.0552950	-.002320	.0940224	.0008047	-.003276	18.87080	.0481875
#2	.0561459	-.002060	.0936895	.0013452	-.002724	18.91001	.0486417
#3	.0571849	-.001134	.0955095	.0029498	-.001031	18.92021	.0512639

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0011686	.0009600	3.458917	.0635441	.0074544	.0551471	31.02448
Stddev	.0000721	.0000533	.010362	.0004881	.0001831	.0002051	.05945
%RSD	6.166037	5.552291	.2995755	.7681515	2.455707	.3719918	.1916310
#1	.0012448	.0009593	3.449301	.0640757	.0076467	.0549441	31.09045
#2	.0011015	.0009071	3.469892	.0634402	.0074342	.0553543	31.00795
#3	.0011594	.0010137	3.457558	.0631162	.0072822	.0551430	30.97505

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.3039263	3.218632	.0218651	.0010596	.7424036	.0469757	.2408749
Stddev	.0012553	.011636	.0000523	.0002110	.0117885	.0014296	.0005640
%RSD	.4130390	.3615157	.2393011	19.91474	1.587880	3.043232	.2341325
#1	.3026352	3.208030	.0218537	.0008258	.7423926	.0468424	.2414763
#2	.3051426	3.216786	.0218194	.0011173	.7541976	.0456175	.2403578
#3	.3040009	3.231081	.0219222	.0012358	.7306206	.0484673	.2407906

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.477232	.0041658	.0010734	.0070126	.6799960	1.578442	1.552906
Stddev	.003999	.0001915	.0000535	.0001263	.0010118	.009416	.005233
%RSD	.2706850	4.596142	4.987935	1.800273	.1487913	.5965654	.3369728
#1	1.477640	.0043512	.0011163	.0070693	.6788523	1.571093	1.557173
#2	1.473045	.0041773	.0010134	.0068679	.6807743	1.575177	1.547068
#3	1.481011	.0039688	.0010905	.0071005	.6803613	1.589057	1.554479

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2819-14 Acquired: 8/13/2025 16:10:37 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.4272766	.0151018	-.011078
Stddev	.0015997	.0010046	.000059
%RSD	.3743919	6.651937	.5339193
#1	.4279249	.0158123	-.011072
#2	.4254545	.0139525	-.011141
#3	.4284504	.0155405	-.011023

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5856.838	105729.6	12562.83	4407.339	7437.104
Stddev	1.171	87.5	42.88	8.296	11.229
%RSD	.0199968	.0827426	.3413558	.1882393	.1509811
#1	5857.671	105705.7	12607.08	4411.773	7439.178
#2	5857.343	105826.5	12559.96	4397.768	7447.151
#3	5855.499	105656.5	12521.46	4412.476	7424.983

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCV05 Acquired: 8/13/2025 16:14:50 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV05 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**5.008184**	**4.535581**	**4.810757**	**5.071942**	**5.093948**	**9.688008**	**9.835879**
Stddev	.012213	.134476	.005069	.016476	.010488	.030487	.175521
%RSD	.2438682	2.964905	.1053730	.3248500	.2058863	.3146899	1.784499
#1	4.994081	4.515866	4.815596	5.055795	5.081950	9.723189	10.03740
#2	5.015170	4.412050	4.805485	5.071302	5.098525	9.671504	9.71639
#3	5.015300	4.678825	4.811188	5.088729	5.101369	9.669331	9.75385

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2345472**	**2.404781**	**24.05975**	**.9910976**	**2.415756**	**1.243462**	**5.093967**
Stddev	.0023443	.003151	.06526	.0049268	.002382	.002336	.046423
%RSD	.9994838	.1310496	.2712580	.4971069	.0986044	.1878590	.9113343
#1	.2372284	2.408164	24.13258	.9859178	2.418156	1.241226	5.041633
#2	.2335290	2.401928	24.00656	.9916503	2.413392	1.245887	5.110087
#3	.2328842	2.404252	24.04010	.9957248	2.415722	1.243274	5.130181

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.382668**	**23.82309**	**2.431815**	**1.250084**	**23.79499**	**2.460924**	**2.514845**
Stddev	.013619	.09548	.001667	.000295	.17171	.011911	.006230
%RSD	.5715897	.4007756	.0685383	.0236224	.7216063	.4840047	.2477289
#1	2.397178	23.93248	2.433195	1.250300	23.60174	2.474645	2.517563
#2	2.380667	23.75649	2.429963	1.249748	23.85318	2.454879	2.507718
#3	2.370161	23.78031	2.432287	1.250205	23.93003	2.453248	2.519255

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**25.01605**	**4.703926**	**5.060931**	**4.822431**	**4.862344**	**5.152760**	**4.623886**
Stddev	.24828	.062604	.008490	.012141	.024646	.067816	.005770
%RSD	.9924974	1.330880	.1677609	.2517579	.5068676	1.316102	.1247822
#1	24.78744	4.775714	5.051870	4.822921	4.890374	5.079557	4.628521
#2	24.98053	4.675387	5.062220	4.810052	4.852593	5.165282	4.625713
#3	25.28018	4.660678	5.068704	4.834319	4.844067	5.213442	4.617424

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCV05 Acquired: 8/13/2025 16:14:50 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV05 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**4.723238**	**4.553126**	**4.856926**
Stddev	.036432	.017246	.022126
%RSD	.7713261	.3787828	.4555532
#1	4.705899	4.573014	4.831990
#2	4.698714	4.544085	4.874209
#3	4.765100	4.542281	4.864580

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5506.087**	**98525.82**	**11696.80**	**4119.269**	**6709.778**
Stddev	12.822	418.78	113.77	17.004	7.032
%RSD	.2328637	.4250489	.9726645	.4127957	.1048091
#1	5514.507	98911.53	11566.18	4125.836	6705.549
#2	5491.331	98585.54	11749.90	4132.010	6705.888
#3	5512.424	98080.38	11774.31	4099.961	6717.896

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCB05 Acquired: 8/13/2025 16:19:00 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB05 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0018468	-.000638	-.000707	.0013979	-.001684	-.012440	-.010541
Stddev	.0015917	.000768	.000424	.0001202	.001345	.010019	.001202
%RSD	86.18538	120.3210	60.04589	8.596568	79.89197	80.53817	11.40620
#1	.0028402	-.000613	-.001001	.0013324	-.003185	-.002493	-.010317
#2	.0000110	.000116	-.000900	.0015366	-.000589	-.012298	-.011839
#3	.0026892	-.001419	-.000220	.0013246	-.001276	-.022529	-.009466

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.000041	-.000169	-.017867	-.000098	-.000078	-.000062	-.008156
Stddev	.000045	.000049	.001508	.000061	.000142	.000070	.005202
%RSD	109.8832	29.05314	8.440845	62.08776	182.6525	114.0563	63.78417
#1	.000008	-.000148	-.019258	-.000137	-.000054	.000010	-.006437
#2	-.000050	-.000226	-.016264	-.000028	.000051	-.000130	-.014001
#3	-.000081	-.000134	-.018078	-.000130	-.000231	-.000064	-.004031

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.000111	-.009397	-.000044	.0007661	.2681295	.0011412	-.000066
Stddev	.000342	.006801	.000158	.0004394	.0134366	.0021803	.000185
%RSD	306.5115	72.37306	360.6661	57.34955	5.011242	191.0505	280.6693
#1	-.000226	-.007377	.000092	.0010490	.2526475	.0035233	-.000274
#2	-.000381	-.016978	-.000006	.0002600	.2767491	-.000756	.000081
#3	.000273	-.003835	-.000218	.0009895	.2749921	.000656	-.000004

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.2972330	.0029028	-.000014	-.000220	-.001009	.0044840	.0013016
Stddev	.0116232	.0002418	.000120	.000148	.000543	.0060030	.0018185
%RSD	3.910477	8.329453	838.8570	67.32381	53.83823	133.8770	139.7177
#1	.3022219	.0028757	-.000070	-.000066	-.001614	.0007944	-.000588
#2	.3055290	.0031570	-.000097	-.000232	-.000846	.0012468	.003040
#3	.2839482	.0026757	.000124	-.000361	-.000566	.0114107	.001453

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCB05 Acquired: 8/13/2025 16:19:00 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB05 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**-.000600**	**-.001000**	**-.000350**
Stddev	.002094	.001362	.000037
%RSD	348.7752	136.2270	10.47710
#1	-.000861	-.000598	-.000368
#2	-.002551	-.002518	-.000308
#3	.001612	.000116	-.000374

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5705.191**	**103393.4**	**11833.18**	**4257.600**	**7542.594**
Stddev	14.305	88.3	45.83	16.592	11.941
%RSD	.2507329	.0854283	.3873068	.3897100	.1583103
#1	5693.370	103425.1	11854.02	4276.758	7528.867
#2	5721.093	103461.6	11864.89	4248.266	7550.580
#3	5701.109	103293.6	11780.63	4247.778	7548.335

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2820-09 Acquired: 8/13/2025 16:23:20 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0628705	-.009917	.1052736	.0077410	-.006199	60.98124
Stddev	.0008748	.002270	.0006586	.0014554	.001212	.03229
%RSD	1.391410	22.89401	.6256464	18.80153	19.55075	.0529510
#1	.0618642	-.007415	.1047135	.0073226	-.006028	60.95144
#2	.0632984	-.010490	.1051081	.0065407	-.007488	60.97674
#3	.0634491	-.011846	.1059992	.0093598	-.005082	61.01555

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1100026	.0043867	.0070576	15.45684	.1446314	.1166744
Stddev	.0006241	.0000752	.0000989	.01080	.0003502	.0001801
%RSD	.5673837	1.714887	1.401612	.0698713	.2421118	.1543278
#1	.1099426	.0044113	.0070394	15.44923	.1443557	.1168752
#2	.1106546	.0044466	.0069690	15.46920	.1445131	.1165273
#3	.1094106	.0043023	.0071643	15.45210	.1450254	.1166207

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0817355	130.2292	.5729521	18.77518	.1260416	.0055557
Stddev	.0004656	.3942	.0015895	.09825	.0003654	.0002949
%RSD	.5696111	.3026802	.2774261	.5233017	.2898635	5.307317
#1	.0822722	129.7986	.5730785	18.88859	.1257706	.0052629
#2	.0814931	130.5723	.5744746	18.72105	.1258972	.0058526
#3	.0814411	130.3167	.5713031	18.71590	.1264571	.0055516

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	19.31055	.2094757	.3121155	9.048743	.1010540	.0430849
Stddev	.07520	.0015439	.0002323	.066356	.0017664	.0002029
%RSD	.3894310	.7370112	.0744417	.7333216	1.748016	.4709123
#1	19.23442	.2111753	.3119546	8.972497	.0996678	.0428907
#2	19.38478	.2081601	.3123819	9.080306	.1004512	.0432955
#3	19.31243	.2090916	.3120101	9.093428	.1030429	.0430685

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2820-09 Acquired: 8/13/2025 16:23:20 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0086497	2.066037	3.121620	2.959936	F 121.2632	.1111684
Stddev	.0005792	.005517	.027765	.014832	.7538	.0006940
%RSD	6.696455	.2670513	.8894402	.5011064	.6215947	.6242998
#1	.0092032	2.060306	3.095266	2.968400	120.7538	.1108597
#2	.0086981	2.066493	3.150608	2.942809	120.9068	.1106823
#3	.0080478	2.071312	3.118987	2.968598	122.1291	.1119632

Elem	Sr4077
Units	ppm
Avg	.0633939
Stddev	.0004224
%RSD	.6663265
#1	.0637370
#2	.0635225
#3	.0629221

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6047.697	110321.6	13630.82	4600.120	6873.599
Stddev	3.427	159.2	38.95	5.532	10.172
%RSD	.0566610	.1443344	.2857165	.1202650	.1479880
#1	6045.536	110488.3	13590.50	4593.965	6871.080
#2	6051.648	110171.0	13633.71	4601.718	6884.795
#3	6045.908	110305.4	13668.23	4604.678	6864.924

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2820-10 Acquired: 8/13/2025 16:27:25 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0458854	-.001297	.0732546	.0058382	-.003323	44.20861
Stddev	.0004193	.000821	.0011578	.0006372	.000980	.09014
%RSD	.9138928	63.34230	1.580467	10.91365	29.49138	.2038912
#1	.0460564	-.001392	.0720997	.0065457	-.004454	44.28220
#2	.0461923	-.002066	.0744152	.0053094	-.002783	44.10807
#3	.0454076	-.000432	.0732490	.0056596	-.002731	44.23556

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0990888	.0026826	.0027625	16.98924	.1498167	.0208848
Stddev	.0003903	.0000304	.0001032	.03780	.0007894	.0001538
%RSD	.3938335	1.132673	3.735938	.2225148	.5269224	.7365924
#1	.0990539	.0027066	.0028136	17.00282	.1506849	.0209178
#2	.0987172	.0026484	.0026438	16.94652	.1496231	.0207171
#3	.0994954	.0026928	.0028303	17.01837	.1491421	.0210194

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0426454	74.09354	.2983274	17.38385	.0503060	.0033076
Stddev	.0006275	.25583	.0003033	.04322	.0001572	.0003918
%RSD	1.471417	.3452821	.1016784	.2486388	.3124705	11.84599
#1	.0432976	73.79909	.2979971	17.43194	.0504705	.0028800
#2	.0420459	74.22011	.2983918	17.37140	.0502901	.0036495
#3	.0425927	74.26141	.2985934	17.34823	.0501573	.0033932

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	25.64488	.2306219	.1164593	11.15268	.1006994	.0401229
Stddev	.11666	.0006904	.0006704	.05336	.0006070	.0001028
%RSD	.4548895	.2993487	.5756505	.4784834	.6028037	.2560815
#1	25.51109	.2301783	.1166251	11.09197	.1002404	.0401720
#2	25.72532	.2302702	.1157216	11.17389	.1013877	.0400049
#3	25.69824	.2314173	.1170313	11.19218	.1004701	.0401919

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2820-10 Acquired: 8/13/2025 16:27:25 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0070478	1.855688	2.851007	1.853259	F 75.00901	.0738318
Stddev	.0008563	.004979	.005048	.002335	.09756	.0002360
%RSD	12.14977	.2683070	.1770541	.1259909	.1300627	.3195877
#1	.0062038	1.852929	2.852895	1.853712	75.11712	.0735634
#2	.0079159	1.852699	2.854839	1.850730	74.98237	.0740064
#3	.0070238	1.861435	2.845288	1.855334	74.92754	.0739256

Elem	Sr4077
Units	ppm
Avg	.1434227
Stddev	.0004725
%RSD	.3294259
#1	.1432447
#2	.1430651
#3	.1439583

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5926.693	108055.4	13441.02	4452.094	6965.546
Stddev	6.802	181.1	63.71	11.458	7.792
%RSD	.1147610	.1675556	.4739774	.2573534	.1118586
#1	5923.217	107848.2	13377.08	4438.903	6956.670
#2	5934.531	108134.4	13504.50	4459.574	6968.708
#3	5922.333	108183.5	13441.48	4457.804	6971.259

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2820-24 Acquired: 8/13/2025 16:31:32 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0100422	-.003091	.0242821	.0002366	-.001959	11.00623	.0339094
Stddev	.0006359	.001479	.0006591	.0009710	.001611	.00729	.0002141
%RSD	6.331958	47.84728	2.714395	410.4521	82.23863	.0662197	.6313969
#1	.0094449	-.001455	.0238627	.0001629	-.000918	10.99782	.0337945
#2	.0107106	-.003485	.0250418	.0012423	-.003814	11.01021	.0341564
#3	.0099711	-.004334	.0239418	-.000695	-.001144	11.01066	.0337772

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0014067	.0015051	1.143664	.0491919	.0071257	.0510101	22.25977
Stddev	.0000339	.0000927	.016326	.0001065	.0001757	.0000684	.06609
%RSD	2.412529	6.161059	1.427526	.2165382	2.466422	.1340307	.2968958
#1	.0013740	.0015654	1.148425	.0493053	.0069298	.0509444	22.30463
#2	.0014417	.0013983	1.157080	.0490939	.0071780	.0510809	22.18388
#3	.0014044	.0015517	1.125487	.0491767	.0072694	.0510051	22.29080

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1152239	2.958108	.0225097	.0012271	.5835596	.0246447	.1482109
Stddev	.0004100	.015245	.0000794	.0005636	.0088615	.0009757	.0009343
%RSD	.3557887	.5153688	.3528426	45.93433	1.518521	3.959066	.6304049
#1	.1154795	2.956077	.0224388	.0014293	.5925514	.0253167	.1471604
#2	.1154411	2.943980	.0225955	.0016618	.5832931	.0250917	.1489491
#3	.1147510	2.974267	.0224949	.0005902	.5748344	.0235256	.1485231

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.383879	.0005063	.0016523	.0053822	.4520159	1.285582	.6670487
Stddev	.044094	.0003015	.0001295	.0001942	.0031559	.009770	.0033619
%RSD	3.186293	59.55271	7.837949	3.608977	.6981948	.7599497	.5039916
#1	1.434617	.0006134	.0015275	.0054269	.4483829	1.280766	.6697656
#2	1.362192	.0007397	.0017860	.0055502	.4540801	1.279154	.6632890
#3	1.354829	.0001659	.0016433	.0051695	.4535846	1.296824	.6680916

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2820-24 Acquired: 8/13/2025 16:31:32 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	5.765592	.0176684	-.009014
Stddev	.018823	.0004329	.000072
%RSD	.3264769	2.450278	.7963738
#1	5.786035	.0180353	-.009094
#2	5.748977	.0177789	-.008955
#3	5.761763	.0171909	-.008993

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5815.306	105094.3	12312.37	4383.901	7552.557
Stddev	13.021	81.8	26.21	9.470	12.621
%RSD	.2239121	.0778669	.2128967	.2160136	.1671055
#1	5821.061	105076.5	12318.74	4392.736	7561.195
#2	5824.458	105183.6	12283.56	4373.904	7558.403
#3	5800.399	105022.9	12334.81	4385.063	7538.073

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2826-01 Acquired: 8/13/2025 16:35:45 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1395862	-.025763	2.398305	.0125785	.0065415	128.8316	1.566728
Stddev	.0013217	.001038	.001958	.0025831	.0020755	.1042	.006777
%RSD	.9468532	4.030769	.0816551	20.53567	31.72774	.0808437	.4325325
#1	.1388828	-.024771	2.396997	.0122183	.0041538	128.7202	1.560332
#2	.1387650	-.026843	2.397360	.0101944	.0075569	128.8481	1.573830
#3	.1411108	-.025676	2.400556	.0153227	.0079137	128.9265	1.566022

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0092384	.0403645	369.4236	2.532358	.1223023	3.432179	280.8822
Stddev	.0000987	.0000262	.9183	.005853	.0003617	.002533	.4802
%RSD	1.068560	.0648431	.2485788	.2311251	.2957257	.0738115	.1709635
#1	.0091935	.0403940	370.0573	2.531227	.1219632	3.433751	280.3289
#2	.0091702	.0403439	369.8429	2.527153	.1226830	3.429256	281.1271
#3	.0093516	.0403558	368.3705	2.538693	.1222607	3.433529	281.1905

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	7.039300	96.94582	6.744927	.0125275	6.570137	.3384918	3.255711
Stddev	.026798	.26020	.003762	.0001446	.023254	.0022527	.000994
%RSD	.3806913	.2684012	.0557809	1.154358	.3539342	.6655139	.0305178
#1	7.011408	96.82999	6.741070	.0126914	6.544278	.3370258	3.255552
#2	7.064851	96.76365	6.748587	.0124730	6.589330	.3373638	3.254806
#3	7.041641	97.24383	6.745122	.0124180	6.576801	.3410856	3.256774

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	22.19676	.2881801	.0084255	.3160914	2.769079	7.537017	6.946138
Stddev	.04321	.0030465	.0001559	.0006798	.010314	.054433	.011706
%RSD	.1946549	1.057164	1.850810	.2150654	.3724642	.7222077	.1685291
#1	22.15471	.2861832	.0082486	.3155884	2.757269	7.516157	6.942225
#2	22.24104	.2916867	.0085428	.3158209	2.776314	7.496099	6.959300
#3	22.19453	.2866705	.0084853	.3168648	2.773652	7.598795	6.936890

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2826-01 Acquired: 8/13/2025 16:35:45 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**15.62017**	**.2315961**	**.8013124**
Stddev	.01530	.0004101	.0043260
%RSD	.0979717	.1770736	.5398634
#1	15.61340	.2311558	.7972390
#2	15.60942	.2319672	.8058531
#3	15.63769	.2316653	.8008452

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5706.696**	**104008.1**	**13626.67**	**4304.267**	**5804.035**
Stddev	8.034	139.7	7.30	10.523	3.066
%RSD	.1407891	.1342996	.0535347	.2444747	.0528283
#1	5699.817	103890.5	13627.18	4292.283	5801.266
#2	5715.526	104162.5	13633.70	4311.997	5807.330
#3	5704.743	103971.3	13619.13	4308.521	5803.508

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2827-01 Acquired: 8/13/2025 16:39:54 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1483785**	**-.074518**	**70.05285**	**.0051202**	**.0227446**	**73.14248**
Stddev	.0024244	.002086	.15429	.0029995	.0012414	.11638
%RSD	1.633961	2.798813	.2202465	58.58230	5.458095	.1591145
#1	.1510508	-.075387	70.11788	.0019906	.0237707	73.22952
#2	.1477646	-.076029	69.87669	.0053999	.0230983	73.18764
#3	.1463200	-.072139	70.16398	.0079701	.0213647	73.01029

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**7.019622**	**.0077333**	**.0335457**	**206.2075**	**2.806505**	**.1333898**
Stddev	.087332	.0000428	.0002606	.3819	.004113	.0005159
%RSD	1.244118	.5537602	.7767909	.1851800	.1465360	.3867898
#1	7.063897	.0077015	.0335720	206.2542	2.807204	.1337388
#2	7.075949	.0077820	.0332729	206.5638	2.802088	.1327972
#3	6.919020	.0077164	.0337921	205.8044	2.810223	.1336335

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.695565**	**485.6378**	**3.347125**	**26.43496**	**.5569049**	**.0737720**
Stddev	.002136	.9572	.013068	.01988	.0011532	.0003967
%RSD	.1259870	.1970914	.3904120	.0751850	.2070772	.5377141
#1	1.693762	486.5389	3.338999	26.43087	.5568312	.0741636
#2	1.697924	485.7416	3.362198	26.45656	.5557903	.0737818
#3	1.695008	484.6330	3.340177	26.41744	.5580932	.0733705

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**22.91282**	**.5266232**	**10.17829**	**7.509590**	**^ *****	**.2211758**
Stddev	.11791	.0029539	.05684	.041509	-----	.0001737
%RSD	.5146216	.5609228	.5584790	.5527472	-----	.0785169
#1	22.84096	.5245638	10.11667	7.490038	^ -----	.2209858
#2	23.04890	.5300077	10.18951	7.557264	^ -----	.2212153
#3	22.84859	.5252981	10.22868	7.481468	^ -----	.2213263

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2827-01 Acquired: 8/13/2025 16:39:54 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.752780	1.483108	3.499438	11.47636	F 90.96765	.0630121
Stddev	.006148	.004796	.010824	.03346	.44028	.0007146
%RSD	.3507803	.3233535	.3093095	.2915863	.4839953	1.134003
#1	1.756280	1.478163	3.511893	11.50771	91.19106	.0632498
#2	1.745681	1.487739	3.494113	11.44112	90.46045	.0635774
#3	1.756380	1.483422	3.492307	11.48025	91.25143	.0622089

Elem	Sr4077
Units	ppm
Avg	.4316986
Stddev	.0029183
%RSD	.6759958
#1	.4297735
#2	.4350563
#3	.4302661

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5336.988	98666.08	12923.86	4046.074	6115.859
Stddev	11.976	6.33	11.74	13.005	16.515
%RSD	.2243917	.0064155	.0908273	.3214215	.2700336
#1	5349.960	98663.96	12910.89	4061.006	6120.348
#2	5334.652	98673.20	12926.95	4037.222	6129.665
#3	5326.353	98661.08	12933.75	4039.996	6097.563

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2827-05 Acquired: 8/13/2025 16:44:11 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1044894	-.022758	8.629621	-.005811	.0215296	93.47949
Stddev	.0021559	.001633	.007328	.003909	.0015632	.02872
%RSD	2.063292	7.176200	.0849176	67.26907	7.260596	.0307256
#1	.1021916	-.023672	8.638079	-.008367	.0197732	93.45464
#2	.1048089	-.023730	8.625195	-.001311	.0227682	93.47289
#3	.1064678	-.020872	8.625588	-.007754	.0220474	93.51094

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	3.833491	.0088923	.0331300	414.0308	.2967711	.1204523
Stddev	.011486	.0001465	.0001906	3.3338	.0044883	.0003603
%RSD	.2996258	1.647250	.5754391	.8052033	1.512380	.2991599
#1	3.822894	.0088246	.0331239	416.7793	.2991695	.1205668
#2	3.831881	.0090604	.0329425	410.3224	.2995506	.1200486
#3	3.845696	.0087920	.0333236	414.9908	.2915931	.1207415

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.168222	295.5230	4.508764	62.94097	.5112842	.0219489
Stddev	.001116	3.4002	.008909	.00810	.0012810	.0002974
%RSD	.0954979	1.150585	.1975916	.0128678	.2505479	1.354777
#1	1.168335	297.3536	4.510427	62.93166	.5114810	.0222285
#2	1.167053	297.6158	4.499141	62.94640	.5099162	.0216365
#3	1.169276	291.5997	4.516724	62.94485	.5124555	.0219818

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	19.65561	.3852155	4.072211	30.64986	.2677562	.0091125
Stddev	.18589	.0012060	.058918	.27890	.0073364	.0001311
%RSD	.9457486	.3130666	1.446835	.9099443	2.739948	1.438322
#1	19.72236	.3841671	4.121845	30.79223	.2699303	.0091275
#2	19.79891	.3849458	4.087688	30.82883	.2737598	.0089746
#3	19.44556	.3865334	4.007099	30.32851	.2595785	.0092355

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2827-05 Acquired: 8/13/2025 16:44:11 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.2312864	6.011005	4.225126	6.646962	F 78.00988	.1138177
Stddev	.0011711	.015091	.074054	.004547	.21543	.0011660
%RSD	.5063224	.2510607	1.752706	.0684006	.2761556	1.024435
#1	.2299414	5.998876	4.259280	6.651283	78.25803	.1130308
#2	.2320800	6.006235	4.275939	6.647385	77.90083	.1132651
#3	.2318378	6.027906	4.140158	6.642219	77.87077	.1151573

Elem	Sr4077
Units	ppm
Avg	1.741626
Stddev	.029660
%RSD	1.703006
#1	1.732582
#2	1.717540
#3	1.774755

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5582.559	102636.8	13253.48	4268.724	5788.391
Stddev	7.768	1201.0	18.65	45.783	5.664
%RSD	.1391463	1.170112	.1407516	1.072521	.0978507
#1	5590.907	102057.7	13271.64	4230.568	5794.514
#2	5575.543	101835.2	13254.43	4256.112	5787.320
#3	5581.228	104017.6	13234.37	4319.490	5783.339

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2829-01 Acquired: 8/13/2025 16:48:29 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0146157	.0127385	.0165349	.0055024	.0057026	55.68469	.0761651
Stddev	.0012553	.0018284	.0007517	.0025820	.0013963	.02899	.0002045
%RSD	8.588339	14.35333	4.546366	46.92525	24.48575	.0520578	.2684381
#1	.0133578	.0148452	.0164574	.0027669	.0059455	55.65340	.0763976
#2	.0158683	.0118054	.0158249	.0078971	.0042008	55.71062	.0760844
#3	.0146210	.0115650	.0173224	.0058430	.0069616	55.69005	.0760133

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0013009	.0010851	81.50139	.1776811	.0272492	.1724574	25.66638
Stddev	.0000699	.0000220	.11623	.0004306	.0000787	.0001102	.10759
%RSD	5.373722	2.026834	.1426071	.2423640	.2889551	.0638851	.4191708
#1	.0012394	.0011012	81.36718	.1773325	.0272602	.1725013	25.56025
#2	.0012863	.0010600	81.56877	.1775482	.0271656	.1725388	25.66353
#3	.0013769	.0010941	81.56822	.1781625	.0273219	.1723320	25.77537

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.4952632	16.34188	.1046040	.0026768	4.287983	.1538768	.2256946
Stddev	.0024369	.03958	.0001021	.0001853	.030102	.0032141	.0007184
%RSD	.4920426	.2422069	.0975707	6.923747	.7020036	2.088722	.3182873
#1	.4926014	16.30231	.1045348	.0028507	4.257639	.1563915	.2264066
#2	.4958037	16.38148	.1047212	.0024818	4.288473	.1549831	.2249700
#3	.4973844	16.34185	.1045560	.0026979	4.317837	.1502557	.2257072

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.5862546	.1866260	.0247643	.0163739	4.455656	5.020977	.2891648
Stddev	.0298464	.0004926	.0002656	.0002707	.012928	.020250	.0025540
%RSD	5.091032	.2639336	1.072437	1.653148	.2901502	.4032991	.8832270
#1	.5889382	.1861484	.0248499	.0160660	4.440747	5.029726	.2870665
#2	.5551570	.1871323	.0244665	.0165743	4.462461	4.997824	.2884194
#3	.6146686	.1865973	.0249766	.0164813	4.463760	5.035381	.2920086

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2829-01 Acquired: 8/13/2025 16:48:29 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	1.108610	.0162724	.1966764
Stddev	.007751	.0001082	.0006815
%RSD	.6991807	.6645950	.3465180
#1	1.099706	.0161786	.1959911
#2	1.112272	.0163907	.1966841
#3	1.113851	.0162478	.1973541

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5611.318	100878.3	12124.14	4274.155	6695.629
Stddev	7.889	190.4	11.32	2.471	14.487
%RSD	.1405990	.1887140	.0933311	.0578234	.2163607
#1	5602.896	101025.5	12136.11	4271.928	6692.235
#2	5612.521	100946.2	12122.68	4276.814	6683.140
#3	5618.537	100663.3	12113.62	4273.723	6711.511

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-09 Acquired: 8/13/2025 16:52:38 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.5068867	-.123210	11.62454	.0582227	.1547743	70.67201
Stddev	.0018936	.004310	.02015	.0037146	.0010104	.19003
%RSD	.3735730	3.498242	.1733766	6.380026	.6528439	.2688954
#1	.5083855	-.121075	11.63033	.0539487	.1536767	70.46385
#2	.5075160	-.120383	11.60213	.0606730	.1556657	70.83620
#3	.5047586	-.128171	11.64117	.0600465	.1549806	70.71599

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	4.635758	.0176286	.0365074	93.91628	.8634016	.1572956
Stddev	.006603	.0001553	.0003493	.01233	.0002484	.0008083
%RSD	.1424317	.8811444	.9568729	.0131321	.0287739	.5138664
#1	4.643357	.0174939	.0362677	93.91809	.8636221	.1581316
#2	4.631428	.0177985	.0369082	93.92762	.8631325	.1565182
#3	4.632488	.0175935	.0363463	93.90315	.8634502	.1572371

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	4.405984	753.9727	3.460085	22.03555	.5543822	.0253858
Stddev	.003754	3.9375	.006010	.02107	.0006017	.0016877
%RSD	.0852065	.5222312	.1736878	.0956221	.1085350	6.648293
#1	4.407393	753.4117	3.466763	22.01132	.5550317	.0250086
#2	4.408831	750.3458	3.455113	22.04956	.5538437	.0239186
#3	4.401730	758.1606	3.458378	22.04576	.5542712	.0272302

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	3.459966	.6084114	2.516628	10.23195	^ *****	.0391904
Stddev	.010184	.0016812	.003120	.01244	-----	.0003148
%RSD	.2943355	.2763205	.1239663	.1215563	-----	.8032400
#1	3.467697	.6066000	2.514874	10.24186	^ -----	.0389056
#2	3.448427	.6099217	2.514780	10.21800	^ -----	.0395284
#3	3.463773	.6087126	2.520230	10.23600	^ -----	.0391372

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-09 Acquired: 8/13/2025 16:52:38 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.8749492	4.626747	F 23.56135	7.183246	16.32204	.1249249
Stddev	.0016167	.006269	.07363	.013380	.05022	.0011741
%RSD	.1847788	.1354896	.3124889	.1862660	.3077005	.9398070
#1	.8734651	4.630581	23.48792	7.182433	16.29213	.1245564
#2	.8747105	4.619512	23.56095	7.170291	16.29398	.1262390
#3	.8766720	4.630146	23.63517	7.197014	16.38003	.1239793

Elem	Sr4077
Units	ppm
Avg	.0758773
Stddev	.0043948
%RSD	5.791960
#1	.0781212
#2	.0786971
#3	.0708135

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5616.504	102408.8	12765.54	4234.089	6008.079
Stddev	3.720	102.0	34.41	14.181	10.395
%RSD	.0662392	.0996101	.2695882	.3349180	.1730158
#1	5612.211	102291.0	12752.64	4223.042	5999.857
#2	5618.779	102470.6	12739.43	4250.080	6019.763
#3	5618.522	102464.7	12804.54	4229.146	6004.617

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-09DUP Acquired: 8/13/2025 16:56:47 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.5170091	-.128514	11.75216	.0551547	.1570451	69.87703
Stddev	.0036280	.003279	.00789	.0018148	.0007550	.36954
%RSD	.7017233	2.551845	.0671467	3.290404	.4807302	.5288456
#1	.5166836	-.132221	11.74346	.0531075	.1567746	70.30331
#2	.5135548	-.125991	11.75885	.0557905	.1564626	69.64723
#3	.5207888	-.127331	11.75416	.0565660	.1578980	69.68056

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	4.643892	.0177295	.0387108	93.37976	.8672155	.1568746
Stddev	.021892	.0001613	.0006854	.44004	.0011664	.0004926
%RSD	.4714078	.9099266	1.770540	.4712416	.1345000	.3140031
#1	4.668391	.0178801	.0394854	93.88787	.8673944	.1564528
#2	4.626247	.0175593	.0381828	93.12365	.8659700	.1567549
#3	4.637039	.0177493	.0384644	93.12775	.8682821	.1574159

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	4.474758	749.9925	3.421493	21.66305	.5564017	.0232006
Stddev	.010289	1.1504	.013257	.14470	.0004962	.0008693
%RSD	.2299280	.1533893	.3874495	.6679395	.0891768	3.746688
#1	4.469558	749.7730	3.436386	21.82569	.5563485	.0223222
#2	4.468107	751.2368	3.410981	21.54859	.5559342	.0240604
#3	4.486609	748.9676	3.417114	21.61487	.5569223	.0232192

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	3.266850	.6103996	2.497082	10.05460	^ *****	.0407402
Stddev	.011192	.0047355	.001938	.04054	-----	.0002222
%RSD	.3426017	.7757958	.0775925	.4032215	-----	.5453071
#1	3.267642	.6150097	2.498009	10.08697	^ -----	.0408028
#2	3.255282	.6055480	2.494855	10.00913	^ -----	.0409243
#3	3.277625	.6106411	2.498381	10.06771	^ -----	.0404934

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-09DUP Acquired: 8/13/2025 16:56:47 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.8943350	4.565187	F 24.17116	7.043078	16.59765	.1246956
Stddev	.0018740	.024225	.10345	.005152	.02265	.0013964
%RSD	.2095384	.5306367	.4279741	.0731565	.1364949	1.119867
#1	.8923032	4.592525	24.25834	7.039131	16.59435	.1231022
#2	.8947062	4.546389	24.05685	7.041196	16.57683	.1257065
#3	.8959956	4.556648	24.19828	7.048907	16.62178	.1252780

Elem	Sr4077
Units	ppm
Avg	.0765440
Stddev	.0045816
%RSD	5.985578
#1	.0813431
#2	.0722163
#3	.0760725

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5605.268	102739.1	12684.22	4233.567	6008.971
Stddev	10.837	48.5	5.99	5.389	6.048
%RSD	.1933422	.0471929	.0472436	.1272853	.1006416
#1	5599.313	102716.4	12678.80	4228.018	6004.667
#2	5617.777	102706.0	12690.66	4238.779	6015.886
#3	5598.714	102794.7	12683.20	4233.903	6006.362

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-09LX5 Acquired: 8/13/2025 17:00:55 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1114476	-.027335	2.440055	.0130022	.0319116	15.68400	1.009649
Stddev	.0029274	.000841	.006831	.0009162	.0015833	.05311	.003096
%RSD	2.626740	3.077920	.2799695	7.046108	4.961421	.3386032	.3066798
#1	.1117946	-.027687	2.447355	.0119811	.0327431	15.65265	1.006260
#2	.1141861	-.026375	2.438994	.0132732	.0329059	15.65402	1.010358
#3	.1083622	-.027943	2.433816	.0137521	.0300858	15.74531	1.012330

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0042110	.0057110	21.52531	.1978020	.0310388	1.034941	169.5526
Stddev	.0000838	.0001279	.08996	.0003938	.0001228	.003069	1.4147
%RSD	1.989236	2.240219	.4179242	.1990693	.3954568	.2965717	.8343679
#1	.0042357	.0055658	21.44552	.1977702	.0310677	1.037541	170.4346
#2	.0041176	.0057603	21.50761	.1982106	.0311445	1.035727	170.3022
#3	.0042796	.0058070	21.62281	.1974250	.0309042	1.031555	167.9208

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.7964026	5.018822	.1126685	.0063075	.8139610	.1380775	.5571640
Stddev	.0036154	.041165	.0008010	.0005217	.0155468	.0023672	.0048166
%RSD	.4539661	.8202128	.7109300	8.270391	1.910022	1.714368	.8644841
#1	.7939115	4.976818	.1135876	.0057634	.8082590	.1356848	.5608768
#2	.7947469	5.020554	.1121195	.0068034	.8315537	.1381293	.5588939
#3	.8005493	5.059094	.1122983	.0063558	.8020702	.1404183	.5517214

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	2.306839	.0302560	.0085856	.1817699	1.038072	5.092558	1.396058
Stddev	.049691	.0074943	.0001600	.0009990	.003361	.053783	.001985
%RSD	2.154087	24.76964	1.863019	.5495862	.3238156	1.056115	.1422033
#1	2.270306	.0366018	.0087703	.1808512	1.035580	5.139319	1.393859
#2	2.363423	.0321784	.0084935	.1828334	1.036741	5.104572	1.397720
#3	2.286788	.0219877	.0084930	.1816251	1.041895	5.033784	1.396594

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-09LX5 Acquired: 8/13/2025 17:00:55 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	3.269684	.0261822	.0125809
Stddev	.000464	.0002129	.0017660
%RSD	.0141975	.8132413	14.03697
#1	3.269280	.0264212	.0117673
#2	3.270191	.0260128	.0113683
#3	3.269581	.0261126	.0146070

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5684.676	102977.7	11933.45	4309.641	6969.526
Stddev	8.228	781.0	73.80	39.468	12.280
%RSD	.1447329	.7584010	.6184355	.9158176	.1761996
#1	5680.160	102495.0	11989.20	4270.941	6960.451
#2	5679.694	102559.3	11961.40	4308.146	6964.629
#3	5694.172	103878.7	11849.76	4349.836	6983.500

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCV06 Acquired: 8/13/2025 17:05:00 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV06 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**5.042629**	**5.384016**	**4.847802**	**5.154760**	**5.108953**	**9.693028**	**9.807771**
Stddev	.009903	.165242	.005493	.024028	.023208	.018718	.106625
%RSD	.1963763	3.069125	.1133074	.4661249	.4542676	.1931088	1.087153
#1	5.033789	5.211400	4.847917	5.130167	5.095204	9.672500	9.898930
#2	5.040768	5.399913	4.842252	5.155935	5.095906	9.697434	9.690522
#3	5.053330	5.540736	4.853236	5.178179	5.135748	9.709150	9.833861

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2383448**	**2.423771**	**24.14399**	**.9884967**	**2.431373**	**1.246068**	**5.073739**
Stddev	.0004951	.002177	.07436	.0013958	.000418	.001468	.012611
%RSD	.2077171	.0898264	.3079790	.1412014	.0171762	.1177902	.2485569
#1	.2377735	2.424343	24.05915	.9890562	2.430893	1.246819	5.078277
#2	.2386125	2.425604	24.19784	.9869079	2.431652	1.244377	5.083452
#3	.2386485	2.421364	24.17499	.9895259	2.431574	1.247009	5.059486

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.399823**	**23.99836**	**2.442756**	**1.255255**	**23.30557**	**2.463389**	**2.518384**
Stddev	.012198	.09311	.000861	.001595	.07036	.006836	.003316
%RSD	.5082716	.3879913	.0352654	.1270445	.3018964	.2775143	.1316689
#1	2.392336	23.93731	2.442593	1.256196	23.36529	2.457657	2.515876
#2	2.413898	24.10553	2.443688	1.253414	23.32340	2.470955	2.522144
#3	2.393235	23.95225	2.441988	1.256155	23.22801	2.461554	2.517133

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**24.75928**	**4.793299**	**5.103397**	**4.897045**	**4.861095**	**5.150910**	**4.658086**
Stddev	.02302	.019725	.013266	.013081	.016477	.009585	.007983
%RSD	.0929938	.4115131	.2599417	.2671207	.3389532	.1860930	.1713830
#1	24.75838	4.772592	5.091379	4.881940	4.847622	5.158183	4.654073
#2	24.78274	4.795438	5.101181	4.904642	4.879465	5.140048	4.652906
#3	24.73671	4.811868	5.117632	4.904552	4.856197	5.154500	4.667280

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCV06 Acquired: 8/13/2025 17:05:00 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV06 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**4.860561**	**4.517107**	**4.895347**
Stddev	.015662	.012428	.025120
%RSD	.3222213	.2751224	.5131321
#1	4.842927	4.503038	4.903653
#2	4.872854	4.526589	4.867126
#3	4.865901	4.521693	4.915260

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5538.876**	**98999.53**	**11604.15**	**4150.276**	**6719.929**
Stddev	9.399	118.63	36.93	10.899	6.276
%RSD	.1697001	.1198249	.3182613	.2626190	.0933876
#1	5539.659	98959.38	11644.58	4144.536	6721.042
#2	5547.860	98906.18	11572.19	4143.445	6725.573
#3	5529.110	99133.02	11595.68	4162.846	6713.171

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCB06 Acquired: 8/13/2025 17:09:12 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB06 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0012380	-.001320	-.000696	-.001443	-.001238	-.020077	-.010116
Stddev	.0007930	.002063	.000280	.001705	.002737	.005241	.001104
%RSD	64.05364	156.3131	40.18541	118.1670	221.0879	26.10447	10.91592
#1	.0021497	-.003419	-.000490	-.000281	-.001795	-.021868	-.009124
#2	.0007091	.000705	-.001014	-.003400	-.003654	-.024188	-.009918
#3	.0008551	-.001244	-.000583	-.000648	.001735	-.014176	-.011306

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0000043	-.000133	-.019127	.0001300	-.000183	.0007722	-.005072
Stddev	.0000301	.000018	.002838	.0001797	.000111	.0000879	.003047
%RSD	699.5130	13.53380	14.83772	138.1861	60.33841	11.38265	60.08081
#1	.0000119	-.000153	-.022371	.0000773	-.000196	.0007548	-.005315
#2	-.000029	-.000120	-.017105	.0003302	-.000067	.0006943	-.007991
#3	.000030	-.000125	-.017904	-.000017	-.000287	.0008675	-.001911

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.000044	-.000266	.0004091	.0004794	.1379824	-.001081	-.000051
Stddev	.000432	.010715	.0002488	.0002574	.0116954	.001203	.000030
%RSD	975.4545	4030.787	60.81928	53.69104	8.476017	111.2646	58.50475
#1	-.000529	-.012628	.0001461	.0007473	.1369927	.000152	-.000020
#2	.000299	.005476	.0006408	.0004570	.1268133	-.002251	-.000080
#3	.000098	.006355	.0004405	.0002339	.1501412	-.001145	-.000052

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.3089144	.0037515	-.000142	-.000066	-.000535	.0077620	.0026199
Stddev	.0070065	.0004349	.000137	.000137	.000499	.0085442	.0030560
%RSD	2.268107	11.59199	96.22786	207.4738	93.19977	110.0776	116.6448
#1	.3127437	.0042331	-.000128	-.000022	-.000821	.0174953	.0010956
#2	.3131718	.0033876	-.000013	-.000220	-.000825	.0042918	.0006258
#3	.3008278	.0036336	-.000286	.000043	.000041	.0014988	.0061382

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCB06 Acquired: 8/13/2025 17:09:12 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB06 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0056333	-.001380	-.000339
Stddev	.0033433	.001669	.000074
%RSD	59.34913	120.8774	21.89610
#1	.0023516	-.000317	-.000254
#2	.0055134	-.000521	-.000389
#3	.0090349	-.003304	-.000375

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5737.472	103539.6	11562.50	4286.944	7564.012
Stddev	6.663	214.1	57.33	17.884	7.887
%RSD	.1161383	.2067639	.4958422	.4171671	.1042707
#1	5741.045	103333.1	11554.88	4272.072	7572.388
#2	5741.587	103525.1	11623.27	4281.971	7562.921
#3	5729.784	103760.5	11509.37	4306.787	7556.727

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-09MS Acquired: 8/13/2025 17:13:32 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.309558**	**1.656942**	**10.79502**	**1.892510**	**.8399824**	**61.98874**
Stddev	.009206	.007912	.08300	.012958	.0063985	.08700
%RSD	.7030122	.4775208	.7688886	.6847247	.7617442	.1403428
#1	1.314431	1.655136	10.82684	1.890284	.8422627	61.90232
#2	1.315303	1.665601	10.85740	1.906437	.8449285	62.07630
#3	1.298939	1.650089	10.70081	1.880808	.8327561	61.98761

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**4.564405**	**.1844582**	**.2323631**	**72.00912**	**.9885516**	**.3387792**
Stddev	.001956	.0003108	.0028258	.07622	.0032993	.0023966
%RSD	.0428640	.1685178	1.216109	.1058510	.3337553	.7074331
#1	4.562912	.1845214	.2335644	71.93269	.9847432	.3403132
#2	4.566620	.1847326	.2343898	72.08514	.9905453	.3400069
#3	4.563682	.1841206	.2291352	72.00954	.9903662	.3360175

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**4.960816**	**811.3024**	**3.482315**	**19.78485**	**.9894247**	**.1042820**
Stddev	.036872	6.9841	.007853	.12924	.0063946	.0040601
%RSD	.7432609	.8608495	.2255197	.6532181	.6462959	3.893347
#1	4.979495	809.1432	3.473320	19.67675	.9940100	.1025386
#2	4.984610	805.6530	3.487807	19.92800	.9921442	.1013848
#3	4.918342	819.1112	3.485819	19.74981	.9821198	.1089226

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**5.759301**	**.7841703**	**2.363669**	**19.61342**	^ *****	**.4233106**
Stddev	.033462	.0014063	.004636	.10988	-----	.0023500
%RSD	.5810167	.1793389	.1961529	.5602447	-----	.5551574
#1	5.735236	.7828704	2.365478	19.48670	^ -----	.4250544
#2	5.745153	.7856631	2.358401	19.67128	^ -----	.4242393
#3	5.797513	.7839774	2.367128	19.68229	^ -----	.4206382

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-09MS Acquired: 8/13/2025 17:13:32 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.689791	3.632354	F 25.90019	12.23448	16.93743	.3037707
Stddev	.013491	.008107	.07643	.07837	.12675	.0019426
%RSD	.7983776	.2231756	.2951040	.6405826	.7483320	.6394825
#1	1.698030	3.627734	25.81855	12.26476	17.01400	.3019185
#2	1.697120	3.641715	25.91197	12.29319	17.00715	.3036012
#3	1.674221	3.627614	25.97005	12.14549	16.79112	.3057925

Elem	Sr4077
Units	ppm
Avg	.0558917
Stddev	.0073372
%RSD	13.12753
#1	.0579323
#2	.0619927
#3	.0477503

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5498.562	100468.3	12559.47	4098.853	6001.027
Stddev	28.649	297.2	19.91	4.672	42.164
%RSD	.5210309	.2958354	.1584946	.1139868	.7026181
#1	5478.306	100809.8	12575.88	4102.800	5971.109
#2	5486.039	100326.5	12537.32	4100.064	5982.721
#3	5531.341	100268.4	12565.20	4093.694	6049.250

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-09MSD Acquired: 8/13/2025 17:17:35 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.294557**	**1.615152**	**10.98460**	**1.887040**	**.8324383**	**62.18777**
Stddev	.003662	.055660	.01741	.008558	.0031502	.65500
%RSD	.2828688	3.446112	.1584887	.4535308	.3784326	1.053260
#1	1.292277	1.615802	10.98301	1.884110	.8358301	61.43783
#2	1.292612	1.559170	10.96804	1.880332	.8318807	62.47782
#3	1.298781	1.670484	11.00275	1.896679	.8296041	62.64767

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**4.382495**	**.1836054**	**.2323633**	**73.99576**	**1.136670**	**.3527649**
Stddev	.049395	.0015159	.0008976	.78465	.002919	.0003648
%RSD	1.127087	.8256081	.3862691	1.060404	.2568189	.1034222
#1	4.326861	.1826156	.2313685	73.14703	1.133320	.3531475
#2	4.399426	.1828501	.2326089	74.14547	1.138024	.3524209
#3	4.421197	.1853505	.2331124	74.69477	1.138667	.3527262

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**4.873529**	**744.0530**	**4.656637**	**22.28705**	**1.041241**	**.0983552**
Stddev	.008628	9.4675	.053335	.25818	.000752	.0037941
%RSD	.1770476	1.272417	1.145349	1.158409	.0722151	3.857538
#1	4.867737	754.8364	4.597304	22.02182	1.040757	.1025373
#2	4.869404	737.1051	4.672012	22.30180	1.040859	.0951339
#3	4.883446	740.2176	4.700596	22.53753	1.042107	.0973944

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**5.546545**	**.8733662**	**2.310044**	**19.62445**	**^ *******	**.4265446**
Stddev	.029133	.0062565	.006141	.11273	-----	.0004053
%RSD	.5252375	.7163649	.2658232	.5744617	-----	.0950223
#1	5.513710	.8674169	2.303445	19.49667	^ -----	.4267091
#2	5.569295	.8727913	2.311097	19.70985	^ -----	.4260829
#3	5.556630	.8798902	2.315590	19.66683	^ -----	.4268417

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-09MSD Acquired: 8/13/2025 17:17:35 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.510096	4.538217	F 22.69122	12.24197	16.15476	.2934906
Stddev	.003566	.050692	.12684	.03448	.05947	.0025572
%RSD	.2361131	1.117004	.5589766	.2816146	.3681109	.8713035
#1	1.511652	4.481235	22.55769	12.28141	16.12926	.2925118
#2	1.506017	4.555112	22.81009	12.22692	16.11229	.2915673
#3	1.512618	4.578304	22.70588	12.21758	16.22272	.2963926

Elem	Sr4077
Units	ppm
Avg	.1099942
Stddev	.0179734
%RSD	16.34028
#1	.0892610
#2	.1195603
#3	.1211615

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5494.737	100713.6	12586.10	4111.181	6004.036
Stddev	2.161	333.3	121.41	24.717	3.159
%RSD	.0393373	.3309117	.9646662	.6012175	.0526181
#1	5492.968	101089.8	12699.15	4139.721	6005.216
#2	5497.147	100455.1	12601.38	4097.088	6000.457
#3	5494.097	100596.0	12457.77	4096.734	6006.435

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-09A Acquired: 8/13/2025 17:21:38 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.238164**	**1.757962**	**12.54815**	**1.843375**	**.7097440**	**70.67279**
Stddev	.118726	.317580	.13202	.281859	.1250662	1.16967
%RSD	9.588897	18.06523	1.052071	15.29040	17.62131	1.655055
#1	1.118491	1.449250	12.41121	1.557728	.5848143	71.76194
#2	1.240081	1.740913	12.55861	1.851107	.7094713	69.43651
#3	1.355920	2.083722	12.67462	2.121288	.8349462	70.81992

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**4.632471**	**.1218876**	**.2327979**	**92.82427**	**1.098159**	**.3583926**
Stddev	.066665	.0096261	.0301571	1.38635	.028184	.0319461
%RSD	1.439077	7.897515	12.95420	1.493522	2.566505	8.913711
#1	4.690819	.1136236	.2020795	94.20972	1.069746	.3255281
#2	4.559813	.1195821	.2339538	91.43701	1.098624	.3603164
#3	4.646781	.1324571	.2623604	92.82609	1.126109	.3893333

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**4.694833**	**746.9136**	**3.432164**	**22.21839**	**1.053565**	**.0661150**
Stddev	.035250	7.0967	.046512	.35186	.076993	.0057721
%RSD	.7508231	.9501354	1.355184	1.583652	7.307897	8.730461
#1	4.662460	740.4359	3.471186	22.45421	.974046	.0594640
#2	4.689651	754.4990	3.380695	21.81395	1.058893	.0690655
#3	4.732387	745.8059	3.444612	22.38701	1.127756	.0698155

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**4.508976**	**.7152030**	**2.599404**	**14.58347**	**^ *****	**.3215056**
Stddev	.211732	.0245520	.014703	.75933	-----	.0643404
%RSD	4.695800	3.432875	.5656350	5.206809	-----	20.01223
#1	4.324269	.7103557	2.582565	13.95002	^ -----	.2568049
#2	4.462609	.6934362	2.605954	14.37518	^ -----	.3222323
#3	4.740049	.7418172	2.609695	15.42521	^ -----	.3854796

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-09A Acquired: 8/13/2025 17:21:38 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.372272	4.601499	F 23.90945	12.51223	16.43532	.1980737
Stddev	.114286	.072106	.27155	.83359	.00747	.0151214
%RSD	8.328216	1.567008	1.135724	6.662195	.0454517	7.634209
#1	1.256868	4.672512	23.81763	11.65716	16.44369	.1887128
#2	1.374543	4.528348	23.69571	12.55698	16.42935	.1899894
#3	1.485406	4.603637	24.21500	13.32254	16.43291	.2155188

Elem	Sr4077
Units	ppm
Avg	.1424456
Stddev	.0235271
%RSD	16.51654
#1	.1497213
#2	.1161401
#3	.1614753

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5620.810	102601.2	12830.37	4220.805	6022.762
Stddev	9.533	496.0	197.68	27.935	11.370
%RSD	.1696006	.4834691	1.540690	.6618498	.1887791
#1	5631.690	103172.4	12630.07	4252.952	6032.775
#2	5613.923	102279.2	13025.31	4202.429	6010.402
#3	5616.817	102351.9	12835.72	4207.033	6025.110

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-03 Acquired: 8/13/2025 17:25:41 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1626718**	**-.044950**	**2.901331**	**.0113233**	**.0076799**	**107.0331**
Stddev	.0009131	.002201	.002485	.0011094	.0011278	.1487
%RSD	.5613091	4.896287	.0856650	9.797345	14.68514	.1389337
#1	.1620674	-.043398	2.901952	.0107605	.0068426	106.9096
#2	.1622257	-.047469	2.898594	.0126013	.0089624	107.1982
#3	.1637222	-.043984	2.903447	.0106082	.0072348	106.9915

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.342180**	**.0100555**	**.0215490**	**91.33607**	**.3554494**	**.1640387**
Stddev	.005733	.0001983	.0002829	.29556	.0009108	.0003430
%RSD	.4271696	1.971818	1.312865	.3235961	.2562466	.2090865
#1	1.336715	.0098752	.0214191	91.16521	.3546814	.1638278
#2	1.348149	.0102679	.0213544	91.67736	.3564557	.1638538
#3	1.341678	.0100234	.0218735	91.16565	.3552112	.1644344

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.415705**	**374.4046**	**4.642225**	**68.82118**	**.4306590**	**.0152839**
Stddev	.002539	.5206	.025842	.17396	.0004056	.0002141
%RSD	.1793361	.1390431	.5566734	.2527693	.0941801	1.400689
#1	1.413366	374.1854	4.630379	68.71613	.4302017	.0151453
#2	1.415344	374.9990	4.671867	69.02198	.4307998	.0155305
#3	1.418405	374.0296	4.624430	68.72543	.4309754	.0151760

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**3.105826**	**.4579401**	**2.394720**	**16.90506**	**.1801404**	**.0186282**
Stddev	.020034	.0040814	.006937	.02943	.0040195	.0002939
%RSD	.6450465	.8912558	.2896717	.1741122	2.231325	1.577554
#1	3.086806	.4605128	2.386730	16.90812	.1840332	.0186956
#2	3.126739	.4600734	2.398224	16.87422	.1760052	.0183065
#3	3.103933	.4532341	2.399207	16.93285	.1803828	.0188825

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-03 Acquired: 8/13/2025 17:25:41 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.2469871	4.236024	F 19.24141	7.884509	6.686284	.2050392
Stddev	.0005437	.015620	.01179	.016669	.019307	.0018937
%RSD	.2201162	.3687427	.0612868	.2114176	.2887548	.9235750
#1	.2469709	4.223382	19.24755	7.867977	6.664633	.2037217
#2	.2464518	4.253486	19.22781	7.884235	6.701712	.2072093
#3	.2475387	4.231205	19.24887	7.901313	6.692506	.2041866

Elem	Sr4077
Units	ppm
Avg	.1316666
Stddev	.0011971
%RSD	.9091903
#1	.1307562
#2	.1330226
#3	.1312210

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5826.036	106127.5	13161.55	4380.642	6080.832
Stddev	1.998	173.7	79.87	6.712	3.922
%RSD	.0342880	.1636373	.6068242	.1532118	.0645058
#1	5828.302	106249.0	13196.96	4387.789	6079.315
#2	5824.531	105928.6	13070.10	4379.665	6077.894
#3	5825.275	106204.9	13217.59	4374.473	6085.287

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-05 Acquired: 8/13/2025 17:29:44 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.2126858	-.057857	4.282957	.0139451	.0335319	121.3003
Stddev	.0019287	.001749	.013107	.0017553	.0031425	.3202
%RSD	.9068265	3.023507	.3060271	12.58732	9.371671	.2639832
#1	.2137682	-.058919	4.272536	.0159577	.0318303	120.9475
#2	.2104591	-.058813	4.278661	.0127306	.0316071	121.3807
#3	.2138302	-.055838	4.297672	.0131471	.0371582	121.5726

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.590818	.0130006	.0380501	110.6571	.5647292	.2828633
Stddev	.005156	.0001815	.0002907	.4541	.0018050	.0004283
%RSD	.3241147	1.396267	.7639067	.4103734	.3196276	.1514286
#1	1.587617	.0128747	.0377186	110.1767	.5640809	.2824898
#2	1.588071	.0129184	.0382611	110.7151	.5667689	.2827692
#3	1.596766	.0132087	.0381707	111.0794	.5633379	.2833308

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	2.691837	460.9497	5.820541	59.12875	.9273159	.0196940
Stddev	.005993	3.9806	.024032	.35432	.0007433	.0020895
%RSD	.2226420	.8635619	.4128816	.5992302	.0801554	10.60966
#1	2.687146	465.3190	5.799810	58.74183	.9264609	.0220597
#2	2.689776	457.5292	5.814931	59.20704	.9278088	.0181007
#3	2.698588	460.0011	5.846881	59.43737	.9276779	.0189216

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	3.517837	.6174950	4.542601	16.84459	^ *****	.0357038
Stddev	.020786	.0030618	.008406	.12361	-----	.0000305
%RSD	.5908689	.4958465	.1850524	.7338042	-----	.0855315
#1	3.541401	.6158658	4.534028	16.98531	^ -----	.0356687
#2	3.502102	.6155922	4.542945	16.79484	^ -----	.0357189
#3	3.510010	.6210269	4.550830	16.75360	^ -----	.0357239

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-05 Acquired: 8/13/2025 17:29:44 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.3751083	4.722835	F 16.55943	9.668054	7.132989	.1928392
Stddev	.0008599	.010028	.06524	.034766	.026547	.0006657
%RSD	.2292346	.2123258	.3939742	.3595925	.3721711	.3452232
#1	.3746766	4.714886	16.63477	9.628463	7.114119	.1928131
#2	.3745498	4.719517	16.52225	9.693602	7.121502	.1921869
#3	.3760985	4.734101	16.52129	9.682096	7.163345	.1935176

Elem	Sr4077
Units	ppm
Avg	.2481124
Stddev	.0045657
%RSD	1.840181
#1	.2428407
#2	.2508022
#3	.2506943

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5823.194	105927.7	13328.79	4359.507	5932.021
Stddev	11.052	238.7	89.02	8.220	12.536
%RSD	.1897862	.2253476	.6678530	.1885546	.2113249
#1	5834.597	105724.1	13430.54	4367.349	5946.405
#2	5822.456	105868.6	13290.55	4360.216	5926.230
#3	5812.531	106190.4	13265.29	4350.955	5923.427

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-07 Acquired: 8/13/2025 17:33:52 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.2874679	-.074721	5.144197	.0215084	.0524381	124.3803
Stddev	.0020538	.002733	.007669	.0030139	.0031983	1.4720
%RSD	.7144624	3.658272	.1490853	14.01263	6.099166	1.183478
#1	.2895074	-.072481	5.135585	.0243156	.0487922	125.1856
#2	.2854000	-.077767	5.150291	.0218863	.0547708	125.2740
#3	.2874963	-.073915	5.146714	.0183235	.0537512	122.6813

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	2.813869	.0184552	.0544729	113.3056	2.632413	.6104101
Stddev	.040298	.0003123	.0003477	1.5592	.007636	.0002514
%RSD	1.432127	1.692026	.6382394	1.376103	.2900609	.0411815
#1	2.834766	.0185176	.0541860	113.9965	2.641016	.6101422
#2	2.839427	.0187316	.0543731	114.4000	2.629786	.6104473
#3	2.767415	.0181164	.0548595	111.5203	2.626438	.6106408

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	4.832450	607.0410	6.708830	75.61339	3.130719	.0245451
Stddev	.010228	2.2118	.102984	.86820	.002805	.0008809
%RSD	.2116464	.3643657	1.535057	1.148210	.0895824	3.588910
#1	4.828448	608.2188	6.756525	75.98023	3.127686	.0244057
#2	4.844074	604.4895	6.779321	76.23796	3.133219	.0237423
#3	4.824829	608.4147	6.590645	74.62198	3.131253	.0254874

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	6.064338	1.176196	6.873469	19.20966	^ *****	.0838391
Stddev	.043165	.013767	.017816	.11828	-----	.0005608
%RSD	.7117840	1.170493	.2591924	.6157240	-----	.6688854
#1	6.111612	1.184914	6.893858	19.34523	^ -----	.0842109
#2	6.027023	1.183350	6.860910	19.15620	^ -----	.0841124
#3	6.054380	1.160325	6.865638	19.12755	^ -----	.0831941

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: Q2832-07 Acquired: 8/13/2025 17:33:52 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.6535547	5.980799	F 35.79728	10.40630	8.317878	.2265325
Stddev	.0022558	.076062	.28656	.03251	.043347	.0016299
%RSD	.3451592	1.271774	.8005136	.3123787	.5211265	.7194833
#1	.6516619	6.018318	36.08968	10.37973	8.271971	.2269559
#2	.6560508	6.030813	35.78524	10.44255	8.323560	.2279090
#3	.6529516	5.893267	35.51693	10.39663	8.358104	.2247328

Elem	Sr4077
Units	ppm
Avg	.1053842
Stddev	.0114699
%RSD	10.88392
#1	.1096276
#2	.1141278
#3	.0923972

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5800.960	105871.8	13566.87	4326.345	5832.947
Stddev	9.265	364.7	163.06	12.800	6.220
%RSD	.1597230	.3444390	1.201909	.2958535	.1066441
#1	5791.743	105456.2	13525.45	4313.945	5826.173
#2	5800.862	106138.4	13428.51	4339.510	5838.402
#3	5810.274	106020.8	13746.64	4325.579	5834.267

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: PB169215BS Acquired: 8/13/2025 17:42:07 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.8098559**	**2.163288**	**.9703870**	**2.078532**	**.8419443**	**2.020039**
Stddev	.0021268	.006551	.0020778	.013014	.0025158	.016636
%RSD	.2626112	.3028119	.2141175	.6261072	.2988040	.8235413
#1	.8076511	2.164087	.9689917	2.066937	.8420725	2.037694
#2	.8100216	2.169403	.9693944	2.076053	.8393669	2.017766
#3	.8118949	2.156375	.9727749	2.092608	.8443936	2.004656

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1892723**	**.2017506**	**.1923842**	**1.033484**	**.4157631**	**.1985200**
Stddev	.0017982	.0016368	.0004945	.021898	.0005109	.0006279
%RSD	.9500428	.8112796	.2570582	2.118841	.1228838	.3162666
#1	.1912906	.2036224	.1918542	1.058335	.4152032	.1978224
#2	.1886852	.2010412	.1924650	1.017015	.4158823	.1986979
#3	.1878411	.2005883	.1928334	1.025102	.4162039	.1990397

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.3164279**	**3.063915**	**.2036323**	**2.018439**	**.5003844**	**.0773759**
Stddev	.0005911	.007467	.0020322	.025195	.0013786	.0002532
%RSD	.1867989	.2436933	.9979807	1.248248	.2755112	.3272504
#1	.3159403	3.059599	.2057227	2.040085	.4989993	.0770885
#2	.3162583	3.072537	.2035106	1.990782	.5003976	.0775663
#3	.3170853	3.059609	.2016637	2.024449	.5017565	.0774728

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.790608**	**.3086457**	**.2094068**	**9.889119**	**.2942874**	**.4272990**
Stddev	.019448	.0019316	.0003615	.040434	.0023908	.0008186
%RSD	.6969171	.6258194	.1726331	.4088703	.8124165	.1915649
#1	2.807110	.3094516	.2096575	9.894204	.2965155	.4266514
#2	2.769166	.3100438	.2089924	9.846383	.2945849	.4270266
#3	2.795549	.3064417	.2095706	9.926769	.2917617	.4282190

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: PB169215BS Acquired: 8/13/2025 17:42:07 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.6897690	.1990584	.9339743	5.416043	F .0019589	.1915965
Stddev	.0027190	.0023859	.0106976	.021028	.0011814	.0013790
%RSD	.3941916	1.198589	1.145383	.3882599	60.31055	.7197359
#1	.6867652	.2015360	.9275689	5.392761	.0010165	.1931535
#2	.6904800	.1988631	.9280301	5.421713	.0015759	.1905292
#3	.6920619	.1967762	.9463240	5.433655	.0032843	.1911067

Elem	Sr4077
Units	ppm
Avg	.1958461
Stddev	.0015679
%RSD	.8005777
#1	.1976450
#2	.1951235
#3	.1947698

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5643.099	101499.1	11437.82	4238.110	7221.228
Stddev	14.629	143.0	102.85	18.419	17.611
%RSD	.2592284	.1408958	.8992056	.4345984	.2438769
#1	5659.783	101347.7	11319.08	4217.177	7241.518
#2	5632.472	101631.9	11495.16	4251.832	7212.258
#3	5637.041	101517.8	11499.21	4245.322	7209.909

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCV07 Acquired: 8/13/2025 17:46:06 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV07 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**5.061751**	**5.358708**	**4.884184**	**5.158125**	**5.121728**	**9.835716**	**9.846023**
Stddev	.006174	.049189	.004860	.013048	.013897	.016428	.125289
%RSD	.1219640	.9179269	.0994963	.2529624	.2713388	.1670220	1.272483
#1	5.061243	5.363418	4.885294	5.146779	5.114275	9.854138	9.978686
#2	5.055847	5.307334	4.888392	5.155214	5.113147	9.822589	9.829669
#3	5.068162	5.405373	4.878865	5.172383	5.137762	9.830419	9.729715

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2412303**	**2.444688**	**24.56440**	**1.005668**	**2.448709**	**1.252828**	**5.139318**
Stddev	.0014978	.003691	.06028	.002626	.002727	.001636	.032814
%RSD	.6208869	.1509783	.2453991	.2611341	.1113816	.1305542	.6384872
#1	.2429523	2.445921	24.63349	1.004971	2.450829	1.252413	5.108507
#2	.2405088	2.447604	24.53718	1.003461	2.449667	1.251440	5.135624
#3	.2402298	2.440538	24.52253	1.008572	2.445632	1.254631	5.173822

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.426538**	**24.43010**	**2.462730**	**1.255634**	**23.63989**	**2.492681**	**2.532925**
Stddev	.008475	.08315	.003602	.001815	.13670	.003999	.004493
%RSD	.3492821	.3403772	.1462586	.1445167	.5782662	.1604231	.1773659
#1	2.436291	24.52612	2.464274	1.256361	23.58539	2.497180	2.538045
#2	2.420961	24.38187	2.465302	1.253569	23.53884	2.491332	2.531089
#3	2.422362	24.38230	2.458613	1.256973	23.79543	2.489532	2.529642

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**24.81823**	**4.789741**	**5.140567**	**4.939621**	**4.920080**	**5.230056**	**4.702327**
Stddev	.14336	.026219	.010365	.008277	.007624	.051728	.003153
%RSD	.5776378	.5473900	.2016318	.1675685	.1549586	.9890547	.0670513
#1	24.73653	4.819353	5.137956	4.936586	4.927466	5.179097	4.698751
#2	24.73439	4.780386	5.131757	4.948988	4.912238	5.228550	4.703520
#3	24.98376	4.769482	5.151987	4.933290	4.920535	5.282520	4.704709

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCV07 Acquired: 8/13/2025 17:46:06 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV07 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**4.856117**	**4.514361**	**4.915100**
Stddev	.008828	.007981	.010800
%RSD	.1817848	.1767899	.2197312
#1	4.847165	4.520595	4.925030
#2	4.864815	4.505366	4.916670
#3	4.856369	4.517121	4.903602

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**5571.628**	**98750.90**	**11525.47**	**4172.309**	**6754.504**
Stddev	9.155	100.03	60.25	10.100	14.252
%RSD	.1643132	.1012949	.5227922	.2420781	.2110032
#1	5565.641	98859.18	11469.48	4166.065	6738.047
#2	5582.167	98731.60	11517.71	4166.900	6762.660
#3	5567.076	98661.93	11589.23	4183.962	6762.805

Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCB07 Acquired: 8/13/2025 17:50:17 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB07 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0007834**	**-.000977**	**-.001403**	**-.000714**	**-.000627**	**-.015122**	**-.009368**
Stddev	.0005179	.001205	.001112	.000675	.000462	.006713	.001339
%RSD	66.10615	123.3274	79.25845	94.55189	73.71355	44.39299	14.28868
#1	.0011536	.000094	-.002684	-.001455	-.000543	-.020681	-.010900
#2	.0001916	-.000744	-.000688	-.000134	-.001126	-.007664	-.008422
#3	.0010050	-.002282	-.000836	-.000554	-.000213	-.017022	-.008784

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0000134**	**-.000105**	**-.009295**	**.0000067**	**-.000172**	**-.000066**	**-.002980**
Stddev	.0000322	.000086	.004347	.0000791	.000073	.000219	.005712
%RSD	240.2001	82.04779	46.76270	1182.281	42.55719	329.5975	191.6678
#1	-.000023	-.000120	-.005527	.0000899	-.000173	-.000231	-.008596
#2	.000038	-.000182	-.008308	-.000068	-.000244	.000182	.002824
#3	.000025	-.000012	-.014050	-.000002	-.000098	-.000151	-.003169

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000295**	**-.016698**	**-.000070**	**.0005424**	**.0497158**	**.0018910**	**-.000262**
Stddev	.000317	.014768	.000167	.0001428	.0062732	.0009997	.000092
%RSD	107.4310	88.44412	239.0154	26.32410	12.61808	52.86805	35.20208
#1	.000034	-.030382	-.000262	.0006393	.0461085	.0008380	-.000366
#2	-.000597	-.001043	.000021	.0006094	.0569594	.0028272	-.000190
#3	-.000321	-.018669	.000032	.0003784	.0460794	.0020079	-.000229

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2415448**	**.0049203**	**.0000744**	**-.000227**	**-.000145**	**.0051313**	**.0012769**
Stddev	.0327008	.0004377	.0001539	.000141	.000456	.0032052	.0011478
%RSD	13.53821	8.895402	206.7526	62.10962	314.4662	62.46361	89.88482
#1	.2461504	.0050151	-.000102	-.000351	.000305	.0087122	.0025234
#2	.2067853	.0044429	.000143	-.000074	-.000133	.0025309	.0010438
#3	.2716987	.0053027	.000182	-.000255	-.000607	.0041507	.0002636


Reviewed By:jaswal
On:8/19/2025 2:31:42 AM
Inst Id :P4
LB :LB136801

Sample Name: CCB07 Acquired: 8/13/2025 17:50:17 Type: Unk
Method: 6010-200.7 NEW(v96) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB07 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0002820	-.004491	-.000254
Stddev	.0019750	.001698	.000037
%RSD	700.4355	37.81408	14.61798
#1	-.001986	-.003235	-.000270
#2	.001212	-.003815	-.000212
#3	.001620	-.006423	-.000281

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5764.274	103250.6	11518.18	4320.515	7589.105
Stddev	3.950	84.7	32.77	5.128	11.655
%RSD	.0685278	.0820157	.2844690	.1186836	.1535706
#1	5767.850	103178.1	11496.66	4325.909	7600.862
#2	5760.034	103343.7	11501.99	4319.933	7577.556
#3	5764.937	103230.0	11555.89	4315.703	7588.897

Reviewed By:
On:
Inst Id :P4
LB :LB136973




As 189.042 {479}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.000108	Re-Slope:	1.000000
A1 (Gain):	0.047247	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999921	Status:	OK.
Std Error of Est:	0.000014		
Predicted MDL:	0.002207		
Predicted MQL:	0.007357		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	-.00011	.000	1
S1	.02000	.01928	-.001	-3.58	.00080	.000	1
S3	2.5000	2.4454	-.055	-2.19	.11531	.000	1
S4	5.0000	5.0512	.051	1.02	.23830	.001	1
S5	10.000	10.033	.033	.330	.47344	.001	1
S2	.80000	.77122	-.029	-3.60	.03629	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Tl 190.856 {477}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000507	Re-Slope:	1.000000
A1 (Gain):	0.076555	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999842	Status:	OK.
Std Error of Est:	0.000044		
Predicted MDL:	0.001681		
Predicted MQL:	0.005605		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00051	.000	1
S1	.04000	.04062	.001	1.56	.00342	.000	1
S3	2.5000	2.4728	-.027	-1.09	.18632	.007	1
S4	5.0000	5.1421	.142	2.84	.38718	.014	1
S5	10.000	9.8933	-.107	-1.07	.74393	.021	1
S2	.80000	.79120	-.009	-1.10	.05996	.001	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Pb 220.353 {453}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000077	Re-Slope:	1.000000
A1 (Gain):	0.194154	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999919	Status:	OK.
Std Error of Est:	0.000045		
Predicted MDL:	0.001282		
Predicted MQL:	0.004274		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00008	.000	1
S1	.01200	.01251	.001	4.27	.00246	.000	1
S3	2.5000	2.4396	-.060	-2.42	.47320	.001	1
S4	5.0000	5.0035	.004	.071	.97046	.001	1
S5	10.000	10.082	.082	.816	1.9553	.003	1
S2	.80000	.77472	-.025	-3.16	.15032	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Se 196.090 {472}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000404	Re-Slope:	1.000000
A1 (Gain):	0.043354	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999919	Status:	OK.
Std Error of Est:	0.000013		
Predicted MDL:	0.003019		
Predicted MQL:	0.010063		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00040	.000	1
S1	.02000	.01821	-.002	-8.94	.00119	.000	1
S3	2.5000	2.4445	-.055	-2.22	.10635	.000	1
S4	5.0000	5.0534	.053	1.07	.21943	.001	1
S5	10.000	10.031	.031	.306	.43515	.001	1
S2	.80000	.77325	-.027	-3.34	.03392	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Sb 206.833 {463}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000712	Re-Slope:	1.000000
A1 (Gain):	0.087185	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999948	Status:	OK.
Std Error of Est:	0.000033		
Predicted MDL:	0.001901		
Predicted MQL:	0.006336		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00071	.000	1
S1	.05000	.04962	-.000	-.762	.00505	.000	1
S3	2.5000	2.4605	-.040	-1.58	.21582	.001	1
S4	5.0000	4.9774	-.023	-.451	.43585	.001	1
S5	10.000	10.083	.083	.828	.88215	.003	1
S2	.80000	.77965	-.020	-2.54	.06888	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Al 396.152 { 85}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.000307	Re-Slope:	1.000000
A1 (Gain):	0.073709	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999989	Status:	OK.
Std Error of Est:	0.000027		
Predicted MDL:	0.010722		
Predicted MQL:	0.035741		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	-.00031	.000	1
S1	.10000	.10423	.004	4.23	.00795	.000	1
S3	5.0000	4.9880	-.012	-.240	.37450	.001	1
S4	10.000	9.9363	-.064	-.637	.74639	.001	1
S5	20.000	20.061	.061	.305	1.5070	.002	1
S2	1.6000	1.6102	.010	.639	.12067	.001	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Ba 493.409 { 68}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.068534	Re-Slope:	1.000000
A1 (Gain):	2.291628	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999968	Status:	OK.
Std Error of Est:	0.001350		
Predicted MDL:	0.000981		
Predicted MQL:	0.003270		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00001	.000	.000	.06855	.000	1
S1	.10000	.09508	-.005	-4.92	.28641	.002	1
S3	5.0000	4.9976	-.002	-.048	11.521	.057	1
S4	10.000	9.9097	-.090	-.903	22.778	.184	1
S5	20.000	20.125	.125	.626	46.188	.327	1
S2	1.6000	1.5724	-.028	-1.72	3.6720	.012	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Be 234.861 {144}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.001364	Re-Slope:	1.000000
A1 (Gain):	6.857220	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999796	Status:	OK.
Std Error of Est:	0.000397		
Predicted MDL:	0.000071		
Predicted MQL:	0.000236		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	-.00137	.000	1
S1	.00600	.00654	.001	8.96	.04316	.000	1
S3	.12500	.12592	.001	.733	.85824	.005	1
S4	.25000	.25559	.006	2.24	1.7436	.005	1
S5	.50000	.49154	-.008	-1.69	3.3539	.010	1
S2	.04000	.04142	.001	3.54	.28140	.002	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Cd 226.502 {449}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.000530	Re-Slope:	1.000000
A1 (Gain):	2.954374	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999854	Status:	OK.
Std Error of Est:	0.000458		
Predicted MDL:	0.000072		
Predicted MQL:	0.000240		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	-.00053	.000	1
S1	.00600	.00573	-.000	-4.56	.01645	.000	1
S3	1.2500	1.2127	-.037	-2.98	3.5847	.009	1
S4	2.5000	2.5162	.016	.649	7.4382	.012	1
S5	5.0000	5.0422	.042	.844	14.906	.017	1
S2	.40000	.37915	-.021	-5.21	1.1204	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Ca 373.690 { 90}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.001357	Re-Slope:	1.000000
A1 (Gain):	0.054101	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999986	Status:	OK.
Std Error of Est:	0.000153		
Predicted MDL:	0.011484		
Predicted MQL:	0.038280		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00005	-.000	.000	.00135	.001	1
S2	4.0000	4.0459	.046	1.15	.22024	.000	1
S3	12.500	12.617	.117	.937	.68396	.001	1
S4	25.000	25.023	.023	.091	1.3551	.002	1
S5	50.000	49.795	-.205	-.411	2.6953	.005	1
S1	2.0000	2.0195	.019	.974	.11061	.001	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Cr 267.716 {126}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000257	Re-Slope:	1.000000
A1 (Gain):	0.184351	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999912	Status:	OK.
Std Error of Est:	0.000018		
Predicted MDL:	0.000338		
Predicted MQL:	0.001127		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00026	.000	1
S1	.01000	.00969	-.000	-3.11	.00204	.000	1
S3	.50000	.49977	-.000	-.046	.09243	.000	1
S4	1.0000	1.0204	.020	2.04	.18844	.001	1
S5	2.0000	1.9792	-.021	-1.04	.36528	.002	1
S2	.16000	.16090	.001	.561	.02993	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Co 228.616 {448}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000009	Re-Slope:	1.000000
A1 (Gain):	1.020308	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999844	Status:	OK.
Std Error of Est:	0.000367		
Predicted MDL:	0.000184		
Predicted MQL:	0.000614		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00001	.000	1
S1	.03000	.02904	-.001	-3.21	.02951	.000	1
S3	1.2500	1.2148	-.035	-2.82	1.2416	.003	1
S4	2.5000	2.4804	-.020	-.785	2.5351	.003	1
S5	5.0000	5.0721	.072	1.44	5.1838	.009	1
S2	.40000	.38370	-.016	-4.08	.39219	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Cu 224.700 {450}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000164	Re-Slope:	1.000000
A1 (Gain):	0.612509	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999973	Status:	OK.
Std Error of Est:	0.000054		
Predicted MDL:	0.000412		
Predicted MQL:	0.001372		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00016	.000	1
S1	.02000	.02138	.001	6.92	.01340	.000	1
S3	.62500	.62880	.004	.608	.38850	.001	1
S4	1.2500	1.2542	.004	.338	.77477	.001	1
S5	2.5000	2.4872	-.013	-.511	1.5364	.004	1
S2	.20000	.20336	.003	1.68	.12574	.001	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Fe 240.488 {140}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000032	Re-Slope:	1.000000
A1 (Gain):	0.002988	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999944	Status:	OK.
Std Error of Est:	0.000002		
Predicted MDL:	0.006544		
Predicted MQL:	0.021813		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00001	.000	.000	.00003	.000	1
S1	.10000	.09610	-.004	-3.90	.00032	.000	1
S3	2.5000	2.4870	-.013	-.521	.00740	.000	1
S4	5.0000	4.9875	-.013	-.250	.01481	.000	1
S5	10.000	10.063	.063	.626	.02984	.000	1
S2	.80000	.76683	-.033	-4.15	.00230	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Mn 257.610 {131}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000268	Re-Slope:	1.000000
A1 (Gain):	0.393545	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999971	Status:	OK.
Std Error of Est:	0.000050		
Predicted MDL:	0.000585		
Predicted MQL:	0.001949		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00027	.000	1
S1	.02000	.02137	.001	6.84	.00868	.000	1
S3	1.2500	1.2649	.015	1.19	.49816	.002	1
S4	2.5000	2.5021	.002	.084	.98515	.003	1
S5	5.0000	4.9743	-.026	-.513	1.9583	.004	1
S2	.40000	.40731	.007	1.83	.16059	.001	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Mg 279.079 {121}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.000018	Re-Slope:	1.000000
A1 (Gain):	0.010746	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999910	Status:	OK.
Std Error of Est:	0.000076		
Predicted MDL:	0.027574		
Predicted MQL:	0.091912		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00018	-.000	.000	-.00002	.000	1
S2	4.0000	4.0137	.014	.342	.04311	.000	1
S3	12.500	12.522	.022	.175	.13455	.000	1
S4	25.000	24.973	-.027	-.107	.26835	.001	1
S5	50.000	49.812	-.188	-.376	.53529	.000	1
S1	2.0000	2.1791	.179	8.96	.02340	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Ni 231.604 {446}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.000566	Re-Slope:	1.000000
A1 (Gain):	0.591930	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999903	Status:	OK.
Std Error of Est:	0.000194		
Predicted MDL:	0.000339		
Predicted MQL:	0.001130		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	-.00056	.000	1
S1	.04000	.03935	-.001	-1.62	.02273	.000	1
S3	1.2500	1.2203	-.030	-2.38	.72176	.002	1
S4	2.5000	2.4912	-.009	-.351	1.4741	.002	1
S5	5.0000	5.0530	.053	1.06	2.9904	.004	1
S2	.40000	.38615	-.014	-3.46	.22801	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Ag 328.068 {103}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.000829	Re-Slope:	1.000000
A1 (Gain):	0.190715	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999995	Status:	OK.
Std Error of Est:	0.000005		
Predicted MDL:	0.000519		
Predicted MQL:	0.001730		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	-.00083	.000	1
S1	.01000	.01018	.000	1.76	.00109	.000	1
S3	.62500	.62029	-.005	-.754	.11689	.000	1
S4	1.2500	1.2524	.002	.189	.23686	.000	1
S5	2.5000	2.5024	.002	.095	.47410	.000	1
S2	.20000	.19981	-.000	-.097	.03709	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Na 589.592 { 57}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.002109	Re-Slope:	1.000000
A1 (Gain):	0.007388	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999365	Status:	OK.
Std Error of Est:	0.000139		
Predicted MDL:	0.018006		
Predicted MQL:	0.060019		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00042	.000	.000	.00211	.000	1
S2	4.0000	3.7403	-.260	-6.49	.02974	.000	1
S3	12.500	12.409	-.091	-.732	.09379	.001	1
S4	25.000	24.191	-.809	-3.23	.18084	.001	1
S5	50.000	51.423	1.42	2.85	.38203	.003	1
S1	2.0000	1.7369	-.263	-13.2	.01494	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




V 292.402 {115}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.000188	Re-Slope:	1.000000
A1 (Gain):	0.122406	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999970	Status:	OK.
Std Error of Est:	0.000022		
Predicted MDL:	0.002848		
Predicted MQL:	0.009493		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	-.00019	.000	1
S1	.04000	.04109	.001	2.72	.00466	.000	1
S3	1.2500	1.2352	-.015	-1.18	.14878	.000	1
S4	2.5000	2.4911	-.009	-.355	.30028	.001	1
S5	5.0000	5.0297	.030	.594	.60655	.002	1
S2	.40000	.39290	-.007	-1.77	.04719	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Zn 206.200 {464}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.002941	Re-Slope:	1.000000
A1 (Gain):	1.040372	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999959	Status:	OK.
Std Error of Est:	0.000220		
Predicted MDL:	0.000145		
Predicted MQL:	0.000482		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00294	.000	1
S1	.04000	.04023	.000	.587	.04480	.000	1
S3	1.2500	1.2670	.017	1.36	1.3211	.004	1
S4	2.5000	2.5141	.014	.564	2.6185	.003	1
S5	5.0000	4.9613	-.039	-.774	5.1645	.015	1
S2	.40000	.40737	.007	1.84	.42676	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Zn 213.856 {158}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.007596	Re-Slope:	1.000000
A1 (Gain):	2.195143	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999973	Status:	OK.
Std Error of Est:	0.000384		
Predicted MDL:	0.000244		
Predicted MQL:	0.000812		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00759	.001	1
S1	.04000	.04058	.001	1.46	.09730	.001	1
S3	1.2500	1.2576	.008	.609	2.7873	.002	1
S4	2.5000	2.5224	.022	.897	5.5829	.008	1
S5	5.0000	4.9661	-.034	-.677	10.985	.010	1
S2	.40000	.40323	.003	.808	.89885	.002	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




K 766.490 { 44}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000810	Re-Slope:	1.000000
A1 (Gain):	0.004068	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999545	Status:	OK.
Std Error of Est:	0.000065		
Predicted MDL:	0.039264		
Predicted MQL:	0.130880		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00037	.000	.000	.00081	.000	1
S2	4.0000	3.7003	-.300	-7.49	.01586	.000	1
S3	12.500	12.261	-.239	-1.91	.05069	.000	1
S4	25.000	24.497	-.503	-2.01	.10047	.001	1
S5	50.000	51.230	1.23	2.46	.20922	.002	1
S1	2.0000	1.8122	-.188	-9.39	.00818	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




B 249.678 {135}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.005257	Re-Slope:	1.000000
A1 (Gain):	0.710174	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999868	Status:	OK.
Std Error of Est:	0.000603		
Predicted MDL:	0.000739		
Predicted MQL:	0.002462		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00001	-.000	.000	.00525	.000	1
S1	.10000	.10700	.007	7.00	.08082	.001	1
S3	2.5000	2.4942	-.006	-.231	1.7702	.010	1
S4	5.0000	5.1057	.106	2.11	3.6184	.008	1
S5	10.000	9.8737	-.126	-1.26	6.9917	.019	1
S2	.80000	.81947	.019	2.43	.58518	.003	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Mo 202.030 {467}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000510	Re-Slope:	1.000000
A1 (Gain):	0.618511	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999998	Status:	OK.
Std Error of Est:	0.000094		
Predicted MDL:	0.000257		
Predicted MQL:	0.000856		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	-.00000	-.000	.000	.00051	.000	1
S1	.20000	.20151	.002	.753	.12514	.000	1
S3	2.5000	2.4922	-.008	-.312	1.5420	.003	1
S4	5.0000	4.9906	-.009	-.188	3.0872	.007	1
S5	10.000	10.014	.014	.140	6.1943	.018	1
S2	.80000	.80177	.002	.221	.49641	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Sn 189.989 {478}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000422	Re-Slope:	1.000000
A1 (Gain):	0.139897	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999948	Status:	OK.
Std Error of Est:	0.000047		
Predicted MDL:	0.000634		
Predicted MQL:	0.002114		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00042	.000	1
S1	.04000	.03749	-.003	-6.28	.00566	.000	1
S3	2.5000	2.4732	-.027	-1.07	.34605	.000	1
S4	5.0000	4.9538	-.046	-.923	.69273	.002	1
S5	10.000	10.089	.089	.891	1.4104	.003	1
S2	.80000	.78643	-.014	-1.70	.11033	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Ti 336.121 {100}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.001472	Re-Slope:	1.000000
A1 (Gain):	0.277447	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999972	Status:	OK.
Std Error of Est:	0.000069		
Predicted MDL:	0.001625		
Predicted MQL:	0.005418		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00147	.000	1
S1	.04000	.03713	-.003	-7.18	.01176	.000	1
S3	2.5000	2.4941	-.006	-.235	.69320	.002	1
S4	5.0000	4.9544	-.046	-.911	1.3756	.003	1
S5	10.000	10.060	.060	.602	2.7916	.004	1
S2	.80000	.79407	-.006	-.742	.22170	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Si 288.158 {117}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000091	Re-Slope:	1.000000
A1 (Gain):	0.002752	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999570	Status:	OK.
Std Error of Est:	0.000009		
Predicted MDL:	0.012666		
Predicted MQL:	0.042220		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00007	.000	.000	.00009	.000	1
S1	.40000	.35700	-.043	-10.7	.00108	.000	1
S3	2.5000	2.4340	-.066	-2.64	.00689	.000	1
S4	5.0000	4.8993	-.101	-2.01	.01378	.000	1
S5	10.000	10.249	.249	2.49	.02872	.000	1
S2	.80000	.76021	-.040	-4.97	.00222	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




P 177.495 {490}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	-0.001087	Re-Slope:	1.000000
A1 (Gain):	0.046617	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999600	Status:	OK.
Std Error of Est:	0.000031		
Predicted MDL:	0.001627		
Predicted MQL:	0.005424		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	-.00109	.000	1
S1	.02000	.01766	-.002	-11.7	-.00026	.000	1
S3	2.5000	2.3980	-.102	-4.08	.11068	.000	1
S4	5.0000	4.9578	-.042	-.844	.22999	.002	1
S5	10.000	10.212	.212	2.12	.47490	.002	1
S2	.80000	.73451	-.065	-8.19	.03315	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




S 182.034 {485}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000177	Re-Slope:	1.000000
A1 (Gain):	0.029531	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999227	Status:	OK.
Std Error of Est:	0.000027		
Predicted MDL:	0.002784		
Predicted MQL:	0.009279		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00001	.000	.000	.00018	.000	1
S1	.02000	.01720	-.003	-14.0	.00066	.000	1
S3	2.5000	2.3541	-.146	-5.84	.06950	.001	1
S4	5.0000	4.9077	-.092	-1.85	.14472	.002	1
S5	10.000	10.318	.318	3.18	.30412	.002	1
S2	.80000	.72265	-.077	-9.67	.02146	.000	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Li 670.784 { 50}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.010427	Re-Slope:	1.000000
A1 (Gain):	0.742140	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999996	Status:	OK.
Std Error of Est:	0.000052		
Predicted MDL:	0.001933		
Predicted MQL:	0.006444		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.01043	.001	1
S5	10.000	10.018	.018	.181	7.4512	.025	1
S4	5.0000	5.0020	.002	.040	3.7256	.008	1
S3	2.5000	2.4862	-.014	-.550	1.8571	.007	1
S1	.02000	.02012	.000	.601	.02562	.001	1
S2	.80000	.79351	-.006	-.811	.59980	.001	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Sr 407.771 { 83}

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000606	Re-Slope:	1.000000
A1 (Gain):	7.474096	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999994	Status:	OK.
Std Error of Est:	0.000596		
Predicted MDL:	0.000102		
Predicted MQL:	0.000340		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00061	.000	1
S1	.02000	.01940	-.001	-3.00	.14632	.001	1
S3	2.5000	2.5192	.019	.769	18.847	.035	1
S4	5.0000	4.9911	-.009	-.179	37.340	.233	1
S5	10.000	9.9870	-.013	-.130	74.716	.102	1
S2	.80000	.80326	.003	.407	6.0100	.017	1

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Y 224.306 {450}*

Date of Fit:	8/26/2025 12:02:29	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000000	Re-Slope:	1.000000
A1 (Gain):	0.000000	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.000000	Status:	Warning Zero Gain
Std Error of Est:	0.000000		
Predicted MDL:	n/a		
Predicted MQL:	n/a		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis



Y 360.073 { 94}*


Date of Fit:	<not fit>	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000000	Re-Slope:	1.000000
A1 (Gain):	0.000000	Y-int:	0.000000

Reviewed By:
On:
Inst Id :P4
LB :LB136973

A2 (Curvature):	0.000000					
n (Exponent):	1.000000					
Correlation:	0.000000		Status:	Warning	Zero Gain	
Std Error of Est:	0.000000					
Predicted MDL:	n/a					
Predicted MQL:	n/a					

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis




Y 371.030 { 91}*

Date of Fit:	<not fit>		Type of Fit:	Linear	Weighting:	1/Conc
A0 (Offset):		0.000000		Re-Slope:	1.000000	
A1 (Gain):		0.000000		Y-int:	0.000000	
A2 (Curvature):		0.000000				
n (Exponent):		1.000000				
Correlation:		0.000000		Status:	Warning	Zero Gain
Std Error of Est:		0.000000				
Predicted MDL:		n/a				
Predicted MQL:		n/a				

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis

Reviewed By:
On:
Inst Id :P4
LB :LB136973




Y 224.306 {150}*

Date of Fit:	<not fit>	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000000	Re-Slope:	1.000000
A1 (Gain):	0.000000	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.000000	Status:	Warning Zero Gain
Std Error of Est:	0.000000		
Predicted MDL:	n/a		
Predicted MQL:	n/a		

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis



In 230.606 {446}*

Date of Fit:	<not fit>	Type of Fit:	Linear	Weighting:	1/Conc

A0 (Offset):	0.000000	Re-Slope:	1.000000
A1 (Gain):	0.000000	Y-int:	0.000000

Reviewed By:
On:
Inst Id :P4
LB :LB136973

A2 (Curvature):		0.000000					
n (Exponent):		1.000000					
Correlation:		0.000000		Status:	Warning	Zero Gain	
Std Error of Est:		0.000000					
Predicted MDL:		n/a					
Predicted MQL:		n/a					
Std. Name	**Stated Conc.**	**Found Conc.**	**Difference**	**% Diff.**	**(S)IR**	**Std Dev**	**Emphasis**

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: S0 Acquired: 8/26/2025 10:49:22 Type: Cal
Method: 6010-200.7 NEW(v130) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934	Be2348
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00011**	**.00051**	**.00008**	**.00040**	**.00071**	**-.00031**	**.06855**	**-.00137**
Stddev	.00013	.00002	.00014	.00013	.00010	.00029	.00027	.00038
%RSD	115.64	3.5165	183.76	31.047	13.630	92.789	.39109	27.857
#1	-.00006	.00053	.00020	.00054	.00068	-.00026	.06830	-.00155
#2	-.00025	.00050	-.00008	.00029	.00082	-.00061	.06852	-.00162
#3	-.00002	.00050	.00011	.00038	.00064	-.00005	.06883	-.00093

Elem	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404	Mn2576	Mg2790
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00053**	**.00135**	**.00026**	**.00001**	**.00016**	**.00003**	**.00027**	**-.00002**
Stddev	.00009	.00051	.00005	.00012	.00012	.00001	.00019	.00033
%RSD	17.894	37.894	21.156	968.32	75.663	33.413	72.394	1647.4
#1	-.00062	.00077	.00031	.00011	.00026	.00004	.00048	-.00004
#2	-.00054	.00154	.00020	.00005	.00020	.00002	.00009	.00032
#3	-.00043	.00175	.00026	-.00012	.00003	.00004	.00023	-.00034

Elem	Ni2316	Ag3280	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**-.00056**	**-.00083**	**.00211**	**-.00019**	**.00759**	**.00081**	**.00525**	**.00051**
Stddev	.00007	.00009	.00015	.00046	.00067	.00016	.00028	.00013
%RSD	12.741	10.357	7.3205	247.15	8.7794	19.962	5.2694	24.690
#1	-.00053	-.00083	.00195	-.00058	.00695	.00071	.00550	.00045
#2	-.00052	-.00074	.00213	.00032	.00828	.00073	.00530	.00042
#3	-.00065	-.00092	.00226	-.00031	.00755	.00100	.00495	.00065

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00042**	**.00147**	**.00009**	**-.00109**	**.00018**	**.01043**	**.00061**
Stddev	.00002	.00043	.00003	.00002	.00004	.00081	.00038
%RSD	4.1960	29.512	28.009	1.5399	21.437	7.7901	61.865
#1	.00043	.00104	.00006	-.00107	.00013	.01080	.00017
#2	.00040	.00147	.00010	-.00110	.00021	.01099	.00084
#3	.00043	.00191	.00011	-.00109	.00019	.00950	.00081

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: S0 Acquired: 8/26/2025 10:49:22 Type: Cal
Method: 6010-200.7 NEW(v130) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6333.6	108640.	10151.	4941.0	7754.8
Stddev	2.8	190.	20.	8.6	16.3
%RSD	.04388	.17500	.19766	.17385	.21059
#1	6333.4	108640.	10173.	4938.4	7748.2
#2	6331.0	108830.	10134.	4950.6	7742.9
#3	6336.5	108450.	10147.	4934.0	7773.4

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: S1 Acquired: 8/26/2025 10:53:47 Type: Cal
Method: 6010-200.7 NEW(v130) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934	Be2348
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00080**	**.00342**	**.00246**	**.00119**	**.00505**	**.00795**	**.28641**	**.04316**
Stddev	.00003	.00014	.00005	.00007	.00013	.00006	.00240	.00035
%RSD	3.9253	4.1358	1.9680	6.1642	2.5275	.80636	.83655	.81860
#1	.00082	.00335	.00241	.00123	.00495	.00796	.28728	.04349
#2	.00082	.00358	.00246	.00124	.00519	.00788	.28825	.04319
#3	.00076	.00333	.00251	.00111	.00500	.00801	.28370	.04279

Elem	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404	Mn2576	Mg2790
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.01645**	**.11061**	**.00204**	**.02951**	**.01340**	**.00032**	**.00868**	**.02340**
Stddev	.00020	.00057	.00004	.00008	.00011	.00002	.00024	.00032
%RSD	1.2291	.51667	1.7674	.27851	.80728	7.1676	2.7082	1.3603
#1	.01626	.11043	.00205	.02959	.01339	.00030	.00891	.02377
#2	.01641	.11015	.00207	.02943	.01351	.00034	.00869	.02320
#3	.01666	.11125	.00200	.02951	.01330	.00031	.00844	.02324

Elem	Ni2316	Ag3280	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.02273**	**.00109**	**.01494**	**.00466**	**.09730**	**.00818**	**.08082**	**.12514**
Stddev	.00019	.00005	.00024	.00010	.00070	.00027	.00061	.00026
%RSD	.84029	4.5018	1.5917	2.1298	.72454	3.3147	.76058	.21159
#1	.02294	.00112	.01484	.00476	.09671	.00787	.08076	.12534
#2	.02257	.00104	.01477	.00456	.09710	.00835	.08147	.12484
#3	.02266	.00113	.01521	.00466	.09808	.00833	.08024	.12525

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.00566**	**.01176**	**.00108**	**-.00026**	**.00066**	**.02562**	**.14632**
Stddev	.00014	.00007	.00005	.00006	.00005	.00056	.00074
%RSD	2.4522	.58242	4.4852	23.518	8.1431	2.2044	.50831
#1	.00557	.01175	.00104	-.00021	.00069	.02595	.14714
#2	.00559	.01171	.00113	-.00024	.00069	.02593	.14568
#3	.00582	.01184	.00106	-.00033	.00060	.02497	.14615

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: S1 Acquired: 8/26/2025 10:53:47 Type: Cal
Method: 6010-200.7 NEW(v130) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6366.6	106470.	9860.7	4843.5	7651.9
Stddev	9.9	500.	40.3	15.9	5.8
%RSD	.15525	.46992	.40845	.32897	.07558
#1	6371.6	106970.	9837.4	4843.0	7658.2
#2	6372.9	106460.	9837.5	4859.6	7650.7
#3	6355.2	105970.	9907.2	4827.8	7646.8

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: S2 Acquired: 8/26/2025 10:58:08 Type: Cal
Method: 6010-200.7 NEW(v130) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934	Be2348	Cd2265
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.03629**	**.05996**	**.15032**	**.03392**	**.06888**	**.12067**	**3.6720**	**.28140**	**1.1204**
Stddev	.00011	.00058	.00013	.00013	.00023	.00080	.0121	.00152	.0005
%RSD	.29297	.96252	.08829	.39096	.33028	.66101	.32910	.53945	.04188
#1	.03618	.06027	.15019	.03394	.06866	.12126	3.6811	.28065	1.1206
#2	.03630	.05929	.15032	.03377	.06885	.12099	3.6765	.28315	1.1199
#3	.03639	.06032	.15045	.03404	.06912	.11976	3.6583	.28040	1.1207

Elem	Ca3736	Cr2677	Co2286	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.22024**	**.02993**	**.39219**	**.12574**	**.00230**	**.16059**	**.04311**	**.22801**	**.03709**
Stddev	.00012	.00004	.00022	.00057	.00002	.00088	.00030	.00037	.00023
%RSD	.05594	.13301	.05545	.45100	.94793	.55026	.70522	.16198	.62839
#1	.22029	.02990	.39203	.12553	.00230	.16153	.04315	.22795	.03684
#2	.22033	.02997	.39210	.12531	.00233	.16048	.04279	.22767	.03730
#3	.22010	.02992	.39244	.12638	.00229	.15977	.04340	.22840	.03714

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.02974**	**.04719**	**.89885**	**.01586**	**.58518**	**.49641**	**.11033**	**.22170**	**.00222**
Stddev	.00013	.00025	.00223	.00006	.00340	.00049	.00033	.00049	.00003
%RSD	.44721	.51984	.24820	.37854	.58096	.09820	.29846	.21932	1.4097
#1	.02981	.04747	.90134	.01593	.58293	.49596	.11024	.22126	.00222
#2	.02983	.04705	.89821	.01583	.58909	.49634	.11005	.22222	.00218
#3	.02959	.04705	.89702	.01583	.58351	.49693	.11069	.22163	.00225

Elem	P_1774	S_1820	Li6707	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.03315**	**.02146**	**.59980**	**6.0100**
Stddev	.00010	.00010	.00088	.0167
%RSD	.29369	.46150	.14594	.27739
#1	.03306	.02141	.59977	6.0245
#2	.03312	.02139	.60069	6.0137
#3	.03325	.02157	.59894	5.9918

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: S2 Acquired: 8/26/2025 10:58:08 Type: Cal
Method: 6010-200.7 NEW(v130) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6326.2	108020.	10206.	4959.7	7476.4
Stddev	12.4	332.	59.	9.5	12.9
%RSD	.19581	.30764	.57742	.19105	.17221
#1	6333.5	107850.	10212.	4953.1	7479.7
#2	6333.2	107810.	10145.	4955.4	7487.2
#3	6311.9	108400.	10262.	4970.5	7462.1

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: S3 Acquired: 8/26/2025 11:02:13 Type: Cal
Method: 6010-200.7 NEW(v130) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934	Be2348	Cd2265
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.11531**	**.18632**	**.47320**	**.10635**	**.21582**	**.37450**	**11.521**	**.85824**	**3.5847**
Stddev	.00018	.00662	.00133	.00027	.00072	.00121	.057	.00452	.0095
%RSD	.15942	3.5510	.28173	.25239	.33420	.32224	.49573	.52631	.26351
#1	.11545	.19146	.47307	.10629	.21563	.37502	11.585	.86022	3.5767
#2	.11510	.18864	.47193	.10612	.21521	.37536	11.502	.85307	3.5823
#3	.11537	.17886	.47459	.10665	.21662	.37312	11.476	.86142	3.5951

Elem	Ca3736	Cr2677	Co2286	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.68396**	**.09243**	**1.2416**	**.38850**	**.00740**	**.49816**	**.13455**	**.72176**	**.11689**
Stddev	.00106	.00021	.0033	.00129	.00002	.00154	.00044	.00188	.00034
%RSD	.15533	.22464	.26796	.33141	.23870	.30947	.32410	.25988	.29417
#1	.68288	.09265	1.2391	.38796	.00741	.49967	.13442	.72057	.11719
#2	.68501	.09224	1.2404	.38757	.00741	.49823	.13503	.72078	.11651
#3	.68400	.09239	1.2454	.38997	.00738	.49658	.13419	.72392	.11697

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.09379**	**.14878**	**2.7873**	**.05069**	**1.7702**	**1.5420**	**.34605**	**.69320**	**.00689**
Stddev	.00079	.00010	.0019	.00008	.0099	.0032	.00048	.00210	.00006
%RSD	.84230	.06751	.06974	.16410	.56095	.20995	.13813	.30292	.85729
#1	.09459	.14884	2.7857	.05070	1.7711	1.5418	.34660	.69513	.00689
#2	.09375	.14866	2.7868	.05077	1.7598	1.5388	.34574	.69353	.00696
#3	.09301	.14884	2.7895	.05060	1.7796	1.5453	.34581	.69096	.00684

Elem	P_1774	S_1820	Li6707	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.11068**	**.06950**	**1.8571**	**18.847**
Stddev	.00043	.00050	.0070	.035
%RSD	.39209	.72191	.37799	.18332
#1	.11106	.07000	1.8645	18.885
#2	.11077	.06951	1.8561	18.839
#3	.11021	.06900	1.8505	18.818

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: S3 Acquired: 8/26/2025 11:02:13 Type: Cal
Method: 6010-200.7 NEW(v130) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6202.5	105810.	10396.	4859.5	7130.0
Stddev	13.1	181.	36.	13.1	9.4
%RSD	.21195	.17096	.34209	.26854	.13208
#1	6201.8	105680.	10391.	4844.6	7129.9
#2	6216.0	105730.	10434.	4868.9	7139.4
#3	6189.7	106020.	10364.	4864.9	7120.6

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: S4 Acquired: 8/26/2025 11:06:20 Type: Cal
Method: 6010-200.7 NEW(v130) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934	Be2348	Cd2265
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.23830	.38718	.97046	.21943	.43585	.74639	22.778	1.7436	7.4382
Stddev	.00089	.01360	.00122	.00119	.00061	.00100	.184	.0053	.0121
%RSD	.37526	3.5124	.12599	.54384	.13941	.13428	.80817	.30109	.16200
#1	.23732	.37482	.96921	.21834	.43563	.74633	22.698	1.7441	7.4252
#2	.23851	.38498	.97050	.21924	.43539	.74542	22.989	1.7486	7.4402
#3	.23907	.40175	.97166	.22071	.43654	.74742	22.647	1.7381	7.4491

Elem	Ca3736	Cr2677	Co2286	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1.3551	.18844	2.5351	.77477	.01481	.98515	.26835	1.4741	.23686
Stddev	.0022	.00068	.0025	.00099	.00007	.00284	.00100	.0021	.00028
%RSD	.16235	.36131	.10037	.12735	.47448	.28790	.37339	.14252	.11725
#1	1.3527	.18815	2.5327	.77413	.01472	.98188	.26722	1.4718	.23683
#2	1.3569	.18922	2.5348	.77427	.01484	.98685	.26911	1.4744	.23715
#3	1.3557	.18796	2.5378	.77590	.01485	.98672	.26873	1.4760	.23660

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.18084	.30028	5.5829	.10047	3.6184	3.0872	.69273	1.3756	.01378
Stddev	.00112	.00059	.0084	.00052	.0083	.0069	.00215	.0027	.00009
%RSD	.62019	.19788	.15059	.51283	.22813	.22460	.30997	.19281	.65265
#1	.17961	.29962	5.5912	.09988	3.6227	3.0804	.69037	1.3730	.01368
#2	.18180	.30078	5.5831	.10085	3.6235	3.0870	.69328	1.3754	.01381
#3	.18112	.30043	5.5744	.10067	3.6088	3.0943	.69455	1.3783	.01386

Elem	P_1774	S_1820	Li6707	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.22999	.14472	3.7256	37.340
Stddev	.00232	.00159	.0078	.233
%RSD	1.0074	1.1006	.21067	.62360
#1	.22765	.14326	3.7167	37.592
#2	.23005	.14449	3.7316	37.297
#3	.23228	.14642	3.7284	37.132

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: S4 Acquired: 8/26/2025 11:06:20 Type: Cal
Method: 6010-200.7 NEW(v130) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6126.4	103930.	10156.	4745.9	6904.1
Stddev	11.1	374.	48.	17.5	9.9
%RSD	.18162	.35980	.47135	.36801	.14341
#1	6118.3	103890.	10125.	4744.9	6895.3
#2	6121.8	103580.	10132.	4729.0	6902.1
#3	6139.1	104320.	10211.	4763.8	6914.8

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: S5 Acquired: 8/26/2025 11:10:32 Type: Cal
Method: 6010-200.7 NEW(v130) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934	Be2348	Cd2265
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.47344**	**.74393**	**1.9553**	**.43515**	**.88215**	**1.5070**	**46.188**	**3.3539**	**14.906**
Stddev	.00089	.02123	.0028	.00123	.00256	.0022	.327	.0101	.017
%RSD	.18840	2.8532	.14455	.28298	.28990	.14867	.70880	.30236	.11304
#1	.47406	.73906	1.9578	.43583	.88511	1.5060	46.145	3.3655	14.925
#2	.47384	.76717	1.9522	.43588	.88067	1.5054	45.884	3.3490	14.895
#3	.47241	.72556	1.9559	.43372	.88068	1.5095	46.535	3.3471	14.897

Elem	Ca3736	Cr2677	Co2286	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**2.6953**	**.36528**	**5.1838**	**1.5364**	**.02984**	**1.9583**	**.53529**	**2.9904**	**.47410**
Stddev	.0054	.00159	.0087	.0038	.00023	.0036	.00040	.0040	.00032
%RSD	.20132	.43584	.16838	.24540	.77134	.18435	.07566	.13449	.06832
#1	2.6895	.36398	5.1932	1.5406	.02958	1.9550	.53532	2.9948	.47386
#2	2.6961	.36705	5.1761	1.5351	.03003	1.9622	.53567	2.9868	.47447
#3	2.7003	.36480	5.1819	1.5334	.02991	1.9576	.53486	2.9898	.47397

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.38203**	**.60655**	**10.985**	**.20922**	**6.9917**	**6.1943**	**1.4104**	**2.7916**	**.02872**
Stddev	.00348	.00186	.010	.00218	.0193	.0176	.0027	.0044	.00029
%RSD	.91013	.30654	.09018	1.0414	.27612	.28411	.18858	.15826	1.0270
#1	.37829	.60476	10.976	.20690	7.0129	6.2090	1.4128	2.7872	.02838
#2	.38516	.60642	10.984	.21123	6.9751	6.1990	1.4110	2.7960	.02892
#3	.38265	.60847	10.996	.20953	6.9870	6.1748	1.4075	2.7918	.02885

Elem	P_1774	S_1820	Li6707	Sr4077
Units	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**.47490**	**.30412**	**7.4512**	**74.716**
Stddev	.00228	.00180	.0247	.102
%RSD	.48090	.59330	.33170	.13657
#1	.47496	.30365	7.4250	74.599
#2	.47715	.30611	7.4741	74.781
#3	.47258	.30259	7.4545	74.770

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: S5 Acquired: 8/26/2025 11:10:32 Type: Cal
Method: 6010-200.7 NEW(v130) Mode: IR Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5985.5	104070.	10464.	4654.6	6563.4
Stddev	14.8	403.	41.	12.7	9.0
%RSD	.24720	.38719	.39477	.27179	.13759
#1	5973.8	104310.	10422.	4660.0	6557.9
#2	5980.5	103600.	10467.	4640.1	6558.4
#3	6002.1	104300.	10504.	4663.6	6573.8

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: ICV01 Acquired: 8/26/2025 11:14:43 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICV01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**3.894584**	**3.774081**	**3.968343**	**3.927521**	**4.053187**	**7.859028**
Stddev	.005916	.027274	.003276	.003711	.004631	.023824
%RSD	.1518956	.7226589	.0825555	.0944987	.1142466	.3031373
#1	3.901250	3.759469	3.972125	3.931581	4.058186	7.832980
#2	3.889957	3.757228	3.966367	3.924303	4.052332	7.879712
#3	3.892546	3.805548	3.966537	3.926680	4.049044	7.864393

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**7.804942**	**.1970261**	**1.924887**	**19.65661**	**.8003681**	**1.963537**
Stddev	.034654	.0007789	.002319	.01685	.0012083	.001503
%RSD	.4439992	.3953385	.1204854	.0857379	.1509653	.0765323
#1	7.838691	.1978425	1.927021	19.65802	.8002157	1.965004
#2	7.769449	.1962910	1.922419	19.63909	.7992433	1.962001
#3	7.806685	.1969448	1.925219	19.67271	.8016454	1.963606

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.008607**	**3.957371**	**1.961886**	**19.56305**	**1.962158**	**.9722079**
Stddev	.001285	.026153	.007300	.06541	.002831	.0031085
%RSD	.1274481	.6608624	.3720972	.3343781	.1442672	.3197406
#1	1.009795	3.931608	1.956353	19.52411	1.965044	.9743549
#2	1.007242	3.956609	1.959145	19.52647	1.959386	.9686432
#3	1.008784	3.983897	1.970160	19.63857	1.962043	.9736256

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**19.59513**	**1.958050**	**1.996460**	**19.69779**	**3.826285**	**3.934074**
Stddev	.08535	.009983	.006052	.08012	.022522	.005402
%RSD	.4355841	.5098478	.3031434	.4067643	.5886121	.1373168
#1	19.53138	1.949863	1.991425	19.61368	3.850751	3.938865
#2	19.56192	1.955116	1.994780	19.70645	3.806416	3.928219
#3	19.69210	1.969172	2.003175	19.77322	3.821687	3.935138

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: ICV01 Acquired: 8/26/2025 11:14:43 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICV01 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	3.840507	3.862977	3.664361	F .0051477	F .0236531	F -.006668
Stddev	.005854	.002801	.029076	.0010780	.0017134	.000864
%RSD	.1524369	.0725060	.7934676	20.94165	7.243823	12.95655
#1	3.847076	3.860052	3.635043	.0046162	.0219220	-.006520
#2	3.835843	3.863242	3.664851	.0044387	.0253482	-.007596
#3	3.838600	3.865635	3.693188	.0063883	.0236891	-.005887

Elem	Sr4077
Units	ppm
Avg	F -.003249
Stddev	.000036
%RSD	1.092476
#1	-.003226
#2	-.003232
#3	-.003290

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6248.981	107085.5	10782.03	4833.319	7144.051
Stddev	9.665	259.1	17.09	24.130	10.652
%RSD	.1546690	.2419788	.1585099	.4992384	.1491062
#1	6250.809	107300.8	10765.24	4848.588	7149.890
#2	6257.602	107157.8	10799.41	4845.869	7150.507
#3	6238.532	106797.9	10781.42	4805.501	7131.756

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: LLICV01 Acquired: 8/26/2025 11:20:39 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: LLICV01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0222122**	**.0398567**	**.0118934**	**.0204705**	**.0502331**	**.1051967**	**.0957843**
Stddev	.0005670	.0024119	.0002586	.0014563	.0017438	.0120907	.0009888
%RSD	2.552629	6.051415	2.174210	7.114330	3.471391	11.49344	1.032280
#1	.0222317	.0426064	.0116087	.0194479	.0516634	.1027089	.0969160
#2	.0227693	.0380992	.0121138	.0198256	.0507454	.0945435	.0950875
#3	.0216358	.0388644	.0119577	.0221379	.0482905	.1183379	.0953495

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0064943**	**.0057497**	**2.058814**	**.0098785**	**.0295421**	**.0217526**	**.0971202**
Stddev	.0000740	.0001264	.025530	.0001014	.0000632	.0001597	.0029684
%RSD	1.139612	2.198697	1.240017	1.026747	.2139015	.7339710	3.056444
#1	.0065163	.0057202	2.070961	.0099571	.0296056	.0217349	.0997956
#2	.0065548	.0056406	2.076002	.0099144	.0294792	.0219203	.0939269
#3	.0064117	.0058883	2.029479	.0097640	.0295414	.0216025	.0976381

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0213685**	**2.172632**	**.0396880**	**.0107575**	**1.799637**	**.0407942**	**.0419514**
Stddev	.0007050	.013035	.0001895	.0001751	.026447	.0014491	.0002211
%RSD	3.299345	.5999776	.4774752	1.627915	1.469596	3.552254	.5269976
#1	.0221701	2.165008	.0399015	.0109468	1.783475	.0404296	.0419863
#2	.0210906	2.187683	.0395398	.0107242	1.785278	.0395622	.0417150
#3	.0208447	2.165204	.0396227	.0106013	1.830158	.0423908	.0421530

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.812458**	**.1110962**	**.2065451**	**.0392545**	**.0383987**	**.3693868**	**.0178293**
Stddev	.012624	.0002635	.0003656	.0007408	.0016698	.0082413	.0001879
%RSD	.6965027	.2372066	.1770152	1.887220	4.348512	2.231078	1.053746
#1	1.797894	.1113377	.2062757	.0394983	.0396492	.3729799	.0176802
#2	1.819201	.1108151	.2069613	.0384225	.0365025	.3599589	.0180403
#3	1.820278	.1111358	.2063984	.0398427	.0390445	.3752215	.0177674

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: LLICV01 Acquired: 8/26/2025 11:20:39 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: LLICV01 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0199467	.0199020	.0198405
Stddev	.0022616	.0018495	.0001732
%RSD	11.33809	9.293227	.8726932
#1	.0208352	.0180314	.0199813
#2	.0216291	.0217297	.0198929
#3	.0173758	.0199451	.0196471

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6376.351	109371.0	10432.90	4892.646	7719.658
Stddev	6.166	183.2	13.57	11.567	15.286
%RSD	.0966972	.1675488	.1300792	.2364069	.1980085
#1	6372.706	109246.0	10444.59	4903.437	7711.557
#2	6383.470	109581.4	10418.02	4880.435	7737.289
#3	6372.877	109285.7	10436.07	4894.067	7710.127

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: ICB01 Acquired: 8/26/2025 11:24:54 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICB01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0023291**	**.0004050**	**.0006538**	**-.001645**	**-.000838**	**.0098669**	**-.001269**
Stddev	.0009860	.0006529	.0002411	.002099	.000252	.0084747	.000560
%RSD	42.33337	161.1960	36.88217	127.5721	30.08221	85.89017	44.11264
#1	.0031421	.0010282	.0008353	.000666	-.000978	.0040261	-.001853
#2	.0026127	.0004609	.0007458	-.003433	-.000547	.0195869	-.001217
#3	.0012324	-.000274	.0003802	-.002168	-.000990	.0059877	-.000737

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0000032**	**.0000364**	**.0006487**	**-.000167**	**-.000145**	**.0001436**	**.0001811**
Stddev	.0000495	.0000391	.0034420	.000188	.000082	.0001066	.0053914
%RSD	1523.145	107.3974	530.5716	112.6972	56.38067	74.22093	2977.851
#1	.0000431	.0000248	.0026038	-.000133	-.000117	.0001474	.0015386
#2	-.000052	.0000044	.0026680	.000002	-.000081	.0000352	.0047639
#3	.000019	.0000799	-.003326	-.000369	-.000237	.0002482	-.005759

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000222**	**-.003105**	**-.000244**	**-.000498**	**-.054306**	**-.000857**	**-.000734**
Stddev	.000486	.028816	.000134	.000472	.013177	.003191	.000199
%RSD	218.4037	928.0086	55.11455	94.78074	24.26449	372.4909	27.04564
#1	-.000271	-.036167	-.000140	-.000211	-.040329	.002810	-.000964
#2	-.000682	.010184	-.000395	-.000240	-.066502	-.003004	-.000615
#3	.000286	.016668	-.000195	-.001042	-.056085	-.002376	-.000625

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.075197**	**.0042250**	**-.000034**	**.0083757**	**-.000622**	**.0166935**	**-.000271**
Stddev	.010194	.0002901	.000287	.0004323	.000640	.0064255	.000564
%RSD	13.55600	6.867428	844.4457	5.161364	103.0050	38.49095	207.8419
#1	-.074918	.0043468	.000297	.0084510	-.000982	.0191148	-.000767
#2	-.085528	.0044344	-.000214	.0079108	.000118	.0094092	.000342
#3	-.065146	.0038938	-.000185	.0087655	-.001001	.0215566	-.000389

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: ICB01 Acquired: 8/26/2025 11:24:54 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICB01 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0009292	-.003549	-.000123
Stddev	.0014027	.001074	.000020
%RSD	150.9458	30.25820	16.57322
#1	.0023839	-.002505	-.000119
#2	-.000415	-.003492	-.000105
#3	.000819	-.004650	-.000145

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6371.352	108540.2	10463.73	4858.148	7820.737
Stddev	8.382	215.2	88.65	25.004	5.874
%RSD	.1315604	.1982379	.8471717	.5146796	.0751054
#1	6374.706	108498.0	10482.51	4843.027	7826.600
#2	6361.812	108349.2	10367.20	4844.407	7814.853
#3	6377.538	108773.3	10541.48	4887.009	7820.758

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CRI01 Acquired: 8/26/2025 11:29:12 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CRI01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0225846	.0419983	.0125047	.0197097	.0508653	.1050243	.0975178
Stddev	.0003952	.0013166	.0005657	.0023705	.0016501	.0036706	.0007030
%RSD	1.749898	3.134881	4.523484	12.02704	3.244065	3.494962	.7208431
#1	.0229177	.0415351	.0121187	.0183490	.0510363	.1044534	.0974472
#2	.0221479	.0409760	.0131540	.0183331	.0524232	.1016727	.0968528
#3	.0226882	.0434840	.0122415	.0224468	.0491363	.1089469	.0982534

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0064045	.0057971	2.064578	.0097813	.0295043	.0217802	.0967880
Stddev	.0000451	.0000351	.010093	.0001066	.0001195	.0006108	.0064560
%RSD	.7037666	.6054228	.4888738	1.089752	.4048522	2.804266	6.670303
#1	.0064373	.0057798	2.071342	.0098981	.0295534	.0217128	.0893557
#2	.0063531	.0058375	2.052977	.0096892	.0293681	.0212060	.1010054
#3	.0064232	.0057740	2.069415	.0097565	.0295914	.0224219	.1000029

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0215073	2.221921	.0397676	.0103118	1.820110	.0380299	.0415651
Stddev	.0003076	.010224	.0003618	.0002536	.023569	.0019230	.0001245
%RSD	1.430214	.4601601	.9098383	2.459051	1.294944	5.056601	.2996600
#1	.0212964	2.226979	.0401449	.0105553	1.802461	.0382698	.0416175
#2	.0218603	2.210154	.0397342	.0100492	1.846876	.0359982	.0416548
#3	.0213653	2.228631	.0394236	.0103310	1.810993	.0398217	.0414229

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.809069	.1072900	.2065888	.0390664	.0396256	.3707451	.0183256
Stddev	.046862	.0006801	.0007864	.0002889	.0011704	.0032230	.0004076
%RSD	2.590392	.6338643	.3806658	.7395109	2.953575	.8693434	2.224024
#1	1.768447	.1080714	.2074551	.0388628	.0400314	.3723199	.0178970
#2	1.860338	.1069665	.2059199	.0393971	.0383063	.3728779	.0183715
#3	1.798423	.1068320	.2063914	.0389393	.0405390	.3670374	.0187083

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CRI01 Acquired: 8/26/2025 11:29:12 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CRI01 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0192800	.0169419	.0197806
Stddev	.0023752	.0007505	.0000272
%RSD	12.31961	4.430085	.1375546
#1	.0165459	.0162902	.0197807
#2	.0208347	.0167730	.0198078
#3	.0204595	.0177625	.0197533

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6368.709	109155.5	10500.46	4922.929	7713.939
Stddev	6.230	396.7	67.97	4.180	5.811
%RSD	.0978273	.3633899	.6473420	.0849162	.0753341
#1	6370.134	109528.5	10441.63	4922.581	7720.647
#2	6374.104	108738.8	10574.87	4918.932	7710.744
#3	6361.890	109199.1	10484.89	4927.271	7710.426

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: ICSA01 Acquired: 8/26/2025 11:33:27 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSA01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0064697**	**-.001441**	**.0032318**	**.0019420**	**-.002270**	**248.9648**	**.0013380**
Stddev	.0010099	.001751	.0020358	.0012706	.004101	.4406	.0001471
%RSD	15.60961	121.5339	62.99224	65.42920	180.6589	.1769859	10.99098
#1	.0066439	-.002291	.0011795	.0019458	.000575	249.0213	.0011768
#2	.0073813	.000573	.0032654	.0006694	-.006970	249.3744	.0013724
#3	.0053841	-.002603	.0052506	.0032107	-.000414	248.4986	.0014648

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0015442**	**.0021087**	**235.5287**	**.0607078**	**.0019381**	**.0076338**	**99.17567**
Stddev	.0000707	.0000180	.1840	.0008243	.0000478	.0009923	.75985
%RSD	4.578226	.8538229	.0781029	1.357738	2.467671	12.99854	.7661667
#1	.0016253	.0021132	235.4049	.0612143	.0019933	.0065777	99.87232
#2	.0015116	.0021239	235.7400	.0611525	.0019095	.0077771	99.28927
#3	.0014957	.0020888	235.4410	.0597567	.0019114	.0085467	98.36541

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0083559**	**250.2836**	**.0058631**	**.0014484**	**.1096759**	**.0034571**	**.0026990**
Stddev	.0001073	.4208	.0002598	.0006001	.0193110	.0030900	.0003828
%RSD	1.284385	.1681245	4.430358	41.43037	17.60735	89.38313	14.18451
#1	.0082379	249.9322	.0061630	.0013114	.0905256	.0002399	.0029993
#2	.0083821	250.7499	.0057146	.0021051	.1291438	.0064019	.0028297
#3	.0084476	250.1687	.0057117	.0009286	.1093583	.0037295	.0022679

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1635780**	**.0495108**	**.0013896**	**-.001387**	**-.003268**	**.0190345**	**.0084839**
Stddev	.0337631	.0027570	.0002789	.000170	.000691	.0181828	.0022689
%RSD	20.64035	5.568531	20.07213	12.26153	21.15116	95.52565	26.74380
#1	.1246927	.0526793	.0015960	-.001536	-.003622	.0235774	.0097410
#2	.1854486	.0481941	.0015005	-.001422	-.003710	.0345152	.0058647
#3	.1805927	.0476589	.0010723	-.001202	-.002471	-.000989	.0098460

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: ICSA01 Acquired: 8/26/2025 11:33:27 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSA01 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	-.009639	F -.031216	.0015828
Stddev	.003169	.000520	.0004063
%RSD	32.87336	1.667186	25.66907
#1	-.005993	-.030669	.0011533
#2	-.011193	-.031704	.0016342
#3	-.011730	-.031276	.0019610

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5630.100	95734.17	10238.54	4351.261	6114.422
Stddev	4.082	396.80	39.14	9.710	4.319
%RSD	.0724985	.4144784	.3823038	.2231516	.0706349
#1	5633.571	95402.51	10282.28	4360.413	6117.254
#2	5631.126	95626.24	10226.52	4341.075	6116.560
#3	5625.603	96173.77	10206.82	4352.295	6109.450

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: ICSAB01 Acquired: 8/26/2025 11:37:38 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSAB01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1071488**	**.1087484**	**.0505699**	**.0513156**	**.5969946**	**249.8381**
Stddev	.0034265	.0003073	.0007304	.0024968	.0022613	.2708
%RSD	3.197908	.2826182	1.444313	4.865557	.3787823	.1083880
#1	.1056540	.1089502	.0498316	.0535374	.5983851	249.6495
#2	.1047237	.1083947	.0512921	.0517958	.5982134	249.7164
#3	.1110688	.1089003	.0505861	.0486136	.5943854	250.1484

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.4899446**	**.4949417**	**.9797164**	**236.4588**	**.5600968**	**.4946421**
Stddev	.0017809	.0031351	.0013093	.3371	.0014879	.0003855
%RSD	.3634825	.6334317	.1336422	.1425574	.2656409	.0779321
#1	.4887733	.4928449	.9794065	236.6003	.5585987	.4942648
#2	.4890665	.4934345	.9785898	236.0741	.5615741	.4946261
#3	.4919940	.4985458	.9811528	236.7021	.5601177	.4950353

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.4860813**	**100.2495**	**.4903575**	**250.3288**	**.9830581**	**.2170106**
Stddev	.0007753	.4251	.0007821	.5484	.0004569	.0008519
%RSD	.1594955	.4240487	.1595019	.2190570	.0464813	.3925730
#1	.4857054	100.3054	.4896098	249.9729	.9828693	.2163434
#2	.4855657	100.6439	.4902928	250.0533	.9827259	.2167182
#3	.4869729	99.7992	.4911700	250.9603	.9835792	.2179703

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0864728**	**.4820516**	**1.038652**	**-.126461**	**.9629270**	**.9812571**
Stddev	.0188330	.0008931	.001413	.034533	.0024948	.0010093
%RSD	21.77906	.1852764	.1360582	27.30760	.2590801	.1028559
#1	.0788262	.4810450	1.039469	-.166184	.9602071	.9822347
#2	.0726658	.4823606	1.039467	-.109620	.9634652	.9813177
#3	.1079264	.4827492	1.037021	-.103579	.9651087	.9802189

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: ICSAB01 Acquired: 8/26/2025 11:37:38 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSAB01 Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.9680235	.9560483	1.000771	F .0091219	F .0016213	F -.033850
Stddev	.0013551	.0047175	.016606	.0042964	.0009585	.000818
%RSD	.1399864	.4934372	1.659308	47.09935	59.12190	2.416538
#1	.9672218	.9524717	.990536	.0125519	.0026255	-.034790
#2	.9695881	.9542784	1.019931	.0105110	.0015223	-.033457
#3	.9672606	.9613948	.991845	.0043029	.0007161	-.033303

Elem	Sr4077
Units	ppm
Avg	F .0015510
Stddev	.0009030
%RSD	58.21887
#1	.0011244
#2	.0009404
#3	.0025883

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5680.441	96694.58	10408.86	4414.042	6178.742
Stddev	13.004	276.05	37.43	12.861	16.172
%RSD	.2289210	.2854851	.3595647	.2913664	.2617386
#1	5675.012	96774.14	10424.37	4417.566	6175.300
#2	5695.279	96387.48	10436.04	4399.786	6196.358
#3	5671.031	96922.10	10366.17	4424.773	6164.568

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: ICSADLX20 Acquired: 8/26/2025 11:41:39 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSA01DLX20 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0001185	.0002361	.0001413	-.000699	-.000619	12.20673	-.001751
Stddev	.0011132	.0024133	.0005793	.003730	.002511	.01212	.000736
%RSD	939.4032	1022.185	410.0745	533.4295	405.3990	.0992650	42.02418
#1	.0013146	-.001635	.0000812	.002684	-.003515	12.21674	-.002000
#2	-.000887	.002960	-.000406	-.004699	.000966	12.21018	-.000923
#3	-.000072	-.000617	.000748	-.000083	.000691	12.19326	-.002331

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0000897	.0001235	12.10111	.0029190	.0000282	.0012425	4.763286
Stddev	.0000563	.0000108	.03139	.0000686	.0000343	.0003536	.005447
%RSD	62.80099	8.748329	.2593904	2.351567	121.4296	28.45748	.1143528
#1	.0001105	.0001354	12.11653	.0028424	.0000168	.0008435	4.766514
#2	.0001326	.0001143	12.12180	.0029749	.0000667	.0013668	4.766347
#3	.0000259	.0001209	12.06499	.0029398	.0000011	.0015171	4.756997

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0005563	12.56389	-.000043	.0000693	-.111648	-.000209	-.001190
Stddev	.0003004	.02232	.000008	.0003429	.015234	.001799	.000231
%RSD	53.99680	.1776645	18.83426	494.4790	13.64442	860.6371	19.42824
#1	.0008726	12.58246	-.000040	-.000073	-.128681	-.001153	-.001307
#2	.0005213	12.53912	-.000036	-.000180	-.099327	.001866	-.000923
#3	.0002749	12.57009	-.000052	.000460	-.106935	-.001340	-.001339

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.321335	.0014971	.0001638	-.001068	-.001285	.0090303	-.000053
Stddev	.055758	.0003386	.0001058	.000492	.001064	.0035510	.001330
%RSD	17.35209	22.61630	64.61166	46.03860	82.85762	39.32339	2502.333
#1	-.318804	.0016221	.0000606	-.000968	-.001840	.0060993	.000409
#2	-.378316	.0017554	.0001587	-.001602	-.001956	.0129791	-.001553
#3	-.266885	.0011138	.0002721	-.000634	-.000057	.0080125	.000984

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: ICSADLX20 Acquired: 8/26/2025 11:41:39 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSA01DLX20 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0018360	-.008240	.0000968
Stddev	.0023937	.001828	.0000856
%RSD	130.3736	22.18930	88.50418
#1	.0033327	-.009531	.0000684
#2	-.000925	-.009041	.0001930
#3	.003100	-.006147	.0000289

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6232.144	105610.4	10393.57	4782.270	7440.230
Stddev	3.829	208.7	40.18	19.676	10.025
%RSD	.0614456	.1975818	.3865533	.4114433	.1347340
#1	6231.283	105576.9	10438.23	4759.744	7440.293
#2	6228.820	105420.5	10360.37	4790.962	7430.174
#3	6236.331	105833.7	10382.11	4796.103	7450.223

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: ICSABDLX20 Acquired: 8/26/2025 11:45:55 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSAB01DLX20 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0053195	.0057690	.0030781	.0029730	.0268374	12.43064	.0240952
Stddev	.0005006	.0006448	.0006871	.0011348	.0005634	.02993	.0001427
%RSD	9.409867	11.17648	22.32346	38.17037	2.099287	.2408059	.5922589
#1	.0053383	.0057845	.0036859	.0020383	.0272572	12.46101	.0242392
#2	.0058105	.0064058	.0023325	.0026449	.0270579	12.42976	.0239538
#3	.0048099	.0051166	.0032160	.0042356	.0261971	12.40116	.0240927

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0258992	.0474067	12.32422	.0283135	.0240120	.0257215	4.824061
Stddev	.0001182	.0001387	.03432	.0001110	.0000913	.0003146	.028189
%RSD	.4561759	.2924886	.2784577	.3920420	.3802443	1.223122	.5843431
#1	.0258916	.0472468	12.34324	.0283549	.0239071	.0259843	4.856610
#2	.0257850	.0474946	12.28460	.0283978	.0240558	.0258072	4.807989
#3	.0260209	.0474786	12.34482	.0281877	.0240732	.0253729	4.807585

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0255591	12.81033	.0482646	.0101548	-.114778	.0230611	.0502622
Stddev	.0001527	.01448	.0002681	.0004834	.008630	.0009049	.0000734
%RSD	.5974747	.1130467	.5555886	4.760208	7.518821	3.923933	.1461401
#1	.0257354	12.82694	.0484535	.0104655	-.116494	.0237391	.0503465
#2	.0254725	12.80371	.0479577	.0104010	-.105419	.0220335	.0502286
#3	.0254693	12.80034	.0483825	.0095978	-.122421	.0234106	.0502116

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.346975	.0489501	.0497580	.0461088	.0452142	.0551672	.0007339
Stddev	.018759	.0003606	.0002550	.0000712	.0008653	.0114003	.0007669
%RSD	5.406475	.7365877	.5125845	.1544474	1.913866	20.66500	104.4878
#1	-.326772	.0490535	.0497825	.0460267	.0461234	.0536615	.0011052
#2	-.350313	.0485491	.0494916	.0461459	.0444007	.0672456	.0012445
#3	-.363842	.0492476	.0499999	.0461538	.0451184	.0445946	-.000148

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: ICSABDLX20 Acquired: 8/26/2025 11:45:55 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: ICSAB01DLX20 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0012198	-.007322	.0001870
Stddev	.0010965	.000897	.0000656
%RSD	89.89425	12.24643	35.05691
#1	.0021757	-.007937	.0001195
#2	.0000228	-.006293	.0001912
#3	.0014608	-.007735	.0002504

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6238.803	106697.4	10472.78	4797.822	7438.389
Stddev	8.310	604.3	44.78	8.794	11.336
%RSD	.1332015	.5663635	.4275958	.1833011	.1523959
#1	6234.998	106239.3	10426.12	4787.715	7436.276
#2	6248.334	106470.6	10515.41	4803.735	7450.632
#3	6233.076	107382.3	10476.83	4802.015	7428.258

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCV01 Acquired: 8/26/2025 11:50:07 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**5.038747**	**5.029405**	**4.992721**	**5.073722**	**4.975816**	**9.950866**	**9.610912**
Stddev	.016341	.087962	.005997	.009261	.012787	.026170	.073164
%RSD	.3242976	1.748962	.1201052	.1825224	.2569878	.2629952	.7612644
#1	5.025439	5.065313	4.993525	5.065583	4.971290	9.941550	9.677381
#2	5.033819	4.929169	4.986363	5.071786	4.965907	9.930628	9.622839
#3	5.056985	5.093733	4.998275	5.083798	4.990251	9.980420	9.532517

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2548067**	**2.511350**	**24.77373**	**1.038106**	**2.476666**	**1.253494**	**5.035431**
Stddev	.0005410	.004843	.05119	.003499	.004056	.002269	.017599
%RSD	.2122999	.1928283	.2066255	.3370583	.1637788	.1810018	.3495104
#1	.2545254	2.513036	24.83106	1.037403	2.477983	1.251812	5.040314
#2	.2544644	2.505890	24.75756	1.035013	2.472115	1.252595	5.050072
#3	.2554304	2.515124	24.73259	1.041904	2.479900	1.256075	5.015905

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.446300**	**24.81528**	**2.485940**	**1.267866**	**24.17918**	**2.479577**	**2.563168**
Stddev	.006516	.06239	.003075	.000706	.05109	.003600	.007255
%RSD	.2663625	.2514331	.1237091	.0557204	.2113045	.1451902	.2830318
#1	2.452422	24.88673	2.486743	1.268596	24.23817	2.483089	2.555644
#2	2.447026	24.77161	2.482542	1.267186	24.15043	2.479748	2.563742
#3	2.439451	24.78748	2.488533	1.267816	24.14895	2.475894	2.570119

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**24.48354**	**5.039807**	**5.015666**	**4.917033**	**4.876764**	**5.005014**	**4.977055**
Stddev	.03220	.004380	.007797	.013408	.004946	.013011	.016009
%RSD	.1315230	.0869148	.1554515	.2726857	.1014288	.2599514	.3216626
#1	24.51745	5.038508	5.011072	4.920870	4.882405	4.990171	4.965706
#2	24.47979	5.044690	5.011257	4.902125	4.874720	5.014445	4.970091
#3	24.45337	5.036223	5.024669	4.928104	4.873168	5.010426	4.995366

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCV01 Acquired: 8/26/2025 11:50:07 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV01 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**4.880136**	**4.893434**	**4.869114**
Stddev	.022609	.011403	.066643
%RSD	.4632883	.2330245	1.368688
#1	4.869830	4.905847	4.929105
#2	4.864516	4.883425	4.797380
#3	4.906062	4.891030	4.880857

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**6078.229**	**101847.7**	**10239.25**	**4581.321**	**6880.633**
Stddev	9.057	57.3	13.51	24.777	16.680
%RSD	.1490116	.0562400	.1319279	.5408214	.2424200
#1	6080.412	101823.5	10246.99	4602.527	6876.136
#2	6085.995	101913.1	10247.11	4587.349	6899.100
#3	6068.279	101806.6	10223.65	4554.086	6866.662

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCB01 Acquired: 8/26/2025 11:54:18 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB01 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0006176**	**-.000316**	**.0012330**	**-.002688**	**-.001407**	**-.006024**	**-.001247**
Stddev	.0012056	.000646	.0013598	.001618	.001769	.001494	.000650
%RSD	195.1988	204.7138	110.2840	60.21187	125.7176	24.79715	52.10370
#1	-.000627	-.000492	.0019182	-.003852	.000428	-.006137	-.000638
#2	.001780	.000400	.0021140	-.000840	-.001547	-.007458	-.001172
#3	.000700	-.000855	-.000333	-.003372	-.003103	-.004477	-.001931

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000036**	**.0000054**	**.0055376**	**-.000230**	**-.000028**	**.0002325**	**-.003010**
Stddev	.000045	.0000405	.0049817	.000291	.000169	.0001626	.005709
%RSD	125.2170	745.3766	89.96175	126.6720	597.7568	69.93893	189.6847
#1	.000016	-.000020	.0008153	-.000009	-.000088	.0002912	.001381
#2	-.000058	.000052	.0107435	-.000121	.000163	.0000487	-.000947
#3	-.000065	-.000016	.0050539	-.000560	-.000159	.0003577	-.009463

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0000236**	**-.001826**	**-.000046**	**.0005950**	**-.123941**	**-.000022**	**-.001106**
Stddev	.0001830	.011111	.000184	.0003311	.017251	.000720	.000141
%RSD	777.0131	608.6755	399.3327	55.65093	13.91879	3242.305	12.75970
#1	-.000036	-.011602	-.000186	.0003369	-.141979	-.000825	-.001269
#2	-.000122	.010258	.000162	.0004798	-.122240	.000567	-.001028
#3	.000229	-.004133	-.000114	.0009683	-.107603	.000191	-.001020

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.321034**	**.0041109**	**.0002390**	**.0000062**	**.0004556**	**-.001839**	**-.000313**
Stddev	.027096	.0003279	.0002213	.0002292	.0011314	.007108	.001297
%RSD	8.440143	7.976149	92.57317	3681.453	248.3040	386.4546	414.0173
#1	-.330160	.0044706	.0001657	-.000086	.0000315	.004561	-.000898
#2	-.342387	.0040335	.0004877	-.000162	-.000402	-.009490	.001173
#3	-.290553	.0038287	.0000637	.000267	.001738	-.000589	-.001215

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCB01 Acquired: 8/26/2025 11:54:18 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB01 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0015867	-.005190	-.000023
Stddev	.0031995	.001337	.000137
%RSD	201.6404	25.75878	605.7610
#1	.0042825	-.004061	-.000087
#2	-.001949	-.006667	.000135
#3	.002427	-.004843	-.000116

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6298.890	105925.8	10327.91	4728.146	7782.296
Stddev	11.833	447.2	34.06	12.273	10.993
%RSD	.1878559	.4222196	.3298124	.2595829	.1412574
#1	6307.807	105999.9	10320.83	4714.915	7786.528
#2	6303.397	105446.1	10297.95	4739.159	7790.544
#3	6285.466	106331.3	10364.96	4730.365	7769.816

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: LR-1 Acquired: 8/26/2025 11:58:36 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.001368**	**.0115265**	**.0162547**	**-.019656**	**.0180270**	**.0553082**	**-.011258**
Stddev	.002369	.0048783	.0018134	.000628	.0008793	.0091397	.000521
%RSD	173.1770	42.32230	11.15615	3.192484	4.877635	16.52504	4.625222
#1	.001084	.0102139	.0176669	-.019913	.0173533	.0522130	-.010855
#2	-.001543	.0074388	.0168874	-.020114	.0190217	.0481180	-.011074
#3	-.003644	.0169267	.0142097	-.018941	.0177061	.0655936	-.011846

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0001699**	**.0004733**	**2433.866**	**.0154361**	**-.000256**	**.0028594**	**.4403304**
Stddev	.0000324	.0000029	37.055	.0005855	.000346	.0004391	.0028997
%RSD	19.08992	.6108984	1.522460	3.793112	135.3350	15.35732	.6585362
#1	.0001860	.0004731	2452.143	.0152709	-.000140	.0023705	.4434947
#2	.0001911	.0004705	2458.230	.0149510	.000018	.0029872	.4378001
#3	.0001325	.0004763	2391.223	.0160865	-.000645	.0032204	.4396964

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0146244**	**2226.632**	**.0113796**	**-.001710**	**2461.046**	**.0075075**	**.0510873**
Stddev	.0001673	9.002	.0004009	.000435	20.513	.0007927	.0001894
%RSD	1.144124	.4043034	3.522805	25.46347	.8334894	10.55847	.3708179
#1	.0148104	2216.515	.0110830	-.001921	2453.238	.0076568	.0511260
#2	.0144863	2233.757	.0118357	-.001999	2484.316	.0082150	.0508815
#3	.0145764	2229.626	.0112202	-.001209	2445.584	.0066508	.0512544

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.799882**	**.0095920**	**.0030581**	**-.005542**	**-.115064**	**-.029845**	**.0105824**
Stddev	.909419	.0005507	.0004172	.001419	.001431	.023899	.0044874
%RSD	9.279894	5.741677	13.64203	25.61067	1.243681	80.07651	42.40460
#1	8.845260	.0098020	.0035156	-.004543	-.114958	-.057424	.0102043
#2	9.898295	.0089671	.0029600	-.007167	-.116545	-.015207	.0062959
#3	10.65609	.0100068	.0026987	-.004916	-.113689	-.016904	.0152469

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: LR-1 Acquired: 8/26/2025 11:58:36 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**.2194363**	**-.376783**	**.0612813**
Stddev	.0033417	.007008	.0000980
%RSD	1.522875	1.859979	.1598454
#1	.2232078	-.379065	.0611734
#2	.2182570	-.382365	.0613059
#3	.2168441	-.368918	.0613646

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**3982.085**	**70923.42**	**9025.904**	**3085.825**	**4093.430**
Stddev	5.923	207.68	27.359	21.848	8.855
%RSD	.1487516	.2928171	.3031162	.7080209	.2163161
#1	3988.489	71034.38	9048.581	3110.146	4103.543
#2	3980.965	70683.84	8995.517	3067.858	4089.681
#3	3976.802	71052.05	9033.614	3079.471	4087.067

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: LR-2 Acquired: 8/26/2025 12:03:15 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1092092**	**-.332027**	**.1015375**	**.0121388**	**-.159663**	**1945.588**	**.0219708**
Stddev	.0060098	.009069	.0098723	.0032177	.006660	30.287	.0001893
%RSD	5.503047	2.731494	9.722772	26.50757	4.171180	1.556676	.8618074
#1	.1097626	-.340421	.1087218	.0118775	-.164294	1962.665	.0217525
#2	.1029419	-.333252	.0902804	.0090598	-.152031	1910.620	.0220697
#3	.1149233	-.322407	.1056103	.0154792	-.162665	1963.480	.0220902

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0463009**	**.1199418**	**13.65306**	**-.042602**	**.0872013**	**.2910311**	**2677.860**
Stddev	.0006260	.0010629	.02130	.000228	.0002441	.0039471	5.158
%RSD	1.352077	.8861493	.1559789	.5344154	.2799132	1.356248	.1926152
#1	.0464407	.1200459	13.66137	-.042840	.0869494	.2889083	2679.809
#2	.0456168	.1209488	13.66894	-.042580	.0872178	.2885996	2681.759
#3	.0468452	.1188308	13.62886	-.042386	.0874367	.2955853	2672.011

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.186870**	**1.027932**	**.0607358**	**^ *****	**2.753588**	**.0471577**	**.1052004**
Stddev	.002344	.006020	.0026423	-----	.028067	.0017080	.0020276
%RSD	1.254562	.5856241	4.350418	-----	1.019291	3.621960	1.927368
#1	-.188540	1.020982	.0579310	^ -----	2.723624	.0467419	.1068602
#2	-.184190	1.031351	.0631781	^ -----	2.779265	.0456959	.1058004
#3	-.187880	1.031464	.0610983	^ -----	2.757874	.0490352	.1029405

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1805.060**	**^ *****	**-.016240**	**.0350511**	**.0038250**	**.5041804**	**.1134616**
Stddev	19.347	-----	.000718	.0024772	.0004196	.0135359	.0048997
%RSD	1.071803	-----	4.422716	7.067534	10.96968	2.684725	4.318337
#1	1826.004	^ -----	-.015895	.0336772	.0033416	.5068752	.1105748
#2	1787.858	^ -----	-.015759	.0335652	.0040380	.5161661	.1191188
#3	1801.317	^ -----	-.017065	.0379108	.0040954	.4894998	.1106912

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: LR-2 Acquired: 8/26/2025 12:03:15 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	-.369958	.2658079	-2.56687
Stddev	.012090	.0009131	.00496
%RSD	3.268030	.3435334	.1931278
#1	-.356027	.2647966	-2.56868
#2	-.377710	.2665720	-2.57067
#3	-.376138	.2660551	-2.56126

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5005.325	83275.62	9643.397	3837.245	4173.849
Stddev	13.589	247.02	56.643	27.505	8.070
%RSD	.2714993	.2966265	.5873716	.7167943	.1933527
#1	4993.097	82999.88	9620.076	3808.660	4172.360
#2	5002.924	83476.69	9707.977	3839.552	4166.628
#3	5019.955	83350.28	9602.139	3863.525	4182.561

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: LR-3 Acquired: 8/26/2025 12:07:43 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0989965**	**.0163850**	**227.8308**	**.0042162**	**.1240869**	**.0887592**	**-.001853**
Stddev	.0013200	.0014220	1.1166	.0039747	.0024317	.0057021	.000342
%RSD	1.333352	8.678885	.4900788	94.27307	1.959697	6.424178	18.46401
#1	.1002594	.0148645	228.9218	.0067222	.1240992	.0869212	-.001677
#2	.0991041	.0176822	226.6903	-.000367	.1216491	.0842029	-.001635
#3	.0976260	.0166084	227.8802	.006293	.1265125	.0951536	-.002248

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.004914**	**-.001641**	**.2310593**	**24.34153**	**.0060246**	**53.15555**	**.0614782**
Stddev	.000069	.000080	.0113610	.06744	.0000806	.01899	.0041665
%RSD	1.397728	4.847806	4.916939	.2770554	1.338114	.0357184	6.777233
#1	-.004994	-.001718	.2425676	24.28341	.0060024	53.15101	.0613858
#2	-.004874	-.001646	.2307588	24.32572	.0059573	53.13924	.0573587
#3	-.004875	-.001559	.2198514	24.41548	.0061139	53.17639	.0656902

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**49.26658**	**-.191912**	**23.81698**	**.0047206**	**.1653596**	**-.038913**	**30.32700**
Stddev	.37653	.035724	.00811	.0005314	.0231797	.002150	.13324
%RSD	.7642762	18.61491	.0340445	11.25763	14.01778	5.525792	.4393493
#1	49.69473	-.170316	23.82558	.0043279	.1386619	-.038414	30.17412
#2	48.98701	-.233147	23.80947	.0045086	.1770583	-.041269	30.41840
#3	49.11800	-.172272	23.81591	.0053254	.1803587	-.037056	30.38848

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.3375625**	**.0009308**	**-.014961**	**.0024782**	**-.045019**	**-.110460**	**.1485144**
Stddev	.0392656	.0005100	.000191	.0001208	.000672	.014627	.0005987
%RSD	11.63211	54.79351	1.274592	4.874353	1.493380	13.24227	.4031482
#1	.2972054	.0015084	-.015156	.0025055	-.045355	-.126519	.1478907
#2	.3756371	.0007418	-.014775	.0023461	-.044245	-.106963	.1490845
#3	.3398448	.0005423	-.014950	.0025830	-.045456	-.097898	.1485682

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: LR-3 Acquired: 8/26/2025 12:07:43 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0671256	-.003530	.0003850
Stddev	.0036813	.001181	.0000712
%RSD	5.484252	33.45919	18.49021
#1	.0628851	-.004338	.0003355
#2	.0689901	-.002175	.0003529
#3	.0695017	-.004079	.0004666

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5914.833	106934.0	10598.57	4585.900	7415.757
Stddev	27.286	325.3	57.47	21.254	36.968
%RSD	.4613075	.3041898	.5422517	.4634653	.4985080
#1	5883.775	107272.1	10533.30	4608.588	7373.124
#2	5934.950	106906.7	10620.83	4566.452	7438.938
#3	5925.774	106623.2	10641.58	4582.662	7435.209

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2832-09DLX10 Acquired: 8/26/2025 12:12:37 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0612712**	**-.010306**	**1.022550**	**.0070367**	**.0144197**	**8.098681**	**.5228869**
Stddev	.0024085	.001412	.005365	.0006244	.0005339	.003286	.0005351
%RSD	3.930936	13.69722	.5246490	8.873185	3.702801	.0405703	.1023305
#1	.0609025	-.008719	1.028503	.0071781	.0144398	8.102444	.5234977
#2	.0638429	-.011421	1.021058	.0063538	.0138760	8.097218	.5226616
#3	.0590684	-.010778	1.018090	.0075784	.0149433	8.096381	.5225013

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0025535**	**.0011574**	**10.09886**	**.0787054**	**.0147163**	**.5902098**	**92.11320**
Stddev	.0000483	.0000350	.01496	.0004562	.0001062	.0019679	.44722
%RSD	1.892188	3.026250	.1481632	.5796389	.7220119	.3334214	.4855111
#1	.0025287	.0011979	10.10802	.0791049	.0148302	.5916535	92.47388
#2	.0026092	.0011370	10.08159	.0788031	.0146989	.5910076	92.25293
#3	.0025226	.0011373	10.10696	.0782083	.0146198	.5879683	91.61280

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.3902514**	**2.247641**	**.0544981**	**.0030884**	**.6601427**	**.0632563**	**.2395813**
Stddev	.0021063	.032742	.0002222	.0006655	.0299888	.0028097	.0003469
%RSD	.5397330	1.456738	.4076469	21.54692	4.542782	4.441819	.1447951
#1	.3916995	2.211108	.0544398	.0025139	.6409492	.0652357	.2394200
#2	.3878351	2.257476	.0547436	.0038175	.6447788	.0644927	.2399795
#3	.3912196	2.274338	.0543109	.0029338	.6947002	.0600403	.2393444

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.542829**	**.0543610**	**.0049539**	**.0820179**	**.5086230**	**2.652838**	**.7560534**
Stddev	.048439	.0021159	.0002066	.0011970	.0004842	.012674	.0037984
%RSD	3.139593	3.892259	4.170268	1.459469	.0951992	.4777481	.5023959
#1	1.490394	.0566906	.0048971	.0830833	.5084739	2.640076	.7603161
#2	1.552188	.0538340	.0051830	.0822479	.5091642	2.653017	.7548163
#3	1.585905	.0525585	.0047817	.0807226	.5082308	2.665422	.7530277

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2832-09DLX10 Acquired: 8/26/2025 12:12:37 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**2.155474**	**.0113388**	**.0123316**
Stddev	.005889	.0006336	.0004217
%RSD	.2732210	5.587755	3.419332
#1	2.156438	.0106240	.0124630
#2	2.149162	.0118311	.0118598
#3	2.160822	.0115614	.0126718

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**6323.928**	**107227.0**	**10851.08**	**4799.509**	**7421.461**
Stddev	8.765	325.1	47.67	28.547	4.700
%RSD	.1386001	.3031543	.4392973	.5947894	.0633285
#1	6321.775	106857.4	10832.87	4768.429	7418.786
#2	6333.569	107355.0	10905.17	4805.536	7426.887
#3	6316.440	107468.6	10815.21	4824.561	7418.708

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2832-09DUPDLX10 Acquired: 8/26/2025 12:16:43 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0641962**	**-.011684**	**1.063500**	**.0063952**	**.0134792**	**8.755853**	**.5654492**
Stddev	.0030138	.001269	.019133	.0010685	.0026480	.019604	.0019167
%RSD	4.694625	10.86009	1.799081	16.70774	19.64531	.2238924	.3389738
#1	.0675699	-.012633	1.082631	.0068459	.0159085	8.760464	.5653435
#2	.0632486	-.010243	1.063504	.0051752	.0106564	8.772740	.5674167
#3	.0617702	-.012176	1.044365	.0071645	.0138728	8.734354	.5635876

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0027569**	**.0011419**	**10.89307**	**.0837960**	**.0154283**	**.6148271**	**98.08410**
Stddev	.0000841	.0002727	.01091	.0005869	.0002832	.0114932	.84850
%RSD	3.051398	23.88012	.1001152	.7004305	1.835687	1.869343	.8650770
#1	.0027743	.0013322	10.88697	.0843273	.0155442	.6258234	98.93814
#2	.0028309	.0012641	10.90566	.0838947	.0156352	.6157635	97.24124
#3	.0026654	.0008295	10.88657	.0831660	.0151055	.6028943	98.07291

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.4203685**	**2.445102**	**.0566257**	**.0030375**	**.6175479**	**.0685609**	**.2585054**
Stddev	.0011232	.014000	.0008522	.0001443	.0089567	.0026573	.0015383
%RSD	.2671875	.5725906	1.505037	4.751961	1.450368	3.875875	.5950779
#1	.4191316	2.437084	.0575512	.0028802	.6151932	.0678183	.2596112
#2	.4213246	2.461268	.0564525	.0031639	.6100038	.0663538	.2591563
#3	.4206495	2.436954	.0558734	.0030685	.6274468	.0715105	.2567486

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.533590**	**.0563992**	**.0048061**	**.0856904**	**.5488217**	**2.825675**	**.7934562**
Stddev	.020310	.0046064	.0003314	.0019482	.0011416	.020308	.0208838
%RSD	1.324325	8.167418	6.896189	2.273575	.2080153	.7187106	2.632007
#1	1.520545	.0600399	.0051242	.0879005	.5476430	2.847662	.8146027
#2	1.523235	.0512206	.0048313	.0849488	.5499222	2.807621	.7929205
#3	1.556990	.0579370	.0044628	.0842219	.5488999	2.821743	.7728454

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2832-09DUPDLX10 Acquired: 8/26/2025 12:16:43 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**2.260712**	**.0131799**	**.0143085**
Stddev	.053262	.0006420	.0008972
%RSD	2.355997	4.870709	6.270630
#1	2.309445	.0137862	.0135338
#2	2.268837	.0125075	.0152916
#3	2.203854	.0132462	.0141000

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**6264.614**	**107559.1**	**10757.77**	**4811.184**	**7352.193**
Stddev	8.799	392.9	72.87	20.052	1.181
%RSD	.1404521	.3653067	.6773944	.4167866	.0160627
#1	6266.978	107106.6	10802.75	4788.036	7351.003
#2	6271.989	107813.8	10673.69	4823.215	7353.364
#3	6254.874	107757.0	10796.87	4822.302	7352.212

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2832-09LDLX50 Acquired: 8/26/2025 12:20:49 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0095409**	**-.002354**	**.1617386**	**.0002835**	**.0011263**	**1.279728**	**.0805323**
Stddev	.0006885	.000449	.0008102	.0003794	.0005014	.006318	.0010233
%RSD	7.216405	19.07250	.5009608	133.8302	44.52380	.4937212	1.270703
#1	.0088748	-.001866	.1625796	.0001352	.0005515	1.272447	.0809731
#2	.0094981	-.002750	.1609631	.0000007	.0014746	1.283779	.0812613
#3	.0102498	-.002446	.1616731	.0007147	.0013526	1.282957	.0793624

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0004066**	**.0002522**	**1.642637**	**.0121159**	**.0023787**	**.0964381**	**13.96310**
Stddev	.0000828	.0000384	.015066	.0001926	.0000449	.0003506	.05566
%RSD	20.35839	15.23888	.9171665	1.589909	1.887287	.3635199	.3986079
#1	.0004229	.0002590	1.626835	.0122150	.0024299	.0963375	13.89925
#2	.0004801	.0002108	1.656839	.0122387	.0023602	.0968280	13.98867
#3	.0003169	.0002867	1.644236	.0118939	.0023461	.0961488	14.00137

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0612017**	**.3484577**	**.0082987**	**.0004330**	**.1348622**	**.0122854**	**.0372471**
Stddev	.0004827	.0148995	.0002076	.0005495	.0185850	.0006194	.0001209
%RSD	.7886723	4.275827	2.501859	126.9019	13.78070	5.041674	.3244585
#1	.0608360	.3504621	.0081305	-.000145	.1146100	.0123628	.0373513
#2	.0617488	.3622535	.0082349	.000949	.1388414	.0128624	.0371146
#3	.0610204	.3326576	.0085308	.000495	.1511353	.0116309	.0372755

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2562672**	**.0023691**	**.0005833**	**.0131212**	**.0789958**	**.3963671**	**.1172115**
Stddev	.0452357	.0009358	.0001699	.0007029	.0015470	.0051757	.0001146
%RSD	17.65179	39.50005	29.11684	5.356749	1.958274	1.305777	.0978027
#1	.2063965	.0020541	.0005890	.0139245	.0774941	.4012480	.1172787
#2	.2677504	.0016314	.0007503	.0126191	.0789090	.3969132	.1170792
#3	.2946547	.0034217	.0004107	.0128201	.0805844	.3909400	.1172767

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2832-09LDLX50 Acquired: 8/26/2025 12:20:49 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.3313397	-.002111	.0023011
Stddev	.0009287	.001356	.0000317
%RSD	.2802893	64.23087	1.375763
#1	.3308484	-.003676	.0023151
#2	.3324108	-.001330	.0023234
#3	.3307598	-.001326	.0022649

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6334.411	109254.2	10469.51	4854.908	7698.394
Stddev	8.409	419.8	15.01	21.244	13.988
%RSD	.1327481	.3842783	.1433299	.4375687	.1816983
#1	6333.002	109014.3	10456.57	4850.527	7688.706
#2	6343.435	109739.0	10466.01	4878.000	7714.431
#3	6326.795	109009.4	10485.96	4836.196	7692.047

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2832-09MSDLX10 Acquired: 8/26/2025 12:25:04 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1334128**	**.1915165**	**1.023775**	**.2022706**	**.0917516**	**7.494568**	**.4858686**
Stddev	.0021542	.0048070	.001824	.0027554	.0016814	.008671	.0012739
%RSD	1.614699	2.509965	.1781866	1.362234	1.832532	.1156926	.2621920
#1	.1310292	.1961168	1.025051	.2023198	.0916496	7.504318	.4871059
#2	.1339886	.1865265	1.024588	.2050011	.0901235	7.487723	.4859387
#3	.1352206	.1919061	1.021685	.1994910	.0934816	7.491663	.4845610

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0227749**	**.0207336**	**9.184896**	**.1121430**	**.0331189**	**.5694290**	**80.33907**
Stddev	.0000914	.0000479	.037626	.0006841	.0001044	.0004514	.08808
%RSD	.4012801	.2312344	.4096524	.6100323	.3151852	.0792740	.1096320
#1	.0228472	.0207521	9.227966	.1127633	.0332116	.5698559	80.44031
#2	.0226722	.0207695	9.168311	.1114093	.0331392	.5694748	80.28004
#3	.0228054	.0206792	9.158413	.1122565	.0330058	.5689565	80.29687

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.3699012**	**2.223516**	**.0984884**	**.0092029**	**.7228026**	**.0879817**	**.2398249**
Stddev	.0001512	.028209	.0002995	.0004527	.0298269	.0040059	.0009224
%RSD	.0408698	1.268653	.3041261	4.918954	4.126556	4.553126	.3845955
#1	.3698359	2.194344	.0988079	.0090592	.6884878	.0900881	.2406267
#2	.3697936	2.250652	.0984433	.0097100	.7425111	.0904949	.2400311
#3	.3700740	2.225552	.0982139	.0088396	.7374091	.0833620	.2388169

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.090368**	**.0658695**	**.0455597**	**.1417958**	**.4751012**	**2.406240**	**1.259991**
Stddev	.047674	.0001239	.0002442	.0014505	.0024196	.002263	.009291
%RSD	2.280674	.1880860	.5359257	1.022914	.5092901	.0940370	.7373597
#1	2.035464	.0658334	.0457695	.1426354	.4778382	2.407124	1.268741
#2	2.121285	.0660075	.0452917	.1401210	.4742186	2.407928	1.250241
#3	2.114356	.0657677	.0456178	.1426311	.4732467	2.403669	1.260991

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2832-09MSDLX10 Acquired: 8/26/2025 12:25:04 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	1.969355	.0297175	.0327032
Stddev	.016373	.0005515	.0001408
%RSD	.8313959	1.855765	.4304247
#1	1.980602	.0298955	.0325988
#2	1.950571	.0301580	.0328633
#3	1.976893	.0290990	.0326475

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6319.827	109036.0	10672.48	4853.299	7423.971
Stddev	2.177	122.6	14.38	16.951	.733
%RSD	.0344539	.1124672	.1347571	.3492771	.0098752
#1	6321.899	109097.2	10658.38	4833.750	7424.412
#2	6317.558	109115.9	10671.95	4863.930	7423.125
#3	6320.024	108894.8	10687.13	4862.217	7424.377

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2832-09MSDDLX10 Acquired: 8/26/2025 12:29:07 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1335950**	**.1903441**	**1.018110**	**.2012190**	**.0915372**	**7.489093**	**.4866824**
Stddev	.0014086	.0066327	.002095	.0019252	.0010462	.010741	.0000854
%RSD	1.054353	3.484613	.2057590	.9567830	1.142883	.1434152	.0175458
#1	.1336558	.1948412	1.020382	.2004881	.0907387	7.486852	.4867745
#2	.1321570	.1934644	1.016255	.2034026	.0927214	7.500777	.4866059
#3	.1349722	.1827266	1.017694	.1997663	.0911514	7.479650	.4866667

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0227184**	**.0205651**	**9.198323**	**.1117690**	**.0332015**	**.5663089**	**80.97301**
Stddev	.0002978	.0001793	.012763	.0001692	.0000630	.0009505	.62106
%RSD	1.310794	.8719927	.1387549	.1513568	.1897267	.1678376	.7669969
#1	.0229384	.0207474	9.185503	.1118981	.0332714	.5673807	80.33406
#2	.0223796	.0205590	9.211028	.1115775	.0331837	.5655686	81.57448
#3	.0228373	.0203889	9.198437	.1118314	.0331492	.5659774	81.01048

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.3697738**	**2.236910**	**.0982672**	**.0092480**	**.7067543**	**.0886091**	**.2408556**
Stddev	.0004174	.005698	.0004601	.0001706	.0264531	.0042204	.0006359
%RSD	.1128695	.2547215	.4681858	1.844742	3.742895	4.762964	.2640322
#1	.3701153	2.241543	.0987970	.0090814	.6763339	.0928211	.2401220
#2	.3693085	2.238639	.0979680	.0094224	.7243549	.0843803	.2412502
#3	.3698974	2.230548	.0980367	.0092403	.7195740	.0886261	.2411946

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.078379**	**.0673889**	**.0448044**	**.1408110**	**.4765654**	**2.423936**	**1.248806**
Stddev	.037934	.0016774	.0002587	.0010183	.0028475	.030073	.012256
%RSD	1.825155	2.489124	.5774254	.7231631	.5974986	1.240685	.9814364
#1	2.035997	.0664670	.0448788	.1413127	.4737920	2.397276	1.258715
#2	2.089991	.0693250	.0450177	.1414812	.4794815	2.456536	1.252603
#3	2.109149	.0663746	.0445166	.1396392	.4764227	2.417996	1.235101

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2832-09MSDDLX10 Acquired: 8/26/2025 12:29:07 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	1.950350	.0308175	.0320205
Stddev	.021756	.0011391	.0004116
%RSD	1.115516	3.696276	1.285406
#1	1.969469	.0299427	.0324939
#2	1.954905	.0321055	.0318200
#3	1.926677	.0304043	.0317475

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6317.164	108606.3	10732.81	4820.177	7429.099
Stddev	2.328	209.5	70.76	12.837	3.705
%RSD	.0368554	.1929427	.6593219	.2663077	.0498693
#1	6319.781	108528.8	10655.62	4834.741	7425.181
#2	6316.388	108446.5	10794.62	4810.511	7429.569
#3	6315.323	108843.5	10748.19	4815.278	7432.546

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2832-09ADLX10 Acquired: 8/26/2025 12:33:11 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1346015**	**.1907544**	**1.016447**	**.2013567**	**.0916561**	**7.440452**	**.4768012**
Stddev	.0009735	.0067767	.002296	.0007969	.0013449	.008950	.0023632
%RSD	.7232548	3.552568	.2258621	.3957499	1.467387	.1202887	.4956279
#1	.1349311	.1970448	1.018752	.2018784	.0924957	7.450473	.4740731
#2	.1353675	.1916400	1.016427	.2017524	.0923678	7.437628	.4781118
#3	.1335060	.1835785	1.014161	.2004395	.0901048	7.433254	.4782186

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0230488**	**.0207214**	**9.106024**	**.1111328**	**.0329763**	**.5668980**	**78.79903**
Stddev	.0001117	.0000339	.036752	.0002584	.0002172	.0005408	.30640
%RSD	.4847649	.1633626	.4036036	.2325253	.6586297	.0953889	.3888432
#1	.0229197	.0207061	9.080498	.1113034	.0330775	.5665515	78.89587
#2	.0231131	.0206980	9.089427	.1112594	.0331244	.5675211	78.45590
#3	.0231135	.0207602	9.148148	.1108355	.0327269	.5666214	79.04531

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.3646428**	**2.199589**	**.0984108**	**.0084195**	**.6318956**	**.0875098**	**.2403117**
Stddev	.0031598	.023311	.0001080	.0003277	.0162541	.0016676	.0003172
%RSD	.8665510	1.059792	.1097780	3.892526	2.572277	1.905574	.1319976
#1	.3610090	2.173016	.0983408	.0081955	.6132270	.0859067	.2403467
#2	.3661753	2.209160	.0983563	.0082673	.6395537	.0892351	.2406100
#3	.3667442	2.216592	.0985352	.0087956	.6429061	.0873875	.2399785

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.010223**	**.0595445**	**.0449997**	**.1407410**	**.4683566**	**2.361827**	**1.250063**
Stddev	.038181	.0007042	.0001069	.0011376	.0003738	.020616	.005763
%RSD	1.899347	1.182708	.2375477	.8082559	.0798175	.8728927	.4610396
#1	1.966376	.0593731	.0450962	.1417085	.4679701	2.374308	1.255627
#2	2.036132	.0589418	.0448848	.1394879	.4687164	2.338031	1.250442
#3	2.028160	.0603186	.0450180	.1410266	.4683832	2.373143	1.244119

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2832-09ADLX10 Acquired: 8/26/2025 12:33:11 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	1.949947	.0303998	.0322599
Stddev	.019161	.0018410	.0004110
%RSD	.9826250	6.056072	1.274156
#1	1.968178	.0285559	.0317854
#2	1.951689	.0304055	.0325072
#3	1.929975	.0322380	.0324871

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6327.124	108689.6	10557.81	4820.411	7447.358
Stddev	10.891	296.0	28.79	16.474	9.903
%RSD	.1721255	.2723096	.2727254	.3417539	.1329773
#1	6337.000	108347.9	10586.64	4820.923	7445.338
#2	6328.929	108856.5	10529.05	4803.687	7458.116
#3	6315.444	108864.5	10557.75	4836.623	7438.621

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: PB169380BL Acquired: 8/26/2025 12:37:15 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0016322**	**.0001019**	**-.000733**	**-.000076**	**-.001710**	**.0008117**	**-.000871**
Stddev	.0021582	.0001486	.000575	.002169	.000264	.0073392	.000297
%RSD	132.2221	145.8671	78.35730	2858.370	15.45816	904.1573	34.09767
#1	.0001591	.0002624	-.000305	-.002532	-.001489	-.006317	-.001122
#2	.0041095	.0000743	-.001386	.000726	-.002003	.000407	-.000543
#3	.0006281	-.000031	-.000508	.001578	-.001638	.008345	-.000947

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000034**	**.0000302**	**.0173652**	**-.000291**	**-.000062**	**.0004754**	**-.003350**
Stddev	.000025	.0000516	.0051028	.000055	.000150	.0002492	.004986
%RSD	74.08553	170.8138	29.38539	18.82187	239.7584	52.41217	148.8163
#1	-.000028	.0000585	.0228560	-.000355	-.000235	.0006863	-.005252
#2	-.000013	-.000029	.0127685	-.000255	.000011	.0005396	.002306
#3	-.000062	.000062	.0164712	-.000265	.000036	.0002004	-.007105

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000070**	**-.003703**	**-.000041**	**.0005400**	**-.027845**	**.0011111**	**-.001489**
Stddev	.000686	.023615	.000165	.0001125	.011790	.0017830	.000354
%RSD	975.7061	637.6604	406.4023	20.82399	42.34171	160.4690	23.76302
#1	-.000278	-.022500	-.000002	.0006307	-.041042	.0030699	-.001886
#2	.000695	-.011413	.000102	.0004142	-.024142	-.000417	-.001376
#3	-.000628	.022803	-.000222	.0005751	-.018351	.000681	-.001206

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.108532**	**-.003365**	**-.000131**	**-.000339**	**.0005093**	**-.000780**	**.0005626**
Stddev	.008403	.000458	.000202	.000365	.0008064	.008333	.0002564
%RSD	7.742272	13.62444	154.0832	107.6509	158.3442	1068.789	45.58295
#1	-.116592	-.003894	-.000194	.000061	.0011320	.008738	.0007566
#2	-.109179	-.003108	-.000294	-.000654	-.000402	-.004312	.0006593
#3	-.099824	-.003092	.000095	-.000425	.000798	-.006764	.0002719

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: PB169380BL Acquired: 8/26/2025 12:37:15 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0011958	-.004907	-.000105
Stddev	.0015031	.000404	.000068
%RSD	125.7064	8.231560	64.42565
#1	.0013619	-.004458	-.000087
#2	-.000384	-.005024	-.000048
#3	.002609	-.005240	-.000180

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6235.576	107769.5	10598.03	4775.914	7715.625
Stddev	4.000	157.9	57.07	13.910	4.729
%RSD	.0641415	.1465421	.5385125	.2912439	.0612893
#1	6231.101	107820.1	10648.86	4789.002	7715.321
#2	6238.803	107895.9	10608.93	4777.431	7711.055
#3	6236.825	107592.4	10536.29	4761.308	7720.498

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCV02 Acquired: 8/26/2025 12:41:34 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV02 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**5.092658**	**5.153713**	**5.011788**	**5.132128**	**5.017742**	**10.16461**	**10.08943**
Stddev	.017534	.057594	.002352	.023211	.013125	.04308	.17977
%RSD	.3443088	1.117519	.0469273	.4522663	.2615712	.4238036	1.781762
#1	5.093019	5.137813	5.011880	5.128512	5.027998	10.11503	9.88185
#2	5.074946	5.105739	5.009391	5.110938	5.002951	10.19291	10.19396
#3	5.110009	5.217586	5.014092	5.156935	5.022277	10.18588	10.19247

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2456886**	**2.525705**	**25.37225**	**1.032073**	**2.487626**	**1.260362**	**5.256663**
Stddev	.0011195	.002004	.03403	.001546	.000761	.001264	.017198
%RSD	.4556670	.0793527	.1341349	.1497779	.0305824	.1003074	.3271732
#1	.2444344	2.523548	25.33296	1.032009	2.487755	1.260765	5.273130
#2	.2460446	2.526056	25.39175	1.030560	2.486809	1.258945	5.258045
#3	.2465869	2.527510	25.39205	1.033650	2.488314	1.261375	5.238816

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.517898**	**25.22337**	**2.498847**	**1.257676**	**26.94354**	**2.537092**	**2.532036**
Stddev	.005062	.05108	.000733	.002301	.10104	.006757	.004891
%RSD	.2010326	.2025266	.0293423	.1829390	.3749965	.2663292	.1931784
#1	2.512746	25.16987	2.499622	1.258061	27.05132	2.529762	2.527899
#2	2.522865	25.22860	2.498164	1.255207	26.92832	2.538441	2.530774
#3	2.518084	25.27164	2.498756	1.259760	26.85097	2.543073	2.537434

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**25.98650**	**4.861814**	**5.062253**	**4.958744**	**5.038166**	**5.226402**	**5.039524**
Stddev	.05461	.013912	.015506	.005617	.002276	.030036	.021717
%RSD	.2101384	.2861408	.3063116	.1132821	.0451808	.5747030	.4309244
#1	26.04381	4.846840	5.054443	4.963000	5.035551	5.231132	5.038565
#2	25.98062	4.864267	5.052205	4.952377	5.039241	5.194281	5.018303
#3	25.93507	4.874337	5.080112	4.960856	5.039705	5.253793	5.061704

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCV02 Acquired: 8/26/2025 12:41:34 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV02 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**4.940466**	**5.163127**	**5.055040**
Stddev	.021267	.005932	.020943
%RSD	.4304570	.1148942	.4143078
#1	4.937956	5.162512	5.060310
#2	4.920565	5.169342	5.072845
#3	4.962876	5.157526	5.031965

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**6081.585**	**104124.6**	**10689.16**	**4658.783**	**6879.222**
Stddev	1.965	283.8	52.08	30.284	9.674
%RSD	.0323166	.2725886	.4871879	.6500308	.1406230
#1	6083.817	104174.9	10738.30	4679.985	6875.596
#2	6080.115	104380.0	10694.62	4672.265	6871.885
#3	6080.822	103819.0	10634.57	4624.100	6890.185

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCB02 Acquired: 8/26/2025 12:45:45 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB02 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000062**	**-.000824**	**-.000332**	**.0007804**	**-.002940**	**.0239641**	**-.001136**
Stddev	.001461	.000924	.000480	.0013467	.003046	.0098336	.001372
%RSD	2359.446	112.0764	144.2403	172.5668	103.6082	41.03455	120.6984
#1	.000907	-.001849	-.000788	.0017700	.000574	.0339303	-.000067
#2	.000649	-.000055	.000168	.0013243	-.004833	.0236933	-.000659
#3	-.001742	-.000569	-.000378	-.000753	-.004562	.0142688	-.002683

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0000154**	**.0000184**	**.0057764**	**-.000414**	**-.000034**	**.0003356**	**.0041823**
Stddev	.0000581	.0000546	.0042661	.000100	.000121	.0002463	.0028763
%RSD	375.9713	296.5663	73.85374	24.01896	351.9508	73.38054	68.77277
#1	.0000153	.0000634	.0065134	-.000526	.000064	.0000540	.0022820
#2	.0000735	-.000042	.0096260	-.000384	-.000169	.0004421	.0027735
#3	-.000043	.000034	.0011898	-.000333	.000002	.0005108	.0074914

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000027**	**-.004690**	**.0000223**	**.0002245**	**-.063062**	**.0000068**	**-.000099**
Stddev	.000295	.021939	.0000771	.0002002	.013145	.0019102	.000149
%RSD	1107.846	467.8098	345.8554	89.19070	20.84447	28265.46	150.4115
#1	-.000010	-.029737	.0000230	.0001015	-.060963	.0015255	.000044
#2	.000259	.004547	-.000055	.0004554	-.077130	-.002138	-.000253
#3	-.000329	.011121	.000099	.0001164	-.051093	.000633	-.000088

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.147793**	**.0007301**	**.0004195**	**.0004071**	**-.001195**	**.0026304**	**.0010088**
Stddev	.041274	.0008001	.0002387	.0005934	.001436	.0091562	.0013432
%RSD	27.92673	109.5879	56.89393	145.7773	120.1905	348.0861	133.1422
#1	-.195361	-.000172	.0006875	-.000267	-.000409	-.007655	.0023062
#2	-.126555	.001008	.0003417	.000850	-.000323	.009893	.0010962
#3	-.121463	.001354	.0002295	.000639	-.002852	.005654	-.000376

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCB02 Acquired: 8/26/2025 12:45:45 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB02 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0028602	-.004742	-.000022
Stddev	.0003770	.000343	.000118
%RSD	13.17943	7.223676	534.6878
#1	.0032069	-.004775	.000105
#2	.0029149	-.004384	-.000042
#3	.0024589	-.005067	-.000129

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6346.587	110207.6	10571.67	4901.287	7807.240
Stddev	9.282	147.0	21.46	27.068	13.075
%RSD	.1462529	.1334107	.2029943	.5522621	.1674707
#1	6339.871	110234.7	10580.64	4884.266	7802.243
#2	6342.710	110339.1	10547.19	4932.500	7797.402
#3	6357.179	110048.9	10587.20	4887.095	7822.077

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: PB169380BS Acquired: 8/26/2025 12:50:06 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.7726398**	**1.976237**	**.9438955**	**1.963387**	**.7871989**	**1.986955**
Stddev	.0040972	.025777	.0027170	.009551	.0026774	.005103
%RSD	.5302877	1.304327	.2878467	.4864642	.3401153	.2568416
#1	.7732965	2.001426	.9438378	1.959941	.7847325	1.986107
#2	.7682539	1.949910	.9412079	1.956037	.7868177	1.992429
#3	.7763690	1.977375	.9466409	1.974183	.7900464	1.982328

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1847673**	**.2119990**	**.1887362**	**1.017458**	**.4071023**	**.1935941**
Stddev	.0002862	.0014604	.0002108	.007017	.0004336	.0004250
%RSD	.1548781	.6888662	.1116902	.6896565	.1065083	.2195174
#1	.1844368	.2105480	.1885002	1.009560	.4066055	.1934871
#2	.1849336	.2134686	.1888026	1.019840	.4074045	.1932328
#3	.1849314	.2119806	.1889058	1.022974	.4072969	.1940623

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.3088525**	**2.814520**	**.1965976**	**1.979500**	**.4859208**	**.0751842**
Stddev	.0005276	.015521	.0012750	.017048	.0007810	.0005948
%RSD	.1708111	.5514710	.6485148	.8612079	.1607295	.7910949
#1	.3083821	2.821935	.1951287	1.964544	.4854388	.0746922
#2	.3087526	2.796682	.1974179	1.975895	.4855017	.0750152
#3	.3094229	2.824944	.1972462	1.998062	.4868219	.0758452

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.548742**	**.2939507**	**.2083801**	**9.017841**	**.3065805**	**.4008029**
Stddev	.019993	.0020296	.0009847	.050611	.0032687	.0011556
%RSD	.7844316	.6904710	.4725670	.5612336	1.066194	.2883233
#1	2.535622	.2916397	.2074533	8.981459	.3028586	.4007362
#2	2.538851	.2947686	.2094140	8.996425	.3089848	.3996821
#3	2.571753	.2954438	.2082731	9.075639	.3078981	.4019904

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: PB169380BS Acquired: 8/26/2025 12:50:06 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.6564671	.1864396	.8491429	5.530900	F .0009931	.1959286
Stddev	.0029470	.0008157	.0046142	.036160	.0010979	.0022291
%RSD	.4489163	.4374972	.5433946	.6537752	110.5429	1.137703
#1	.6569262	.1855119	.8445246	5.540839	.0018135	.1937976
#2	.6533176	.1870443	.8537530	5.490810	-.000254	.1957439
#3	.6591577	.1867625	.8491511	5.561050	.001420	.1982442

Elem	Sr4077
Units	ppm
Avg	.1873843
Stddev	.0010447
%RSD	.5575350
#1	.1864255
#2	.1872297
#3	.1884977

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6299.922	107460.2	10232.74	4763.279	7554.005
Stddev	.774	255.0	68.91	14.150	3.555
%RSD	.0122821	.2373073	.6734192	.2970641	.0470616
#1	6299.099	107587.9	10311.84	4779.514	7549.940
#2	6300.032	107626.2	10185.73	4756.758	7555.543
#3	6300.635	107166.6	10200.64	4753.566	7556.532

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: PB169379BL Acquired: 8/26/2025 12:54:06 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0008019**	**-.000042**	**.0000769**	**.0010752**	**-.003295**	**.0039226**	**-.001135**
Stddev	.0012251	.000789	.0004045	.0019804	.001249	.0065374	.001117
%RSD	152.7833	1892.494	526.1957	184.1927	37.89935	166.6622	98.44485
#1	.0021831	-.000828	.0001232	-.000641	-.001995	.0079125	-.002425
#2	-.000153	-.000046	.0004563	.000624	-.004485	.0074772	-.000477
#3	.000376	.000749	-.000349	.003242	-.003404	-.003622	-.000503

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000028**	**-.000007**	**.0538111**	**-.000303**	**.0000035**	**.0000336**	**-.004572**
Stddev	.000064	.000047	.0116823	.000224	.0000119	.0001987	.003906
%RSD	230.6177	662.6828	21.70976	73.82845	338.0192	591.9553	85.42399
#1	.000044	.000042	.0672120	-.000129	.0000033	.0002438	-.009043
#2	-.000081	-.000010	.0484475	-.000555	-.000008	-.000151	-.001823
#3	-.000046	-.000053	.0457738	-.000224	.000015	.000008	-.002851

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000040**	**.0198690**	**-.000128**	**.0000459**	**-.089465**	**.0004776**	**-.001346**
Stddev	.000226	.0126848	.000197	.0002460	.019987	.0015679	.000168
%RSD	566.2794	63.84205	153.7271	536.4735	22.34102	328.2729	12.51043
#1	.000145	.0091015	-.000326	.0002898	-.112161	-.001230	-.001530
#2	-.000291	.0338521	.000069	-.000202	-.074487	.000812	-.001199
#3	.000027	.0166533	-.000128	.000050	-.081748	.001852	-.001310

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.166647**	**-.001921**	**-.000227**	**.0003770**	**-.001233**	**.0008839**	**-.000093**
Stddev	.039224	.000513	.000144	.0004192	.000707	.0120916	.001247
%RSD	23.53695	26.68904	63.61217	111.1817	57.32743	1367.936	1341.141
#1	-.122013	-.002323	-.000148	.0004740	-.000564	-.009282	.001328
#2	-.182306	-.002098	-.000140	.0007393	-.001164	-.002321	-.000604
#3	-.195622	-.001344	-.000394	-.000082	-.001973	.014255	-.001003

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: PB169379BL Acquired: 8/26/2025 12:54:06 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0007753	-.005232	-.000042
Stddev	.0009955	.000589	.000014
%RSD	128.4131	11.26699	34.43548
#1	.0015358	-.005817	-.000030
#2	-.000352	-.005241	-.000058
#3	.001141	-.004638	-.000037

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6350.579	109152.9	10572.89	4821.790	7797.830
Stddev	12.018	216.9	67.25	13.456	12.496
%RSD	.1892364	.1986769	.6360598	.2790661	.1602495
#1	6364.059	109304.7	10648.18	4836.057	7810.816
#2	6340.987	109249.5	10518.78	4819.984	7796.782
#3	6346.690	108904.6	10551.71	4809.328	7785.890

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: PB169379BS Acquired: 8/26/2025 12:58:24 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.7553036**	**1.910178**	**.9241075**	**1.916471**	**.7888550**	**1.955618**
Stddev	.0037034	.024094	.0012289	.000558	.0016105	.014882
%RSD	.4903229	1.261346	.1329854	.0291369	.2041607	.7610009
#1	.7524225	1.895892	.9231702	1.917045	.7870294	1.966503
#2	.7594809	1.937996	.9254988	1.916437	.7894608	1.961691
#3	.7540073	1.896646	.9236534	1.915930	.7900747	1.938659

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1857503**	**.2006286**	**.1846319**	**1.063155**	**.3990438**	**.1891341**
Stddev	.0014780	.0009647	.0003533	.004341	.0013484	.0002275
%RSD	.7956954	.4808559	.1913626	.4083021	.3379084	.1202699
#1	.1873239	.2008904	.1846351	1.064285	.4001348	.1889477
#2	.1843914	.1995600	.1849836	1.058361	.3994602	.1890671
#3	.1855354	.2014354	.1842770	1.066819	.3975363	.1893876

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.3011982**	**2.827612**	**.1945271**	**1.955898**	**.4751213**	**.0732735**
Stddev	.0000889	.025039	.0012256	.028269	.0007810	.0003254
%RSD	.0295183	.8855263	.6300579	1.445323	.1643764	.4440579
#1	.3012880	2.850521	.1946672	1.940732	.4756391	.0736429
#2	.3011102	2.831434	.1932374	1.938449	.4755019	.0731485
#3	.3011965	2.800882	.1956767	1.988514	.4742230	.0730293

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.650614**	**.2916802**	**.2033413**	**9.146020**	**.2894521**	**.4025476**
Stddev	.007512	.0019851	.0011422	.047077	.0020853	.0004398
%RSD	.2834094	.6805863	.5617331	.5147234	.7204276	.1092467
#1	2.643860	.2925864	.2046270	9.169971	.2896267	.4026593
#2	2.658704	.2894037	.2029535	9.176305	.2872849	.4020628
#3	2.649277	.2930506	.2024435	9.091783	.2914445	.4029208

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: PB169379BS Acquired: 8/26/2025 12:58:24 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.6584640	.1882133	.8898250	5.397527	F .0027923	.1916489
Stddev	.0020992	.0013050	.0127907	.030155	.0013185	.0004560
%RSD	.3188013	.6933346	1.437438	.5586824	47.21829	.2379332
#1	.6573406	.1896918	.8814752	5.412491	.0043133	.1918046
#2	.6608859	.1877260	.9045504	5.417273	.0020897	.1911354
#3	.6571656	.1872222	.8834495	5.362817	.0019740	.1920066

Elem	Sr4077
Units	ppm
Avg	.1860182
Stddev	.0009008
%RSD	.4842390
#1	.1870443
#2	.1856523
#3	.1853579

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6305.691	108118.0	10538.12	4807.112	7567.582
Stddev	2.621	360.3	20.48	30.611	5.936
%RSD	.0415642	.3332240	.1943545	.6367795	.0784361
#1	6303.989	107850.3	10518.06	4772.228	7561.357
#2	6308.710	107976.1	10558.99	4819.615	7568.212
#3	6304.376	108527.6	10537.30	4829.491	7573.179

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2941-01 Acquired: 8/26/2025 13:02:25 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0424263**	**-.022229**	**.2660582**	**.0014950**	**-.012429**	**104.0658**	**.4160364**
Stddev	.0015743	.001327	.0014195	.0014399	.000923	.0996	.0011181
%RSD	3.710602	5.971766	.5335336	96.31719	7.425232	.0956700	.2687534
#1	.0442441	-.021727	.2667861	-.000023	-.011652	103.9858	.4172184
#2	.0415300	-.021225	.2669662	.001665	-.013449	104.0343	.4158952
#3	.0415049	-.023734	.2644224	.002842	-.012185	104.1773	.4149955

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0077971**	**.0090434**	**55.23238**	**.2419337**	**.1025341**	**.3128604**	**238.7035**
Stddev	.0001247	.0002697	.11481	.0009806	.0000886	.0009432	2.3302
%RSD	1.599580	2.982292	.2078691	.4053212	.0864394	.3014624	.9761883
#1	.0077549	.0088755	55.18865	.2429614	.1024391	.3121966	240.6788
#2	.0076990	.0089001	55.14587	.2418314	.1025485	.3124446	239.2980
#3	.0079375	.0093545	55.36263	.2410082	.1026146	.3139400	236.1336

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.460028**	**42.97527**	**.1784299**	**.0040417**	**7.249352**	**.4034289**	**1.003132**
Stddev	.003071	.13344	.0003504	.0008868	.097582	.0007193	.002774
%RSD	.1248366	.3104967	.1963971	21.94151	1.346073	.1783051	.2765146
#1	2.459929	42.88552	.1786664	.0047096	7.322501	.4035379	1.004089
#2	2.457008	42.91168	.1780273	.0043799	7.287001	.4040875	1.000007
#3	2.463147	43.12861	.1785959	.0030355	7.138555	.4026613	1.005301

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**13.10657**	**.1460872**	**.0150623**	**.0244586**	**4.643025**	**2.623924**	**9.544387**
Stddev	.17211	.0118635	.0001430	.0004802	.006049	.012783	.043074
%RSD	1.313166	8.120822	.9491081	1.963318	.1302902	.4871648	.4513008
#1	13.26595	.1545497	.0150679	.0239916	4.649967	2.623159	9.562748
#2	13.12968	.1511850	.0149166	.0244333	4.638884	2.637072	9.495176
#3	12.92407	.1325269	.0152023	.0249510	4.640223	2.611540	9.575239

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2941-01 Acquired: 8/26/2025 13:02:25 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**2.782683**	**.1531178**	**.1365634**
Stddev	.020774	.0003758	.0011636
%RSD	.7465391	.2454183	.8520536
#1	2.800262	.1530485	.1358455
#2	2.759759	.1527814	.1359388
#3	2.788027	.1535233	.1379059

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**7294.385**	**126439.9**	**13331.72**	**5624.858**	**6643.106**
Stddev	10.267	483.6	93.37	19.783	8.847
%RSD	.1407521	.3824470	.7003339	.3517039	.1331701
#1	7294.515	125935.2	13385.86	5625.127	6645.938
#2	7284.054	126485.5	13385.38	5644.506	6633.190
#3	7304.586	126899.1	13223.91	5604.943	6650.190

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2947-01 Acquired: 8/26/2025 13:06:28 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.2055666	-.037089	2.238149	.0240847	.0181807	143.4216
Stddev	.0031623	.001591	.004753	.0025050	.0030767	.1585
%RSD	1.538326	4.290669	.2123588	10.40067	16.92293	.1105044
#1	.2036485	-.038254	2.233443	.0223303	.0152735	143.5828
#2	.2092165	-.037737	2.238059	.0269535	.0178657	143.2659
#3	.2038349	-.035276	2.242947	.0229704	.0214027	143.4159

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	2.419068	.0165306	.0206753	142.1906	1.507854	.2200004
Stddev	.009145	.0001381	.0001627	.0814	.003129	.0000611
%RSD	.3780508	.8356828	.7868453	.0572303	.2075328	.0277861
#1	2.427675	.0166554	.0204881	142.2582	1.511431	.2199753
#2	2.409466	.0163822	.0207557	142.1003	1.505621	.2199558
#3	2.420064	.0165541	.0207822	142.2134	1.506510	.2200701

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	4.263393	352.6083	4.478921	29.24440	2.247830	.0136777
Stddev	.006506	.4157	.009929	.05403	.003950	.0006863
%RSD	.1526092	.1178789	.2216771	.1847489	.1757351	5.017717
#1	4.255880	352.4124	4.488527	29.18371	2.243378	.0129831
#2	4.267209	352.3267	4.468699	29.26222	2.249193	.0136947
#3	4.267088	353.0856	4.479538	29.28727	2.250918	.0143554

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	3.350101	.4231722	3.634081	13.67921	.2368234	.1280400
Stddev	.013939	.0038564	.003181	.03724	.0015414	.0001358
%RSD	.4160637	.9113125	.0875451	.2722623	.6508715	.1060592
#1	3.335710	.4257771	3.637276	13.67658	.2355422	.1280230
#2	3.363538	.4249974	3.634054	13.64335	.2385339	.1281834
#3	3.351056	.4187420	3.630913	13.71770	.2363940	.1279134

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2947-01 Acquired: 8/26/2025 13:06:28 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.2103827	3.806408	5.376234	F 13.03354	F 11.49808	.1383702
Stddev	.0011525	.011954	.003186	.04828	.06993	.0017206
%RSD	.5478033	.3140486	.0592673	.3704188	.6081667	1.243500
#1	.2100661	3.819220	5.379482	12.98881	11.41834	.1377851
#2	.2116604	3.795555	5.376107	13.08472	11.54895	.1370184
#3	.2094216	3.804448	5.373113	13.02708	11.52694	.1403071

Elem	Sr4077
Units	ppm
Avg	.5475977
Stddev	.0041746
%RSD	.7623493
#1	.5506467
#2	.5428398
#3	.5493067

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6832.896	118164.9	12789.32	5219.892	6427.807
Stddev	12.016	409.3	31.60	12.134	10.947
%RSD	.1758564	.3463389	.2470516	.2324529	.1703044
#1	6846.768	118471.1	12803.41	5229.320	6440.352
#2	6826.204	118323.6	12811.42	5224.154	6422.878
#3	6825.716	117700.1	12753.13	5206.202	6420.192

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2947-03 Acquired: 8/26/2025 13:10:28 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1275089**	**-.033877**	**1.289526**	**.0036097**	**.0039293**	**248.7254**
Stddev	.0008638	.002231	.002802	.0032109	.0006139	1.2240
%RSD	.6774522	6.586627	.2172635	88.95220	15.62281	.4921202
#1	.1265402	-.034160	1.289107	.0071246	.0040562	248.8851
#2	.1277873	-.031518	1.286958	.0008306	.0044698	247.4293
#3	.1281992	-.035954	1.292514	.0028738	.0032619	249.8617

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.980328**	**.0211563**	**.0259245**	**76.22744**	**.9833508**	**.1833397**
Stddev	.056731	.0004290	.0003099	2.06679	.0021302	.0004101
%RSD	2.864713	2.027945	1.195519	2.711343	.2166244	.2236562
#1	2.044941	.0216298	.0257993	78.61349	.9833656	.1835078
#2	1.957364	.0207933	.0256968	75.07580	.9812132	.1828723
#3	1.938680	.0210458	.0262775	74.99305	.9854735	.1836390

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.9084591**	**390.9125**	**4.106071**	**31.28519**	**1.033078**	**.0062726**
Stddev	.0022753	3.4334	.105178	.80164	.001610	.0016963
%RSD	.2504547	.8783129	2.561513	2.562376	.1558275	27.04343
#1	.9082068	392.9305	4.226059	32.20945	1.032053	.0065220
#2	.9063205	392.8590	4.062337	30.86709	1.032246	.0078304
#3	.9108500	386.9482	4.029817	30.77903	1.034933	.0044654

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.471644**	**.5033125**	**3.258474**	**7.759395**	**.2232814**	**.0639829**
Stddev	.028618	.0135113	.004304	.090119	.0179052	.0003165
%RSD	1.157860	2.684465	.1320721	1.161413	8.019106	.4947100
#1	2.485617	.5189126	3.262939	7.816027	.2048579	.0643387
#2	2.490592	.4953389	3.258132	7.806684	.2406187	.0638775
#3	2.438724	.4956858	3.254352	7.655475	.2243676	.0637325

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2947-03 Acquired: 8/26/2025 13:10:28 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.2011667	7.003729	4.312873	F 20.05426	4.668262	.1051399
Stddev	.0001317	.189376	.065455	.05475	.011047	.0028765
%RSD	.0654470	2.703936	1.517676	.2730021	.2366451	2.735870
#1	.2013115	7.222329	4.380500	20.09296	4.680870	.1078105
#2	.2010542	6.899294	4.308289	19.99162	4.660277	.1055150
#3	.2011345	6.889563	4.249830	20.07820	4.663640	.1020943

Elem	Sr4077
Units	ppm
Avg	.0333060
Stddev	.0098389
%RSD	29.54083
#1	.0445382
#2	.0262126
#3	.0291673

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	7170.337	122955.9	12802.00	5473.267	6452.874
Stddev	8.481	374.4	216.22	11.612	8.537
%RSD	.1182800	.3044592	1.688940	.2121532	.1322948
#1	7161.275	123380.7	12560.56	5483.939	6447.771
#2	7178.084	122674.5	12977.77	5460.902	6462.730
#3	7171.651	122812.3	12867.66	5474.960	6448.122

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2947-03DUP Acquired: 8/26/2025 13:14:35 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1191905**	**-.033825**	**1.478568**	**.0100426**	**.0214930**	**235.8803**
Stddev	.0007949	.002959	.000281	.0022460	.0017752	2.5768
%RSD	.6668907	8.748631	.0190058	22.36537	8.259227	1.092436
#1	.1195555	-.036665	1.478888	.0078174	.0208919	233.3103
#2	.1197373	-.030759	1.478451	.0100013	.0234907	238.4639
#3	.1182787	-.034051	1.478364	.0123090	.0200965	235.8667

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.944371**	**.0207291**	**.0233100**	**77.26457**	**1.734536**	**.2107402**
Stddev	.004307	.0001521	.0000856	.04133	.005826	.0004395
%RSD	.2214855	.7337246	.3674571	.0534963	.3358665	.2085650
#1	1.946215	.0207349	.0234023	77.29091	1.728316	.2106438
#2	1.939449	.0205741	.0232331	77.28586	1.735425	.2103569
#3	1.947448	.0208781	.0232947	77.21693	1.739865	.2112200

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.743008**	**397.7859**	**5.116056**	**29.98127**	**1.724660**	**.0148275**
Stddev	.003833	1.0451	.005119	.10149	.000700	.0004779
%RSD	.2198941	.2627209	.1000553	.3385202	.0405730	3.222907
#1	1.739743	396.6245	5.110693	29.89631	1.724051	.0146922
#2	1.742054	398.0828	5.116586	30.09366	1.724506	.0153584
#3	1.747228	398.6504	5.120889	29.95385	1.725424	.0144318

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.753098**	**.4913207**	**3.438908**	**7.742598**	**.3314747**	**.1150419**
Stddev	.020207	.0021334	.014624	.032342	.0068930	.0003628
%RSD	.7339863	.4342194	.4252372	.4177161	2.079479	.3153965
#1	2.744262	.4937196	3.436981	7.758406	.3260057	.1149251
#2	2.738814	.4906064	3.425344	7.705392	.3292012	.1147518
#3	2.776219	.4896361	3.454400	7.763995	.3392172	.1154487

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2947-03DUP Acquired: 8/26/2025 13:14:35 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.3123207	6.963046	4.786819	F 18.94265	4.863261	.1173400
Stddev	.0013430	.001590	.010030	.15773	.041999	.0013353
%RSD	.4299908	.0228299	.2095428	.8326737	.8636006	1.138022
#1	.3136367	6.964182	4.792756	19.10858	4.906042	.1174593
#2	.3123732	6.961229	4.775238	18.92473	4.861651	.1159490
#3	.3109523	6.963726	4.792462	18.79465	4.822090	.1186117

Elem	Sr4077
Units	ppm
Avg	.0050658
Stddev	.0016136
%RSD	31.85174
#1	.0068823
#2	.0037985
#3	.0045167

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6943.798	119023.1	12889.55	5336.012	6358.350
Stddev	11.880	262.0	24.79	35.095	5.387
%RSD	.1710950	.2200996	.1922951	.6577039	.0847277
#1	6955.729	119319.1	12910.61	5366.211	6363.313
#2	6943.696	118929.0	12862.23	5344.314	6359.117
#3	6931.968	118821.2	12895.81	5297.510	6352.620

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2947-03LX5 Acquired: 8/26/2025 13:18:42 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0305625**	**-.007562**	**.2850222**	**-.000231**	**-.000453**	**64.35027**	**.5026010**
Stddev	.0005569	.001601	.0033647	.002350	.001095	.23721	.0014817
%RSD	1.822167	21.17397	1.180510	1016.506	241.8842	.3686253	.2948112
#1	.0300859	-.008379	.2888137	-.002895	.000761	64.07984	.5015801
#2	.0311746	-.005717	.2838612	.000655	-.001366	64.44775	.5043005
#3	.0304269	-.008589	.2823917	.001547	-.000754	64.52320	.5019224

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0055161**	**.0036263**	**20.12340**	**.2583747**	**.0383701**	**.2419292**	**100.8709**
Stddev	.0001152	.0000958	.08296	.0014736	.0006308	.0041343	.1272
%RSD	2.088730	2.640713	.4122524	.5703365	1.643885	1.708884	.1260690
#1	.0055398	.0037331	20.03260	.2594077	.0389835	.2458740	101.0099
#2	.0056176	.0035480	20.19525	.2566872	.0384036	.2422851	100.7605
#3	.0053909	.0035979	20.14234	.2590292	.0377233	.2376284	100.8422

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.092352**	**8.287617**	**.2226536**	**.0019053**	**.5278939**	**.1325956**	**.8510411**
Stddev	.005083	.010829	.0027382	.0002371	.0081691	.0030092	.0036042
%RSD	.4652826	.1306704	1.229798	12.44651	1.547489	2.269450	.4235078
#1	1.088550	8.275625	.2254918	.0016567	.5198081	.1307848	.8551782
#2	1.098125	8.296682	.2224412	.0021291	.5361438	.1360693	.8485807
#3	1.090382	8.290545	.2200278	.0019303	.5277299	.1309328	.8493643

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.714809**	**.0512504**	**.0161410**	**.0421746**	**1.830510**	**1.163195**	**4.308620**
Stddev	.028619	.0016390	.0003455	.0008818	.003364	.007181	.060448
%RSD	1.668920	3.197965	2.140275	2.090813	.1838001	.6173111	1.402960
#1	1.743629	.0526985	.0164480	.0427114	1.826939	1.171486	4.378419
#2	1.686396	.0494711	.0162081	.0426554	1.833620	1.159003	4.273906
#3	1.714402	.0515815	.0157669	.0411569	1.830970	1.159096	4.273535

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2947-03LX5 Acquired: 8/26/2025 13:18:42 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.9509768	.0217419	.0086197
Stddev	.0113583	.0010959	.0003713
%RSD	1.194385	5.040347	4.307752
#1	.9621301	.0207798	.0081915
#2	.9394240	.0229348	.0088526
#3	.9513762	.0215112	.0088150

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6544.878	112491.0	11283.07	5026.377	7252.531
Stddev	16.231	164.8	46.92	24.130	21.253
%RSD	.2479964	.1465268	.4158757	.4800738	.2930471
#1	6527.503	112380.9	11335.36	5012.704	7228.311
#2	6547.480	112680.5	11244.63	5054.239	7261.216
#3	6559.650	112411.5	11269.21	5012.189	7268.067

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2947-03MS Acquired: 8/26/2025 13:22:47 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.7646316	1.786047	3.095649	1.523182	.3140639	273.7819
Stddev	.0015586	.009679	.004650	.006258	.0024241	.3985
%RSD	.2038417	.5419429	.1502140	.4108470	.7718509	.1455477
#1	.7644621	1.775359	3.091405	1.523428	.3147876	273.3340
#2	.7631646	1.794222	3.094924	1.529313	.3113604	273.9147
#3	.7662680	1.788561	3.100620	1.516804	.3160438	274.0970

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	2.900236	.1762842	.2234219	107.9996	1.963456	.4236491
Stddev	.004930	.0009816	.0001558	.1555	.003129	.0004457
%RSD	.1700020	.5568546	.0697375	.1440047	.1593546	.1051985
#1	2.900524	.1769754	.2233353	107.9963	1.963172	.4235131
#2	2.895168	.1751605	.2233286	107.8457	1.960479	.4232873
#3	2.905016	.1767165	.2236018	108.1567	1.966717	.4241470

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.820170	414.7897	5.118498	39.94516	2.081041	.0803272
Stddev	.001482	1.2723	.011861	.12120	.002055	.0006245
%RSD	.0814145	.3067288	.2317221	.3034260	.0987342	.7773959
#1	1.821750	413.3399	5.116857	39.86250	2.079823	.0796098
#2	1.818810	415.7201	5.107544	39.88869	2.079886	.0807487
#3	1.819951	415.3093	5.131094	40.08430	2.083413	.0806231

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	5.688303	.7552411	4.633689	19.15015	.5986639	.4446433
Stddev	.020895	.0045162	.008016	.02633	.0065578	.0006504
%RSD	.3673328	.5979754	.1729922	.1375171	1.095403	.1462773
#1	5.665675	.7508726	4.632813	19.13713	.5911076	.4453827
#2	5.692367	.7598917	4.642106	19.18046	.6020161	.4441595
#3	5.706868	.7549590	4.626146	19.13287	.6028680	.4443877

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2947-03MS Acquired: 8/26/2025 13:22:47 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	1.139137	7.356092	7.723124	F 27.50236	6.257913	.3025107
Stddev	.001374	.013341	.010553	.05445	.001004	.0008976
%RSD	.1206533	.1813656	.1366430	.1979856	.0160488	.2967291
#1	1.138295	7.356546	7.729425	27.45800	6.258340	.3029843
#2	1.140723	7.342530	7.729007	27.48595	6.256766	.3014754
#3	1.138393	7.369202	7.710941	27.56313	6.258633	.3030723

Elem	Sr4077
Units	ppm
Avg	.2755048
Stddev	.0020219
%RSD	.7338890
#1	.2771328
#2	.2732416
#3	.2761401

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	7145.858	121703.6	13330.24	5442.774	6246.891
Stddev	6.969	231.2	49.46	12.983	.981
%RSD	.0975202	.1899913	.3710065	.2385274	.0157031
#1	7138.026	121963.5	13321.40	5455.479	6247.912
#2	7148.175	121626.8	13383.53	5443.312	6245.956
#3	7151.373	121520.6	13285.81	5429.530	6246.805

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2947-03MSD Acquired: 8/26/2025 13:26:50 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.7519962**	**1.815586**	**2.464576**	**1.551182**	**.3209983**	**274.5876**
Stddev	.0015034	.030644	.007377	.015684	.0002872	1.0670
%RSD	.1999285	1.687845	.2993406	1.011131	.0894593	.3885830
#1	.7531444	1.843929	2.473021	1.568120	.3212020	273.9513
#2	.7525497	1.819760	2.459388	1.537160	.3206698	275.8194
#3	.7502944	1.783069	2.461318	1.548267	.3211231	273.9919

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.517627**	**.1780825**	**.2237906**	**87.90761**	**1.209791**	**.4209261**
Stddev	.011333	.0006494	.0004548	.19214	.007042	.0005729
%RSD	.4501479	.3646494	.2032375	.2185700	.5820868	.1361030
#1	2.514146	.1785932	.2242871	87.94416	1.213116	.4215750
#2	2.530292	.1783027	.2236906	88.07885	1.214554	.4204902
#3	2.508442	.1773517	.2233940	87.69982	1.201702	.4207131

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.223822**	**417.7606**	**5.008854**	**35.21812**	**1.473069**	**.0779378**
Stddev	.002435	2.5702	.023859	.03442	.000762	.0012766
%RSD	.1989705	.6152429	.4763438	.0977221	.0517073	1.637990
#1	1.226517	417.6307	5.012973	35.25591	1.473295	.0783860
#2	1.221779	420.3933	5.030386	35.20989	1.473691	.0789298
#3	1.223172	415.2577	4.983204	35.18857	1.472219	.0764975

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**5.750075**	**.7553214**	**3.393620**	**17.90393**	**.5317687**	**.3826648**
Stddev	.049298	.0013617	.006955	.10685	.0048663	.0011314
%RSD	.8573520	.1802744	.2049570	.5967918	.9151131	.2956732
#1	5.728490	.7566372	3.385935	17.86449	.5265061	.3839090
#2	5.806484	.7554089	3.399483	18.02489	.5361057	.3823879
#3	5.715250	.7539181	3.395442	17.82240	.5326943	.3816976

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2947-03MSD Acquired: 8/26/2025 13:26:50 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.9086954	7.714925	8.214331	F 27.28646	4.955391	.3003511
Stddev	.0007587	.035490	.037973	.10387	.030893	.0015342
%RSD	.0834975	.4600129	.4622760	.3806709	.6234232	.5108105
#1	.9091829	7.708112	8.218994	27.36632	4.965418	.2986421
#2	.9090821	7.753327	8.249756	27.32402	4.980024	.3016097
#3	.9078212	7.683336	8.174241	27.16903	4.920729	.3008016

Elem	Sr4077
Units	ppm
Avg	.2257354
Stddev	.0009582
%RSD	.4244713
#1	.2246605
#2	.2264999
#3	.2260458

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	7138.107	121045.9	13250.78	5421.248	6264.518
Stddev	8.052	573.2	57.24	18.461	5.748
%RSD	.1128097	.4735462	.4319505	.3405363	.0917520
#1	7128.824	120835.3	13206.38	5435.892	6259.303
#2	7143.206	120607.8	13230.58	5400.510	6263.570
#3	7142.292	121694.6	13315.38	5427.342	6270.681

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCV03 Acquired: 8/26/2025 13:30:55 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV03 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**5.131213**	**5.217672**	**5.055123**	**5.182727**	**5.058389**	**10.16836**	**10.00084**
Stddev	.025187	.060381	.018454	.027355	.021451	.04908	.13315
%RSD	.4908549	1.157243	.3650479	.5278175	.4240687	.4826433	1.331414
#1	5.102157	5.197009	5.034645	5.156572	5.033668	10.21283	10.12047
#2	5.144657	5.285673	5.070462	5.211142	5.069406	10.17655	10.02468
#3	5.146826	5.170336	5.060263	5.180467	5.072092	10.11571	9.85738

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2548506**	**2.548480**	**25.25415**	**1.041179**	**2.510144**	**1.271485**	**5.082583**
Stddev	.0018049	.009125	.11227	.003183	.008794	.006279	.029937
%RSD	.7082193	.3580749	.4445684	.3057524	.3503309	.4938672	.5890024
#1	.2567162	2.538012	25.36069	1.041326	2.500197	1.264254	5.049230
#2	.2547222	2.554759	25.26484	1.044287	2.516884	1.274634	5.107128
#3	.2531132	2.552669	25.13691	1.037925	2.513352	1.275567	5.091390

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.489331**	**25.18477**	**2.521775**	**1.264303**	**24.63962**	**2.525608**	**2.525506**
Stddev	.011251	.20314	.009487	.003177	.10459	.014940	.009138
%RSD	.4519851	.8065820	.3762168	.2512690	.4244955	.5915256	.3618090
#1	2.502262	25.28544	2.510824	1.267018	24.52305	2.542042	2.527035
#2	2.483957	25.31791	2.527498	1.265082	24.67051	2.521935	2.533782
#3	2.481775	24.95096	2.527004	1.260809	24.72528	2.512848	2.515700

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**24.74972**	**5.034533**	**5.116992**	**5.003092**	**4.999734**	**5.092652**	**5.077512**
Stddev	.06266	.033548	.021272	.029180	.021949	.022348	.030756
%RSD	.2531589	.6663578	.4157053	.5832308	.4389961	.4388360	.6057287
#1	24.69004	5.070419	5.095237	4.970644	5.021986	5.087955	5.056478
#2	24.81498	5.029225	5.137744	5.027179	4.999116	5.116975	5.112810
#3	24.74416	5.003956	5.117995	5.011451	4.978102	5.073025	5.063247

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCV03 Acquired: 8/26/2025 13:30:55 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV03 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**4.987379**	**5.002821**	**4.994234**
Stddev	.030048	.019015	.042908
%RSD	.6024875	.3800776	.8591545
#1	4.960060	5.022428	5.009945
#2	5.019563	5.001575	4.945685
#3	4.982515	4.984460	5.027072

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**6059.168**	**104572.2**	**10379.83**	**4702.694**	**6860.099**
Stddev	17.233	216.2	62.69	25.367	16.968
%RSD	.2844056	.2067022	.6039610	.5394242	.2473438
#1	6079.028	104573.6	10309.84	4699.227	6879.662
#2	6048.164	104355.3	10398.82	4679.239	6851.270
#3	6050.313	104787.6	10430.83	4729.617	6849.367

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCB03 Acquired: 8/26/2025 13:35:07 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB03 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0012895**	**-.000063**	**.0003247**	**-.001424**	**-.001140**	**.0054012**	**-.002776**
Stddev	.0011480	.001998	.0011542	.001971	.001089	.0086932	.000724
%RSD	89.02689	3178.407	355.4394	138.4493	95.48229	160.9501	26.08475
#1	.0000020	-.000582	-.000561	-.003344	-.000170	.0008500	-.003407
#2	.0022063	.002143	-.000094	.000595	-.000933	-.000072	-.002936
#3	.0016601	-.001750	.001630	-.001522	-.002317	.015425	-.001985

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0000678**	**-.000017**	**.0000846**	**-.000359**	**-.000052**	**.0001582**	**-.003035**
Stddev	.0000749	.000035	.0058087	.000180	.000158	.0003323	.001928
%RSD	110.3819	200.1830	6867.854	50.13504	305.9853	209.9957	63.51691
#1	-.000015	-.000051	-.000678	-.000563	.000130	-.000225	-.001978
#2	.000087	.000018	.006237	-.000223	-.000132	.000345	-.005260
#3	.000131	-.000019	-.005305	-.000291	-.000153	.000355	-.001867

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0000734**	**-.015306**	**-.000056**	**.0003147**	**-.113055**	**.0005576**	**-.001460**
Stddev	.0007062	.009776	.000212	.0003391	.021343	.0014959	.000231
%RSD	961.9250	63.86719	375.8540	107.7512	18.87852	268.2787	15.81805
#1	.0006156	-.004469	-.000300	.0000459	-.137694	.0020907	-.001394
#2	.0003298	-.023460	.000085	.0006956	-.101195	-.000898	-.001716
#3	-.000725	-.017990	.000046	.0002025	-.100275	.000480	-.001269

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.257523**	**.0019236**	**.0003858**	**-.000050**	**-.000981**	**.0011282**	**.0002295**
Stddev	.042189	.0007413	.0000340	.000154	.001304	.0066454	.0017443
%RSD	16.38272	38.53718	8.823165	307.1850	132.9351	589.0022	760.1501
#1	-.305141	.0025000	.0004064	-.000208	-.001705	-.006535	.0004243
#2	-.242625	.0010874	.0004046	.000100	-.001763	.004624	.0018682
#3	-.224804	.0021833	.0003466	-.000043	.000525	.005296	-.001604

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCB03 Acquired: 8/26/2025 13:35:07 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB03 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0000461	-.002578	-.000054
Stddev	.0008968	.001197	.000077
%RSD	1947.278	46.40671	143.1437
#1	.0003160	-.003890	-.000003
#2	.0007768	-.001546	-.000143
#3	-.000955	-.002299	-.000016

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6462.364	112138.3	10477.32	4946.096	7951.601
Stddev	12.436	226.0	106.59	8.217	8.783
%RSD	.1924314	.2015112	1.017302	.1661395	.1104531
#1	6460.885	112242.1	10569.30	4955.092	7943.809
#2	6475.473	111879.1	10502.16	4938.985	7961.118
#3	6450.734	112293.8	10360.51	4944.210	7949.873

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2947-03A Acquired: 8/26/2025 13:39:26 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.7685968**	**1.850529**	**2.208018**	**1.606134**	**.6537588**	**247.1725**
Stddev	.0039585	.046226	.004294	.008371	.0032593	3.1258
%RSD	.5150292	2.498006	.1944748	.5211654	.4985485	1.264621
#1	.7659329	1.870188	2.211130	1.610489	.6560583	243.9476
#2	.7731455	1.883677	2.203119	1.611429	.6551892	247.3810
#3	.7667119	1.797723	2.209803	1.596483	.6500288	250.1888

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.093368**	**.1743694**	**.2190629**	**74.98845**	**1.303787**	**.3823204**
Stddev	.033925	.0019345	.0002054	1.29234	.003502	.0004021
%RSD	1.620616	1.109432	.0937667	1.723381	.2686005	.1051786
#1	2.056489	.1723159	.2190749	73.54459	1.307474	.3827167
#2	2.100368	.1746349	.2188517	75.38390	1.303384	.3823318
#3	2.123248	.1761575	.2192620	76.03685	1.300505	.3819127

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.131896**	**405.9326**	**4.151550**	**32.00642**	**1.517364**	**.0757193**
Stddev	.002307	2.5616	.072082	.61266	.000497	.0010670
%RSD	.2038099	.6310294	1.736260	1.914165	.0327715	1.409095
#1	1.134528	403.2829	4.072397	31.31030	1.517815	.0746108
#2	1.130226	408.3959	4.168837	32.24536	1.516831	.0758080
#3	1.130934	406.1189	4.213417	32.46360	1.517445	.0767391

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**5.183610**	**.7274054**	**3.436047**	**16.89900**	**.5318277**	**.3926000**
Stddev	.027644	.0124430	.014060	.11123	.0040298	.0004278
%RSD	.5332932	1.710603	.4091857	.6582015	.7577248	.1089540
#1	5.153490	.7131399	3.444820	16.81836	.5307099	.3930937
#2	5.207822	.7360215	3.443490	17.02589	.5362984	.3923397
#3	5.189519	.7330547	3.419830	16.85276	.5284748	.3923668

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2947-03A Acquired: 8/26/2025 13:39:26 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.8707039	6.998775	5.617552	F 25.57281	4.665341	.2690750
Stddev	.0039113	.115965	.080089	.11963	.020283	.0028817
%RSD	.4492093	1.656933	1.425685	.4677914	.4347648	1.070972
#1	.8707659	6.871960	5.533705	25.65432	4.680213	.2657711
#2	.8745838	7.024952	5.693261	25.62863	4.673574	.2703839
#3	.8667620	7.099415	5.625690	25.43547	4.642235	.2710699

Elem	Sr4077
Units	ppm
Avg	.1672069
Stddev	.0080165
%RSD	4.794346
#1	.1591016
#2	.1673876
#3	.1751315

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	7050.487	119872.7	13231.94	5314.398	6322.323
Stddev	11.382	388.1	145.54	10.381	8.897
%RSD	.1614299	.3237986	1.099891	.1953386	.1407220
#1	7038.889	119562.9	13377.95	5317.897	6312.331
#2	7050.931	119747.0	13230.98	5322.577	6325.250
#3	7061.639	120308.1	13086.88	5302.719	6329.387

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2949-01 Acquired: 8/26/2025 13:43:31 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0095326	-.003509	.0308737	.0007926	-.001597	40.69523	.1203366
Stddev	.0027769	.000892	.0004740	.0029719	.001750	.05372	.0003953
%RSD	29.13054	25.42005	1.535173	374.9375	109.5981	.1320117	.3285352
#1	.0078178	-.004174	.0304978	.0006130	-.001739	40.69017	.1207785
#2	.0080436	-.003858	.0307171	.0038502	-.003272	40.64422	.1200164
#3	.0127365	-.002496	.0314061	-.002085	.000220	40.75130	.1202149

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.0016529	.0009096	39.98383	.1129686	.0216112	.2029984	55.15035
Stddev	.0000190	.0000347	.14563	.0002627	.0001914	.0005928	.13705
%RSD	1.146644	3.815941	.3642173	.2325005	.8856876	.2920429	.2485041
#1	.0016718	.0008720	40.13992	.1132719	.0216045	.2035683	55.30664
#2	.0016339	.0009163	39.85162	.1128139	.0218059	.2023850	55.05067
#3	.0016529	.0009405	39.95994	.1128200	.0214232	.2030418	55.09375

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.4870371	15.29210	.0483527	.0026291	4.120888	.0938293	.1379819
Stddev	.0038370	.04936	.0001339	.0001761	.021357	.0010887	.0004209
%RSD	.7878323	.3227725	.2770140	6.698168	.5182533	1.160296	.3050443
#1	.4907788	15.34860	.0484929	.0027796	4.145199	.0927224	.1383883
#2	.4831113	15.25732	.0482260	.0026722	4.105148	.0938667	.1380095
#3	.4872212	15.27040	.0483393	.0024354	4.112318	.0948989	.1375478

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	2.364825	.0470143	.0046105	.0090443	1.714907	3.022015	1.249778
Stddev	.006316	.0011592	.0000556	.0003546	.006282	.014313	.002793
%RSD	.2671016	2.465627	1.206064	3.921008	.3662902	.4736316	.2234753
#1	2.367997	.0483495	.0045816	.0087210	1.722111	3.020130	1.249660
#2	2.368927	.0464282	.0046746	.0089884	1.710574	3.008738	1.252628
#3	2.357551	.0462651	.0045753	.0094236	1.712037	3.037178	1.247046

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2949-01 Acquired: 8/26/2025 13:43:31 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**3.158720**	**.0147803**	**.0919943**
Stddev	.005420	.0012414	.0006650
%RSD	.1715992	8.398763	.7229070
#1	3.161434	.0161849	.0926187
#2	3.162247	.0138303	.0912950
#3	3.152479	.0143258	.0920690

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**6242.699**	**107582.6**	**10999.90**	**4769.577**	**7201.277**
Stddev	3.280	85.6	25.71	7.632	3.415
%RSD	.0525397	.0795317	.2337268	.1600087	.0474262
#1	6243.510	107656.5	10980.33	4762.300	7204.906
#2	6245.497	107488.8	11029.02	4768.912	7198.126
#3	6239.090	107602.4	10990.36	4777.520	7200.800

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2940-01 Acquired: 8/26/2025 13:47:37 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.001385**	**-.001973**	**.0209749**	**.0027847**	**-.000516**	**.0963317**	**.0030679**
Stddev	.001191	.000975	.0003599	.0005001	.001657	.0077849	.0005298
%RSD	86.01376	49.44494	1.715979	17.95913	321.3402	8.081305	17.26918
#1	-.002316	-.002887	.0211854	.0022281	-.002231	.1020531	.0030509
#2	-.001797	-.000946	.0211800	.0029299	.001076	.0874665	.0025468
#3	-.000043	-.002085	.0205593	.0031962	-.000392	.0994753	.0036060

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000203**	**-.000007**	**5.183958**	**.0006184**	**.0077896**	**.1793693**	**2.608124**
Stddev	.000021	.000043	.004947	.0003058	.0000814	.0003339	.008472
%RSD	10.43436	616.3112	.0954354	49.45142	1.044907	.1861314	.3248406
#1	-.000210	-.000026	5.178399	.0006609	.0077839	.1796369	2.605122
#2	-.000180	-.000038	5.187878	.0009007	.0078737	.1794760	2.601561
#3	-.000220	.000042	5.185596	.0002935	.0077112	.1789952	2.617688

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0237525**	**2.019017**	**.0121898**	**.0001910**	**17.09727**	**.0001556**	**.0802379**
Stddev	.0002683	.012295	.0002021	.0003851	.06063	.0006443	.0002446
%RSD	1.129735	.6089632	1.658084	201.6232	.3546219	414.1773	.3048667
#1	.0235708	2.026629	.0122523	.0006226	17.12794	-.000025	.0805194
#2	.0236259	2.004833	.0123533	.0000679	17.02743	.000871	.0801180
#3	.0240607	2.025591	.0119638	-.000118	17.13644	-.000379	.0800764

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.7441315**	**.0066321**	**.0003891**	**-.000786**	**-.000135**	**6.719789**	**.0096037**
Stddev	.0038863	.0002520	.0000878	.000168	.001183	.030956	.0015234
%RSD	.5222646	3.799587	22.56239	21.42732	877.2994	.4606630	15.86269
#1	.7404763	.0067663	.0004826	-.000653	-.001379	6.755113	.0088619
#2	.7437045	.0063415	.0003083	-.000975	-.000000	6.697394	.0113560
#3	.7482137	.0067887	.0003765	-.000729	.000975	6.706861	.0085934

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2940-01 Acquired: 8/26/2025 13:47:37 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	6.363131	-.009592	.0252916
Stddev	.011405	.000591	.0001019
%RSD	.1792332	6.161026	.4030591
#1	6.356068	-.010078	.0251740
#2	6.376288	-.008934	.0253475
#3	6.357035	-.009765	.0253534

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6242.851	107138.8	10749.10	4691.008	7573.839
Stddev	1.867	238.9	37.31	25.458	7.239
%RSD	.0298988	.2229416	.3470973	.5426994	.0955769
#1	6241.753	106876.9	10780.50	4667.036	7581.151
#2	6241.795	107194.7	10707.86	4717.729	7573.690
#3	6245.007	107344.7	10758.95	4688.258	7566.675

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2937-01 Acquired: 8/26/2025 13:51:50 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0038544**	**.0001994**	**.0133034**	**.0035205**	**.0000236**	**3.355455**
Stddev	.0023344	.0006257	.0010844	.0027932	.0016034	.034443
%RSD	60.56565	313.7531	8.151364	79.34297	6804.334	1.026487
#1	.0015980	.0003083	.0120679	.0055477	-.000860	3.332412
#2	.0037055	-.000474	.0137443	.0003343	-.000944	3.395049
#3	.0062597	.000763	.0140978	.0046794	.001874	3.338902

Elem	Ba4934	Be2348	Cd2265	Ca3736	Cr2677	Co2286
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.3784552**	**.0001247**	**.0003253**	**115.5536**	**.0045190**	**.0103504**
Stddev	.0041289	.0000139	.0000160	1.3361	.0000883	.0001623
%RSD	1.090977	11.15093	4.916202	1.156269	1.954464	1.568203
#1	.3771904	.0001232	.0003276	115.0176	.0046205	.0103286
#2	.3830685	.0001117	.0003400	117.0744	.0044595	.0102001
#3	.3751067	.0001394	.0003082	114.5687	.0044770	.0105225

Elem	Cu2247	Fe2404	Mn2576	Mg2790	Ni2316	Ag3280
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0183002**	**10.37044**	**.3355254**	**44.85592**	**.0124472**	**.0007027**
Stddev	.0003987	.01086	.0027763	.54212	.0001986	.0002898
%RSD	2.178823	.1047045	.8274362	1.208576	1.595487	41.24409
#1	.0182309	10.36009	.3344253	44.54045	.0126765	.0005790
#2	.0179407	10.36948	.3386832	45.48190	.0123279	.0004953
#3	.0187291	10.38174	.3334678	44.54542	.0123374	.0010339

Elem	Na5895	V_2924	Zn2138	K_7664	B_2496	Mo2020
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**859.7894**	**.0156666**	**.1052062**	**17.89980**	**.1489861**	**.0075971**
Stddev	14.4298	.0018298	.0009874	.04431	.0020348	.0000959
%RSD	1.678291	11.67948	.9384977	.2475495	1.365741	1.262386
#1	876.4498	.0159268	.1043480	17.88063	.1478352	.0077077
#2	851.6585	.0173523	.1062853	17.95047	.1513355	.0075370
#3	851.2597	.0137206	.1049851	17.86831	.1477877	.0075466

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2937-01 Acquired: 8/26/2025 13:51:50 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	Sn1899	Ti3361	Si2881	P_1774	S_1820	Li6707
Units	ppm	ppm	ppm	ppm	ppm	ppm
Avg	-.003368	.1146633	F 13.31791	F 11.52546	F 27.27796	-.018998
Stddev	.000425	.0025414	.01102	.00896	.07825	.001088
%RSD	12.63107	2.216444	.0827272	.0777479	.2868572	5.725098
#1	-.003036	.1123458	13.31304	11.52276	27.30227	-.019412
#2	-.003220	.1173812	13.33052	11.51816	27.19045	-.019818
#3	-.003847	.1142630	13.31016	11.53546	27.34117	-.017764

Elem	Sr4077
Units	ppm
Avg	.4473321
Stddev	.0044314
%RSD	.9906266
#1	.4455852
#2	.4523707
#3	.4440404

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	5617.717	95419.08	10783.34	4259.903	6073.227
Stddev	8.243	31.22	102.39	12.834	2.425
%RSD	.1467405	.0327230	.9495072	.3012663	.0399279
#1	5626.952	95395.58	10798.35	4274.526	6075.940
#2	5611.104	95407.15	10674.28	4254.672	6072.471
#3	5615.094	95454.51	10877.40	4250.511	6071.270

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2937-02 Acquired: 8/26/2025 13:56:07 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.001502**	**-.001601**	**.0054779**	**.0019182**	**-.000327**	**.0835361**	**.0032070**
Stddev	.001575	.000882	.0008868	.0024628	.001978	.0029906	.0006381
%RSD	104.8308	55.07648	16.18918	128.3883	605.0479	3.580044	19.89579
#1	.000269	-.001058	.0065011	.0005042	.001826	.0804742	.0038249
#2	-.002031	-.001126	.0050006	.0004885	-.002065	.0836842	.0032456
#3	-.002745	-.002618	.0049319	.0047620	-.000743	.0864499	.0025505

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000155**	**.0000413**	**5.038296**	**.0001486**	**.0083835**	**.0152729**	**2.480684**
Stddev	.000006	.0000357	.013298	.0003117	.0001174	.0000559	.013303
%RSD	3.826300	86.33464	.2639413	209.6759	1.400817	.3657581	.5362654
#1	-.000149	.0000058	5.038085	-.000146	.0083631	.0152192	2.468209
#2	-.000161	.0000771	5.025104	.000475	.0085098	.0153307	2.494683
#3	-.000154	.0000410	5.051698	.000117	.0082776	.0152687	2.479161

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.5317183**	**1.907311**	**.0101399**	**-.000200**	**14.59767**	**-.002736**	**.0987019**
Stddev	.0014842	.020107	.0001239	.000274	.06183	.002261	.0000381
%RSD	.2791373	1.054182	1.222166	136.9054	.4235788	82.64412	.0386198
#1	.5334303	1.927950	.0102084	.000058	14.53604	-.005327	.0987340
#2	.5307945	1.887783	.0099969	-.000488	14.65970	-.001723	.0986598
#3	.5309300	1.906199	.0102145	-.000170	14.59726	-.001158	.0987119

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**1.260000**	**.0082989**	**.0001626**	**-.001240**	**-.000674**	**6.324025**	**.0182495**
Stddev	.044064	.0003224	.0001557	.000368	.001305	.018168	.0008731
%RSD	3.497111	3.884802	95.79098	29.68675	193.5895	.2872788	4.784171
#1	1.310640	.0079292	.0000892	-.000878	-.001451	6.310461	.0192523
#2	1.230406	.0084458	.0003415	-.001614	.000832	6.316947	.0178378
#3	1.238953	.0085215	.0000571	-.001227	-.001403	6.344666	.0176584

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2937-02 Acquired: 8/26/2025 13:56:07 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	6.233865	-.007350	.0261358
Stddev	.018037	.001846	.0001088
%RSD	.2893438	25.12062	.4160869
#1	6.229729	-.007961	.0260997
#2	6.253611	-.005275	.0262580
#3	6.218255	-.008812	.0260496

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6279.609	109495.2	10563.20	4793.939	7631.983
Stddev	6.600	83.6	27.45	16.108	2.272
%RSD	.1051093	.0763112	.2598336	.3360152	.0297652
#1	6276.128	109538.4	10534.23	4777.028	7631.811
#2	6287.222	109398.9	10566.57	4809.102	7634.335
#3	6275.479	109548.3	10588.81	4795.688	7629.802

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2937-02DUP Acquired: 8/26/2025 14:00:19 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000596**	**-.002214**	**.0049360**	**.0020785**	**-.001762**	**.0781920**	**.0034385**
Stddev	.001032	.000849	.0003808	.0024685	.000670	.0071503	.0010338
%RSD	173.0723	38.35452	7.714733	118.7630	38.03905	9.144492	30.06634
#1	.000501	-.001505	.0052642	.0009214	-.002106	.0711790	.0023760
#2	-.000741	-.001981	.0050252	.0049129	-.000989	.0854720	.0034985
#3	-.001548	-.003155	.0045185	.0004011	-.002190	.0779251	.0044411

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.000078**	**.0000267**	**4.857713**	**.0000741**	**.0078464**	**.0146240**	**2.478676**
Stddev	.000034	.0000308	.023152	.0001874	.0001462	.0002563	.009032
%RSD	43.08358	115.4163	.4766050	252.7399	1.863300	1.752486	.3644004
#1	-.000117	.0000055	4.838193	.0002485	.0076790	.0145025	2.484747
#2	-.000058	.0000126	4.851654	-.000124	.0079112	.0149184	2.482985
#3	-.000060	.0000620	4.883292	.000098	.0079491	.0144510	2.468296

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0607430**	**1.855025**	**.0092709**	**.0000800**	**15.00823**	**.0007140**	**.0935749**
Stddev	.0002362	.017220	.0002628	.0001008	.09450	.0016345	.0004951
%RSD	.3888469	.9282684	2.834671	126.0232	.6296702	228.9266	.5291331
#1	.0604912	1.835344	.0092885	.0001171	14.97931	.0024942	.0930370
#2	.0609596	1.862408	.0089998	-.000034	15.11381	.0003668	.0936759
#3	.0607782	1.867321	.0095245	.000157	14.93157	-.000719	.0940117

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.7903331**	**.0068000**	**.0001088**	**-.001172**	**.0004642**	**6.249253**	**.0179201**
Stddev	.0459867	.0001775	.0001904	.000485	.0010856	.011753	.0013048
%RSD	5.818647	2.610862	175.0350	41.36916	233.8748	.1880727	7.281203
#1	.7975598	.0065963	-.000015	-.000681	-.000763	6.260672	.0169976
#2	.8322785	.0069222	.000013	-.001651	.000857	6.249894	.0173498
#3	.7411609	.0068814	.000328	-.001185	.001299	6.237192	.0194130

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2937-02DUP Acquired: 8/26/2025 14:00:19 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	5.702188	-.008668	.0252538
Stddev	.009665	.002512	.0001201
%RSD	.1694962	28.97866	.4755178
#1	5.691162	-.005773	.0251267
#2	5.706210	-.009960	.0253654
#3	5.709193	-.010271	.0252692

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6423.166	112394.1	11161.86	4908.062	7804.507
Stddev	6.625	266.2	7.06	16.998	7.222
%RSD	.1031467	.2368548	.0632317	.3463315	.0925389
#1	6416.528	112168.5	11158.79	4903.263	7798.126
#2	6429.778	112326.2	11169.94	4893.979	7803.048
#3	6423.191	112687.8	11156.86	4926.943	7812.348

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2937-02LX5 Acquired: 8/26/2025 14:04:34 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0007292**	**-.000893**	**.0020178**	**-.001217**	**-.001705**	**.0172173**	**-.001684**
Stddev	.0012387	.000317	.0006366	.002552	.001451	.0036898	.000690
%RSD	169.8814	35.45488	31.54867	209.6266	85.10043	21.43104	40.94436
#1	.0008874	-.001257	.0018081	-.002001	-.002175	.0194799	-.001425
#2	.0018811	-.000736	.0027328	.001634	-.000077	.0129594	-.002466
#3	-.000581	-.000685	.0015125	-.003285	-.002863	.0192124	-.001162

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0000003**	**-.000008**	**1.170618**	**-.000342**	**.0018068**	**.0034959**	**.5562005**
Stddev	.0000124	.000031	.013334	.000102	.0001339	.0001421	.0077620
%RSD	4430.183	368.7887	1.139032	29.87341	7.412413	4.064235	1.395534
#1	-.000012	-.000029	1.183658	-.000245	.0018784	.0035957	.5498808
#2	.000001	-.000023	1.157009	-.000448	.0018897	.0035587	.5648644
#3	.000013	.000027	1.171187	-.000332	.0016523	.0033332	.5538564

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1221812**	**.4261452**	**.0021284**	**.0004684**	**3.316560**	**-.001027**	**.0210016**
Stddev	.0006306	.0211801	.0002260	.0002357	.043318	.001657	.0003064
%RSD	.5161441	4.970157	10.61891	50.30425	1.306099	161.3686	1.459083
#1	.1226478	.4020790	.0023860	.0007397	3.315128	-.001964	.0206515
#2	.1224321	.4344084	.0020361	.0003145	3.360576	.000886	.0212213
#3	.1214638	.4419481	.0019632	.0003512	3.273976	-.002002	.0211319

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1600623**	**-.000825**	**-.000158**	**.0000038**	**.0003269**	**1.415999**	**.0031138**
Stddev	.0344243	.000326	.000104	.0002733	.0011683	.023377	.0003874
%RSD	21.50682	39.52196	65.87876	7240.404	357.4030	1.650943	12.44193
#1	.1237383	-.000666	-.000264	-.000207	.0003157	1.433759	.0034280
#2	.1922050	-.001199	-.000155	-.000094	-.000836	1.424724	.0026809
#3	.1642436	-.000608	-.000056	.000312	.001501	1.389514	.0032324

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2937-02LX5 Acquired: 8/26/2025 14:04:34 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	1.406671	-.002815	.0058039
Stddev	.017623	.001309	.0000316
%RSD	1.252802	46.51677	.5450946
#1	1.425299	-.004285	.0058046
#2	1.404450	-.002389	.0058351
#3	1.390264	-.001772	.0057719

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6374.771	111443.2	10802.55	4901.372	7787.689
Stddev	7.315	379.4	37.09	32.342	5.423
%RSD	.1147478	.3404735	.3433488	.6598533	.0696331
#1	6382.130	111343.9	10788.45	4911.151	7782.193
#2	6367.501	111123.3	10844.62	4865.269	7787.839
#3	6374.681	111862.5	10774.57	4927.696	7793.035

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2937-02MS Acquired: 8/26/2025 14:08:51 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.7884252**	**1.904372**	**.9298804**	**1.951801**	**.7980505**	**2.090875**	**.1919853**
Stddev	.0032828	.042806	.0007640	.008888	.0007251	.004908	.0004463
%RSD	.4163724	2.247786	.0821583	.4553870	.0908627	.2347267	.2324586
#1	.7867042	1.881622	.9290034	1.941781	.7979740	2.087828	.1922407
#2	.7863607	1.877744	.9302362	1.954885	.7988108	2.096537	.1914699
#3	.7922106	1.953749	.9304015	1.958736	.7973666	2.088261	.1922451

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2040330**	**.1877842**	**5.979608**	**.4189600**	**.2038979**	**.3212342**	**5.248330**
Stddev	.0004252	.0002923	.024801	.0019722	.0001400	.0002012	.032253
%RSD	.2083866	.1556740	.4147534	.4707246	.0686647	.0626228	.6145335
#1	.2044955	.1877925	5.959269	.4166829	.2037467	.3211805	5.213960
#2	.2039443	.1874878	6.007237	.4201232	.2039238	.3210653	5.277935
#3	.2036592	.1880723	5.972318	.4200739	.2040231	.3214567	5.253097

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2627534**	**3.874880**	**.4947271**	**.0721428**	**16.29092**	**.3027012**	**.3041804**
Stddev	.0013131	.007935	.0003991	.0002433	.03238	.0022052	.0008560
%RSD	.4997391	.2047882	.0806769	.3372794	.1987780	.7285196	.2814144
#1	.2620495	3.867699	.4951064	.0718907	16.25637	.3002203	.3051687
#2	.2619422	3.873541	.4947642	.0721615	16.29581	.3034448	.3036742
#3	.2642683	3.883399	.4943108	.0723762	16.32058	.3044386	.3036982

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**9.968497**	**.2858793**	**.4362292**	**.6708479**	**.1949910**	**6.923383**	**5.398275**
Stddev	.058881	.0003342	.0006903	.0007679	.0008704	.026744	.029998
%RSD	.5906708	.1168964	.1582529	.1144596	.4463981	.3862829	.5556895
#1	9.915251	.2859454	.4354361	.6707273	.1939870	6.900887	5.376749
#2	10.03173	.2861755	.4365565	.6716690	.1954522	6.952953	5.385537
#3	9.95851	.2855170	.4366951	.6701476	.1955338	6.916310	5.432540

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2937-02MS Acquired: 8/26/2025 14:08:51 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	6.053367	.1889530	.2149189
Stddev	.024459	.0012244	.0008437
%RSD	.4040494	.6480130	.3925659
#1	6.029506	.1894471	.2149117
#2	6.052213	.1898532	.2157662
#3	6.078383	.1875587	.2140788

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6220.053	108585.1	10233.62	4777.944	7461.630
Stddev	7.159	313.0	43.83	19.423	5.627
%RSD	.1151019	.2882545	.4283198	.4065105	.0754166
#1	6219.269	108769.9	10218.00	4759.139	7457.118
#2	6213.318	108223.7	10199.73	4776.764	7459.836
#3	6227.572	108761.8	10283.12	4797.930	7467.935

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2937-02MSD Acquired: 8/26/2025 14:12:49 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.7674565**	**1.799443**	**.9024690**	**1.896332**	**.7823184**	**2.063145**	**.1936740**
Stddev	.0018324	.012181	.0013666	.003793	.0030762	.003838	.0006569
%RSD	.2387659	.6769127	.1514306	.2000282	.3932097	.1860467	.3391932
#1	.7673007	1.801032	.9033470	1.891958	.7850465	2.059255	.1943991
#2	.7657069	1.786546	.9008944	1.898708	.7789844	2.063250	.1931186
#3	.7693618	1.810752	.9031655	1.898331	.7829244	2.066929	.1935042

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.1894229**	**.1824670**	**5.990213**	**.4059931**	**.1985423**	**.3129151**	**5.457507**
Stddev	.0005469	.0001788	.025013	.0009280	.0001692	.0005291	.005865
%RSD	.2887335	.0979724	.4175698	.2285672	.0852171	.1690833	.1074670
#1	.1900464	.1825499	6.011253	.4050654	.1985529	.3132931	5.452325
#2	.1890243	.1822618	5.962556	.4059924	.1983680	.3123104	5.463874
#3	.1891981	.1825892	5.996830	.4069214	.1987059	.3131417	5.456321

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2580143**	**3.833503**	**.4816097**	**.0706351**	**17.89512**	**.2972997**	**.2989562**
Stddev	.0011600	.024940	.0006179	.0004531	.04483	.0012228	.0017762
%RSD	.4495735	.6505806	.1282915	.6414176	.2505063	.4113087	.5941369
#1	.2588668	3.852002	.4820989	.0706607	17.84466	.2969083	.2974532
#2	.2566934	3.805140	.4809154	.0701698	17.91034	.2963204	.3009163
#3	.2584828	3.843367	.4818149	.0710749	17.93035	.2986702	.2984993

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**10.35805**	**.2674840**	**.4263543**	**.6556990**	**.1922554**	**7.328459**	**5.242374**
Stddev	.02994	.0003940	.0008440	.0026589	.0004029	.004327	.008508
%RSD	.2890061	.1472805	.1979496	.4055082	.2095563	.0590461	.1622922
#1	10.32367	.2672044	.4273172	.6526300	.1918638	7.326980	5.232569
#2	10.37836	.2679346	.4257429	.6571547	.1922339	7.325066	5.247813
#3	10.37210	.2673131	.4260028	.6573121	.1926687	7.333332	5.246739

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2937-02MSD Acquired: 8/26/2025 14:12:49 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**5.999047**	**.1840704**	**.2138573**
Stddev	.024148	.0021072	.0003760
%RSD	.4025323	1.144761	.1758084
#1	5.971283	.1843594	.2141384
#2	6.010689	.1818337	.2134302
#3	6.015168	.1860181	.2140033

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**6233.432**	**108288.5**	**10724.10**	**4752.044**	**7483.587**
Stddev	9.423	327.2	22.39	26.678	9.766
%RSD	.1511705	.3021819	.2088022	.5614075	.1305013
#1	6223.252	108662.1	10709.57	4779.516	7472.896
#2	6235.196	108052.9	10749.88	4726.237	7485.824
#3	6241.849	108150.5	10712.84	4750.379	7492.040

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2937-02A Acquired: 8/26/2025 14:16:47 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.7792972	1.885151	.9193083	1.938632	.7940629	2.106745	.2047715
Stddev	.0035194	.038638	.0006036	.004493	.0031140	.012306	.0010604
%RSD	.4516123	2.049614	.0656569	.2317717	.3921605	.5841201	.5178452
#1	.7808977	1.929766	.9188174	1.934618	.7948828	2.111717	.2044436
#2	.7752620	1.862979	.9191251	1.937791	.7966850	2.092731	.2039138
#3	.7817319	1.862707	.9199822	1.943486	.7906210	2.115787	.2059571

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.1882241	.1850815	6.018239	.4107932	.2006124	.3172318	5.688486
Stddev	.0011057	.0001052	.021491	.0003939	.0002258	.0006624	.060581
%RSD	.5874567	.0568339	.3570941	.0958877	.1125344	.2088104	1.064971
#1	.1889760	.1851458	6.042726	.4104311	.2006225	.3169181	5.693639
#2	.1887418	.1851386	6.002513	.4107359	.2008329	.3167846	5.625494
#3	.1869545	.1849601	6.009477	.4112126	.2003817	.3179928	5.746326

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	.5099395	3.887288	.4901588	.0696736	19.18033	.3108215	.3028944
Stddev	.0013627	.030088	.0001361	.0004057	.10150	.0032205	.0007485
%RSD	.2672205	.7739992	.0277714	.5822615	.5291999	1.036133	.2471192
#1	.5105292	3.913183	.4902938	.0698006	19.15517	.3144881	.3030789
#2	.5083813	3.894398	.4901612	.0700006	19.09377	.3084508	.3035334
#3	.5109079	3.854282	.4900216	.0692196	19.29204	.3095257	.3020709

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	11.19036	.2674131	.4279025	.6614455	.1983795	7.496269	5.381291
Stddev	.10419	.0017419	.0004033	.0019562	.0009475	.068497	.015797
%RSD	.9310281	.6513773	.0942578	.2957529	.4776249	.9137501	.2935536
#1	11.23380	.2688134	.4274524	.6622779	.1994703	7.488598	5.392731
#2	11.07148	.2679634	.4282311	.6592108	.1979069	7.431930	5.363267
#3	11.26579	.2654626	.4280239	.6628480	.1977612	7.568279	5.387876

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: Q2937-02A Acquired: 8/26/2025 14:16:47 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**6.056837**	**.1991208**	**.2216070**
Stddev	.016756	.0019252	.0004760
%RSD	.2766379	.9668301	.2148141
#1	6.073625	.2001035	.2221236
#2	6.040114	.1969026	.2211860
#3	6.056772	.2003563	.2215116

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**6172.610**	**108269.8**	**10961.54**	**4749.329**	**7418.037**
Stddev	12.710	261.5	34.79	1.511	6.766
%RSD	.2059089	.2415251	.3173745	.0318059	.0912152
#1	6186.347	108252.1	10933.56	4749.300	7423.095
#2	6170.214	108539.8	10950.56	4747.833	7420.665
#3	6161.269	108017.7	11000.49	4750.854	7410.350

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCV04 Acquired: 8/26/2025 14:20:44 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV04 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**5.103610**	**5.073365**	**4.997521**	**5.158756**	**5.049329**	**10.11860**	**10.04551**
Stddev	.006613	.104250	.002651	.011618	.008643	.00683	.04936
%RSD	.1295804	2.054849	.0530549	.2252056	.1711624	.0675021	.4913580
#1	5.109849	4.985360	4.994890	5.147611	5.059293	10.11078	10.04206
#2	5.104304	5.188497	4.997481	5.170795	5.043877	10.12338	9.99797
#3	5.096677	5.046238	5.000192	5.157861	5.044815	10.12163	10.09651

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.2462281**	**2.518342**	**25.08160**	**1.036604**	**2.484044**	**1.262454**	**5.234644**
Stddev	.0009166	.002738	.05539	.001792	.001825	.001180	.033493
%RSD	.3722626	.1087226	.2208230	.1728906	.0734581	.0934376	.6398375
#1	.2470521	2.515184	25.04905	1.038037	2.482405	1.263140	5.209753
#2	.2452408	2.519800	25.14555	1.034594	2.483716	1.261092	5.272724
#3	.2463914	2.520044	25.05019	1.037179	2.486010	1.263131	5.221455

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**2.489923**	**24.85122**	**2.494785**	**1.266855**	**26.08243**	**2.527856**	**2.542302**
Stddev	.008610	.09393	.002434	.001368	.14116	.002717	.003136
%RSD	.3457926	.3779861	.0975614	.1079598	.5412058	.1074872	.1233482
#1	2.487222	24.76466	2.492699	1.265428	25.99350	2.524951	2.545827
#2	2.499559	24.95110	2.494196	1.266983	26.24519	2.530336	2.541256
#3	2.482987	24.83790	2.497459	1.268155	26.00859	2.528280	2.539823

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**25.80210**	**4.872695**	**5.103493**	**4.940598**	**4.996005**	**5.234014**	**5.005258**
Stddev	.13871	.015414	.005211	.006422	.019813	.025962	.021223
%RSD	.5375864	.3163326	.1021109	.1299882	.3965791	.4960250	.4240190
#1	25.67759	4.888155	5.102120	4.937366	4.986661	5.204082	4.998133
#2	25.95161	4.857327	5.109253	4.947995	5.018762	5.247540	5.029126
#3	25.77709	4.872604	5.099105	4.936434	4.982592	5.250420	4.988513

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCV04 Acquired: 8/26/2025 14:20:44 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCV04 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	**4.915495**	**5.047142**	**5.023871**
Stddev	.032360	.013788	.008519
%RSD	.6583203	.2731904	.1695693
#1	4.903301	5.040970	5.019825
#2	4.952180	5.062939	5.033659
#3	4.891004	5.037519	5.018128

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	**6114.675**	**104932.9**	**10823.11**	**4674.163**	**6948.322**
Stddev	6.801	132.3	19.91	19.140	1.880
%RSD	.1112286	.1261201	.1839963	.4094835	.0270627
#1	6108.056	104973.1	10800.54	4692.076	6947.341
#2	6121.645	104785.1	10838.22	4653.996	6950.490
#3	6114.324	105040.5	10830.56	4676.416	6947.134

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCB04 Acquired: 8/26/2025 14:24:56 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB04 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0014848**	**-.000202**	**.0005327**	**-.002142**	**-.001477**	**-.005007**	**-.002387**
Stddev	.0007286	.001777	.0008779	.003741	.000264	.008165	.000908
%RSD	49.07176	877.3627	164.8101	174.6865	17.86270	163.0500	38.05111
#1	.0018297	.001342	.0003790	-.002334	-.001770	-.007143	-.001915
#2	.0006478	.000194	.0014772	.001692	-.001258	-.011892	-.001813
#3	.0019768	-.002144	-.000258	-.005783	-.001403	.004013	-.003435

Elem	Be2348	Cd2265	Ca3736	Cr2677	Co2286	Cu2247	Fe2404
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0000152**	**.0000388**	**-.000988**	**-.000509**	**-.000014**	**-.000057**	**.0032434**
Stddev	.0000204	.0000302	.000939	.000031	.000124	.000405	.0006927
%RSD	134.1016	77.87557	95.03572	6.043658	855.0599	704.2224	21.35623
#1	.0000086	.0000045	-.002040	-.000497	-.000041	-.000320	.0027103
#2	-.000001	.0000614	-.000691	-.000485	-.000123	.000409	.0040263
#3	.000038	.0000504	-.000234	-.000543	.000121	-.000261	.0029936

Elem	Mn2576	Mg2790	Ni2316	Ag3280	Na5895	V_2924	Zn2138
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**.0000639**	**.0200344**	**.0000338**	**.0005075**	**-.041121**	**.0010945**	**-.000797**
Stddev	.0004166	.0041592	.0001122	.0002748	.027017	.0009175	.000145
%RSD	651.9660	20.76012	331.8744	54.15828	65.69993	83.82936	18.18900
#1	.0001198	.0161592	-.000038	.0006027	-.061652	.0000387	-.000632
#2	-.000378	.0244288	-.000023	.0001977	-.051197	.0016989	-.000902
#3	.000450	.0195152	.000163	.0007221	-.010515	.0015460	-.000857

Elem	K_7664	B_2496	Mo2020	Sn1899	Ti3361	Si2881	P_1774
Units	ppm	ppm	ppm	ppm	ppm	ppm	ppm
Avg	**-.122182**	**.0014863**	**.0003147**	**.0002028**	**-.000569**	**-.004744**	**-.001041**
Stddev	.049368	.0005867	.0000512	.0007478	.000715	.007097	.001392
%RSD	40.40536	39.47356	16.26343	368.7789	125.4688	149.6060	133.7207
#1	-.178393	.0020951	.0002572	-.000304	-.001394	-.011203	-.002439
#2	-.102286	.0014393	.0003315	.001062	-.000176	.002854	.000345
#3	-.085866	.0009245	.0003553	-.000149	-.000138	-.005883	-.001029

Reviewed By:
On:
Inst Id :P4
LB :LB136973

Sample Name: CCB04 Acquired: 8/26/2025 14:24:56 Type: Unk
Method: 6010-200.7 NEW(v130) Mode: CONC Corr. Factor: 1.000000
User: Jaswal Custom ID1: CCB04 Custom ID2: Custom ID3:
Comment:

Elem	S_1820	Li6707	Sr4077
Units	ppm	ppm	ppm
Avg	.0070765	-.003189	-.000007
Stddev	.0031055	.001968	.000075
%RSD	43.88388	61.72213	1104.505
#1	.0043942	-.004823	.000044
#2	.0104787	-.003739	-.000093
#3	.0063567	-.001004	.000029

Int. Std.	Y_2243	Y_3600	Y_3710	Y_2243	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	6378.262	111528.7	10786.87	4926.168	7863.257
Stddev	8.925	96.0	34.60	18.198	19.696
%RSD	.1399346	.0860596	.3207408	.3694164	.2504801
#1	6385.717	111444.2	10747.29	4920.744	7882.817
#2	6368.371	111509.0	10801.95	4911.298	7843.428
#3	6380.697	111633.1	10811.37	4946.461	7863.527

Alliance TECHNICAL GROUP

Soil/Sludge Metals Preparation Sheet

PB169215

SOP ID : M3050B-Digestion-20

SDG No : N/A

Matrix : SOIL

Pippete ID: ICP A

Balance ID : M SC-2

Filter paper ID : N/A

pH Strip ID : N/A

Hood ID : #3

Block ID: 1. HOT BLOCK #5 2. N/A

Start Digest Date: 08/12/2025 **Time :** 13:05 **Temp :** 97 °C

End Digest Date: 08/12/2025 **Time :** 11:15 **Temp :** 98 °C

Digestion tube ID: M6054

Block thermometer ID: MET-DIG. #5

Dig Technician Signature: SLG

Supervisor Signature: JSP

Temp : 1. 96°C 2. N/A

Standared Name	MLS USED	STD REF. # FROM LOG
LFS-1	1.00	M6180
LFS-2	1.00	M6181
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	5.00	M6158
1:1 HNO3	10.00	MP86404
30% H2O2	3.00	M6170
Conc. HCL	10.00	M6151
PTFE Boiling Stones	N/A	M5581
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK#5 CELL#35 97C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
~~08/12/2025~~ 1215	SLG MET DIG	JSP MetLab
08/12/2025	Preparation Group	Analysis Group

Alliance TECHNICAL GROUP

Soil/Sludge Metals Preparation Sheet

PB169215

Lab Sample ID	Client Sample ID	pH	Initial Weight (g)	Final Vol (ml)	Color Before	Color After	Texture	Artifact	Comment	Prep Pos
PB169215BL	PBS215	N/A	2.00	100	Colorless	Colorless	Fine	N/A	N/A	1
PB169215BS	LCS215	N/A	2.00	100	Colorless	Coloress	Fine	N/A	M6180,M6181	2
Q2829-01	SILICA	N/A	2.36	100	White	Yellow	Medium	N/A	N/A	3
Q2830-01	BIN0009-DRIVEWAY-TP-SOUTH EAST	N/A	2.10	100	Brown	Yellow	Medium	N/A	N/A	4
Q2830-03	BIN0009-DRIVEWAY-TP-WEST	N/A	2.25	100	Reddish Brown	Yellow	Medium	N/A	N/A	5
Q2830-05	BIN0009-DRIVEWAY-TP-WESTSIDE	N/A	2.27	100	Reddish Brown	Yellow	Medium	N/A	N/A	6
Q2830-07	BIN0009-DRIVEWAY-TP-EASTSIDE	N/A	2.40	100	Brown	Yellow	Medium	N/A	N/A	7
Q2831-01	VNJ-238	N/A	2.20	100	Brown	Yellow	Medium	N/A	N/A	8
Q2832-01	TG-S01	N/A	2.14	100	Dark Brown	Yellow	Medium	N/A	N/A	9
Q2832-03	TG-S02	N/A	2.12	100	Dark Brown	Yellow	Medium	N/A	N/A	10
Q2832-05	TG-S03	N/A	2.20	100	Dark Brown	Yellow	Medium	N/A	N/A	11
Q2832-07	TG-S04	N/A	2.21	100	Dark Brown	Yellow	Medium	N/A	N/A	12
Q2832-09	TG-S05	N/A	2.35	100	Black	Yellow	Medium	N/A	N/A	13
Q2832-09DUP	TG-S05DUP	N/A	2.30	100	Black	Yellow	Medium	N/A	N/A	14
Q2832-09MS	TG-S05MS	N/A	2.31	100	Black	Yellow	Medium	N/A	M6180,M6181	15
Q2832-09MSD	TG-S05MSD	N/A	2.32	100	Black	Yellow	Medium	N/A	M6180,M6181	16

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PB169215 **WorkList ID :** 191224 **Department :** Digestion **Date :** 08-12-2025 12:23:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2829-01	SILICA	Solid	Metals Group3	Cool 4 deg C	VIVA01	D21	08/11/2025	6010D
Q2830-01	BIN0009-DRIVEWAY-TP-SOUT	Solid	Metals ICP-TAL	Cool 4 deg C	PSEG03	D31	08/11/2025	6010D
Q2830-03	BIN0009-DRIVEWAY-TP-WEST	Solid	Metals ICP-TAL	Cool 4 deg C	PSEG03	D31	08/11/2025	6010D
Q2830-05	BIN0009-DRIVEWAY-TP-WEST	Solid	Metals ICP-TAL	Cool 4 deg C	PSEG03	D31	08/11/2025	6010D
Q2830-07	BIN0009-DRIVEWAY-TP-EAST	Solid	Metals ICP-TAL	Cool 4 deg C	PSEG03	D31	08/11/2025	6010D
Q2831-01	VNJ-238	Solid	Metals ICP-TAL	Cool 4 deg C	PSEG03	D21	08/11/2025	6010D
Q2832-01	TG-S01	Solid	Metals ICP-TAL	Cool 4 deg C	PORT06	J21	08/11/2025	6010D
Q2832-03	TG-S02	Solid	Metals ICP-TAL	Cool 4 deg C	PORT06	J21	08/11/2025	6010D
Q2832-05	TG-S03	Solid	Metals ICP-TAL	Cool 4 deg C	PORT06	J21	08/11/2025	6010D
Q2832-07	TG-S04	Solid	Metals ICP-TAL	Cool 4 deg C	PORT06	J21	08/11/2025	6010D
Q2832-09	TG-S05	Solid	Metals ICP-TAL	Cool 4 deg C	PORT06	J21	08/12/2025	6010D

Date/Time 08/12/25 12:40
Raw Sample Received by: SLSmeltzer
Raw Sample Relinquished by: [signature]

Date/Time 08/12/25 13:40
Raw Sample Received by: [signature]
Raw Sample Relinquished by: SLSmeltzer

PERCENT SOLID

Supervisor: rubina
Analyst: jignesh
Date: 8/12/2025

OVENTEMP IN Celsius(°C): 108
Time IN: 17:10
In Date: 08/11/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:22
Out Date: 08/12/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC: LB136776

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2817-01	LEAD-PAINT-CHIPS	1	1.00	1.00	2.00	2.00	100.0	LEAD-PAINT-CHIPS
Q2818-01	B-2-5-1	2	1.15	10.59	11.74	10.42	87.5	
Q2818-02	B-3-5-2	3	1.16	10.66	11.82	10.66	89.1	
Q2823-01	SU-04-081125	4	1.13	10.70	11.83	11.11	93.3	
Q2823-02	SU-04-081125-E2	5	1.14	11.10	12.24	11.34	91.9	
Q2826-01	WC1	6	1.18	10.81	11.99	10.7	88.1	
Q2827-01	TP-8	7	1.16	11.41	12.57	7.33	54.1	
Q2827-02	TP-8-EPH	8	1.16	10.21	11.37	7.68	63.9	
Q2827-03	TP-8-VOC	9	1.18	10.87	12.05	9.22	74.0	
Q2827-05	TP-9	10	1.19	10.61	11.8	9.66	79.8	
Q2827-06	TP-9-EPH	11	1.16	10.06	11.22	8.93	77.2	
Q2827-07	TP-9-VOC	12	1.15	10.48	11.63	9.59	80.5	
Q2828-01	POWDER	13	1.19	10.32	11.51	11.45	99.4	
Q2829-01	SILICA	14	1.00	1.00	2.00	2.00	100.0	silica
Q2830-01	BIN0009-DRIVEWAY-TP-SOUTH-EAST	15	1.18	10.57	11.75	10.37	86.9	
Q2830-02	BIN0009-DRIVEWAY-TP-SOUTH-EAST	16	1.18	10.03	11.21	9.83	86.2	
Q2830-03	BIN0009-DRIVEWAY-TP-WEST	17	1.18	10.74	11.92	10.37	85.6	
Q2830-04	BIN0009-DRIVEWAY-TP-WEST	18	1.15	10.59	11.74	10.21	85.6	
Q2830-05	BIN0009-DRIVEWAY-TP-WESTSIDE	19	1.19	10.46	11.65	9.66	81.0	
Q2830-06	BIN0009-DRIVEWAY-TP-WESTSIDE	20	1.15	10.94	12.09	9.36	75.0	
Q2830-07	BIN0009-DRIVEWAY-TP-EASTSIDE	21	1.18	10.48	11.66	9.22	76.7	
Q2830-08	BIN0009-DRIVEWAY-TP-EASTSIDE	22	1.19	10.66	11.85	9.67	79.5	
Q2831-01	VNJ-238	23	1.18	10.59	11.77	11.04	93.1	
Q2831-02	VNJ-238-VOC	24	1.19	10.62	11.81	11.04	92.7	

PERCENT SOLID

Supervisor: rubina
Analyst: jignesh
Date: 8/12/2025

OVENTEMP IN Celsius(°C): 108
Time IN: 17:10
In Date: 08/11/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:22
Out Date: 08/12/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID-OVEN

QC: LB136776

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

VB136776

WorkList Name : %1-081125 **WorkList ID :** 191189 **Department :** Wet-Chemistry **Date :** 08-11-2025 08:22:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2817-01	LEAD-PAINT-CHIPS	Solid	Percent Solids	Cool 4 deg C	HOME01	D31	08/11/2025	Chemtech -SO
Q2818-01	B-2-5-1	Solid	Percent Solids	Cool 4 deg C	EARTH03	J21	08/11/2025	Chemtech -SO
Q2818-02	B-3-5-2	Solid	Percent Solids	Cool 4 deg C	EARTH03	J21	08/11/2025	Chemtech -SO
Q2823-01	SU-04-081125	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	08/11/2025	Chemtech -SO
Q2823-02	SU-04-081125-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	D31	08/11/2025	Chemtech -SO
Q2826-01	WC1	Solid	Percent Solids	Cool 4 deg C	GENV01	D31	08/11/2025	Chemtech -SO
Q2827-01	TP-8	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2827-02	TP-8-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2827-03	TP-8-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2827-05	TP-9	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2827-06	TP-9-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2827-07	TP-9-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2828-01	POWDER	Solid	Percent Solids	Cool 4 deg C	VIVA01	D21	08/11/2025	Chemtech -SO
Q2829-01	SILICA	Solid	Percent Solids	Cool 4 deg C	VIVA01	D21	08/11/2025	Chemtech -SO
Q2830-01	BIN009-DRIVEWAY-TP-SOUTH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2830-02	BIN009-DRIVEWAY-TP-SOUTH	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2830-03	BIN009-DRIVEWAY-TP-WEST	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2830-04	BIN009-DRIVEWAY-TP-WEST	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2830-05	BIN009-DRIVEWAY-TP-WESTS	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2830-06	BIN009-DRIVEWAY-TP-WESTS	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2830-07	BIN009-DRIVEWAY-TP-EASTS	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO

Date/Time 08.11.25 15:10

Raw Sample Received by: [signature]

Raw Sample Relinquished by: [signature]

Date/Time 08.11.25 17:20

Raw Sample Received by: [signature]

Raw Sample Relinquished by: [signature]

WORKLIST(Hardcopy Internal Chain)

VB136776

WorkList Name : %1-081125 **WorkList ID :** 191189 **Department :** Wet-Chemistry **Date :** 08-11-2025 08:22:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2830-08	BIN009-DRIVEWAY-TP-EASTS	Solid	Percent Solids	Cool 4 deg C	PSEG03	D31	08/11/2025	Chemtech -SO
Q2831-01	VNJ-238	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	08/11/2025	Chemtech -SO
Q2831-02	VNJ-238-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	D21	08/11/2025	Chemtech -SO

Date/Time 08.11.25 15:10
Raw Sample Received by:
Raw Sample Relinquished by:

Date/Time 08.11.25 17:20
Raw Sample Received by:
Raw Sample Relinquished by:

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QCBatch ID # LB136801

Review By	Janvi	Review On	8/15/2025 10:10:36 AM
Supervise By	jaswal	Supervise On	8/19/2025 2:31:42 AM

STD. NAME	STD REF.#
ICAL Standard	MP86640,MP86656,MP86653,MP86643,MP86642,MP86641
ICV Standard	MP86752,MP86656
CCV Standard	MP86646
ICSA Standard	MP86651,MP86753
CRI Standard	MP86656
LCS Standard	
Chk Standard	MP86647,MP86648

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	08/13/25 11:56		Jaswal	OK
2	S1	S1	CAL2	08/13/25 12:00		Jaswal	OK
3	S2	S2	CAL3	08/13/25 12:04		Jaswal	OK
4	S3	S3	CAL4	08/13/25 12:08		Jaswal	OK
5	S4	S4	CAL5	08/13/25 12:12		Jaswal	OK
6	S5	S5	CAL6	08/13/25 12:17		Jaswal	OK
7	ICV01	ICV01	ICV	08/13/25 12:21		Jaswal	OK
8	LLICV01	LLICV01	LLICV	08/13/25 12:28		Jaswal	OK
9	ICB01	ICB01	ICB	08/13/25 12:33		Jaswal	OK
10	CRI01	CRI01	CRDL	08/13/25 12:37		Jaswal	OK
11	ICSA01	ICSA01	ICSA	08/13/25 12:41		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	08/13/25 12:45		Jaswal	OK
13	ICSADL	ICSADL	ICSA	08/13/25 12:49		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	08/13/25 12:54		Jaswal	OK
15	CCV01	CCV01	CCV	08/13/25 12:58		Jaswal	OK
16	CCB01	CCB01	CCB	08/13/25 13:03		Jaswal	OK
17	Q2828-01	POWDER	SAM	08/13/25 13:07		Jaswal	OK
18	Q2828-01DUP	POWDERDUP	DUP	08/13/25 13:12		Jaswal	OK

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QCBatch ID # LB136801

Review By	Janvi	Review On	8/15/2025 10:10:36 AM
Supervise By	jaswal	Supervise On	8/19/2025 2:31:42 AM

STD. NAME	STD REF.#
ICAL Standard	MP86640,MP86656,MP86653,MP86643,MP86642,MP86641
ICV Standard	MP86752,MP86656
CCV Standard	MP86646
ICSA Standard	MP86651,MP86753
CRI Standard	MP86656
LCS Standard	
Chk Standard	MP86647,MP86648

19	Q2828-01L	POWDERL	SD	08/13/25 13:16		Jaswal	OK
20	Q2828-01MS	POWDERMS	MS	08/13/25 13:20		Jaswal	OK
21	Q2828-01MSD	POWDERMSD	MSD	08/13/25 13:25		Jaswal	OK
22	Q2828-01A	POWDERA	PS	08/13/25 13:30	0.1 ml of each M6004 and M6017 in 10ml sample	Jaswal	OK
23	CCV02	CCV02	CCV	08/13/25 13:42		Jaswal	OK
24	CCB02	CCB02	CCB	08/13/25 13:46		Jaswal	OK
25	Q2830-01	BIN0009-DRIVEWAY-	SAM	08/13/25 13:50		Jaswal	OK
26	Q2830-03	BIN0009-DRIVEWAY-	SAM	08/13/25 13:55		Jaswal	OK
27	Q2830-05	BIN0009-DRIVEWAY-	SAM	08/13/25 13:59		Jaswal	OK
28	Q2830-07	BIN0009-DRIVEWAY-	SAM	08/13/25 14:03		Jaswal	OK
29	Q2831-01	VNJ-238	SAM	08/13/25 14:07		Jaswal	OK
30	Q2818-01	B-2-5-1	SAM	08/13/25 14:11		Jaswal	OK
31	Q2818-02	B-3-5-2	SAM	08/13/25 14:16		Jaswal	OK
32	Q2823-01	SU-04-081125	SAM	08/13/25 14:20		Jaswal	OK
33	Q2831-03	VNJ-238	SAM	08/13/25 14:24		Jaswal	OK
34	Q2831-03DUP	VNJ-238DUP	DUP	08/13/25 14:28		Jaswal	OK
35	CCV03	CCV03	CCV	08/13/25 14:33		Jaswal	OK
36	CCB03	CCB03	CCB	08/13/25 14:37		Jaswal	OK
37	Q2831-03L	VNJ-238L	SD	08/13/25 14:41		Jaswal	OK

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QCBatch ID # LB136801

Review By	Janvi	Review On	8/15/2025 10:10:36 AM
Supervise By	jaswal	Supervise On	8/19/2025 2:31:42 AM

STD. NAME	STD REF.#
ICAL Standard	MP86640,MP86656,MP86653,MP86643,MP86642,MP86641
ICV Standard	MP86752,MP86656
CCV Standard	MP86646
ICSA Standard	MP86651,MP86753
CRI Standard	MP86656
LCS Standard	
Chk Standard	MP86647,MP86648

38	Q2831-03MS	VNJ-238MS	MS	08/13/25 14:46		Jaswal	OK
39	Q2831-03MSD	VNJ-238MSD	MSD	08/13/25 14:50		Jaswal	OK
40	Q2831-03A	VNJ-238A	PS	08/13/25 14:54	0.1 ml of each M6004 and M6017 in 10ml sample	Jaswal	OK
41	PB169214BL	PB169214BL	MB	08/13/25 14:58		Jaswal	OK
42	PB169214BS	PB169214BS	LCS	08/13/25 15:02		Jaswal	OK
43	PB169211BL	PB169211BL	MB	08/13/25 15:06		Jaswal	OK
44	PB169215BL	PB169215BL	MB	08/13/25 15:15		Jaswal	OK
45	PB169211BS	PB169211BS	LCS	08/13/25 15:19		Jaswal	OK
46	CCV04	CCV04	CCV	08/13/25 15:23		Jaswal	OK
47	CCB04	CCB04	CCB	08/13/25 15:27		Jaswal	OK
48	PB169200TB	PB169200TB	MB	08/13/25 15:31		Jaswal	OK
49	Q2808-04	TP-7	SAM	08/13/25 15:36		Jaswal	OK
50	Q2826-02	WC1	SAM	08/13/25 15:40		Jaswal	OK
51	Q2827-04	TP-8	SAM	08/13/25 15:45		Jaswal	OK
52	Q2827-08	TP-9	SAM	08/13/25 15:49		Jaswal	OK
53	Q2808-01	TP-7	SAM	08/13/25 15:53		Jaswal	OK
54	Q2817-01	LEAD-PAINT-CHIPS	SAM	08/13/25 15:58		Jaswal	OK
55	Q2819-06	11M-S	SAM	08/13/25 16:02		Jaswal	OK
56	Q2819-11	84SB-W	SAM	08/13/25 16:06		Jaswal	OK

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QCBatch ID # LB136801

Review By	Janvi	Review On	8/15/2025 10:10:36 AM
Supervise By	jaswal	Supervise On	8/19/2025 2:31:42 AM

STD. NAME	STD REF.#
ICAL Standard	MP86640,MP86656,MP86653,MP86643,MP86642,MP86641
ICV Standard	MP86752,MP86656
CCV Standard	MP86646
ICSA Standard	MP86651,MP86753
CRI Standard	MP86656
LCS Standard	
Chk Standard	MP86647,MP86648

57	Q2819-14	17M-W	SAM	08/13/25 16:10		Jaswal	OK
58	CCV05	CCV05	CCV	08/13/25 16:14		Jaswal	OK
59	CCB05	CCB05	CCB	08/13/25 16:19		Jaswal	OK
60	Q2820-09	705R-S	SAM	08/13/25 16:23		Jaswal	OK
61	Q2820-10	SOIL-DUP-2	SAM	08/13/25 16:27		Jaswal	OK
62	Q2820-24	22M-S	SAM	08/13/25 16:31		Jaswal	OK
63	Q2826-01	WC1	SAM	08/13/25 16:35		Jaswal	OK
64	Q2827-01	TP-8	SAM	08/13/25 16:39		Jaswal	OK
65	Q2827-05	TP-9	SAM	08/13/25 16:44		Jaswal	OK
66	Q2829-01	SILICA	SAM	08/13/25 16:48		Jaswal	OK
67	Q2832-09	TG-S05	SAM	08/13/25 16:52		Jaswal	OK
68	Q2832-09DUP	TG-S05DUP	DUP	08/13/25 16:56	Si High	Jaswal	Dilution
69	Q2832-09L	TG-S05L	SD	08/13/25 17:00		Jaswal	OK
70	CCV06	CCV06	CCV	08/13/25 17:05		Jaswal	OK
71	CCB06	CCB06	CCB	08/13/25 17:09		Jaswal	OK
72	Q2832-09MS	TG-S05MS	MS	08/13/25 17:13	Si High	Jaswal	Dilution
73	Q2832-09MSD	TG-S05MSD	MSD	08/13/25 17:17	Si High	Jaswal	Dilution
74	Q2832-09A	TG-S05A	PS	08/13/25 17:21	0.1 ml of each M6004 and M6017 in 10ml sample	Jaswal	OK
75	Q2832-03	TG-S02	SAM	08/13/25 17:25		Jaswal	OK

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QCBatch ID # LB136801

Review By	Janvi	Review On	8/15/2025 10:10:36 AM
Supervise By	jaswal	Supervise On	8/19/2025 2:31:42 AM

STD. NAME	STD REF.#
ICAL Standard	MP86640,MP86656,MP86653,MP86643,MP86642,MP86641
ICV Standard	MP86752,MP86656
CCV Standard	MP86646
ICSA Standard	MP86651,MP86753
CRI Standard	MP86656
LCS Standard	
Chk Standard	MP86647,MP86648

76	Q2832-05	TG-S03	SAM	08/13/25 17:29		Jaswal	OK
77	Q2832-07	TG-S04	SAM	08/13/25 17:33		Jaswal	OK
78	PB169215BS	PB169215BS	LCS	08/13/25 17:42		Jaswal	OK
79	CCV07	CCV07	CCV	08/13/25 17:46		Jaswal	OK
80	CCB07	CCB07	CCB	08/13/25 17:50		Jaswal	OK

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QCBatch ID # LB136973

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard	
ICV Standard	
CCV Standard	
ICSA Standard	
CRI Standard	
LCS Standard	
Chk Standard	

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	08/26/25 10:49		Jaswal	OK
2	S1	S1	CAL2	08/26/25 10:53		Jaswal	OK
3	S2	S2	CAL3	08/26/25 10:58		Jaswal	OK
4	S3	S3	CAL4	08/26/25 11:02		Jaswal	OK
5	S4	S4	CAL5	08/26/25 11:06		Jaswal	OK
6	S5	S5	CAL6	08/26/25 11:10		Jaswal	OK
7	ICV01	ICV01	ICV	08/26/25 11:14		Jaswal	OK
8	LLICV01	LLICV01	LLICV	08/26/25 11:20		Jaswal	OK
9	ICB01	ICB01	ICB	08/26/25 11:24		Jaswal	OK
10	CRI01	CRI01	CRDL	08/26/25 11:29		Jaswal	OK
11	ICSA01	ICSA01	ICSA	08/26/25 11:33		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	08/26/25 11:37		Jaswal	OK
13	ICSADL	ICSADL	ICSA	08/26/25 11:41		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	08/26/25 11:45		Jaswal	OK
15	CCV01	CCV01	CCV	08/26/25 11:50		Jaswal	OK
16	CCB01	CCB01	CCB	08/26/25 11:54		Jaswal	OK
17	LR-1	LR-1	HIGH STD	08/26/25 11:58		Jaswal	OK
18	LR-2	LR-2	HIGH STD	08/26/25 12:03		Jaswal	OK

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QCBatch ID # LB136973

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard	
ICV Standard	
CCV Standard	
ICSA Standard	
CRI Standard	
LCS Standard	
Chk Standard	

19	LR-3	LR-3	HIGH STD	08/26/25 12:07		Jaswal	OK
20	Q2832-09DL	TG-S05DL	SAM	08/26/25 12:12	NOT Required.	Jaswal	Not Ok
21	Q2832-09DUPDL	TG-S05DUPDL	DUP	08/26/25 12:16	10x for Si	Jaswal	Confirms
22	Q2832-09LDL	TG-S05LDL	SD	08/26/25 12:20	NOT Required.	Jaswal	Not Ok
23	Q2832-09MSDL	TG-S05MSDL	MS	08/26/25 12:25	10x for Si	Jaswal	Confirms
24	Q2832-09MSDDL	TG-S05MSDDL	MSD	08/26/25 12:29	10x for Si	Jaswal	Confirms
25	Q2832-09ADL	TG-S05ADL	PS	08/26/25 12:33	NOT Required.	Jaswal	Not Ok
26	PB169380BL	PB169380BL	MB	08/26/25 12:37		Jaswal	OK
27	CCV02	CCV02	CCV	08/26/25 12:41		Jaswal	OK
28	CCB02	CCB02	CCB	08/26/25 12:45		Jaswal	OK
29	PB169380BS	PB169380BS	LCS	08/26/25 12:50		Jaswal	OK
30	PB169379BL	PB169379BL	MB	08/26/25 12:54		Jaswal	OK
31	PB169379BS	PB169379BS	LCS	08/26/25 12:58		Jaswal	OK
32	Q2941-01	SOIL-COMP(1-2)	SAM	08/26/25 13:02		Jaswal	OK
33	Q2947-01	291	SAM	08/26/25 13:06		Jaswal	OK
34	Q2947-03	235	SAM	08/26/25 13:10		Jaswal	OK
35	Q2947-03DUP	235DUP	DUP	08/26/25 13:14		Jaswal	OK
36	Q2947-03L	235L	SD	08/26/25 13:18		Jaswal	OK
37	Q2947-03MS	235MS	MS	08/26/25 13:22		Jaswal	OK
38	Q2947-03MSD	235MSD	MSD	08/26/25 13:26		Jaswal	OK

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: P4

Daily Analysis Runlog For Sequence/QCBatch ID # LB136973

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard	
ICV Standard	
CCV Standard	
ICSA Standard	
CRI Standard	
LCS Standard	
Chk Standard	

39	CCV03	CCV03	CCV	08/26/25 13:30		Jaswal	OK
40	CCB03	CCB03	CCB	08/26/25 13:35		Jaswal	OK
41	Q2947-03A	235A	PS	08/26/25 13:39		Jaswal	OK
42	Q2949-01	367	SAM	08/26/25 13:43		Jaswal	OK
43	Q2940-01	COMP(1-6)	SAM	08/26/25 13:47		Jaswal	OK
44	Q2937-01	RW8-SP100-2025082	SAM	08/26/25 13:51		Jaswal	OK
45	Q2937-02	RW8-SP303-2025082	SAM	08/26/25 13:56		Jaswal	OK
46	Q2937-02DUP	RW8-SP303-2025082	DUP	08/26/25 14:00		Jaswal	OK
47	Q2937-02L	RW8-SP303-2025082	SD	08/26/25 14:04		Jaswal	OK
48	Q2937-02MS	RW8-SP303-2025082	MS	08/26/25 14:08		Jaswal	OK
49	Q2937-02MSD	RW8-SP303-2025082	MSD	08/26/25 14:12		Jaswal	OK
50	Q2937-02A	RW8-SP303-2025082	PS	08/26/25 14:16		Jaswal	OK
51	CCV04	CCV04	CCV	08/26/25 14:20		Jaswal	OK
52	CCB04	CCB04	CCB	08/26/25 14:24		Jaswal	OK

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Prep Standard - Chemical Standard Summary

Order ID : Q2829

Test : Metals Group3

Prepbatch ID : PB169215,

Sequence ID/Qc Batch ID: LB136801,LB136801,LB136973,

Standard ID :
MP86404,MP86640,MP86641,MP86642,MP86643,MP86646,MP86647,MP86648,MP86651,MP86653,MP86655,MP86656,MP86752,MP86753,

Chemical ID :
M5467,M5581,M5658,M5697,M5748,M5798,M5799,M5800,M5801,M5805,M5814,M5816,M5820,M5962,M5969,M5970,M5979,M5985,M5998,M6021,M6023,M6027,M6028,M6030,M6032,M6058,M6076,M6077,M6086,M6128,M6137,M6140,M6142,M6144,M6145,M6146,M6151,M6158,M6159,M6162,M6163,M6164,M6165,M6170,M6171,M6174,M6176,M6177,M6178,M6179,M6180,M6181,M6182,M6184,W3112,

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
169	1:1HNO3	MP86404	07/17/2025	11/16/2025	Sagar Kanani	METALS_SCALE_2 (M SC-2)	METALS_PIPETTE_1 (ICP A)	Sarabjit Jaswal 07/17/2025

FROM 1250.00000ml of M6162 + 1250.00000ml of W3112 = Final Quantity: 2500.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
902	ICP AES CAL BLK (SO/ICB/CCB)	MP86640	08/06/2025	08/27/2025	Janvi Patel	None	METALS_PIPETTE_1 (ICP A)	Sarabjit Jaswal 08/07/2025

FROM 125.00000ml of M6151 + 2350.00000ml of W3112 + 25.00000ml of M6162 = Final Quantity: 500.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
907	ICP AES STD S (S5)	MP86641	08/06/2025	08/27/2025	Janvi Patel	None	METALS_PIPETTE_1 (ICP A)	Sarabjit Jaswal 08/07/2025

FROM 5.00000ml of M5467 + 5.00000ml of M5820 + 5.00000ml of M5969 + 5.00000ml of M5970 + 5.00000ml of M5998 + 5.00000ml of M6077 + 5.00000ml of M6146 + 5.00000ml of M6174 + 5.00000ml of M6176 + 455.00000ml of MP86640 = Final Quantity: 500.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
910	ICP AES STD S4	MP86642	08/06/2025	08/27/2025	Janvi Patel	None	METALS_PIPETTE_1 (ICP A)	Sarabjit Jaswal 08/07/2025

FROM 50.00000ml of MP86640 + 50.00000ml of MP86641 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
909	ICP AES STD S3	MP86643	08/06/2025	08/27/2025	Janvi Patel	None	METALS_PIPETTE_1 (ICP A)	Sarabjit Jaswal 08/07/2025

FROM 25.00000ml of MP86641 + 75.00000ml of MP86640 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
911	ICP AES CCV SOLN	MP86646	08/06/2025	08/27/2025	Janvi Patel	None	METALS_PIPETTE_1 (ICP A)	Sarabjit Jaswal 08/07/2025

FROM 50.00000ml of MP86640 + 50.00000ml of MP86641 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
919	ICP AES INTERNAL STD	MP86647	08/06/2025	08/27/2025	Janvi Patel	None	METALS_PIPETTE_1 (ICP A)	Sarabjit Jaswal 08/07/2025

FROM 1.00000ml of M6179 + 10.00000ml of M5985 + 1969.00000ml of W3112 + 20.00000ml of M6162 = Final Quantity: 2000.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
903	ICP AES RINSE SOLN	MP86648	08/06/2025	08/27/2025	Janvi Patel	None	METALS_PIPETTE_1 (ICP A)	Sarabjit Jaswal 08/07/2025

FROM 200.00000ml of M6162 + 9800.00000ml of W3112 = Final Quantity: 10000.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
904	ICP AES ICSA SOLN	MP86651	08/06/2025	08/27/2025	Janvi Patel	None	METALS_PIPETTE_1 (ICP A)	Sarabjit Jaswal 08/07/2025

FROM 5.00000ml of M6182 + 45.00000ml of MP86640 = Final Quantity: 50.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3913	ICP AES STD S2	MP86653	08/06/2025	08/27/2025	Janvi Patel	None	METALS_PIPETTE_1 (ICP A)	Sarabjit Jaswal 08/07/2025

FROM 4.00000ml of MP86641 + 46.00000ml of MP86640 = Final Quantity: 50.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
2950	ICP AES S1/CRI STOCK STD	MP86655	08/06/2025	08/27/2025	Janvi Patel	None	METALS_PIPETTE_1 (ICP A)	Sarabjit Jaswal 08/07/2025

FROM 0.00000ml of M5748 + 0.00000ml of M5805 + 0.03000ml of M5798 + 0.03000ml of M6028 + 0.05000ml of M5658 + 0.05000ml of M6030 + 0.05000ml of M6140 + 0.05000ml of M6159 + 0.06000ml of M6177 + 0.10000ml of M5697 + 0.10000ml of M5801 + 0.10000ml of M5816 + 0.10000ml of M5820 + 0.10000ml of M5962 + 0.10000ml of M5970 + 0.10000ml of M6128 + 0.10000ml of M6176 + 0.15000ml of M5800 + 0.20000ml of M5799 + 0.20000ml of M5979 + 0.20000ml of M6021 + 0.20000ml of M6023 + 0.20000ml of M6027 + 0.20000ml of M6145 + 0.25000ml of M5467 + 0.25000ml of M6146 + 0.50000ml of M5814 + 0.50000ml of M6032 + 1.00000ml of M6086 + 1.00000ml of M6142 + 1.00000ml of M6144 + 1.00000ml of M6171 + 1.00000ml of M6178 + 2.00000ml of M6137 + 89.03000ml of MP86640 = Final Quantity: 100.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
2951	ICP AES S1/CRI WORK STD	MP86656	08/06/2025	08/27/2025	Janvi Patel	None	METALS_PIPETTE_1 (ICP A)	Sarabjit Jaswal 08/07/2025

FROM 1.00000ml of MP86655 + 49.00000ml of MP86640 = Final Quantity: 50.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
4163	NEW ICV-060925	MP86752	08/06/2025	08/27/2025	Janvi Patel	None	METALS_PIPETTE_1 (ICP A)	Sarabjit Jaswal 08/14/2025

FROM 0.20000ml of M6058 + 0.40000ml of M6163 + 0.40000ml of M6164 + 0.40000ml of M6165 + 48.60000ml of MP86640 = Final Quantity: 50.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3494	ICP AES ICSAB SOLN-1	MP86753	08/06/2025	08/27/2025	Janvi Patel	None	METALS_PIPETTE_1 (ICP A)	Sarabjit Jaswal 08/14/2025

FROM 0.05000ml of M6076 + 5.00000ml of M6182 + 5.00000ml of M6184 + 39.95000ml of MP86640 = Final Quantity: 100.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57058 / Cerium, 1000PPM, 100ML	020623	02/06/2026	03/06/2023 / bin	03/01/2023 / bin	M5467

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	26397-103 / PTFE BOILING STONES	W126678	03/20/2026	03/20/2025 /	06/12/2023 / jaswal	M5581

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	58024 / Chromium, Cr, 500 ml, 1000 PPM	060523	06/05/2026	08/28/2023 / jaswal	08/25/2023 / jaswal	M5658

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	58029 / Cu, 1000 PPM, 500 ml	102523	10/25/2026	04/03/2024 / jaswal	10/27/2023 / jaswal	M5697

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Nickel (Ni) 1000PPM	091223	09/12/2026	01/02/2024 / bin	12/20/2023 / jaswal	M5748

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57004 / Be, 1000 PPM, 125 ml	102523	10/25/2026	02/09/2024 / bin	02/09/2024 / bin	M5798

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57050 / Sn, 1000 PPM, 125 ml	071123	07/11/2026	02/09/2024 / bin	02/09/2024 / bin	M5799

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57027 / CO, 1000 PPM, 125 ml	091923	09/19/2026	05/31/2024 / bin	02/09/2024 / bin	M5800

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57033 / As, 1000 PPM, 125 ml	111323	11/13/2026	02/09/2024 / bin	02/09/2024 / bin	M5801

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	58120 / Ca, 10000 PPM, 500 ml	122223	12/22/2026	04/17/2025 / Janvi	01/03/2024 / jaswal	M5805

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57005 / B, 1000 PPM, 125 ml	071123	07/11/2026	03/26/2024 / Sohil	01/03/2024 / jaswal	M5814

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57016 / S, 1000 PPM, 125 ml	122923	12/29/2026	05/20/2024 / Jaswal	02/09/2024 / jaswal	M5816

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57015 / P, 1000 PPM, 125 ml	091123	09/11/2026	05/01/2024 / jaswal	02/09/2024 / jaswal	M5820

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57034 / Se, 1000 PPM, 125 ml	060624	06/06/2027	07/02/2024 / Jaswal	06/14/2024 / Jaswal	M5962

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57016 / S, 1000 PPM, 125 ml	122923	03/31/2026	03/31/2025 / kareem	06/11/2024 / Jaswal	M5969

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57003 / Li, 1000 PPM, 125 ml	061224	06/21/2027	07/01/2024 / Jaswal	07/01/2024 / Jaswal	M5970

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	CGTI1-1 / TITANIUM 125mL 1000ug/mL	T2-TI719972	08/05/2026	08/06/2025 / Janvi	02/22/2024 / Jaswal	M5979

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	CGIN10-5 / INDIUM 1 x 500 ml	U2-IN729349	02/21/2028	10/08/2024 / Jaswal	06/14/2024 / Jaswal	M5985

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	CLPP-CAL-3 / CLP CAL SOLUTION #3, 125mL	T2-MEB727800	07/30/2026	07/30/2025 / Janvi	02/22/2024 / kareem	M5998

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57023 / V, 1000 PPM, 125 ml	062424	06/24/2027	09/28/2024 / jaswal	08/05/2024 / Jaswal	M6021

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57081 / Tl, 1000 PPM, 125 ml	0624724	06/27/2027	08/05/2024 / kareem	08/05/2024 / Jaswal	M6023

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57028 / Ni, 1000 PPM, 125 ml	062024	06/20/2027	06/04/2025 / Janvi	08/05/2024 / Jaswal	M6027

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57048 / Cd, 1000 PPM, 125 ml	070124	07/01/2027	08/05/2024 / kareem	08/05/2024 / Jaswal	M6028

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57047 / Ag, 1000 PPM, 125 ml	122823	12/28/2026	08/05/2024 / kareem	08/05/2024 / Jaswal	M6030

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57056 / Ba, 1000 PPM, 125 ml	010924	01/09/2027	01/14/2025 / Jaswal	08/05/2024 / Jaswal	M6032

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	CHEM-QC-4 / CHEM-QC-4, Second Source, 1000 ug/ml, B, Mo, Si, Sn, Ti	V2-MEB746173	01/29/2026	01/29/2025 / JANVI	08/22/2024 / Jaswal	M6058

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	Z9651Q / CHEM-CLP-4/.25L	V2-MEB746762	01/01/2026	01/01/2025 / kareem	09/19/2024 / kareem	M6076

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	Z9651Q / CHEM-CLP-4/.25L	V2-MEB746762	09/06/2029	01/23/2025 / kareem	09/19/2024 / kareem	M6077

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	58120 / Calcium, 500 ml, 10000 PPM	082324	08/23/2027	03/06/2025 / JANVI	10/14/2024 / jaswal	M6086

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	58025 / Mn, 1000 PPM, 500 ml	101124	10/11/2027	01/13/2025 / kareem	01/13/2025 / kareem	M6128

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	CGSI1-1 / SILICON 125mL 1000ug/mL	V2-SI744713	07/10/2029	01/14/2025 / Jaswal	10/03/2024 / Jaswal	M6137

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	58126 / Fe, 10000 PPM, 500 ml	011025	01/10/2028	06/25/2025 / Janvi	01/13/2025 / Jaswal	M6140

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	58119 / K, 10000 PPM, 500 ml	103024	10/30/2027	05/06/2025 / JANVI	01/13/2025 / Jaswal	M6142

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	58111 / Na, 10000 PPM, 500 ml	072424	07/24/2027	01/23/2025 / kareem	01/13/2025 / Jaswal	M6144

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	58030 / Zinc, Zn, 500 ml, 1000 PPM	121724	12/17/2027	02/04/2025 /	01/13/2025 / Jaswal	M6145

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57051 / Sb, 1000 PPM, 125 ml	071724	07/17/2027	01/31/2025 / kareem	10/18/2024 / kareem	M6146

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9530-33 / Hydrochloric Acid, Instra-Analyzed (cs/6x2.5L)	22G2862015	02/17/2026	02/18/2025 / Sagar	01/15/2025 / Sagar	M6151

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9598-34 / Nitric Acid, Instra-Analyzed (cs/4x2.5L)	24D1062002	09/09/2025	03/10/2025 / Eman	02/02/2025 / Sagar	M6158

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	58113 / Al, 10000 PPM, 500 ml	011325	03/18/2026	03/18/2025 / kareem	02/09/2025 / kareem	M6159

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9598-34 / Nitric Acid, Instra-Analyzed (cs/4x2.5L)	24H0162012	11/27/2025	05/27/2025 / Sagar	04/27/2025 / Sagar	M6162

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	QCP-CICV-1-125ML / EPA CLP ICP Verification Standard1	V2-MEB744107	06/08/2026	06/09/2025 / jaswal	06/06/2025 / jaswal	M6163

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	QCP-CICV-2-125ML / EPA CLP ICP Verification Standard2	U2-MEB733713	06/08/2026	06/09/2025 / jaswal	06/09/2025 / jaswal	M6164

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	QCP-CICV-3-125ML / EPA CLP ICP Verification Standard3	V2-MEB749572	06/08/2026	06/09/2025 / jaswal	06/09/2025 / jaswal	M6165

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-2186-03 / Hydrogen Peroxide (cs/4x4L)	24D1961001	10/16/2025	06/25/2025 /	05/20/2025 / Sagar	M6170

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	58112 / Mg, 10000 PPM, 500 ml	011525	01/15/2028	07/15/2025 /	02/13/2025 / Janvi	M6171

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	CLPP-CAL-1 / CLP CAL SOLUTION #1, 125mL	V2-MEB742428	07/31/2026	07/31/2025 / Janvi	05/05/2025 / Janvi	M6174

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57038 / Sr, 1000 PPM, 125 ml	092724	09/27/2027	08/06/2025 / Janvi	08/06/2025 / Janvi	M6176

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Lead (Pb) 1000PPM	031525	03/15/2028	08/06/2025 / Janvi	08/06/2025 / Janvi	M6177

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	57042 / Mo, 1000 PPM, 125 ml	080528	08/05/2028	08/06/2025 / Janvi	08/06/2025 / Janvi	M6178

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	CGY10-1 / YTTRIUM 125mL 10,000ug/mL	V2-Y740548	08/05/2026	08/06/2025 / Janvi	06/02/2024 / Janvi	M6179

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	WW-LFS-1 / Laboratory Fortified Stock Solution 1, 125 ml	W2-MEB752149	02/05/2026	08/06/2025 / Janvi	07/22/2025 / Janvi	M6180

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Inorganic Ventures	WW-LFS-2 / Laboratory Fortified Stock Solution 2, 125 ml	V2-MEB7433480	02/05/2026	08/06/2025 / Janvi	07/22/2025 / Janvi	M6181

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
EPA	PART A / ICSA (ICP) STOCK SOLN	ICSA-1211	02/05/2026	08/06/2025 / Janvi	03/06/2024 / Janvi	M6182

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
EPA	PART B / ICSAB (ICP) STOCK SOLN	ICSB-0710	02/05/2026	08/06/2025 / Janvi	03/06/2024 / Janvi	M6184

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

M6032

R: 8/5/24

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com


CERTIFIED WEIGHT REPORT:

Part Number:	**57056**		**Lot #**		
Lot Number:	**010924**	**Solvent:**	24002546	Nitric Acid	
Description:	**Barium (Ba)**				
		2%	40.0 (mL)	Nitric Acid	
Expiration Date:	010927				
Recommended Storage:	Ambient (20 °C)				
Nominal Concentration (µg/mL):	**1000**				
NIST Test Number:	6UTB	5E-05	Balance Uncertainty		
Weight shown below was diluted to (mL):	2000.02	0.058	Flask Uncertainty		

Giovanni Esposito

Formulated By:	Giovanni Esposito	010924
Reviewed By:	Pedro L. Rentas	010924

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Barium nitrate (Ba)	IN023	BAD022019A1	1000	99.999	0.10	52.3	3.82417	3.82441	**1000.1**	**2.0**	10022-31-8	0.5 mg/m3	orl-rat 355 mg/kg	3104a



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)

Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	T	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

Certified by:



* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

R: 01/03/24 M5804, M5805

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

Part Number: 58120
Lot Number: 122223
Description: Calcium (Ca)

Expiration Date: 122226
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 10000
NIST Test Number: 6UTB
Weight shown below was diluted to (mL): 4000.0 0.06 Flask Uncertainty


		Lot #	
Solvent:	24002546	Nitric Acid	
2%	80.0 (mL)	Nitric Acid	

5E-05 Balance Uncertainty

Giovanni Esposito

Formulated By:	Giovanni Esposito	122223
Reviewed By:	Pedro L. Rentas	122223

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Calcium carbonate (Ca)	IN014	CAD03202383	10000	99.999	0.10	39.9	100.2525	100.2670	**10001.4**	**20.0**	471-34-1	5 mg/m3	orl-rat >2000mg/kg	3109a



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)

Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	T	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	30	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

Revd 10/14/2024 M6085/M6086/M6087

CERTIFIED WEIGHT REPORT:


Part Number:	**58120**		**Solvent:**	24002546	Nitric Acid
Lot Number:	**082324**				
Description:	**Calcium (Ca)**		2%	80.0	Nitric Acid
Expiration Date:	082327			(mL)	
Recommended Storage:	Ambient (20 °C)				
Nominal Concentration (µg/mL):	**10000**				
NIST Test Number:	6UTB	5E-05	Balance Uncertainty		
Weight shown below was diluted to (mL):	4000.1	0.15	Flask Uncertainty		

Lot #

Giovanni Esposito

Formulated By:	Giovanni Esposito	082324
Reviewed By:	Pedro L. Rentas	082324

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Calcium carbonate (Ca)	IN014	CAD03203B3	10000	99.999	0.10	39.9	100.2537	100.2656	**10001.2**	**20.0**	471-34-1	5 mg/m3	ori-rat >2000mg/kg	3109a



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	T	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	30	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

R 8/5/24

M6028

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com


CERTIFIED WEIGHT REPORT:

Part Number:	**57048**		**Lot #**		
Lot Number:	**070124**	**Solvent:**	24002546	Nitric Acid	
Description:	**Cadmium (Cd)**	2%	40.0 (mL)	Nitric Acid	
Expiration Date:	070127				
Recommended Storage:	Ambient (20 °C)				
Nominal Concentration (µg/mL):	**1000**				
NIST Test Number:	6UTB	5E-05	Balance Uncertainty		
Weight shown below was diluted to (mL):	2000.07	0.100	Flask Uncertainty		

Aleah O'Brady

Formulated By:	Aleah O'Brady	070124
Reviewed By:	Pedro L. Rentas	070124

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Cadmium nitrate tetrahydrate (Cd)	IN024	CDM092021A1	1000	99.999	0.10	36.5	5.4797	5.4804	**1000.1**	**2.0**	10022-68-1	0.01 mg/m3	orl-rat 60.2mg/kg	3108



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	T	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

M5467 R: 03/01/23 (LA)

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

Part Number: 57058
Lot Number: 020623
Description: Cerium (Ce)

Expiration Date: 020626
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 1000
NIST Test Number: 6UTB


Weight shown below was diluted to (mL): 1000.12 0.058 Flask Uncertainty

		Lot #	
Solvent:	21110221	Nitric Acid	
2%	20.0 (mL)	Nitric Acid	

5E-05 Balance Uncertainty

Formulated By:	Lawrence Barry	020623
Reviewed By:	Pedro L. Rentas	020623

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Cerium nitrate hexahydrate (Ce)	IN146	Z512CEB1	1000	99.999	0.10	32.8	3.04919	3.04921	**1000.0**	**2.0**	10294-41-4	NA	NA	NA



Instrumental Analysis by Inductively Coupled Plasma Mass Spectroscopy (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	T	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T)= Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

RD: 06/12/2023, M5581, M5582, M5583, M5584, M5585

JENSEN INERT PRODUCTS

MATERIAL CERTIFICATE OF COMPLIANCE

DATE: JUNE 12, 2023

CUSTOMER:	PCI SCIENTIFIC SUPPLY, INC
PURCHASE ORDER NO.	6054931
CATALOG NO.	BOI5021-450L
PRODUCT DESCRIPTION:	BOILING STONES, TFE, 454GMS
QUANTITY:	10 EACH
LOT NO.	W126678
SPECIFICATION (S):	Made from Virgin PTFE Resin

We certify that we have complied with the terms and conditions of the above Purchase Order and the Part Specifications in the manufacturing of the above product.

Laura Valencia

Laura Valencia
Quality Assurance Inspector

F:U:J:CF:PCISCI:COC-68118-BOI5021-061223

3773 NW 126th Avenue • Coral Springs, Florida 33065 • Fax (954) 344-2008 • 1-800 446-3781
Email: jensen@jenseninert.com Visit our Web Site at www.jenseninert.com

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

M5658 R ! 8/25/23


ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

		Lot #	Solvent:	
Part Number:	**58024**	21110221	Nitric Acid	
Lot Number:	**060523**			
Description:	**Chromium (Cr)**			
		2.0%	40.0 (mL)	Nitric Acid
Expiration Date:	060526			
Recommended Storage:	Ambient (20 °C)			
Nominal Concentration (µg/mL):	**1000**			
NIST Test Number:	6UTB	5E-05	Balance Uncertainty	
Volume shown below was diluted to (mL):	2000.02	0.058	Flask Uncertainty	

Formulated By:	Lawrence Barry	060523
Reviewed By:	Pedro L. Rentas	060523

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Nominal Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Chromium(III) nitrate nonahydrate (Cr)	58124	071122	0.1000	200.0	0.084	1000	10000.1	**1000.0**	**2.2**	7789-02-8	0.5 mg(Cr)/m3	orl-rat 3250 mg/kg	3112a



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	T	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T)= Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

M5697 R: 10/27/23


ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

		Lot #	Solvent:	
Part Number:	**58029**	24002546	Nitric Acid	
Lot Number:	**102523**			
Description:	**Copper (Cu)**			
		2.0%	40.0 (mL)	Nitric Acid
Expiration Date:	102526			
Recommended Storage:	Ambient (20 °C)			
Nominal Concentration (µg/mL):	**1000**			
NIST Test Number:	6UTB	5E-05	Balance Uncertainty	
Volume shown below was diluted to (mL):	2000.02	0.058	Flask Uncertainty	

Formulated By:	Benson Chan	102523
Reviewed By:	Pedro L. Rentas	102523

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Nominal Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Copper(II) nitrate trihydrate (Cu)	58129	100223	0.1000	200.0	0.084	1000	10000.1	**1000.0**	**2.2**	10031-43-3	1 mg/m3	orl-rat 794 mg/kg	3114



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)

Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	T	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

Certified by:

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

R: 12/20/23 M5748


ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

		Lot #	Solvent:	
Part Number:	**57028**	24002546	Nitric Acid	
Lot Number:	**091223**			
Description:	**Nickel (Ni)**			
		2.0%	40.0 (mL)	Nitric Acid
Expiration Date:	091226			
Recommended Storage:	Ambient (20 °C)			
Nominal Concentration (µg/mL):	**1000**			
NIST Test Number:	6UTB	5E-05	Balance Uncertainty	
Volume shown below was diluted to (mL):	2000.02	0.058	Flask Uncertainty	

Formulated By:	Lawrence Barry	091223
Reviewed By:	Pedro L. Rentas	091223

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Nominal Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Nickel(II) nitrate hexahydrate (Ni)	58128	062023	0.1000	200.0	0.084	1000	10000.4	**1000.0**	**2.2**	13478-00-7	1 mg/m3	orl-rat 1620 mg/kg	3136



Instrumental Analysis by Inductively Coupled Plasma Mass Spectroscopy (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)

Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	T	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

R: 02/09/24 M5798


ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

		Lot #	Solvent:	
Part Number:	**57004**	24002546	Nitric Acid	
Lot Number:	**102523**			
Description:	**Beryllium (Be)**	2.0%	40.0 (mL)	Nitric Acid
Expiration Date:	102526			
Recommended Storage:	Ambient (20 °C)			
Nominal Concentration (µg/mL):	**1000**			
NIST Test Number:	6UTB	5E-05	Balance Uncertainty	
Volume shown below was diluted to (mL):	2000.02	0.058	Flask Uncertainty	

Formulated By:	Benson Chan	102523
Reviewed By:	Pedro L. Rentas	102523

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Nominal Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Beryllium nitrate (Be)	58104	091423	0.1000	200.0	0.084	1000	10001.5	**1000.0**	**2.2**	13597-99-4	0.2ug/m3	intrvns-rat 3.16mg/kg	NA



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	T	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

R? 02/09/24 IV15799

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

Part Number: 57050
Lot Number: 071123
Description: Tin (Sn)


Expiration Date: 071126
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 1000
NIST Test Number: 6UTB
Weight shown below was diluted to (mL): 499.93

	Lot #	
Solvents:	21110221	Nitric Acid
	22D0562008	Hydrochloric acid
2%	10.0	Nitric Acid
6%	30.0	Hydrochloric acid
	(mL)	

5E-05 Balance Uncertainty
0.058 Flask Uncertainty

Formulated By:	Benson Chan	071123
Reviewed By:	Pedro L. Rentas	071123

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Ammonium hexafluorostannate(IV) (Sn)	IN010	SND042023A1	1000	99.999	0.10	44.2	1.13107	1.13286	**1001.6**	**2.0**	16919-24-7	7 mg/m3	NA	3161a



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<500	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	T	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

Certified by:

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

R: 02/09/24 IM5800 BH


ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

		Lot #	Solvent:		
Part Number:	**57027**	24002546	Nitric Acid		
Lot Number:	**091923**				
Description:	**Cobalt (Co)**	2.0%	40.0 (mL)	Nitric Acid	
Expiration Date:	091926				
Recommended Storage:	Ambient (20 °C)				
Nominal Concentration (µg/mL):	**1000**				
NIST Test Number:	6UTB	5E-05	Balance Uncertainty		
Volume shown below was diluted to (mL):	2000.02	0.058	Flask Uncertainty		

Formulated By:	Lawrence Barry	091923
Reviewed By:	Pedro L. Rentas	091923

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Nominal Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Cobalt(II) nitrate hexahydrate (Co)	58127	050923	0.1000	200.0	0.084	1000	10000.0	**1000.0**	**2.2**	10026-22-9	0.02 mg/m3	orl-rat 691 mg/kg	3113



Instrumental Analysis by Inductively Coupled Plasma Mass Spectroscopy (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	T	Gé	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T)= Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

R: 02/09/24 IM5801 BH


ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED **WEIGHT** REPORT:

		Lot #	Solvent:		
Part Number:	**57033**	24002546	Nitric Acid		
Lot Number:	**111323**				
Description:	**Arsenic (As)**	2.0%	80.0 (mL)	Nitric Acid	
Expiration Date:	111326				
Recommended Storage:	Ambient (20 °C)				
Nominal Concentration (µg/mL):	**1000**				
NIST Test Number:	6UTB	5E-05	Balance Uncertainty		
Volume shown below was diluted to (mL):	4000.0	0.06	Flask Uncertainty		

Formulated By:	Lawrence Barry	111323
Reviewed By:	Pedro L. Rentas	111323

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Nominal Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Arsenic (As)	58133	020522	0.1000	400.0	0.084	1000	10001.0	**1000.0**	**2.0**	7440-38-2	0.5 mg/m3	orl-rat 500 mg/kg	3103a



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)

Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	T	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge*	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

A: 02/09/24 M.5814


ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

Part Number:	**57005**		**Solvent:**	MKBQ8597V	Ammonium hydroxide
Lot Number:	**071123**				
Description:	**Boron (B)**		2.0%	40.0 (mL)	Ammonium hydroxide
Expiration Date:	071126				
Recommended Storage:	Ambient (20 °C)				
Nominal Concentration (µg/mL):	**1000**				
NIST Test Number:	6UTB	5E-05	Balance Uncertainty		
Weight shown below was diluted to (mL):	1999.48	0.058	Flask Uncertainty		

Formulated By:	Benson Chan	071123
Reviewed By:	Pedro L. Rentas	071123

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Boric acid (B)	IN018	BV092016A1	1000	99.9999	0.10	17.3	11.55772	11.56201	**1000.4**	**2.0**	10043-35-3	2 mg/m3	orl-rat 2660 mg/kg	3107



Instrumental Analysis by Inductively Coupled Plasma Mass Spectroscopy (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)

Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	T	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

Certified by:

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

R: 02/09/24 M5816


ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

Part Number:	**57016**	**Solvent:**	**Lot #** 122923	ASTM Type 1 Water
Lot Number:	**122923**			
Description:	**Sulfur (S)**			
Expiration Date:	122926			
Recommended Storage:	Ambient (20 °C)			
Nominal Concentration (µg/mL):	**1000**			
NIST Test Number:	6UTB	5E-05	Balance Uncertainty	
Weight shown below was diluted to (mL):	4000.0	0.06	Flask Uncertainty	

Formulated By:	Benson Chan	122923
Reviewed By:	Pedro L. Rentas	122923

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Ammonium sulfate (S)	IN117	SLBR7225V	1000	99.9	0.10	24.3	16.4979	16.4980	**1000.0**	**2.0**	7783-20-2	NA	orl-rat 4250mg/kg	3181



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	T	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

Certified by:



* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

R: 02/09/24 M5820


ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

Part Number:	**57015**		**Lot #**		
Lot Number:	**091123**	**Solvent:**	24002546	Nitric Acid	
Description:	**Phosphorous (P)**				
		2%	40.0 (mL)	Nitric Acid	
Expiration Date:	091126				
Recommended Storage:	Ambient (20 °C)				
Nominal Concentration (µg/mL):	**1000**				
NIST Test Number:	6UTB	5E-05	Balance Uncertainty		
Weight shown below was diluted to (mL):	2000.02	0.058	Flask Uncertainty		

Formulated By:	Lawrence Barry	091123
Reviewed By:	Pedro L. Rentas	091123

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Ammonium dihydrogen phosphate (P)	IN008	PV082019A1	1000	99.999	0.10	27.5	7.2729	7.2730	**1000.0**	**2.0**	7722-76-1	5 mg/m3	orl-rat >2000mg/k	3186



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	T	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

M5962 R! 06/14/24.


ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

		Lot #	Solvent:	
Part Number:	**57034**	24002546	Nitric Acid	
Lot Number:	**060624**			
Description:	**Selenium (Se)**			
		2.0%	40.0 (mL)	Nitric Acid
Expiration Date:	060627			
Recommended Storage:	Ambient (20 °C)			
Nominal Concentration (µg/mL):	**1000**			
NIST Test Number:	6UTB	5E-05	Balance Uncertainty	
Volume shown below was diluted to (mL):	2000.07	0.100	Flask Uncertainty	

Formulated By:	Benson Chan	060624
Reviewed By:	Pedro L. Rentas	060624

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Nominal Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Selenium (Se)	58134	071223	0.1000	200.0	0.084	1000	10002.5	**1000.0**	**2.2**	7782-49-2	0.2 mg/m3	orl-rat 6700 mg/kg	3149



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																		
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	T	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

Certified by:



* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

M5969 R: 06/11/24

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com


CERTIFIED WEIGHT REPORT:

Part Number:	**57016**	**Solvent:**	122923	ASTM Type 1 Water
Lot Number:	**122923**			
Description:	**Sulfur (S)**			
Expiration Date:	122926			
Recommended Storage:	Ambient (20 °C)			
Nominal Concentration (µg/mL):	**1000**			
NIST Test Number:	6UTB	5E-05	Balance Uncertainty	
Weight shown below was diluted to (mL):	4000.0	0.06	Flask Uncertainty	

Lot #

Formulated By:	Benson Chan	122923
Reviewed By:	Pedro L. Rentas	122923

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Ammonium sulfate (S)	IN117	SLBR7225V	1000	99.9	0.10	24.3	16.4979	16.4980	**1000.0**	**2.0**	7783-20-2	NA	orl-rat 4250mg/kg	3181



Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

M5970, M5971 R: 7/01/24

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com


CERTIFIED WEIGHT REPORT:

		Lot #	Solvent:
Part Number:	**57003**	24002546	Nitric Acid
Lot Number:	**062124**		
Description:	**Lithium (Li)**		
		2.0%	5.0 (mL) Nitric Acid
Expiration Date:	062127		
Recommended Storage:	Ambient (20 °C)		
Nominal Concentration (µg/mL):	**1000**		
NIST Test Number:	6UTB	5E-05 Balance Uncertainty	
Volume shown below was diluted to (mL):	250.11	0.016 Flask Uncertainty	

Giovanni Esposito
Formulated By: Giovanni Esposito 062124

Pedro Rentas
Reviewed By: Pedro L. Rentas 062124

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Nominal Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Lithium nitrate (Li)	58103	070622	0.1000	25.0	0.004	1000	10000.4	**1000.0**	**2.0**	7790-69-4	5 mg/m3	orl-rat 1426 mg/kg	NA



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)

Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	T	Ni	<0.02	Pr	<0.02	Sc	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

Certified by:

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

M5978, M5979

INORGANIC VENTURES
Refine your results. Redefine your industry.

Certificate of Analysis

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

R: 2/22/24

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).

ACCREDITED Testing Laboratory Certificate 883.01 ACCREDITED Reference Material Producer Certificate 883.02

2.0 PRODUCT DESCRIPTION

Product Code:	Single Analyte Custom Grade Solution
Catalog Number:	CGTI1
Lot Number:	T2-TI719972
Matrix:	2% (v/v) HNO_3 tr. HF
Value / Analyte(s):	1 000 µg/mL ea: Titanium
Starting Material:	Ti Metal
Starting Material Lot#:	2094
Starting Material Purity:	99.9975%

3.0 CERTIFIED VALUES AND UNCERTAINTIES

Certified Value: 1002 ± 5 µg/mL

Density: 1.012 g/mL (measured at 20 ± 4 °C)

Assay Information:

Assay Method #1 **1002 ± 4 µg/mL**
ICP Assay NIST SRM 3162a Lot Number: 130925

- The Calculated Value is a value calculated from the weight of a starting material that has been certified directly vs. a National Institute of Standards and Technology (NIST) SRM/RM. See Sec 4.2 for balance traceability.

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$X_{CRM/RM} = \Sigma(w_i)(X_i)$

X_i = mean of Assay Method i with standard uncertainty $u_{char\ i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance

$w_i = (1/u_{char\ i}^2) / (\Sigma(1/u_{char\ i}^2))$

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\ (u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char} = [\Sigma(w_i)^2\ (u_{char\ i})^2]^{½}$ where $u_{char\ i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$X_{CRM/RM} = (X_a)\ (u_{char\ a})$

X_a = mean of Assay Method A with

$u_{char\ a}$ = the standard uncertainty of characterization Method A

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\ (u^2_{char\ a} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char\ a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

CRM/RMs are tested for trace metallic impurities by Axial ICP-OES and ICP-MS. The result from the most sensitive method for each element, is reported below. Solutions tested by ICP-MS were analyzed in an ULPA-Filtered Clean Room. An ULPA-Filter is 99.9985% efficient for the removal of particles down to 0.3 µm.

M	Ag	<	0.000536	M	Eu	<	0.000268	O	Na	<	0.032670	M	Se		0.001204	O	Zn	<	0.003267
O	Al		0.000872	O	Fe		0.003225	O	Nb	<	0.043560	O	Si		0.004735	O	Zr	<	0.043560
M	As	<	0.008586	M	Ga	<	0.000268	M	Nd	<	0.000268	M	Sm	<	0.000268				
M	Au	<	0.004577	M	Gd	<	0.000268	O	Ni	<	0.010890	M	Sn		0.000096				
O	B	<	0.008929	M	Ge	<	0.002146	M	Os	<	0.000269	O	Sr		0.000096				
M	Ba	<	0.002683	M	Hf		0.002161	O	P	<	0.054450	M	Ta		0.010560				
M	Be	<	0.005366	M	Hg	<	0.003231	M	Pb	<	0.001073	M	Tb	<	0.000268				
M	Bi	<	0.001609	M	Ho	<	0.000268	M	Pd	<	0.000268	M	Te	<	0.001341				
O	Ca		0.000676	M	In	<	0.002683	M	Pr	<	0.000268	M	Th	<	0.053663				
M	Cd	<	0.000268	M	Ir	<	0.000269	M	Pt	<	0.000536	s	Ti	<					
M	Ce	<	0.000268	M	K		0.001172	M	Rb	<	0.000268	M	Tl	<	0.000268				
M	Co	<	0.004293	M	La	<	0.000268	M	Re	<	0.000268	M	Tm	<	0.000268				
M	Cr		0.000752	O	Li	<	0.027225	M	Rh	<	0.000268	M	U	<	0.000268				
M	Cs	<	0.000268	M	Lu	<	0.000268	M	Ru	<	0.000269	M	V	<	0.019855				
O	Cu	<	0.010890	O	Mg	<	0.005445	i	S	<		M	W		0.000473				
M	Dy	<	0.000268	O	Mn	<	0.003267	M	Sb	<	0.006976	M	Y	<	0.002146				
M	Er	<	0.000268	M	Mo		0.000774	O	Sc	<	0.004900	M	Yb	<	0.000536				

M - Checked by ICP-MS O - Checked by ICP-OES i - Spectral Interference
n - Not Checked For s - Solution Standard Element

6.0 INTENDED USE

- For the calibration of analytical instruments and validation of analytical methods as appropriate.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit **www.inorganicventures.com/TCT**

Atomic Weight; Valence; Coordination Number; Chemical Form in Solution - 47.87 +4 6 Ti(F)6-2

Chemical Compatibility - Soluble in concentrated HCl, HF, H3PO4 H2SO4 and HNO3. Avoid neutral to basic media. Unstable at ppm levels with metals that would pull F- away (i.e. Do not mix with Alkaline or Rare Earths or high levels of transition elements unless they are fluorinated). Stable with most inorganic anions with a tendency to hydrolyze forming the hydrated oxide in all dilute acids except HF.

Stability - 2-100 ppb levels stable (Alone or mixed with all other metals) as the Ti(F)6-2 for months in 1% HNO3 / LDPE container. 1-10,000 ppm single element solutions as the Ti(F)6-2 chemically stable for years in 2-5% HNO3 / trace HF in an LDPE container.

Ti Containing Samples (Preparation and Solution) - Metal (Soluble in H2O / HF caution -powder reacts violently); Oxide - low temperature history anatase or rutile (Dissolved by heating in 1:1:1 H2O / HF / H2SO4); Oxide - high temperature history (~800EC) brookite (fuse in Pt0 with K2S2O7); Ores (fuse in Pt0 with KF + K2S2O7 - no KF if silica not present); Organic Matrices (Dry ash at 450EC in Pt0 and dissolve by heating with 1:1:1 H2O / HF / H2SO4 or fuse ash with pyrosulfate if oxide is as plastic pigment and likely in brookite crystalline form).

Atomic Spectroscopic Information (ICP-OES D.L.s are given as radial/axial view):

Technique/Line	Estimated D.L.	Order	Interferences (underlined indicates severe)
ICP-MS 48 amu	14 ppt	N/A	32S16O, 32S14N, 14N16O18O, 14N17N2, 36Ar12C, 48Ca, [96X=2 (where X = Zr, Mo, Ru)]
ICP-OES 323.452 nm	0.0054 / 0.00092 µg/mL	1	Ce, Ar, Ni
ICP-OES 334.941 nm	0.0038 / 0.000028 µg/mL	1	Nb, Ta, Cr, U
ICP-OES 336.121 nm	0.0053 / 0.000034 µg/mL	1	W, Mo, Co

HF Note: This standard should not be prepared or stored in glass.

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

June 17, 2022

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- June 17, 2027

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Approved By:

Thomas Kozikowski
Manager, Quality Control



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director



INORGANIC VENTURES™
Refine your results. Redefine your industry.

Certificate of Analysis

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

MS985
R: 6/14/24

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).

ACCREDITED Testing Laboratory Certificate 863.01
ACCREDITED Reference Material Producer Certificate 863.02

2.0 PRODUCT DESCRIPTION

Product Code:	Single Analyte Custom Grade Solution
Catalog Number:	CGIN10
Lot Number:	U2-IN729349
Matrix:	5% (v/v) HNO3
Value / Analyte(s):	10 000 µg/mL ea: Indium
Starting Material:	Indium Metal
Starting Material Lot#:	2511
Starting Material Purity:	99.9995%

3.0 CERTIFIED VALUES AND UNCERTAINTIES

Certified Value: 10022 ± 30 µg/mL

Density: 1.044 g/mL (measured at 20 ± 4 °C)

Assay Information:

Assay Method #1	**10021 ± 56 µg/mL** ICP Assay NIST SRM 3124a Lot Number: 110516
Assay Method #2	**10035 ± 25 µg/mL** EDTA NIST SRM 928 Lot Number: 928
Assay Method #3	**10001 ± 33 µg/mL** Calculated NIST SRM Lot Number: See Sec. 4.2

- The Calculated Value is a value calculated from the weight of a starting material that has been certified directly vs. a National Institute of Standards and Technology (NIST) SRM/RM. See Sec 4.2 for balance traceability.

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$X_{CRM/RM} = \Sigma(w_i)(X_i)$

X_i = mean of Assay Method i with standard uncertainty $u_{char\ i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance:

$w_i = (1/u_{char\ i})^2 / (\Sigma(1/(u_{char\ i})^2)$

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\ (u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char} = [\Sigma((w_i)^2\ (u_{char\ i})^2)]^{½}$ where $u_{char\ i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$X_{CRM/RM} = (X_a)\ (u_{char\ a})$

X_a = mean of Assay Method A with

$u_{char\ a}$ = the standard uncertainty of characterization Method A

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\ (u^2_{char\ a} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char\ a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

CRM/RMs are tested for trace metallic impurities by Axial ICP-OES and ICP-MS. The result from the most sensitive method for each element, is reported below. Solutions tested by ICP-MS were analyzed in an ULPA-Filtered Clean Room. An ULPA-Filter is 99.9985% efficient for the removal of particles down to 0.3 µm.

M	Ag	<	0.000760	M	Eu	<	0.000760	O	Na		0.012771	M	Se	<	0.023000	M	Zn	<	0.006100
M	Al		0.003385	O	Fe		0.004462	M	Nb	<	0.000760	O	Si		0.024619	M	Zr	<	0.000760
M	As	<	0.004600	M	Ga	<	0.000760	M	Nd	<	0.000760	M	Sm	<	0.000760				
M	Au	<	0.002300	M	Gd	<	0.000760	O	Ni	<	0.005100	M	Sn	<	0.000760				
O	B		0.003692	M	Ge	<	0.001600	M	Os	<	0.000760	O	Sr	<	0.000610				
M	Ba	<	0.001600	M	Hf	<	0.000760	n	P	<		M	Ta	<	0.000760				
O	Be	<	0.000130	M	Hg	<	0.003100	M	Pb		0.001400	M	Tb	<	0.000760				
M	Bi	<	0.000760	M	Ho	<	0.000760	M	Pd	<	0.001600	M	Te	<	0.000760				
O	Ca		0.004616	s	In	<		M	Pr	<	0.000760	M	Th	<	0.000760				
M	Cd	<	0.000760	M	Ir	<	0.000760	M	Pt	<	0.000760	O	Ti	<	0.001100				
M	Ce	<	0.000760	O	K		0.007078	M	Rb	<	0.000760	M	Tl	<	0.000760				
M	Co	<	0.000760	M	La	<	0.000760	M	Re	<	0.000760	M	Tm	<	0.000760				
O	Cr	<	0.001300	O	Li	<	0.000130	M	Rh	<	0.000760	M	U	<	0.000760				
M	Cs	<	0.000760	M	Lu	<	0.000760	M	Ru	<	0.000760	M	V	<	0.001600				
M	Cu	<	0.003800	O	Mg		0.000707	n	S	<		M	W	<	0.001600				
M	Dy	<	0.000760	O	Mn		0.000149	M	Sb	<	0.000760	M	Y	<	0.000760				
M	Er	<	0.000760	M	Mo	<	0.002300	M	Sc	<	0.000760	M	Yb	<	0.000760				

M - Checked by ICP-MS O - Checked by ICP-OES i - Spectral Interference
n - Not Checked For s - Solution Standard Element

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures Terms and Conditions of Sale, https://www.inorganicventures.com/terms-and-conditions-sale. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit **www.inorganicventures.com/TCT**

Atomic Weight; Valence; Coordination Number; Chemical Form in Solution - 114.82 +3 6 In(H2O)6+3

Chemical Compatibility -Soluble in HCl, HNO3, and H2SO4. Avoid neutral and basic media. Stable with most metals and inorganic anions. The oxalate, sulfide, carbonate, hydroxide and phosphate are insoluble in water.

Stability - 2-100 ppb levels stable for months in 1% HNO3 / LDPE container. 1-10,000 ppm solutions chemically stable for years in 2-5% HNO3 / LDPE container.

In Containing Samples (Preparation and Solution) -Metal (Best dissolved in HCl / HNO3); Oxide (Soluble in mineral acids); Ores (Carbonate fusion in Pt0 followed by HCl dissolution); Organic Matrices (Sulfuric/peroxide digestion or dry ash and dissolution in dilute HCl).

Atomic Spectroscopic Information (ICP-OES D.L.s are given as radial/axial view):

Technique/Line	**Estimated D.L.**	**Order**	**Interferences** (underlined indicates severe)
ICP-MS 115 amu	1 ppt	n/a	115Sn, 99Ru16O
ICP-OES 158.583 nm	0.05 / 0.002 µg/mL	1	
ICP-OES 230.606 nm	0.1 / 0.03 µg/mL	1	Ni, Os
ICP-OES 325.609 nm	0.2 / 0.05 µg/mL	1	Mn, Mo, Th

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

February 21, 2023

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- February 21, 2028

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Approved By:

Thomas Kozikowski
Manager, Quality Control



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director



INORGANIC VENTURES
Refine your results. Redefine your industry.

Certificate of Analysis

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

R: 2/22/24
MS-997

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).

ACCREDITED Testing Laboratory Certificate 883.01
ACCREDITED Reference Material Producer Certificate 883.02

2.0 PRODUCT DESCRIPTION

Product Code:	Multi Analyte Custom Grade Solution
Catalog Number:	CLPP-CAL-3
Lot Number:	T2-MEB727800
Matrix:	7% (v/v) HNO3
Value / Analyte(s):	1 000 µg/mL ea: Arsenic, Lead, Selenium, Thallium, 500 µg/mL ea: Cadmium

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ANALYTE	CERTIFIED VALUE	ANALYTE	CERTIFIED VALUE
Arsenic, As	**1 000 ± 7 µg/mL**	**Cadmium, Cd**	**500.0 ± 2.2 µg/mL**
Lead, Pb	**1 000 ± 4 µg/mL**	**Selenium, Se**	**1 000 ± 6 µg/mL**
Thallium, Tl	**1 000 ± 7 µg/mL**		

Density: 1.042 g/mL (measured at 20 ± 4 °C)

Assay Information:

ANALYTE	METHOD	NIST SRM#	SRM LOT#
As	ICP Assay	3103a	100818
As	Calculated		See Sec. 4.2
Cd	ICP Assay	3108	130116
Cd	EDTA	928	928
Pb	ICP Assay	3128	101026
Pb	EDTA	928	928
Se	ICP Assay	3149	100901
Tl	ICP Assay	3158	151215

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$X_{CRM/RM} = \Sigma(w_i)(X_i)$

X_i = mean of Assay Method i with standard uncertainty $u_{char\ i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance:

$w_i = (1/u_{char\ i}^2) / (\Sigma(1/(u_{char\ i}^2))$

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{1/2}$

k = coverage factor = 2

$u_{char} = [\Sigma((w_i)^2\,(u_{char\ i}^2)]^{1/2}$ where $u_{char\ i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$X_{CRM/RM} = (X_a)\,(u_{char\ a})$

X_a = mean of Assay Method A with

$u_{char\ a}$ = the standard uncertainty of characterization Method A

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char\ a} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{1/2}$

k = coverage factor = 2

$u_{char\ a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

N/A

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures Terms and Conditions of Sale, https://www.inorganicventures.com/terms-and-conditions-sale. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit **www.inorganicventures.com/TCT**

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

December 21, 2022

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- December 21, 2027

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Approved By:

Thomas Kozikowski
Manager, Quality Control



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director



INORGANIC VENTURES
Refine your results. Redefine your industry.

Certificate of Analysis

R! 08/22/24 M6058, M6059

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).

ACCREDITED Testing Laboratory Certificate 883.01
ACCREDITED Reference Material Producer Certificate 883.02

2.0 PRODUCT DESCRIPTION

Product Code:	Multi Analyte Custom Grade Solution
Catalog Number:	CHEM-CLP-4
Lot Number:	V2-MEB746172
Matrix:	3% (v/v) HNO_3 3% (v/v) HF
Value / Analyte(s):	1 000 µg/mL ea: Boron, Molybdenum, Silicon, Tin, Titanium

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ANALYTE	CERTIFIED VALUE	ANALYTE	CERTIFIED VALUE
Boron, B	1 000 ± 5 µg/mL	Molybdenum, Mo	1 000 ± 5 µg/mL
Silicon, Si	1 000 ± 7 µg/mL	Tin, Sn	1 000 ± 5 µg/mL
Titanium, Ti	1 000 ± 6 µg/mL		

Density: 1.032 g/mL (measured at 20 ± 4 °C)

Assay Information:

ANALYTE	METHOD	NIST SRM#	SRM LOT#
B	ICP Assay	3107	190605
B	Calculated		See Sec. 4.2
Mo	ICP Assay	traceable to 3134	U2-MO739068
Si	ICP Assay	Traceable to 3150	S2-SI702546
Sn	ICP Assay	3161a	140917
Ti	ICP Assay	traceable to 3162a	T2-TI725816

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$X_{CRM/RM} = \Sigma(w_i)(X_i)$

X_i = mean of Assay Method i with standard uncertainty $u_{char\ i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance:

$w_i = (1/u_{char\ i})^2 / (\Sigma(1/(u_{char\ i})^2)$

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{1/2}$

k = coverage factor = 2

$u_{char} = [\Sigma((w_i)^2\,(u_{char\ i})^2)]^{1/2}$ where $u_{char\ i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$X_{CRM/RM} = (X_a)\,(u_{char\ a})$

X_a = mean of Assay Method A with

$u_{char\ a}$ = the standard uncertainty of characterization Method A

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char\ a} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{1/2}$

k = coverage factor = 2

$u_{char\ a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

N/A

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures Terms and Conditions of Sale, https://www.inorganicventures.com/terms-and-conditions-sale. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit **www.inorganicventures.com/TCT**

HF Note: This standard should not be prepared or stored in glass.

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

August 12, 2024

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- August 12, 2029

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Approved By:

Joseph Burns
Custom VS Manager



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director



INORGANIC VENTURES™
Refine your results. Redefine your industry.

Certificate of Analysis

M6074
M6075
M6076
M6077

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

EXP.: 9/6/2029

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).

ACCREDITED Testing Laboratory Certificate 883.01
ACCREDITED Reference Material Producer Certificate 883.02

2.0 PRODUCT DESCRIPTION

Product Code:	Multi Analyte Custom Grade Solution
Catalog Number:	CHEM-CLP-4
Lot Number:	V2-MEB746762
Matrix:	3% (v/v) HNO3 3% (v/v) HF
Value / Analyte(s):	1 000 µg/mL ea: Boron, Molybdenum, Silicon, Tin, Titanium

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ANALYTE	CERTIFIED VALUE	ANALYTE	CERTIFIED VALUE
Boron, B	**1 000 ± 5 µg/mL**	**Molybdenum, Mo**	**1 000 ± 5 µg/mL**
Silicon, Si	**1 000 ± 7 µg/mL**	**Tin, Sn**	**1 000 ± 5 µg/mL**
Titanium, Ti	**1 000 ± 6 µg/mL**		

Density: 1.033 g/mL (measured at 20 ± 4 °C)

Assay Information:

ANALYTE	METHOD	NIST SRM#	SRM LOT#
B	ICP Assay	3107	190605
B	Calculated		See Sec. 4.2
Mo	ICP Assay	traceable to 3134	U2-MO739068
Si	ICP Assay	Traceable to 3150	S2-SI702546
Sn	ICP Assay	3161a	140917
Ti	ICP Assay	traceable to 3162a	T2-TI725816

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$X_{CRM/RM} = \Sigma(w_i)(X_i)$

X_i = mean of Assay Method i with standard uncertainty $u_{char\ i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance:

$w_i = (1/u_{char\ i})^2 / (\Sigma(1/(u_{char\ i})^2)$

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char} = [\Sigma((w_i)^2\,(u_{char\ i})^2)]^{½}$ where $u_{char\ i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$X_{CRM/RM} = (X_a)\,(u_{char\ a})$

X_a = mean of Assay Method A with

$u_{char\ a}$ = the standard uncertainty of characterization Method A

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char\ a} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char\ a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

N/A

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures Terms and Conditions of Sale, https://www.inorganicventures.com/terms-and-conditions-sale. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit **www.inorganicventures.com/TCT**

HF Note: This standard should not be prepared or stored in glass.

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

September 06, 2024

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- September 06, 2029

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Approved By:

Joseph Burns
Custom VS Manager



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director



INORGANIC VENTURES™
Refine your results. Redefine your industry.

Certificate of Analysis

M6074
M6075
M6076
M6077

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

EXP.: 9/6/2029

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).

ACCREDITED Testing Laboratory Certificate 883.01
ACCREDITED Reference Material Producer Certificate 883.02

2.0 PRODUCT DESCRIPTION

Product Code:	Multi Analyte Custom Grade Solution
Catalog Number:	CHEM-CLP-4
Lot Number:	V2-MEB746762
Matrix:	3% (v/v) HNO3 3% (v/v) HF
Value / Analyte(s):	1 000 µg/mL ea: Boron, Molybdenum, Silicon, Tin, Titanium

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ANALYTE	CERTIFIED VALUE	ANALYTE	CERTIFIED VALUE
Boron, B	**1 000 ± 5 µg/mL**	**Molybdenum, Mo**	**1 000 ± 5 µg/mL**
Silicon, Si	**1 000 ± 7 µg/mL**	**Tin, Sn**	**1 000 ± 5 µg/mL**
Titanium, Ti	**1 000 ± 6 µg/mL**		

Density: 1.033 g/mL (measured at 20 ± 4 °C)

Assay Information:

ANALYTE	METHOD	NIST SRM#	SRM LOT#
B	ICP Assay	3107	190605
B	Calculated		See Sec. 4.2
Mo	ICP Assay	traceable to 3134	U2-MO739068
Si	ICP Assay	Traceable to 3150	S2-SI702546
Sn	ICP Assay	3161a	140917
Ti	ICP Assay	traceable to 3162a	T2-TI725816

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$X_{CRM/RM} = \Sigma(w_i)(X_i)$

X_i = mean of Assay Method i with standard uncertainty $u_{char\ i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance:

$w_i = (1/u_{char\ i})^2 / (\Sigma(1/(u_{char\ i})^2)$

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char} = [\Sigma((w_i)^2\,(u_{char\ i})^2)]^{½}$ where $u_{char\ i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$X_{CRM/RM} = (X_a)\,(u_{char\ a})$

X_a = mean of Assay Method A with

$u_{char\ a}$ = the standard uncertainty of characterization Method A

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char\ a} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char\ a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

N/A

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures Terms and Conditions of Sale, https://www.inorganicventures.com/terms-and-conditions-sale. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit **www.inorganicventures.com/TCT**

HF Note: This standard should not be prepared or stored in glass.

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

September 06, 2024

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- September 06, 2029

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Approved By:

Joseph Burns
Custom VS Manager



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director



Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

R->1/13/28

M6128

Part Number:	**58025**	**Solvent:**	**Lot #** 24002546	Nitric Acid
Lot Number:	**101124**			
Description:	**Manganese (Mn)**	2%	80.0 (mL)	Nitric Acid
Expiration Date:	101127			
Recommended Storage:	Ambient (20 °C)			
Nominal Concentration (µg/mL):	**1000**			
NIST Test Number:	6UTB	5E-05	Balance Uncertainty	
Weight shown below was diluted to (mL):	4000.2	0.10	Flask Uncertainty	


Giovanni Esposito

Formulated By:	Giovanni Esposito	101124

Pedro Rentas

Reviewed By:	Pedro L. Rentas	101124

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	**Actual Conc. (µg/mL)**	Expanded Uncertainty +/- (µg/mL)	**SDS Information** (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Manganese(II) nitrate hydrate (Mn)	IN031	MNM092020A1	1000	99.999	0.10	20.8	19.2322	19.2344	**1000.1**	**2.0**	15710-66-4	5 mg/m3	orl-rat >300mg/kg	3132



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	T	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

Certified by:



* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

INORGANIC VENTURES™
Refine your results. Redefine your industry.

Certificate of Analysis

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

M6137 R→10/3/24

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).

ACCREDITED Testing Laboratory Certificate 883.01
ACCREDITED Reference Material Producer Certificate 883.02

2.0 PRODUCT DESCRIPTION

Product Code:	Single Analyte Custom Grade Solution
Catalog Number:	CGSI1
Lot Number:	V2-SI744713
Matrix:	tr. HNO3 tr. HF
Value / Analyte(s):	1 000 µg/mL ea: Silicon
Starting Material:	Silica
Starting Material Lot#:	1771
Starting Material Purity:	99.9981%

3.0 CERTIFIED VALUES AND UNCERTAINTIES

Certified Value: 999 ± 6 µg/mL

Density: 1.003 g/mL (measured at 20 ± 4 °C)

Assay Information:

Assay Method #1	**999 ± 5 µg/mL** ICP Assay NIST SRM Traceable to 3150 Lot Number: S2-SI702546
Assay Method #2	**1000 ± 7 µg/mL** Calculated NIST SRM Lot Number: See Sec. 4.2

- The Calculated Value is a value calculated from the weight of a starting material that has been certified directly vs. a National Institute of Standards and Technology (NIST) SRM/RM. See Sec 4.2 for balance traceability.

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$X_{CRM/RM} = \Sigma(w_i)(X_i)$

X_i = mean of Assay Method i with standard uncertainty $u_{char\ i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance:

$w_i = (1/u_{char\ i})^2 / (\Sigma(1/(u_{char\ i})^2)$

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{1/2}$

k = coverage factor = 2

$u_{char} = [\Sigma((w_i)^2 (u_{char\ i})^2)]^{1/2}$ where $u_{char\ i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$X_{CRM/RM} = (X_a)(u_{char\ a})$

X_a = mean of Assay Method A with

$u_{char\ a}$ = the standard uncertainty of characterization Method A

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char\ a} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{1/2}$

k = coverage factor = 2

$u_{char\ a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

CRM/RMs are tested for trace metallic impurities by Axial ICP-OES and ICP-MS. The result from the most sensitive method for each element, is reported below. Solutions tested by ICP-MS were analyzed in an ULPA-Filtered Clean Room. An ULPA-Filter is 99.9985% efficient for the removal of particles down to 0.3 µm.

M	Ag	<	0.000310	M	Eu	<	0.000310	O	Na		0.001656	M	Se	<	0.022000	M	Zn	<	0.002500
M	Al		0.010787	M	Fe	<	0.027000	M	Nb	<	0.001300	s	Si	<		O	Zr	<	0.001900
M	As	<	0.001900	M	Ga	<	0.001300	M	Nd	<	0.000310	M	Sm	<	0.000310				
M	Au	<	0.000910	M	Gd	<	0.000310	M	Ni	<	0.005500	M	Sn		0.000096				
M	B		0.016180	M	Ge	<	0.001900	M	Os	<	0.000610	O	Sr		0.000092				
M	Ba		0.000096	M	Hf		0.000423	i	P	<		M	Ta		0.002542				
O	Be	<	0.000570	M	Hg	<	0.000610	M	Pb	<	0.000310	M	Tb	<	0.000310				
M	Bi	<	0.000310	M	Ho	<	0.000610	M	Pd	<	0.000610	M	Te	<	0.000910				
O	Ca		0.011557	M	In	<	0.000310	M	Pr	<	0.000310	M	Th	<	0.001900				
M	Cd	<	0.000310	M	Ir	<	0.000310	M	Pt	<	0.000310	M	Ti		0.001078				
M	Ce	<	0.000610	O	K		0.000577	M	Rb	<	0.009100	M	Tl	<	0.000310				
M	Co	<	0.001600	M	La	<	0.000310	M	Re	<	0.000310	M	Tm	<	0.000310				
M	Cr	<	0.010000	O	Li	<	0.000460	M	Rh	<	0.000310	M	U	<	0.000310				
M	Cs	<	0.000310	M	Lu	<	0.000310	M	Ru	<	0.000310	O	V	<	0.001300				
M	Cu	<	0.002500	O	Mg		0.001348	O	S	<	0.570000	M	W	<	0.001900				
M	Dy	<	0.000310	M	Mn	<	0.002500	M	Sb	<	0.000310	M	Y	<	0.000310				
M	Er	<	0.000310	M	Mo	<	0.000310	O	Sc	<	0.000590	M	Yb	<	0.000310				

M - Checked by ICP-MS O - Checked by ICP-OES i - Spectral Interference
n - Not Checked For s - Solution Standard Element

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures Terms and Conditions of Sale, https://www.inorganicventures.com/terms-and-conditions-sale. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit **www.inorganicventures.com/TCT**

Atomic Weight; Valence; Coordination Number; Chemical Form in Solution - 28.09 +4 6 Si(OH)x(F)y2-

Chemical Compatibility -Soluble in HCl, HF, H3PO4 H2SO4 and HNO3 as the Si(OH)x(F)y2-. Avoid neutral to basic media. Unstable at ppm levels with metals that would pull F- away (i.e. Do not mix with Alkaline or Rare Earths, or high levels of transition elements unless they are fluorinated. Stable with most inorganic anions with a tendency to hydrolyze forming silicic acid (silicic acid is soluble up to ∼100 ppm in water) in all dilute acids except HF.

Stability - 2-100 ppb levels - stability unknown - (alone or mixed with all other metals) as the Si(OH)x(F)y2-. 1-10,000 ppm single element solutions as the Si(OH)x(F)y2- chemically stable for years in 2-5 % HNO3 / trace HF in a LDPE container.

Si Containing Samples (Preparation and Solution) -Metal (Soluble in 1:1:1 H2O / HF / HNO3); Oxide - SiO2, amorphic (dissolve by heating in 1:1:1 H2O / HF / HNO3); Oxide - quartz (fuse in Pt0 with Na2CO3); Geological Samples(fuse in Pt0with Na2CO3 followed by HCl solution of the fuseate); Organic Matrices containing silicates and non volatile silicon compounds (dry ash at 4500C in Pt0 and dissolve by gently warming with 1:1:1 H2O / HF / H2SO4 or fuse / ash with Na2CO3 and dissolve fuseate with HCl / H2O); Silicone Oils - dimethyl silicones depolymerize to form volatile monomer units when heated (Measure directly in alcoholic KOH / xylene mixture where sample is treated first with the KOH at 60-1000C to "unzip" the Si- O-Si polymeric structure or digest with conc. H2SO4 / H2O2 followed by cooling and dissolution of the dehydrated silica with HF.) Note that the direct analysis of silicone oils in an organic solvent will result in false high results due to high vapor pressure of volatile monomer units like hexamethylcyclotrisiloxane. The KOH forms the K2+Si(CH3)2O= salt which is not volatile at room temperature.

Atomic Spectroscopic Information (ICP-OES D.L.s are given as radial/axial view):

Technique/Line	**Estimated D.L.**	**Order**	**Interferences** (underlined indicates severe)
ICP-MS 28 amu	4000 - 8000 ppt	N/A	N2, 12C16O
ICP-OES 212.412 nm	0.02/0.01 µg/mL	1	Hf, Os, Mo, Ta
ICP-OES 251.611 nm	0.012/0.003 µg/mL	1	Ta, U, Zn, Th
ICP-OES 288.158 nm	0.03/0.004 µg/mL	1	Ta, Ce, Cr, Cd, Th

HF Note: This standard should not be prepared or stored in glass.

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

July 10, 2024

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- **July 10, 2029**

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Prepared By:

Uyen Truong
Custom Processing Supervisor



Certificate Approved By:

Muzzammil Khan
Stock Laboratory Supervisor



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director



Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:


Part Number: 58126 R->1|B|2S
Lot Number: 011025
Description: Iron (Fe) M6140
Expiration Date: 011028
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 10000
NIST Test Number: 6UTB
Weight shown below was diluted to (mL): 2000.07 ± 0.100 Flask Uncertainty

		Lot #	
Solvent:	24012496	Nitric Acid	
	2%	40.0 (mL)	Nitric Acid

5E-05 Balance Uncertainty

Formulated By:	Benson Chan	011025
Reviewed By:	Pedro L. Rentas	011025

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Iron(III) nitrate nonahydrate (Fe)	IN028	FED082023A1	10000	99.999	0.10	13.0	153.8534	######	**10000.2**	**20.0**	7782-61-8	1 mg/m3	orl-rat 3250mg/kg	3126A



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)

Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	T	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T)= Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

R-> 1/13/25

Part Number: 57119
Lot Number: 103024
Description: Potassium (K)

M6141
M6142
M6143

Expiration Date: 103027
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 10000
NIST Test Number: 6UTB

		Lot #	
Solvent:	24002546	Nitric Acid	
2%	80.0 (mL)	Nitric Acid	


5E-05 Balance Uncertainty

Weight shown below was diluted to (mL): 4000.1 0.15 Flask Uncertainty

Giovanni Esposito

Formulated By:	Giovanni Esposito	103024
Reviewed By:	Pedro L. Rentas	103024

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Potassium nitrate (K)	IN034	KD062022A1	10000	99.999	0.10	37.7	106.1040	######	**10001.1**	**20.0**	7757-79-1	5 mg/m3	orl-rat 3750 mg/kg	3141a



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	T	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com


CERTIFIED WEIGHT REPORT:

R->1/13/25

Part Number:	**58111**		**Lot #**		
Lot Number:	**072424**	**Solvent:**	24002546	Nitric Acid	
Description:	**Sodium (Na)**				
		MG144 2%	80.0 (mL)	Nitric Acid	
Expiration Date:	072427				
Recommended Storage:	Ambient (20 °C)				
Nominal Concentration (µg/mL):	**10000**				
NIST Test Number:	6UTB	5E-05	Balance Uncertainty		
Weight shown below was diluted to (mL):	4000.2	0.10	Flask Uncertainty		

Formulated By:	Benson Chan 072424
Reviewed By:	Pedro L. Rentas 072424

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Sodium nitrate (Na)	IN036	NAV012015I1	10000	99.999	0.10	26.9	148.7096	######	**10000.0**	**20.0**	7631-99-4	5 mg/m3	orl-rat 3430 mg/kg	3152a



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Sc	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	T	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com


CERTIFIED WEIGHT REPORT:

Part Number:	**58030**	R → 1/13/25	**Solvent:**	**Lot #** 24012496	Nitric Acid
Lot Number:	**121724**				
Description:	**Zinc (Zn)**	M6145	2%	40.0 (mL)	Nitric Acid
Expiration Date:	121727				
Recommended Storage:	Ambient (20 °C)				
Nominal Concentration (µg/mL):	**1000**				
NIST Test Number:	6UTB	5E-05 Balance Uncertainty			
Weight shown below was diluted to (mL):	2000.1	0.10 Flask Uncertainty			

Aleah O'Brady

Formulated By:	Aleah O'Brady	121724
Reviewed By:	Pedro L. Rentas	121724

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Zinc nitrate hexahydrate (Zn)	IN016	ZNE032021A1	1000	99.999	0.10	24.3	8.2308	8.2311	**1000.0**	**2.0**	10196-18-6	1 mg/m3	orl-rat 1190mg/kg	3168



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	T
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

avantor™ | J.T.Baker®

M6151

R-> 1/15/25

Material No.: 9530-33
Batch No.: 22G2862015
Manufactured Date: 2022-06-15
Retest Date: 2027-06-14
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid-base titrn)	36.5 – 38.0 %	37.9 %
ACS – Color (APHA)	≤ 10	5
ACS – Residue after Ignition	≤ 3 ppm	< 1 ppm
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.191
ACS – Bromide (Br)	≤ 0.005 %	< 0.005 %
ACS – Extractable Organic Substances	≤ 5 ppm	< 1 ppm
ACS – Free Chlorine (as Cl_2)	≤ 0.5 ppm	< 0.5 ppm
Phosphate (PO_4)	≤ 0.05 ppm	< 0.03 ppm
Sulfate (SO_4)	≤ 0.5 ppm	< 0.3 ppm
Sulfite (SO_3)	≤ 0.8 ppm	0.3 ppm
Ammonium (NH_4)	≤ 3 ppm	< 1 ppm
Trace Impurities – Arsenic (As)	≤ 0.010 ppm	< 0.003 ppm
Trace Impurities – Aluminum (Al)	≤ 10.0 ppb	1.3 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 3.0 ppb
Trace Impurities – Barium (Ba)	≤ 1.0 ppb	0.2 ppb
Trace Impurities – Beryllium (Be)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Bismuth (Bi)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Boron (B)	≤ 20.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 1.0 ppb	< 0.3 ppb
Trace Impurities – Calcium (Ca)	≤ 50.0 ppb	163.0 ppb
Trace Impurities – Chromium (Cr)	≤ 1.0 ppb	0.7 ppb
Trace Impurities – Cobalt (Co)	≤ 1.0 ppb	< 0.3 ppb
Trace Impurities – Copper (Cu)	≤ 1.0 ppb	< 0.1 ppb
Trace Impurities – Gallium (Ga)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Germanium (Ge)	≤ 3.0 ppb	< 2.0 ppb
Trace Impurities – Gold (Au)	≤ 4.0 ppb	0.6 ppb
Heavy Metals (as Pb)	≤ 100 ppb	< 50 ppb
Trace Impurities – Iron (Fe)	≤ 15 ppb	6 ppb

>>> Continued on page 2 >>>

Hydrochloric Acid, 36.5-38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

avantor™ | J.T.Baker®

Material No.: 9530-33
Batch No.: 22G2862015

Test	Specification	Result
Trace Impurities - Lead (Pb)	≤ 1.0 ppb	< 0.5 ppb
Trace Impurities - Lithium (Li)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities - Magnesium (Mg)	≤ 10.0 ppb	2.9 ppb
Trace Impurities - Manganese (Mn)	≤ 1.0 ppb	< 0.4 ppb
Trace Impurities - Mercury (Hg)	≤ 0.5 ppb	0.1 ppb
Trace Impurities - Molybdenum (Mo)	≤ 10.0 ppb	< 3.0 ppb
Trace Impurities - Nickel (Ni)	≤ 4.0 ppb	< 0.3 ppb
Trace Impurities - Niobium (Nb)	≤ 1.0 ppb	0.8 ppb
Trace Impurities - Potassium (K)	≤ 9.0 ppb	< 2.0 ppb
Trace Impurities - Selenium (Se), For Information Only		< 1.0 ppb
Trace Impurities - Silicon (Si)	≤ 100.0 ppb	< 10.0 ppb
Trace Impurities - Silver (Ag)	≤ 1.0 ppb	0.5 ppb
Trace Impurities - Sodium (Na)	≤ 100.0 ppb	2.3 ppb
Trace Impurities - Strontium (Sr)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities - Tantalum (Ta)	≤ 1.0 ppb	1.6 ppb
Trace Impurities - Thallium (Tl)	≤ 5.0 ppb	< 2.0 ppb
Trace Impurities - Tin (Sn)	≤ 5.0 ppb	4.0 ppb
Trace Impurities - Titanium (Ti)	≤ 1.0 ppb	1.5 ppb
Trace Impurities - Vanadium (V)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities - Zinc (Zn)	≤ 5.0 ppb	0.8 ppb
Trace Impurities - Zirconium (Zr)	≤ 1.0 ppb	0.3 ppb

>>> Continued on page 3 >>>

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

avantor™ | J.T.Baker®

Material No.: 9530-33
Batch No.: 22G2862015

Test	Specification	Result

For Laboratory,Research,or Manufacturing Use
Product Information (not specifications):
Appearance (clear, fuming liquid)
Meets ACS Specifications
Storage Condition: Store below 25 °C.

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

James T Ethier

Jamie Ethier
Vice President Global Quality

Nitric Acid 69%
CMOS

avantor™ | J.T.Baker®

R-02102/2025

M-6158

Material No.: 9606-03
Batch No.: 24D1062002
Manufactured Date: 2024-03-26
Retest Date: 2029-03-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (HNO_3)	69.0 - 70.0 %	69.7 %
Appearance	Passes Test	Passes Test
Color (APHA)	≤ 10	5
Residue after Ignition	≤ 2 ppm	1 ppm
Chloride (Cl)	≤ 0.08 ppm	< 0.03 ppm
Phosphate (PO_4)	≤ 0.10 ppm	< 0.03 ppm
Sulfate (SO_4)	≤ 0.2 ppm	< 0.2 ppm
Trace Impurities - Aluminum (Al)	≤ 40.0 ppb	< 1.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 2.0 ppb
Trace Impurities - Barium (Ba)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities - Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities - Cadmium (Cd)	≤ 50 ppb	< 1 ppb
Trace Impurities - Calcium (Ca)	≤ 50.0 ppb	2.3 ppb
Trace Impurities - Chromium (Cr)	≤ 30.0 ppb	< 1.0 ppb
Trace Impurities - Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Germanium (Ge)	≤ 20 ppb	< 10 ppb
Trace Impurities - Gold (Au)	≤ 20 ppb	< 5 ppb
Heavy Metals (as Pb)	≤ 100 ppb	100 ppb
Trace Impurities - Iron (Fe)	≤ 40.0 ppb	< 1.0 ppb
Trace Impurities - Lead (Pb)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities - Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities - Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Nickel (Ni)	≤ 20.0 ppb	< 5.0 ppb

>>> Continued on page 2 >>>

Nitric Acid 69%
CMOS

avantor™ | J.T.Baker®

Material No.: 9606-03
Batch No.: 24D1062002

Test	Specification	Result
Trace Impurities - Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities - Potassium (K)	≤ 50 ppb	16 ppb
Trace Impurities - Silicon (Si)	≤ 50 ppb	< 10 ppb
Trace Impurities - Silver (Ag)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities - Sodium (Na)	≤ 150.0 ppb	< 5.0 ppb
Trace Impurities - Strontium (Sr)	≤ 30.0 ppb	< 1.0 ppb
Trace Impurities - Tantalum (Ta)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities - Thallium (Tl)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities - Tin (Sn)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities - Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Zinc (Zn)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities - Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count - 0.5 µm and greater	≤ 60 par/ml	10 par/ml
Particle Count - 1.0 µm and greater	≤ 10 par/ml	3 par/ml

>>> Continued on page 3 >>>

Nitric Acid 69%
CMOS

avantor™ | J.T.Baker®

Material No.: 9606-03
Batch No.: 24D1062002

Test	Specification	Result

For Microelectronic Use

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production

Nitric Acid 69%
CMOS

avantor™ | J.T.Baker®

m-5162

R. Date :- 04/27/2025

Material No.: 9606-03
Batch No.: 24H0162012
Manufactured Date: 2024-06-28
Retest Date: 2029-06-27
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (HNO_3)	69.0 – 70.0 %	69.7 %
Appearance	Passes Test	Passes Test
Color (APHA)	≤ 10	5
Residue after Ignition	≤ 2 ppm	< 1 ppm
Chloride (Cl)	≤ 0.08 ppm	0.03 ppm
Phosphate (PO_4)	≤ 0.10 ppm	< 0.03 ppm
Sulfate (SO_4)	≤ 0.2 ppm	< 0.2 ppm
Trace Impurities – Aluminum (Al)	≤ 40.0 ppb	< 1.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 2.0 ppb
Trace Impurities – Barium (Ba)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Bismuth (Bi)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	0.1 ppb
Trace Impurities – Cadmium (Cd)	≤ 50 ppb	< 1 ppb
Trace Impurities – Calcium (Ca)	≤ 50.0 ppb	0.3 ppb
Trace Impurities – Chromium (Cr)	≤ 30.0 ppb	0.1 ppb
Trace Impurities – Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Germanium (Ge)	≤ 20 ppb	< 1 ppb
Trace Impurities – Gold (Au)	≤ 20 ppb	< 1 ppb
Heavy Metals (as Pb)	≤ 100 ppb	< 50 ppb
Trace Impurities – Iron (Fe)	≤ 40.0 ppb	< 1.0 ppb
Trace Impurities – Lead (Pb)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities – Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Nickel (Ni)	≤ 20.0 ppb	< 1.0 ppb

>>> Continued on page 2 >>>

Nitric Acid 69%
CMOS

avantor™ | J.T.Baker®

Material No.: 9606-03
Batch No.: 24H0162012

Test	Specification	Result
Trace Impurities - Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities - Potassium (K)	≤ 50 ppb	< 1 ppb
Trace Impurities - Silicon (Si)	≤ 50 ppb	1 ppb
Trace Impurities - Silver (Ag)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities - Sodium (Na)	≤ 150.0 ppb	< 1.0 ppb
Trace Impurities - Strontium (Sr)	≤ 30.0 ppb	< 1.0 ppb
Trace Impurities - Tantalum (Ta)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Thallium (Tl)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Tin (Sn)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities - Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Zinc (Zn)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities - Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count - 0.5 µm and greater	≤ 60 par/ml	13 par/ml
Particle Count - 1.0 µm and greater	≤ 10 par/ml	5 par/ml

>>> Continued on page 3 >>>

Nitric Acid 69%
CMOS

avantor™ | J.T.Baker®

Material No.: 9606-03
Batch No.: 24H0162012

Test	Specification	Result

For Microelectronic Use

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production

INORGANIC VENTURES™
Refine your results. Redefine your industry.

Certificate of Analysis

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).

ACCREDITED
Testing Laboratory
Certificate 883.01

ACCREDITED
Reference Material Producer
Certificate 883.02

2.0 PRODUCT DESCRIPTION

Product Code:	Multi Analyte Custom Grade Solution
Catalog Number:	QCP-CICV-1
Lot Number:	V2-MEB744107
Matrix:	7% (v/v) HNO3
Value / Analyte(s):	2 500 µg/mL ea: Calcium, Potassium, Magnesium, Sodium, 1 000 µg/mL ea: Aluminum, Barium, 500 µg/mL ea: Iron, 250 µg/mL ea: Nickel, Vanadium, Zinc, Cobalt, Manganese, 125 µg/mL ea: Silver, Copper, 100 µg/mL ea: Chromium, 25 µg/mL ea: Beryllium

Second Source: Whenever possible, this solution was manufactured from a second set of concentrates in our manufacturing facility.

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ANALYTE	CERTIFIED VALUE	ANALYTE	CERTIFIED VALUE
Aluminum, Al	**1 000 ± 4 µg/mL**	**Barium, Ba**	**1 000 ± 6 µg/mL**
Beryllium, Be	**24.98 ± 0.12 µg/mL**	**Calcium, Ca**	**2 500 ± 8 µg/mL**
Chromium, Cr	**99.9 ± 0.6 µg/mL**	**Cobalt, Co**	**250.2 ± 1.2 µg/mL**
Copper, Cu	**125.0 ± 0.5 µg/mL**	**Iron, Fe**	**500.0 ± 2.2 µg/mL**
Magnesium, Mg	**2 500 ± 11 µg/mL**	**Manganese, Mn**	**249.9 ± 1.1 µg/mL**
Nickel, Ni	**250.0 ± 1.2 µg/mL**	**Potassium, K**	**2 500 ± 11 µg/mL**
Silver, Ag	**125.0 ± 0.6 µg/mL**	**Sodium, Na**	**2 500 ± 11 µg/mL**
Vanadium, V	**250.0 ± 1.1 µg/mL**	**Zinc, Zn**	**249.9 ± 1.1 µg/mL**

Density: 1.081 g/mL (measured at 20 ± 4 ºC)

Assay Information:

ANALYTE	METHOD	NIST SRM#	SRM LOT#
Ag	ICP Assay	3151	160729
Ag	Volhard	999c	999c
Al	ICP Assay	3101a	140903
Al	EDTA	928	928
Ba	ICP Assay	3104a	140909
Ba	Gravimetric		See Sec. 4.2
Be	ICP Assay	3105a	090514
Be	Calculated		See Sec. 4.2
Ca	ICP Assay	3109a	130213
Ca	EDTA	928	928
Co	ICP Assay	3113	190630
Co	EDTA	928	928
Cr	ICP Assay	3112a	170630
Cu	ICP Assay	3114	120618
Cu	EDTA	928	928
Fe	ICP Assay	3126a	140812
Fe	EDTA	928	928
K	ICP Assay	3141a	140813
K	Gravimetric		See Sec. 4.2
Mg	ICP Assay	3131a	140110
Mg	EDTA	928	928
Mn	ICP Assay	3132	050429
Mn	EDTA	928	928
Na	ICP Assay	3152a	200413
Na	Gravimetric		See Sec. 4.2
Ni	ICP Assay	3136	120619
Ni	EDTA	928	928
V	ICP Assay	3165	160906
V	EDTA	928	928
Zn	ICP Assay	3168a	120629
Zn	EDTA	928	928

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$X_{CRM/RM} = \Sigma(w_i)(X_i)$

X_i = mean of Assay Method i with standard uncertainty $u_{char\ i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance:

$w_i = (1/u_{char\ i})^2 / (\Sigma(1/(u_{char\ i})^2)$

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char} = [\Sigma((w_i)^2\,(u_{char\ i})^2)]^{½}$ where $u_{char\ i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$X_{CRM/RM} = (X_a)\,(u_{char\ a})$

X_a = mean of Assay Method A with

$u_{char\ a}$ = the standard uncertainty of characterization Method A

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char\ a} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char\ a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

N/A

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures Terms and Conditions of Sale, https://www.inorganicventures.com/terms-and-conditions-sale. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit **www.inorganicventures.com/TCT**

Note: This solution contains Silver (Ag), please refer to our Sample Preparation Guide for more information (https://www.inorganicventures.com/sample-preparation-guide/samples-containing-silver)

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

May 22, 2024

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- **May 22, 2029**

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS


Certificate Prepared By:

Justin Dirico
Stock Processing Supervisor



Certificate Approved By:

Jodie Wall
Stock VSM Coordinator



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director



Certificate of Analysis

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).

ACCREDITED Testing Laboratory Certificate 883.01
ACCREDITED Reference Material Producer Certificate 883.02

2.0 PRODUCT DESCRIPTION

Product Code:	Multi Analyte Custom Grade Solution
Catalog Number:	QCP-CICV-2
Lot Number:	U2-MEB733713
Matrix:	3% (w/v) Tartaric acid 1% (v/v) HNO3
Value / Analyte(s):	500 µg/mL ea: Antimony

Second Source: Whenever possible, this solution was manufactured from a second set of concentrates in our manufacturing facility.

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ANALYTE	CERTIFIED VALUE	ANALYTE	CERTIFIED VALUE
Antimony, Sb	**500.0 ± 2.8 µg/mL**		

Density: 1.017 g/mL (measured at 20 ± 4 ºC)

Assay Information:

ANALYTE	METHOD	NIST SRM#	SRM LOT#
Sb	ICP Assay	3102a	140911

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$X_{CRM/RM} = \Sigma(w_i)(X_i)$

X_i = mean of Assay Method i with standard uncertainty $u_{char\ i}$
w_i = the weighting factors for each method calculated using the inverse square of the variance:
$w_i = (1/u_{char\ i})^2 / (\Sigma(1/(u_{char\ i})^2)$

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2
$u_{char} = [\Sigma((w_i)^2\ (u_{char\ i})^2)]^{½}$ where $u_{char\ i}$ are the errors from each characterization method
u_{bb} = bottle to bottle homogeneity standard uncertainty
u_{lts} = long term stability standard uncertainty (storage)
u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$X_{CRM/RM} = (X_a)(u_{char\ a})$

X_a = mean of Assay Method A with
$u_{char\ a}$ = the standard uncertainty of characterization Method A

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char\ a} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2
$u_{char\ a}$ = the errors from characterization
u_{bb} = bottle to bottle homogeneity standard uncertainty
u_{lts} = long term stability standard uncertainty (storage)
u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

N/A

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures Terms and Conditions of Sale, https://www.inorganicventures.com/terms-and-conditions-sale. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit **www.inorganicventures.com/TCT**

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

June 01, 2023

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- **June 01, 2028**

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity


- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Approved By:

Thomas Kozikowski
Manager, Quality Control



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director



Certificate of Analysis

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).

ACCREDITED
Testing Laboratory
Certificate 883.01

ACCREDITED
Reference Material Producer
Certificate 883.02

2.0 PRODUCT DESCRIPTION

Product Code:	Multi Analyte Custom Grade Solution
Catalog Number:	QCP-CICV-3
Lot Number:	V2-MEB749572
Matrix:	7% (v/v) HNO3
Value / Analyte(s):	500 µg/mL ea: Arsenic, Lead, Selenium, Thallium, 250 µg/mL ea: Cadmium

Second Source: Whenever possible, this solution was manufactured from a second set of concentrates in our manufacturing facility.

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ANALYTE	CERTIFIED VALUE	ANALYTE	CERTIFIED VALUE
Arsenic, As	**500.0 ± 3.1 µg/mL**	**Cadmium, Cd**	**250.1 ± 1.1 µg/mL**
Lead, Pb	**500.0 ± 2.3 µg/mL**	**Selenium, Se**	**500.0 ± 3.2 µg/mL**
Thallium, Tl	**500.0 ± 3.0 µg/mL**		

Density: 1.040 g/mL (measured at 20 ± 4 ºC)

Assay Information:

ANALYTE	METHOD	NIST SRM#	SRM LOT#
As	ICP Assay	3103a	100818
Cd	ICP Assay	3108	130116
Cd	EDTA	928	928
Pb	ICP Assay	3128	101026
Pb	EDTA	928	928
Se	ICP Assay	3149	100901
Tl	ICP Assay	3158	151215

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$X_{CRM/RM} = \Sigma(w_i)(X_i)$

X_i = mean of Assay Method i with standard uncertainty $u_{char\ i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance:

$w_i = (1/u_{char\ i})^2 / (\Sigma(1/(u_{char\ i})^2)$

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char} = [\Sigma((w_i)^2\,(u_{char\ i})^2)]^{½}$ where $u_{char\ i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$X_{CRM/RM} = (X_a)\,(u_{char\ a})$

X_a = mean of Assay Method A with

$u_{char\ a}$ = the standard uncertainty of characterization Method A

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char\ a} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char\ a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

N/A

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures Terms and Conditions of Sale, https://www.inorganicventures.com/terms-and-conditions-sale. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit **www.inorganicventures.com/TCT**

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

January 02, 2025

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- **January 02, 2030**

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS


Certificate Prepared By:

Justin Dirico
Stock Processing Supervisor



Certificate Approved By:

Jodie Wall
Stock VSM Coordinator



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director



Hydrogen Peroxide, 30%
CMOS
(Stabilized)

M-6170

R.Date . 8-05/20/2025

avantor™ | J.T.Baker®

Material No.: 2190-03
Batch No.: 24D1961001

Test	Specification	Result
Trace Impurities – Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Nickel (Ni)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Niobium (Nb)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Potassium (K)	≤ 600.0 ppb	176.7 ppb
Trace Impurities – Silicon (Si)	≤ 100.0 ppb	< 10.0 ppb
Trace Impurities – Silver (Ag)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Sodium (Na)	≤ 100.0 ppb	< 5.0 ppb
Trace Impurities – Strontium (Sr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Tantalum (Ta)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Thallium (Tl)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities – Tin (Sn)	190.0 – 500.0 ppb	272.3 ppb
Trace Impurities – Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Zinc (Zn)	≤ 50 ppb	< 1 ppb
Trace Impurities – Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count – 0.2 µm and greater	≤ 1175 par/ml	202 par/ml
Particle Count – 0.5 µm and greater	≤ 100 par/ml	33 par/ml

>>> Continued on page 3 >>>

Hydrogen Peroxide, 30%
CMOS
(Stabilized)

avantor™ | J.T.Baker®

Material No.: 2190-03
Batch No.: 24D1961001

Test	Specification	Result

For Microelectronic Use

Country of Origin: USA
Packaging Site: Paris Mfg Ctr & DC

Michelle Bales
Sr. Manager, Quality Assurance

Hydrogen Peroxide, 30%
CMOS
(Stabilized)

avantor™ | J.T.Baker®

Material No.: 2190-03
Batch No.: 24D1961001
Manufactured Date: 2024-04-17
Expiration Date: 2025-10-16
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (H_2O_2)	30.0 - 32.0 %	31.6 %
Color (APHA)	≤ 10	< 5
Free Acid (µeq/g)	≤ 0.2	0.1
Residue after Evaporation	≤ 10 ppm	2 ppm
Ammonium (NH_4)	≤ 3 ppm	< 3 ppm
Chloride (Cl)	≤ 0.2 ppm	< 0.2 ppm
Nitrate (NO_3)	≤ 2 ppm	< 2 ppm
Phosphate (PO_4)	≤ 1 ppm	1 ppm
Sulfate (SO_4)	≤ 3 ppm	< 3 ppm
Trace Impurities - Aluminum (Al)	≤ 70.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities - Barium (Ba)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities - Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities - Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities - Cadmium (Cd)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Calcium (Ca)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities - Chromium (Cr)	≤ 20.0 ppb	1.5 ppb
Trace Impurities - Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Gallium (Ga)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities - Germanium (Ge)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities - Gold (Au)	≤ 10.0 ppb	< 5.0 ppb
Heavy Metals (as Pb)	≤ 500.0 ppb	< 250.0 ppb
Trace Impurities - Iron (Fe)	≤ 50.0 ppb	4.6 ppb
Trace Impurities - Lead (Pb)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities - Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities - Magnesium (Mg)	≤ 10.0 ppb	< 1.0 ppb

>>> Continued on page 2 >>>

R -> 2/13/25

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

Part Number: 58112
Lot Number: 011525
Description: Magnesium (Mg)

M 6171

	Lot #	
Solvent:	24012496	Nitric Acid
2%	40.0 (mL)	Nitric Acid


Expiration Date: 011528
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 10000
NIST Test Number: 6UTB

5E-05	Balance Uncertainty	
0.100	Flask Uncertainty	

Weight shown below was diluted to (mL): 2000.07

Formulated By:	Benson Chan	011525
Reviewed By:	Pedro L. Rentas	011525

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Magnesium nitrate hexahydrate (Mg)	IN030	MGD052023A1	10000	99.999	0.10	8.51	234.9183	234.9195	**10000.1**	**20.0**	13446-18-9	NA	orl-rat 5440 mg/kg	3131a



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)

Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	T	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

Certified by:



* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

INORGANIC VENTURES
Refine your results. Redefine your industry.

Certificate of Analysis

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

M6174
R-> 5/5/24

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).

ACCREDITED Testing Laboratory Certificate 883.01
ACCREDITED Reference Material Producer Certificate 883.02

2.0 PRODUCT DESCRIPTION

Product Code:	Multi Analyte Custom Grade Solution
Catalog Number:	CLPP-CAL-1
Lot Number:	V2-MEB742428
Matrix:	5% (v/v) HNO3
Value / Analyte(s):	5 000 µg/mL ea:
	Calcium, Potassium,
	Magnesium, Sodium,
	2 000 µg/mL ea:
	Aluminum, Barium,
	1 000 µg/mL ea:
	Iron,
	500 µg/mL ea:
	Nickel, Vanadium,
	Zinc, Cobalt,
	Manganese,
	250 µg/mL ea:
	Silver, Copper,
	200 µg/mL ea:
	Chromium,
	50 µg/mL ea:
	Beryllium

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ANALYTE	CERTIFIED VALUE	ANALYTE	CERTIFIED VALUE
Aluminum, Al	2 000 ± 7 µg/mL	Barium, Ba	2 000 ± 9 µg/mL
Beryllium, Be	50.03 ± 0.30 µg/mL	Calcium, Ca	5 000 ± 16 µg/mL
Chromium, Cr	200.0 ± 1.1 µg/mL	Cobalt, Co	500.0 ± 2.3 µg/mL
Copper, Cu	250.0 ± 1.1 µg/mL	Iron, Fe	1 000 ± 4 µg/mL
Magnesium, Mg	5 000 ± 22 µg/mL	Manganese, Mn	499.9 ± 2.2 µg/mL
Nickel, Ni	500.1 ± 2.3 µg/mL	Potassium, K	5 000 ± 22 µg/mL
Silver, Ag	250.0 ± 1.1 µg/mL	Sodium, Na	5 000 ± 22 µg/mL
Vanadium, V	500.0 ± 2.2 µg/mL	Zinc, Zn	500.0 ± 2.2 µg/mL

Density: 1.117 g/mL (measured at 20 ± 4 °C)

Assay Information:

ANALYTE	METHOD	NIST SRM#	SRM LOT#
Ag	ICP Assay	3151	160729
Ag	Volhard	999c	999c
Al	ICP Assay	3101a	140903
Al	EDTA	928	928
Ba	ICP Assay	3104a	140909
Ba	Gravimetric		See Sec. 4.2
Be	ICP Assay	3105a	090514
Ca	ICP Assay	3109a	130213
Ca	EDTA	928	928
Co	ICP Assay	3113	190630
Co	EDTA	928	928
Cr	ICP Assay	3112a	170630
Cu	ICP Assay	3114	120618
Cu	EDTA	928	928
Fe	ICP Assay	3126a	140812
Fe	EDTA	928	928
K	ICP Assay	3141a	140813
K	Gravimetric		See Sec. 4.2
Mg	ICP Assay	3131a	140110
Mg	EDTA	928	928
Mn	ICP Assay	3132	050429
Mn	EDTA	928	928
Na	ICP Assay	3152a	200413
Na	Gravimetric		See Sec. 4.2
Ni	ICP Assay	3136	120619
Ni	EDTA	928	928
V	ICP Assay	3165	160906
V	EDTA	928	928
Zn	ICP Assay	3168a	120629
Zn	EDTA	928	928

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$X_{CRM/RM} = \Sigma(w_i)(X_i)$

X_i = mean of Assay Method i with standard uncertainty $u_{char\ i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance:

$w_i = (1/u_{char\ i})^2 / (\Sigma(1/u_{char\ i})^2)$

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\ (u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char} = [\Sigma((w_i)^2\ (u_{char\ i})^2)]^{½}$ where $u_{char\ i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$X_{CRM/RM} = (X_a)\ (u_{char\ a})$

X_a = mean of Assay Method A with

$u_{char\ a}$ = the standard uncertainty of characterization Method A

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\ (u^2_{char\ a} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char\ a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

N/A

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures Terms and Conditions of Sale, https://www.inorganicventures.com/terms-and-conditions-sale. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit **www.inorganicventures.com/TCT**

Note: This solution contains Silver (Ag), please refer to our Sample Preparation Guide for more information (https://www.inorganicventures.com/sample-preparation-guide/samples-containing-silver)

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

March 22, 2024

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- March 22, 2029

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Prepared By:

Uyen Truong
Custom Processing Supervisor



Certificate Approved By:

Jodie Wall
Stock VSM Coordinator



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director



Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com


CERTIFIED WEIGHT REPORT:

R->8/6/29

Part Number:	57038		**Lot #**		
Lot Number:	092724	**Solvent:**	24002546	Nitric Acid	
Description:	Strontium (Sr)				
		M6176 2%	40.0 (mL)	Nitric Acid	
Expiration Date:	092727				
Recommended Storage:	Ambient (20 °C)				
Nominal Concentration (µg/mL):	1000				
NIST Test Number:	6UTB	5E-05	Balance Uncertainty		
Weight shown below was diluted to (mL):	2000.07	0.100	Flask Uncertainty		

Formulated By:	Benson Chan	092724
Reviewed By:	Pedro L. Rentas	092724

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Strontium nitrate (Sr)	IN017	SRZ022018A1	1000	99.997	0.10	41.2	4.85470	4.85481	**1000.0**	**2.0**	10042-76-9	NA	orl-rat >2000mg/kg	3153a



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)

Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	T	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com


CERTIFIED WEIGHT REPORT:

R→ 8/6/25
M6177

		Lot #	Solvent:		
Part Number:	57042	MKBQ8597V	Ammonium hydroxide		
Lot Number:	080525	080525	ASTM Type 1 Water		
Description:	Molybdenum (Mo)	0.5%	10.0 (mL)	Ammonium hydroxide	
Expiration Date:	080528				
Recommended Storage:	Ambient (15 - 30 °C)				
Nominal Concentration (µg/mL):	1000				
NIST Test Number:	6UTB	5E-05	Balance Uncertainty		
Volume shown below was diluted to (mL):	1999.41	0.058	Flask Uncertainty		

Formulated By:	Benson Chan	080525
Reviewed By:	Pedro L. Rentas	080525

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Nominal Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Ammonium molybdate (Mo)	58142	112322	0.1000	200.0	0.084	1000	10001.4	**1000.5**	**2.2**	13106-76-8	5 mg(Mo)/m3	orl-rat 333 mg/kg	3134



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)

Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	T	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

Certified by:



* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware, and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated by an ISO17025 certified organization with weights traceable through NIST to the SI kilogram (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).
Rev 1.0, 2/25/2025

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

R>8/6/28
M6178

Part Number:	**57082**		**Lot #**		
Lot Number:	**031525**	**Solvent:**	24014457	Nitric Acid	
Description:	**Lead (Pb)**		2%	40.0 (mL)	Nitric Acid
Expiration Date:	031528				
Recommended Storage:	Ambient (20 °C)				
Nominal Concentration (µg/mL):	**1000**				
NIST Test Number:	6UTB	5E-05	Balance Uncertainty		
Weight shown below was diluted to (mL):	2000.07	0.10	Flask Uncertainty		


Giovanni Esposito

Formulated By:	Giovanni Esposito	031525

Pedro Rentas

Reviewed By:	Pedro L. Rentas	031525

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Lead(II) nitrate (Pb)	IN029	PBD122016A1	1000	99.999	0.10	62.5	3.20015	3.20067	**1000.2**	**2.0**	10099-74-8	0.05 mg/m3	intrvns-rat 93 mg/kg	3128



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Sc	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	T	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.



* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

INORGANIC VENTURES™
Refine your results. Redefine your industry.

Certificate of Analysis

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

M6179
R -> 6/2/24

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).

ACCREDITED Testing Laboratory Certificate 883.01
ACCREDITED Reference Material Producer Certificate 883.02

2.0 PRODUCT DESCRIPTION

Product Code:	Single Analyte Custom Grade Solution
Catalog Number:	CGY10
Lot Number:	V2-Y740548
Matrix:	2% (v/v) HNO3
Value / Analyte(s):	10 000 µg/mL ea: Yttrium
Starting Material:	Yttrium Oxide
Starting Material Lot#:	2661 and 06230520YL
Starting Material Purity:	99.9984%

3.0 CERTIFIED VALUES AND UNCERTAINTIES

Certified Value: 10000 ± 30 µg/mL

Density: 1.032 g/mL (measured at 20 ± 4 °C)

Assay Information:

Assay Method #1	**10011 ± 25 µg/mL** EDTA NIST SRM 928 Lot Number: 928
Assay Method #2	**9997 ± 50 µg/mL** ICP Assay NIST SRM 3167a Lot Number: 190730
Assay Method #3	**9984 ± 31 µg/mL** Calculated NIST SRM Lot Number: See Sec. 4.2

- The Calculated Value is a value calculated from the weight of a starting material that has been certified directly vs. a National Institute of Standards and Technology (NIST) SRM/RM. See Sec 4.2 for balance traceability.

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$X_{CRM/RM} = \Sigma(w_i)(X_i)$

X_i = mean of Assay Method i with standard uncertainty $u_{char\ i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance:

$w_i = (1/u_{char\ i})^2 / (\Sigma(1/(u_{char\ i})^2)$

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{1/2}$

k = coverage factor = 2

$u_{char} = [\Sigma((w_i)^2\,(u_{char\ i})^2)]^{1/2}$ where $u_{char\ i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$X_{CRM/RM} = (X_a)\,(u_{char\ a})$

X_a = mean of Assay Method A with

$u_{char\ a}$ = the standard uncertainty of characterization Method A

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char\ a} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{1/2}$

k = coverage factor = 2

$u_{char\ a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

CRM/RMs are tested for trace metallic impurities by Axial ICP-OES and ICP-MS. The result from the most sensitive method for each element, is reported below. Solutions tested by ICP-MS were analyzed in an ULPA-Filtered Clean Room. An ULPA-Filter is 99.9985% efficient for the removal of particles down to 0.3 µm.

M	Ag	<	0.004600	M	Eu		0.009037	M	Na		0.086360	M	Se	<	0.005200	M	Zn		0.030125
M	Al		0.014862	O	Fe		0.002410	M	Nb	<	0.000570	O	Si		0.024100	O	Zr	<	0.002600
M	As	<	0.003500	M	Ga	<	0.000570	M	Nd		0.000923	M	Sm		0.000461				
M	Au	<	0.001700	M	Gd	<	0.003500	M	Ni	<	0.005700	M	Sn	<	0.002300				
O	B		0.002209	M	Ge	<	0.005200	M	Os	<	0.001200	M	Sr	<	0.004600				
O	Ba	<	0.002500	M	Hf	<	0.000570	n	P	<		M	Ta	<	0.000570				
O	Be	<	0.001400	M	Hg	<	0.000570	M	Pb		0.005020	M	Tb		0.001044				
M	Bi	<	0.003500	M	Ho		0.009037	M	Pd	<	0.005100	M	Te	<	0.002300				
O	Ca		0.009841	M	In	<	0.002300	M	Pr	<	0.002300	M	Th	<	0.000570				
M	Cd	<	0.000570	M	Ir	<	0.000570	M	Pt	<	0.000570	M	Ti	<	0.003500				
M	Ce	<	0.002300	O	K		0.018677	M	Rb	<	0.000570	M	Tl	<	0.000570				
M	Co	<	0.000570	M	La		0.000461	M	Re	<	0.000570	M	Tm	<	0.003500				
M	Cr	<	0.004000	O	Li	<	0.009300	M	Rh	<	0.008000	M	U	<	0.000570				
M	Cs	<	0.000570	M	Lu		0.000582	M	Ru	<	0.000570	M	V		0.001265				
M	Cu		0.002610	O	Mg		0.001486	n	S	<		M	W	<	0.002300				
M	Dy		0.003815	M	Mn		0.000582	M	Sb		0.005422	s	Y	<					
M	Er		0.003615	M	Mo	<	0.005700	M	Sc	<	0.001200	M	Yb		0.001827				

M - Checked by ICP-MS O - Checked by ICP-OES i - Spectral Interference
n - Not Checked For s - Solution Standard Element

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures Terms and Conditions of Sale, https://www.inorganicventures.com/terms-and-conditions-sale. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit **www.inorganicventures.com/TCT**

Atomic Weight; Valence; Coordination Number; Chemical Form in Solution - 88.91 +3 6 Y(OH)(H2O)x+2

Chemical Compatibility -Soluble in HCl, H2SO4 and HNO3. Avoid HF, H3PO4 and neutral to basic media. Stable with most metals and inorganic anions forming an insoluble carbonate, oxide, oxalate, and fluoride. Avoid mixing with elements / solutions containing moderate amounts of fluoride.

Stability - 2-100 ppb levels stable for months in 1% HNO3 / LDPE container. 1-10,000 ppm solutions chemically stable for years in 2-5% HNO3 / LDPE container.

Y Containing Samples (Preparation and Solution) - Metal (Soluble in acids); Oxide (Dissolve by heating in H2O/ HNO3); Ores (Carbonate fusion in Pt0 followed by HCl dissolution); Organic Matrices (Dry ash and dissolve in 1:1 H2O / HCl or HNO3).

Atomic Spectroscopic Information (ICP-OES D.L.s are given as radial/axial view):

Technique/Line	**Estimated D.L.**	**Order**	**Interferences** (underlined indicates severe)
ICP-MS 89 amu	0.8 ppt	N/A	73Ge16O, 178Hf+2
ICP-OES 360.073 nm	0.005 / 0.000036 µg/mL	1	Ce, Th
ICP-OES 371.030 nm	0.004 / 0.00007 µg/mL	1	Ce
ICP-OES 377.433 nm	0.005 / 0.0009 µg/mL	1	Ta, Th

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

February 20, 2024

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- February 20, 2029

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Prepared By:

Uyen Truong
Custom Processing Supervisor



Certificate Approved By:

Muzzammil Khan
Stock Laboratory Supervisor



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director



INORGANIC VENTURES™
Refine your results. Redefine your industry.

Certificate of Analysis

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).

ACCREDITED Testing Laboratory Certificate 883.01
ACCREDITED Reference Material Producer Certificate 883.02

2.0 PRODUCT DESCRIPTION

Product Code:	Multi Analyte Custom Grade Solution
Catalog Number:	WW-LFS-1
Lot Number:	W2-MEB752149
Matrix:	5% (v/v) HNO3

M6180
R-7/22/23

Value / Analyte(s):	1 000 µg/mL ea: Potassium,	
	600 µg/mL ea: Phosphorus,	
	300 µg/mL ea: Sodium,	Iron,
	200 µg/mL ea: Magnesium, Cerium, Thallium,	Aluminum, Selenium,
	100 µg/mL ea: Lead,	Calcium,
	80 µg/mL ea: Arsenic,	
	70 µg/mL ea: Mercury,	
	50 µg/mL ea: Nickel,	
	40 µg/mL ea: Chromium,	
	30 µg/mL ea: Copper, Vanadium,	Boron,
	20 µg/mL ea: Zinc, Barium, Cadmium, Manganese,	Strontium, Beryllium, Cobalt, Lithium,
	7.5 µg/mL ea: Silver	

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ANALYTE	CERTIFIED VALUE	ANALYTE	CERTIFIED VALUE
Aluminum, Al	200.2 ± 0.7 µg/mL	Arsenic, As	80.1 ± 0.5 µg/mL
Barium, Ba	20.02 ± 0.12 µg/mL	Beryllium, Be	20.02 ± 0.11 µg/mL
Boron, B	30.02 ± 0.15 µg/mL	Cadmium, Cd	20.04 ± 0.09 µg/mL
Calcium, Ca	100.1 ± 0.3 µg/mL	Cerium, Ce	200.2 ± 1.0 µg/mL
Chromium, Cr	40.02 ± 0.26 µg/mL	Cobalt, Co	20.03 ± 0.09 µg/mL
Copper, Cu	30.03 ± 0.13 µg/mL	Iron, Fe	300.3 ± 1.3 µg/mL
Lead, Pb	100.1 ± 0.5 µg/mL	Lithium, Li	20.03 ± 0.09 µg/mL
Magnesium, Mg	200.2 ± 0.9 µg/mL	Manganese, Mn	19.99 ± 0.09 µg/mL
Mercury, Hg	70.0 ± 0.3 µg/mL	Nickel, Ni	50.05 ± 0.22 µg/mL
Phosphorus, P	600.5 ± 2.9 µg/mL	Potassium, K	1 001 ± 4 µg/mL
Selenium, Se	200.2 ± 1.1 µg/mL	Silver, Ag	7.52 ± 0.03 µg/mL
Sodium, Na	300.3 ± 1.3 µg/mL	Strontium, Sr	20.02 ± 0.10 µg/mL
Thallium, Tl	200.2 ± 1.0 µg/mL	Vanadium, V	30.02 ± 0.13 µg/mL
Zinc, Zn	20.05 ± 0.09 µg/mL		

Density: 1.037 g/mL (measured at 20 ± 5 °C)

Assay Information:

ANALYTE	METHOD	NIST SRM#	SRM LOT#
Ag	ICP Assay	3151	160729
Ag	Volhard	999c	999c
Al	ICP Assay	3101a	140903
Al	EDTA	928	928
As	ICP Assay	3103a	100818
B	ICP Assay	3107	190605
B	Calculated		See Sec. 4.2
Ba	ICP Assay	3104a	140909
Ba	Gravimetric		See Sec. 4.2
Be	ICP Assay	3105a	090514
Ca	ICP Assay	3109a	130213
Ca	EDTA	928	928
Cd	ICP Assay	3108	130116
Cd	EDTA	928	928
Ce	ICP Assay	3110	160830
Ce	EDTA	928	928
Co	ICP Assay	3113	190630
Co	EDTA	928	928
Cr	ICP Assay	3112a	170630
Cu	ICP Assay	3114	120618
Cu	EDTA	928	928
Fe	ICP Assay	3126a	140812
Fe	EDTA	928	928
Hg	ICP Assay	3133	160921
Hg	EDTA	928	928
K	ICP Assay	3141a	140813
K	Gravimetric		See Sec. 4.2
Li	ICP Assay	3129a	100714
Li	Gravimetric		See Sec. 4.2
Mg	ICP Assay	3131a	140110
Mg	EDTA	928	928
Mn	ICP Assay	3132	050429
Mn	EDTA	928	928
Na	ICP Assay	3152a	200413
Na	Gravimetric		See Sec. 4.2
Ni	ICP Assay	3136	120619
Ni	EDTA	928	928
P	ICP Assay	3139a	060717
P	Acidimetric	84L	84L
Pb	ICP Assay	3128	101026
Pb	EDTA	928	928
Se	ICP Assay	3149	100901
Sr	EDTA	928	928
Sr	ICP Assay	Traceable to 3153a	K2-SR650985
Tl	ICP Assay	3158	151215
V	ICP Assay	3165	160906
V	EDTA	928	928
Zn	ICP Assay	3168a	120629
Zn	EDTA	928	928

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$X_{CRM/RM} = \Sigma(w_i)(X_i)$

X_i = mean of Assay Method i with standard uncertainty $u_{char\ i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance:

$w_i = (1/u_{char\ i})^2 / (\Sigma(1/(u_{char\ i})^2)$

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char} = [\Sigma((w_i)^2\,(u_{char\ i})^2)]^{½}$ where $u_{char\ i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$X_{CRM/RM} = (X_a)\,(u_{char\ a})$

X_a = mean of Assay Method A with

$u_{char\ a}$ = the standard uncertainty of characterization Method A

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char\ a} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char\ a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

N/A

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures Terms and Conditions of Sale, https://www.inorganicventures.com/terms-and-conditions-sale. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 25° C to minimize the effects of transpiration. Use at 20° ± 5° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit **www.inorganicventures.com/TCT**

Note: This solution contains Silver (Ag), please refer to our Sample Preparation Guide for more information (https://www.inorganicventures.com/sample-preparation-guide/samples-containing-silver)

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

April 08, 2025

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- April 08, 2029

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Prepared By:

Justin Dirico
Stock Processing Supervisor



Certificate Approved By:

Jodie Wall
Stock VSM Coordinator



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director



INORGANIC VENTURES
Refine your results. Redefine your industry.

Certificate of Analysis

300 Technology Drive
Christiansburg, VA 24073 USA
inorganicventures.com

P: 800-669-6799/540-585-3030
F: 540-585-3012
info@inorganicventures.com

M6181
R -> 7/22/28

1.0 ACCREDITATION / REGISTRATION

INORGANIC VENTURES is accredited to ISO 17034, "General Requirements for the Competence of Reference Material Producers" and ISO/IEC 17025, "General Requirements for the Competence of Testing and Calibration Laboratories". Inorganic Ventures is also an ISO 9001 registered manufacturer (QSR Certificate Number QSR-1034).

ACCREDITED Testing Laboratory Certificate 883.01
ACCREDITED Reference Material Producer Certificate 883.02

2.0 PRODUCT DESCRIPTION

Product Code:	Multi Analyte Custom Grade Solution
Catalog Number:	WW-LFS-2
Lot Number:	V2-MEB743480
Matrix:	5% (v/v) HNO3 tr. HF
Value / Analyte(s):	200 µg/mL ea: Silica, 80 µg/mL ea: Antimony, 70 µg/mL ea: Tin, 40 µg/mL ea: Molybdenum, 20 µg/mL ea: Titanium

3.0 CERTIFIED VALUES AND UNCERTAINTIES

ANALYTE	CERTIFIED VALUE	ANALYTE	CERTIFIED VALUE
Antimony, Sb	80.1 ± 0.5 µg/mL	Molybdenum, Mo	40.05 ± 0.22 µg/mL
Silica, SiO2	200.3 ± 1.4 µg/mL	Tin, Sn	70.1 ± 0.4 µg/mL
Titanium, Ti	20.03 ± 0.12 µg/mL		

Density: 1.025 g/mL (measured at 20 ± 4 °C)

Assay Information:

ANALYTE	METHOD	NIST SRM#	SRM LOT#
Mo	ICP Assay	3134	130418
Sb	ICP Assay	3102a	140911
SiO2	Calculated		See Sec. 4.2
Sn	ICP Assay	3161a	140917
Ti	ICP Assay	traceable to 3162a	T2-TI725816

The following equations are used in the calculation of the certified value and the uncertainty. Reported uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k = 2.

Characterization of CRM/RM by Two or More Methods

Certified Value, $X_{CRM/RM}$, where two or more methods of characterization are used is the weighted mean of the results:

$X_{CRM/RM} = \Sigma(w_i)(X_i)$

X_i = mean of Assay Method I with standard uncertainty $u_{char\ i}$

w_i = the weighting factors for each method calculated using the inverse square of the variance:

$w_i = (1/u_{char\ i})^2 / (\Sigma\{1/(u_{char\ i})^2\})$

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char} = [\Sigma((w_i)^2\,(u_{char\ i})^2)]^{½}$ where $u_{char\ i}$ are the errors from each characterization method

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

Characterization of CRM/RM by One Method

Certified Value, $X_{CRM/RM}$, where one method of characterization is used is the mean of individual results:

$X_{CRM/RM} = (X_a)\,(u_{char\ a})$

X_a = mean of Assay Method A with

$u_{char\ a}$ = the standard uncertainty of characterization Method A

CRM/RM Expanded Uncertainty (±) = $U_{CRM/RM} = k\,(u^2_{char\ a} + u^2_{bb} + u^2_{lts} + u^2_{ts})^{½}$

k = coverage factor = 2

$u_{char\ a}$ = the errors from characterization

u_{bb} = bottle to bottle homogeneity standard uncertainty

u_{lts} = long term stability standard uncertainty (storage)

u_{ts} = transport stability standard uncertainty

4.0 TRACEABILITY TO NIST

- This product is traceable to NIST via an unbroken chain of comparisons. The uncertainties for each certified value are reported, taking into account the SRM/RM uncertainty error and the measurement, weighing and volume dilution errors. In rare cases where no NIST SRM/RM are available, the term 'in-house std.' is specified.

4.1 Thermometer Calibration

- All thermometers are NIST traceable through thermometers that are calibrated by an accredited calibration laboratory.

4.2 Balance Calibration

- All analytical balances are calibrated by an accredited calibration laboratory and procedure. The weights used for testing are annually compared to master weights and are traceable to NIST.

4.3 Glassware Calibration

- An in-house procedure is used to calibrate all Class A glassware used in the manufacturing and quality control of CRM/RMs.

5.0 TRACE METALLIC IMPURITIES (TMI) DETERMINED BY ICP-MS AND ICP-OES (µg/mL)

N/A

6.0 INTENDED USE

6.1 This standard is intended for the calibration of analytical instruments and validation of analytical methods as appropriate. This CRM may be used in connection with EPA Methods 6010, 6020 (all versions), Standard Methods 3120 B and USP <232> / ICH Q3D.

6.2 For products attaining traceability through Inorganic Ventures' Primary Certified Reference Materials (PCRM™) see the Limited License to Use PCRM™ in the Inorganic Ventures Terms and Conditions of Sale, https://www.inorganicventures.com/terms-and-conditions-sale. The Terms and Conditions contain information on the use of materials traceable to PCRM™ certified reference materials. This Limited License agreement is especially pertinent for laboratories accredited under ISO:17034.

7.0 INSTRUCTIONS FOR THE CORRECT USE OF THIS REFERENCE MATERIAL

7.1 Storage and Handling Recommendations

- Store between approximately 4° - 30° C while in sealed TCT bag.

- While stored in the sealed TCT bag, transpiration of this CRM/RM is negligible. After opening the sealed TCT bag transpiration of the CRM/RM will occur, resulting in a gradual increase in the analyte concentration(s). It is the responsibility of the user to account for this effect. When the bottle is weighed both before and after being placed in storage, the mass difference observed will be a measure of transpiration mass loss.

- After opening the sealed TCT bag, keep cap tightly sealed when not in use and store between 4° - 24° C to minimize the effects of transpiration. Use at 20° ± 4° C to minimize volumetric dilution error when using the reported density. Do not pipette from the container. Do not return removed aliquots to container.

- For more information, visit **www.inorganicventures.com/TCT**

HF Note: This standard should not be prepared or stored in glass.

8.0 HAZARDOUS INFORMATION

- Please refer to the Safety Data Sheet for information regarding this CRM/RM.

9.0 HOMOGENEITY

- This solution was mixed according to an in-house procedure and is guaranteed to be homogeneous. Homogeneity data indicate that the end user should take a minimum sample size of 0.2 mL to assure homogeneity.

10.0 QUALITY STANDARD DOCUMENTATION

10.1 ISO 9001 Quality Management System Registration

- QSR Certificate Number QSR-1034

10.2 ISO/IEC 17025 "General Requirements for the Competence of Testing and Calibration Laboratories"

- Chemical Testing - Accredited / A2LA Certificate Number 883.01

10.3 ISO 17034 "General Requirements for the Competence of Reference Material Producers"

- Reference Material Producer - Accredited / A2LA Certificate Number 883.02

Inorganic Ventures, 300 Technology Drive, Christiansburg, Va. 24073, USA; Telephone: 800.669.6799; 540.585.3030, Fax: 540.585.3012; inorganicventures.com; info@inorganicventures.com

11.0 CERTIFICATION, LOT EXPIRATION AND PERIOD OF VALIDITY

11.1 Certification Issue Date

May 07, 2024

- The certification is valid within the measurement uncertainty specified provided the CRM/RM is stored and handled in accordance with instructions given in Sec 7.1. This certification is nullified if instructions in Sec 7.1 are not followed or if the CRM/RM is damaged, contaminated, or otherwise modified.

11.2 Lot Expiration Date

- May 07, 2029

- The date after which this CRM/RM should not be used.

- The lot expiration date reflects the period of time that the stability of a CRM/RM can be supported by long term stability studies conducted on properly stored and handled CRM/RMs. Lot expiration is limited primarily by transpiration (loss of water from the solution) and infrequently by chemical stability.

11.3 Period of Validity

- Sealed TCT Bag Open Date: _____

- This CRM/RM should not be used longer than one year (or six months in the case of a 30 mL bottle) from the date of opening the aluminized bag or after the date given in Sec. 11.2, whichever comes first. This is contingent upon the CRM/RM being stored and handled in accordance with the instructions given in Sec. 7.1.

12.0 NAMES AND SIGNATURES OF CERTIFYING OFFICERS

Certificate Prepared By:

Justin Dirico
Stock Processing Supervisor



Certificate Approved By:

Thomas Kozikowski
Stock VS Manager



Certifying Officer:

Paul Gaines
Chairman / Senior Technical Director



R→03/06/24

QATS LABORATORY INORGANIC REFERENCE MATERIAL INTERFERENCE CHECK SAMPLE SET FOR ICP-AES (ICSA WITH ICSB)

NOTE: These instructions are for advisory purposes only. If any apparent conflict exists between these instructions and the analytical protocol or your contract, disregard these instructions.

APPLICATION: For use with the CLP SFAM01.0 SOW and revisions.

CAUTION: Read instructions carefully before opening bottle(s) and proceeding with the analyses.

M6182

Contains Heavy Metals
HAZARDOUS MATERIAL

Safety Data Sheets
Available Upon Request

(A) *SAMPLE DESCRIPTION*

Enclosed is a set of one (1) or more bottles of Aqueous Reference Material, each composed of metals at various concentrations and prepared with nitrate salts and oxy-acids of the respective elements in a 5% nitric acid matrix. **For the reference material source in reporting ICSA and ICSAB mixture use "USEPA". For the reference material lot number for the ICSA use "ICSA-1211" and for the ICSAB mixture use "ICSA-1211+ICSB-0710".**

CAUTION: The bottle(s) should be protected from light during storage to ensure the stability of silver which is contained in the ICSB solution. The bottle(s) should be stored at room temperature. **Do not allow the solution(s) to freeze.**

(B) *BREAKAGE OR MISSING ITEMS*

Check the contents of the shipment carefully for any broken, leaking, or missing items. Check that the seal is intact on each bottle. Refer to the enclosed chain of custody record. Report any problems to Mr. Keith Strout, APTIM Federal Services, LLC, at (702) 895-8722. If requested, return the chain-of-custody record with appropriate annotations and signatures to the address provided below.

QUALITY ASSURANCE TECHNICAL SUPPORT LABORATORY
APTIM Federal Services, LLC
2700 Chandler Avenue - Building C
Las Vegas, NV 89120

(C) *ANALYSIS OF SAMPLES*

The interference check sample set is to be used to verify inter-element and background correction factors of inductively-coupled plasma (ICP) spectrometers. This reference material set consists of two (2) concentrated solutions. The ICSA solution contains the four (4) interferent elements: Al, Ca, Fe, and Mg. The ICSB solution contains the analytes: Ag, As, Sb, Ba, Be,

The Quality Assurance Technical Support (QATS) contract is operated by APTIM Federal Services, LLC.

Cd, Co, Cr, Cu, Mn, Ni, Pb, Tl, Se, V, and Zn. This instruction sheet provides the nominal values for ICP-AES Part A and Part B target analytes when diluted as directed.

Using Class "A" glassware, preparation and analysis must be performed according to the following instructions:

ICSA-1211, Interferents: Pipet 10 mL of the ICSA solution into a 100 mL volumetric flask and dilute to volume with 2% v/v HNO_3. Analyze this ICSA solution by ICP-AES.

ICSB-0710, Analytes, mixed with ICSA-1211, Interferents: Pipet 10 mL of the ICSA solution and 10 mL of the ICSB solution into a 100 mL volumetric flask and dilute to volume with 2% v/v HNO_3. Analyze this ICSAB solution by ICP-AES.

(D) *"CERTIFIED VALUE" CONCENTRATIONS OF QATS ICP-AES ICS SOLUTION(S)*

The "Certified Value" concentrations of the elements, listed in Table 1 below, were derived from statistically pooled analysis results from the following sources, if available: QATS Laboratory, CLP laboratories, Quarterly Blind (QB)/Proficiency Testing (PT) events, CLP pre-award events, and external referee laboratories.

Table 1. "CERTIFIED VALUES" FOR INTERFERENCE CHECK SAMPLE ICP-AES ICSA-1211, AND ICSA-1211 MIXED WITH ICSB-0710

Element	CRQL	Part A (µg/L)	Low Limit (µg/L)	High Limit (µg/L)	Part A +Part B (µg/L)	Low Limit (µg/L)	High Limit (µg/L)
Al	200	255000	216000	294000	247000	209000	285000
Sb	60	(0.0)	-60.0	60.0	618	525	711
As	10	(0.0)	-10.0	10.0	104	88.4	120
Ba	200	(6.0)	-194	206	(537)	337	737
Be	5.0	(0.0)	-5.0	5.0	495	420	570
Cd	5.0	(1.0)	-4.0	6.0	972	826	1120
Ca	5000	245000	208000	282000	235000	199000	271000
Cr	10	(52.0)	42.0	62.0	542	460	624
Co	50	(0.0)	-50.0	50.0	476	404	548
Cu	25	(2.0)	-23.0	27.0	511	434	588
Fe	100	101000	85600	116500	99300	84400	114500
Pb	10	(0.0)	-10.0	10.0	(49.0)	39.0	59.0
Mg	5000	255000	216000	294000	248000	210000	286000
Mn	15	(7.0)	-8.0	22.0	507	430	584
Ni	40	(2.0)	-38.0	42.0	954	810	1100
Se	35	(0.0)	-35.0	35.0	(46.0)	11.0	81.0
Ag	10	(0.0)	-10.0	10.0	201	170	232
Tl	25	(0.0)	-25.0	25.0	(108)	83.0	133
V	50	(0.0)	-50.0	50.0	491	417	565
Zn	60	(0.0)	-60.0	60.0	952	809	1095

The acceptance ranges for all analytes in parentheses in the above table were determined using the listed certified value ± 1 times the associated CLP SOW CRQL. The acceptance ranges for all other analytes were determined using the certified value ± 15 percent of the listed certified value.

R→03/06/24

QATS LABORATORY INORGANIC REFERENCE MATERIAL INTERFERENCE CHECK SAMPLE SET FOR ICP-AES (ICSA WITH ICSB)

NOTE: These instructions are for advisory purposes only. If any apparent conflict exists between these instructions and the analytical protocol or your contract, disregard these instructions.

APPLICATION: For use with the CLP SFAM01.0 SOW and revisions.

CAUTION: Read instructions carefully before opening bottle(s) and proceeding with the analyses.

Contains Heavy Metals
HAZARDOUS MATERIAL

Safety Data Sheets
Available Upon Request

M6184

(A) *SAMPLE DESCRIPTION*

Enclosed is a set of one (1) or more bottles of Aqueous Reference Material, each composed of metals at various concentrations and prepared with nitrate salts and oxy-acids of the respective elements in a 5% nitric acid matrix. **For the reference material source in reporting ICSA and ICSAB mixture use "USEPA". For the reference material lot number for the ICSA use "ICSA-1211" and for the ICSAB mixture use "ICSA-1211+ICSB-0710".**

CAUTION: The bottle(s) should be protected from light during storage to ensure the stability of silver which is contained in the ICSB solution. The bottle(s) should be stored at room temperature. **Do not allow the solution(s) to freeze.**

(B) *BREAKAGE OR MISSING ITEMS*

Check the contents of the shipment carefully for any broken, leaking, or missing items. Check that the seal is intact on each bottle. Refer to the enclosed chain of custody record. Report any problems to Mr. Keith Strout, APTIM Federal Services, LLC, at (702) 895-8722. If requested, return the chain-of-custody record with appropriate annotations and signatures to the address provided below.

QUALITY ASSURANCE TECHNICAL SUPPORT LABORATORY
APTIM Federal Services, LLC
2700 Chandler Avenue - Building C
Las Vegas, NV 89120

(C) *ANALYSIS OF SAMPLES*

The interference check sample set is to be used to verify inter-element and background correction factors of inductively-coupled plasma (ICP) spectrometers. This reference material set consists of two (2) concentrated solutions. The ICSA solution contains the four (4) interferent elements: Al, Ca, Fe, and Mg. The ICSB solution contains the analytes: Ag, As, Sb, Ba, Be,

The Quality Assurance Technical Support (QATS) contract is operated by APTIM Federal Services, LLC.

Cd, Co, Cr, Cu, Mn, Ni, Pb, Tl, Se, V, and Zn. This instruction sheet provides the nominal values for ICP-AES Part A and Part B target analytes when diluted as directed.

Using Class "A" glassware, preparation and analysis must be performed according to the following instructions:

ICSA-1211, Interferents: Pipet 10 mL of the ICSA solution into a 100 mL volumetric flask and dilute to volume with 2% v/v HNO_3. Analyze this ICSA solution by ICP-AES.

ICSB-0710, Analytes, mixed with ICSA-1211, Interferents: Pipet 10 mL of the ICSA solution and 10 mL of the ICSB solution into a 100 mL volumetric flask and dilute to volume with 2% v/v HNO_3. Analyze this ICSAB solution by ICP-AES.

(D) *"CERTIFIED VALUE" CONCENTRATIONS OF QATS ICP-AES ICS SOLUTION(S)*

The "Certified Value" concentrations of the elements, listed in Table 1 below, were derived from statistically pooled analysis results from the following sources, if available: QATS Laboratory, CLP laboratories, Quarterly Blind (QB)/Proficiency Testing (PT) events, CLP pre-award events, and external referee laboratories.

Table 1. "CERTIFIED VALUES" FOR INTERFERENCE CHECK SAMPLE ICP-AES ICSA-1211, AND ICSA-1211 MIXED WITH ICSB-0710

Element	CRQL	Part A (µg/L)	Low Limit (µg/L)	High Limit (µg/L)	Part A +Part B (µg/L)	Low Limit (µg/L)	High Limit (µg/L)
Al	200	255000	216000	294000	247000	209000	285000
Sb	60	(0.0)	-60.0	60.0	618	525	711
As	10	(0.0)	-10.0	10.0	104	88.4	120
Ba	200	(6.0)	-194	206	(537)	337	737
Be	5.0	(0.0)	-5.0	5.0	495	420	570
Cd	5.0	(1.0)	-4.0	6.0	972	826	1120
Ca	5000	245000	208000	282000	235000	199000	271000
Cr	10	(52.0)	42.0	62.0	542	460	624
Co	50	(0.0)	-50.0	50.0	476	404	548
Cu	25	(2.0)	-23.0	27.0	511	434	588
Fe	100	101000	85600	116500	99300	84400	114500
Pb	10	(0.0)	-10.0	10.0	(49.0)	39.0	59.0
Mg	5000	255000	216000	294000	248000	210000	286000
Mn	15	(7.0)	-8.0	22.0	507	430	584
Ni	40	(2.0)	-38.0	42.0	954	810	1100
Se	35	(0.0)	-35.0	35.0	(46.0)	11.0	81.0
Ag	10	(0.0)	-10.0	10.0	201	170	232
Tl	25	(0.0)	-25.0	25.0	(108)	83.0	133
V	50	(0.0)	-50.0	50.0	491	417	565
Zn	60	(0.0)	-60.0	60.0	952	809	1095

The acceptance ranges for all analytes in parentheses in the above table were determined using the listed certified value ± 1 times the associated CLP SOW CRQL. The acceptance ranges for all other analytes were determined using the certified value ± 15 percent of the listed certified value.

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

R 8/05/24

M6027


ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

		Lot #	Solvent:	
Part Number:	**57028**			
Lot Number:	**062024**	24002546	Nitric Acid	
Description:	**Nickel (Ni)**			
		2.0%	40.0 (mL)	Nitric Acid
Expiration Date:	062027			
Recommended Storage:	Ambient (20 °C)			
Nominal Concentration (µg/mL):	**1000**			
NIST Test Number:	6UTB	5E-05	Balance Uncertainty	
Volume shown below was diluted to (mL):	2000.07	0.100	Flask Uncertainty	

Formulated **By:**	Benson Chan	062024
Reviewed By:	Pedro L. Rentas	062024

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Nominal Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Nickel(II) nitrate hexahydrate (Ni)	58128	062023	0.1000	200.0	0.084	1000	10000.4	**1000.0**	2.2	13478-00-7	1 mg/m3	orl-rat 1620 mg/kg	3136



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	T	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

M6030

R: 8/5/24


ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

Part Number:	**57047**		**Lot #**		
Lot Number:	**122823**	**Solvent:**	24002546	Nitric Acid	
Description:	**Silver (Ag)**	2%	80.0 (mL)	Nitric Acid	
Expiration Date:	122826				
Recommended Storage:	Ambient (20 °C)				
Nominal Concentration (µg/mL):	**1000**				
NIST Test Number:	6UTB	5E-05	Balance Uncertainty		
Weight shown below was diluted to (mL):	4000.30	0.058	Flask Uncertainty		

Formulated By:	Benson Chan	122823
Reviewed By:	Pedro L. Rentas	122823

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Silver nitrate (Ag)	IN035	J0612AGA1	1000.0	99.9999	0.10	63.7	6.27992	6.27998	**1000.0**	**2.0**	7761-88-8	10 ug/m3	NA	3151



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	T	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T)= Target analyte

Physical Characterization:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

Certified by:

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

R: 10/18/24

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

CERTIFIED WEIGHT REPORT:

MG146

		Lot #	Solvent:	
Part Number:	**57051**	24002546	Nitric Acid	
Lot Number:	**071724**			
Description:	**Antimony (Sb)**			
		2.0%	40.0 (mL)	Nitric Acid
Expiration Date:	071727			
Recommended Storage:	Ambient (20 °C)			
Nominal Concentration (µg/mL):	**1000**			
NIST Test Number:	6UTB	5E-05	Balance Uncertainty	
Volume shown below was diluted to (mL):	2000.26	0.058	Flask Uncertainty	


Giovanni Esposito

Formulated By:	Giovanni Esposito	071724

Pedro Rentas

Reviewed By:	Pedro L. Rentas	071724

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Nominal Conc. (µg/mL)	Initial Conc. (µg/mL)	**Final Conc. (µg/mL)**	Expanded Uncertainty +/- (µg/mL)	**SDS Information** (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Antimony (Sb)	58151	060324	0.1000	200.0	0.084	1000	10001.4	**1000.0**	**2.2**	7440-36-0	0.5 mg/m3	orl-rat 7000 mg/kg	3102a



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	T	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

Certified by:



* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

R ! 8/5/24

M6023

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com


CERTIFIED WEIGHT REPORT:

Part Number:	**57081**		**Lot #**		
Lot Number:	**062724**	**Solvent:**	24002546	Nitric Acid	
Description:	**Thallium (Tl)**				
		2%	40.0 (mL)	Nitric Acid	
Expiration Date:	062727				
Recommended Storage:	Ambient (20 °C)				
Nominal Concentration (µg/mL):	**1000**				
NIST Test Number:	6UTB	5E-05	Balance Uncertainty		
Weight shown below was diluted to (mL):	2000.1	0.10	Flask Uncertainty		

Aleah O'Brady

Formulated By:	Aleah O'Brady	062724
Reviewed By:	Pedro L. Rentas	062724

Compound	RM#	Lot Number	Nominal Conc. (µg/mL)	Purity (%)	Uncertainty Purity (%)	Assay (%)	Target Weight (g)	Actual Weight (g)	Actual Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Thallium nitrate (Tl)	IN037	BCCF4399	1000	99.999	0.10	77.0	2.5975	2.5977	**1000.1**	**2.0**	10102-45-1	0.1 mg/m3	orl-mus 15mg/kg	3158



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)																	
Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	T	V	<0.02
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

Certified by:

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

M6021

ANAB ISO 17034 Accredited
AR-1539 Certificate Number
https://Absolutestandards.com

Bc. 8/5/24

CERTIFIED WEIGHT REPORT:

		Lot #	Solvent:	
Part Number:	**57023**			
Lot Number:	**062424**	24002546	Nitric Acid	
Description:	**Vanadium (V)**			
		2.0%	40.0 (mL)	Nitric Acid
Expiration Date:	062427			
Recommended Storage:	Ambient (20 °C)			
Nominal Concentration (µg/mL):	**1000**			
NIST Test Number:	6UTB	5E-05	Balance Uncertainty	
Volume shown below was diluted to (mL):	2000.3	0.06	Flask Uncertainty	


Aleah O'Brady

Formulated By:	Aleah O'Brady	062424

Pedro Rentas

Reviewed By:	Pedro L. Rentas	062424

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Uncertainty Pipette (mL)	Nominal Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty +/- (µg/mL)	SDS Information (Solvent Safety Info. On Attached pg.) CAS#	OSHA PEL (TWA)	LD50	NIST SRM
1. Ammonium metavanadate (V)	58123	021224	0.1000	200.0	0.084	1000	10000.3	**1000.0**	**2.2**	7803-55-6	0.05 mg/m3	orl-rat 58.1mg/kg	3165



Instrumental Analysis by Inductively Coupled Plasma Mass Spectrometry (ICP-MS):

Trace Metals Verification by ICP-MS (µg/mL)

Al	<0.02	Cd	<0.02	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	<0.02	Pr	<0.02	Se	<0.2	Tb	<0.02	W	<0.02
Sb	<0.02	Ca	<0.2	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	<0.2	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	<0.01	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	<0.02	V	T
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	<0.02	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	<0.01	Cr	<0.02	Ga	<0.02	Fe	<0.2	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	<0.02	Ge	<0.02	La	<0.02	Mo	<0.02	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	<0.02
B	<0.02	Cu	<0.02	Au	<0.02	Pb	<0.02	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	<0.02	Zr	<0.02

(T) = Target analyte

Physical Characterization:

Certified by:

Homogeneity: No heterogeneity was observed in the preparation of this standard.

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Purified acids, 18.2 megohm deionized water, calibrated Class A glassware and the highest purity raw materials are used in the preparation of all standards.
* All standard containers are meticulously cleaned prior to use.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certifed (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, D.C. (1994).



SHIPPING DOCUMENTS

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 Fax: (908) 788-9222
www.chemtech.net

CHAIN OF CUSTODY RECORD

Alliance Project Number: Q2829

COC Number:

CLIENT INFORMATION

COMPANY: Viva nova
ADDRESS:
CITY: STATE: ZIP:
ATTENTION: Goyal Sanjeev
PHONE: FAX:

PROJECT INFORMATION

PROJECT NAME:
PROJECT #: LOCATION:
PROJECT MANAGER:
E-MAIL:
PHONE: FAX:

BILLING INFORMATION

BILL TO: PO#
ADDRESS:
CITY: STATE: ZIP:
ATTENTION: PHONE:

DATA TURNAROUND INFORMATION

FAX: STD DAYS*
HARD COPY: DAYS*
EDD DAYS*
* TO BE APPROVED BY ALLIANCE
STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

- ☐ RESEULTS ONLY
- ☐ RESULTS + QC
- ☐ New Jersey REDUCED
- ☐ New Jersey CLP
- ☐ EDD Format
- ☐ USEPA CLP
- ☐ New York State ASP "B"
- ☐ New York State ASP "A"
- ☐ Other

ANALYSIS

	1	2	3	4	5	6	7	8	9
Silica									

PRESERVATIVES

	1	2	3	4	5	6	7	8	9

COMMENTS

<-- Specify Preservatives
A-HCl B-HNO3
C-H2SO4 D-NaOH
E-ICE F-Other

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP	SAMPLE TYPE GRAB	SAMPLE COLLECTION DATE	SAMPLE COLLECTION TIME	# of Bottles
1.	Silica						
2.							
3.							
4.							
5.							
6.							
7.							
8.							
9.							
10.							

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER	DATE/TIME	RECEIVED BY
1.		1.
RELINQUISHED BY 2.	DATE/TIME	RECEIVED BY 2.
RELINQUISHED BY 3.	DATE/TIME	RECEIVED FOR LAB BY 3.

Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp_____
☐ Ice in Cooler?:_____
MeOH extraction requires an additional 4oz. Jar for percent solid
Comments:

Page_____of_____

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight
ALLIANCE: ☐ Picked Up ☐ Overnight

Shipment Complete ☐ YES ☐ NO

WHITE - ALLIANCE COPYFOR RETURN TO CLIENT **YELLOW** - ALLIANCE COPY **PINK** - SAMPLER COPY

284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312