



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Viva Nova Inc

SDG No.: Q2829

Contract: VIVA01

Lab Code: ACE

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Silicon	357	400	89	65 - 135	P	08/13/2025	12:37	LB136801
CRI01	Silicon	371	400	93	65 - 135	P	08/26/2025	11:29	LB136973