

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Viva Nova Inc

SDG No.: Q2829

Contract: VIVA01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169215BS Silicon	mg/Kg	46.7	46.7		100	80 - 120	P