



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Portal Partners Tri-Venture

Project Name: Amtrak Sawtooth Bridges 2025

Project # N/A

Order ID # Q2832

Test Name: SVOC-TCL BNA -20

A. Number of Samples and Date of Receipt:

5 Solid samples were received on 08/12/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-TCL BNA -20. This data package contains results for SVOC-TCL BNA -20.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_P using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA. The analysis of SVOC-TCL BNA -20 was based on method 8270E and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The MS {Q2820-17MS} with File ID: BP025415.D recoveries met the requirements for all compounds except for 3,3-Dichlorobenzidine[50%], 3-Nitroaniline[60%] and 4-Chloroaniline[47%], these compounds did not meet the NJDKQP criteria but met the in-house criteria due to matrix interference.

The MSD {Q2820-17MSD} with File ID: BP025416.D recoveries met the requirements for all compounds except for 3,3-Dichlorobenzidine[55%], 3-Nitroaniline[65%] and 4-Chloroaniline[50%], these compounds did not meet the NJDKQP criteria but met the in-house criteria due to matrix interference.

The RPD were met for all analysis.

The Blank Spike for {PB169228BS} with File ID: BP025407.D met requirements for all compounds except for 3,3-Dichlorobenzidine[65%], 4-Chloroaniline[65%], this compound did not meet the NJDKQP criteria but met the in-house criteria.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

The Blank analysis did not indicate the presence of lab contamination.
The Initial Calibration met the requirements.

The Continuous Calibration File ID BP025455.D met the requirements except for Benzaldehyde and Hexachlorocyclopentadiene. Associated samples does not have hit for these compounds, therefor no further corrective action was taken.

The Continuous Calibration File ID BP025528.D met the requirements except for Benzaldehyde. Associated samples does not have hit for these compounds, therefor no further corrective action was taken.

The Continuous Calibration File ID BP025544.D met the requirements except for 2,4-Dinitrophenol and Benzaldehyde. Associated samples does not have hit for these compounds, therefor no further corrective action was taken.

The Tuning criteria met requirements.

Samples TG-S01, TG-S02, TG-S03, TG-S04 and TG-S05 were diluted due to viscous and concentrated matrix.

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The soil samples results are based on a dry weight basis.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____