

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Portal Partners Tri-Venture	SDG No.: Q2832
Contract: PORT06	Lab Code: ACE
ICS Source: EPA	Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	240000	255000	94	216000	294000	08/13/2025	12:41	LB136801
	Antimony	-4.82			-50	50	08/13/2025	12:41	LB136801
	Arsenic	13.5			-20	20	08/13/2025	12:41	LB136801
	Barium	-0.47	6.0	8	-94	106	08/13/2025	12:41	LB136801
	Beryllium	1.31			-6	6	08/13/2025	12:41	LB136801
	Cadmium	2.33	1.0	233	-5	7	08/13/2025	12:41	LB136801
	Calcium	229000	245000	94	208000	282000	08/13/2025	12:41	LB136801
	Chromium	58.2	52.0	112	42	62	08/13/2025	12:41	LB136801
	Cobalt	1.85			-30	30	08/13/2025	12:41	LB136801
	Copper	4.82	2.0	241	-18	22	08/13/2025	12:41	LB136801
	Iron	99200	101000	98	85600	116500	08/13/2025	12:41	LB136801
	Lead	-6.38			-12	12	08/13/2025	12:41	LB136801
	Magnesium	243000	255000	95	216000	294000	08/13/2025	12:41	LB136801
	Manganese	8.17	7.0	117	-13	27	08/13/2025	12:41	LB136801
	Nickel	5.22	2.0	261	-38	42	08/13/2025	12:41	LB136801
	Potassium	-4.10			0	0	08/13/2025	12:41	LB136801
	Selenium	2.91			-20	20	08/13/2025	12:41	LB136801
	Silver	2.33			-10	10	08/13/2025	12:41	LB136801
	Sodium	3.06			0	0	08/13/2025	12:41	LB136801
	Thallium	2.91			-40	40	08/13/2025	12:41	LB136801
	Vanadium	3.65			-40	40	08/13/2025	12:41	LB136801
	Zinc	2.20			-40	40	08/13/2025	12:41	LB136801
ICSAB01	Aluminum	247000	247000	100	209000	285000	08/13/2025	12:45	LB136801
	Antimony	588	618	95	525	711	08/13/2025	12:45	LB136801
	Arsenic	111	104	107	88.4	120	08/13/2025	12:45	LB136801
	Barium	492	537	92	437	637	08/13/2025	12:45	LB136801
	Beryllium	487	495	98	420	570	08/13/2025	12:45	LB136801
	Cadmium	994	972	102	826	1120	08/13/2025	12:45	LB136801
	Calcium	236000	235000	100	199000	271000	08/13/2025	12:45	LB136801
	Chromium	558	542	103	460	624	08/13/2025	12:45	LB136801
	Cobalt	498	476	105	404	548	08/13/2025	12:45	LB136801
	Copper	486	511	95	434	588	08/13/2025	12:45	LB136801
	Iron	102000	99300	103	84400	114500	08/13/2025	12:45	LB136801
	Lead	42.2	49.0	86	37	61	08/13/2025	12:45	LB136801
	Magnesium	250000	248000	101	210000	286000	08/13/2025	12:45	LB136801
	Manganese	493	507	97	430	584	08/13/2025	12:45	LB136801
	Nickel	992	954	104	810	1100	08/13/2025	12:45	LB136801
	Potassium	42.9			0	0	08/13/2025	12:45	LB136801
	Selenium	50.5	46.0	110	26	66	08/13/2025	12:45	LB136801
	Silver	216	201	108	170	232	08/13/2025	12:45	LB136801
	Sodium	-3.58			0	0	08/13/2025	12:45	LB136801
	Thallium	113	108	105	68	148	08/13/2025	12:45	LB136801

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Portal Partners Tri-Venture	SDG No.: Q2832
Contract: PORT06	Lab Code: ACE
ICS Source: EPA	Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	479	491	98	417	565	08/13/2025	12:45	LB136801
	Zinc	1040	952	109	809	1095	08/13/2025	12:45	LB136801
ICSA01	Aluminum	251000	255000	98	216000	294000	08/14/2025	12:27	LB136837
	Antimony	-1.10			-50	50	08/14/2025	12:27	LB136837
	Arsenic	7.57			-20	20	08/14/2025	12:27	LB136837
	Barium	0.097	6.0	2	-94	106	08/14/2025	12:27	LB136837
	Beryllium	1.36			-6	6	08/14/2025	12:27	LB136837
	Cadmium	2.91	1.0	291	-5	7	08/14/2025	12:27	LB136837
	Calcium	237000	245000	97	208000	282000	08/14/2025	12:27	LB136837
	Chromium	60.3	52.0	116	42	62	08/14/2025	12:27	LB136837
	Cobalt	1.84			-30	30	08/14/2025	12:27	LB136837
	Copper	7.21	2.0	360	-18	22	08/14/2025	12:27	LB136837
	Iron	99800	101000	99	85600	116500	08/14/2025	12:27	LB136837
	Lead	-5.21			-12	12	08/14/2025	12:27	LB136837
	Magnesium	251000	255000	98	216000	294000	08/14/2025	12:27	LB136837
	Manganese	7.95	7.0	114	-13	27	08/14/2025	12:27	LB136837
	Nickel	5.40	2.0	270	-38	42	08/14/2025	12:27	LB136837
	Potassium	130			0	0	08/14/2025	12:27	LB136837
	Selenium	-3.06			-20	20	08/14/2025	12:27	LB136837
	Silver	1.48			-10	10	08/14/2025	12:27	LB136837
	Sodium	111			0	0	08/14/2025	12:27	LB136837
	Thallium	-3.36			-40	40	08/14/2025	12:27	LB136837
	Vanadium	2.16			-40	40	08/14/2025	12:27	LB136837
	Zinc	3.33			-40	40	08/14/2025	12:27	LB136837
ICSAB01	Aluminum	249000	247000	101	209000	285000	08/14/2025	12:38	LB136837
	Antimony	593	618	96	525	711	08/14/2025	12:38	LB136837
	Arsenic	110	104	106	88.4	120	08/14/2025	12:38	LB136837
	Barium	485	537	90	437	637	08/14/2025	12:38	LB136837
	Beryllium	496	495	100	420	570	08/14/2025	12:38	LB136837
	Cadmium	989	972	102	826	1120	08/14/2025	12:38	LB136837
	Calcium	236000	235000	100	199000	271000	08/14/2025	12:38	LB136837
	Chromium	566	542	104	460	624	08/14/2025	12:38	LB136837
	Cobalt	494	476	104	404	548	08/14/2025	12:38	LB136837
	Copper	490	511	96	434	588	08/14/2025	12:38	LB136837
	Iron	98600	99300	99	84400	114500	08/14/2025	12:38	LB136837
	Lead	42.6	49.0	87	37	61	08/14/2025	12:38	LB136837
	Magnesium	252000	248000	102	210000	286000	08/14/2025	12:38	LB136837
	Manganese	487	507	96	430	584	08/14/2025	12:38	LB136837
	Nickel	987	954	104	810	1100	08/14/2025	12:38	LB136837
	Potassium	-0.28			0	0	08/14/2025	12:38	LB136837
	Selenium	46.6	46.0	101	26	66	08/14/2025	12:38	LB136837
	Silver	216	201	108	170	232	08/14/2025	12:38	LB136837

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Portal Partners Tri-Venture

SDG No.: Q2832

Contract: PORT06

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Sodium	-6.45			0	0	08/14/2025	12:38	LB136837
	Thallium	109	108	101	68	148	08/14/2025	12:38	LB136837
	Vanadium	479	491	98	417	565	08/14/2025	12:38	LB136837
	Zinc	1040	952	109	809	1095	08/14/2025	12:38	LB136837