

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2832

Contract: PORT06

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169215BS							
Aluminum	mg/Kg	100	101		101	80 - 120	P
Antimony	mg/Kg	40.0	42.1		105	80 - 120	P
Arsenic	mg/Kg	40.0	40.5		101	80 - 120	P
Barium	mg/Kg	10.0	9.46		95	80 - 120	P
Beryllium	mg/Kg	10.0	10.1		101	80 - 120	P
Cadmium	mg/Kg	10.0	9.62		96	80 - 120	P
Calcium	mg/Kg	50.0	51.7	J	103	80 - 120	P
Chromium	mg/Kg	20.0	20.8		104	80 - 120	P
Cobalt	mg/Kg	10.0	9.93		99	80 - 120	P
Copper	mg/Kg	15.0	15.8		105	80 - 120	P
Iron	mg/Kg	150	153		102	80 - 120	P
Lead	mg/Kg	50.0	48.5		97	80 - 120	P
Magnesium	mg/Kg	100	101		101	80 - 120	P
Manganese	mg/Kg	10.0	10.2		102	80 - 120	P
Nickel	mg/Kg	25.0	25.0		100	80 - 120	P
Potassium	mg/Kg	500	494		99	80 - 120	P
Selenium	mg/Kg	100	104		104	80 - 120	P
Silver	mg/Kg	3.8	3.87		102	80 - 120	P
Sodium	mg/Kg	150	140		93	80 - 120	P
Thallium	mg/Kg	100	108		108	80 - 120	P
Vanadium	mg/Kg	15.0	15.4		103	80 - 120	P
Zinc	mg/Kg	10.0	10.5		105	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Portal Partners Tri-Venture

SDG No.: Q2832

Contract: PORT06

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169231BS Mercury	mg/Kg	0.27	0.24		88	80 - 120	CV