

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Portal Partners Tri-Venture</u>	SDG No.: <u>Q2832</u>
Contract: <u>PORT06</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	7.57			-20	20	08/14/2025	12:27	LB136837
	Barium	0.097	6.0	2	-94	106	08/14/2025	12:27	LB136837
	Cadmium	2.91	1.0	291	-5	7	08/14/2025	12:27	LB136837
	Chromium	60.3	52.0	116	42	62	08/14/2025	12:27	LB136837
	Lead	-5.21			-12	12	08/14/2025	12:27	LB136837
	Selenium	-3.06			-20	20	08/14/2025	12:27	LB136837
	Silver	1.48			-10	10	08/14/2025	12:27	LB136837
ICSAB01	Arsenic	110	104	106	88.4	120	08/14/2025	12:38	LB136837
	Barium	485	537	90	437	637	08/14/2025	12:38	LB136837
	Cadmium	989	972	102	826	1120	08/14/2025	12:38	LB136837
	Chromium	566	542	104	460	624	08/14/2025	12:38	LB136837
	Lead	42.6	49.0	87	37	61	08/14/2025	12:38	LB136837
	Selenium	46.6	46.0	101	26	66	08/14/2025	12:38	LB136837
	Silver	216	201	108	170	232	08/14/2025	12:38	LB136837