

Continuing Calibration Report for SequenceID : FC082225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069707.D

Aliphatic C16-C21	7360354.000	60.000	9.997	13.364	122672.567	133561.231	8.153
Aliphatic C21-C28	8628089.000	80.000	13.365	17.031	107851.113	118307.822	8.839
Aliphatic C9-C12	7904665.000	60.000	3.298	6.594	131744.417	144679.273	8.940
Aliphatic C12-C16	5271855.000	40.000	6.595	9.996	131796.375	143435.206	8.114
Aliphatic C28-C40	9905788.000	120.000	17.032	22.001	82548.233	95315.213	13.394
Aliphatic EPH	39070751.000	360.000	3.298	22.001	108529.864	120373.027	9.839