

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: PORT06

Lab Code: ACE

SDG NO.: Q2832

Continuing Calib Date: 08/13/2025

Initial Calibration Date(s): 08/01/2025 08/01/2025

Continuing Calib Time: 15:54

Initial Calibration Time(s): 12:05 20:28

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1016-1       | (1) | 5.81       | 5.81      | 5.71      | 5.91  | 0.00       |
| Aroclor-1016-2       | (2) | 5.83       | 5.83      | 5.73      | 5.93  | 0.00       |
| Aroclor-1016-3       | (3) | 5.89       | 5.90      | 5.80      | 6.00  | 0.01       |
| Aroclor-1016-4       | (4) | 5.99       | 5.99      | 5.89      | 6.09  | 0.00       |
| Aroclor-1016-5       | (5) | 6.28       | 6.29      | 6.19      | 6.39  | 0.01       |
| Aroclor-1260-1       | (1) | 7.40       | 7.40      | 7.30      | 7.50  | 0.00       |
| Aroclor-1260-2       | (2) | 7.65       | 7.66      | 7.56      | 7.76  | 0.01       |
| Aroclor-1260-3       | (3) | 8.01       | 8.01      | 7.91      | 8.11  | 0.00       |
| Aroclor-1260-4       | (4) | 8.24       | 8.24      | 8.14      | 8.34  | 0.00       |
| Aroclor-1260-5       | (5) | 8.57       | 8.57      | 8.47      | 8.67  | 0.01       |
| Tetrachloro-m-xylene |     | 4.66       | 4.66      | 4.56      | 4.76  | 0.00       |
| Decachlorobiphenyl   |     | 10.44      | 10.44     | 10.34     | 10.54 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: PORT06

Lab Code: ACE

SDG NO.: Q2832

Continuing Calib Date: 08/13/2025

Initial Calibration Date(s): 08/01/2025 08/01/2025

Continuing Calib Time: 15:54

Initial Calibration Time(s): 12:05 20:28

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.90       | 4.91      | 4.81      | 5.01 | 0.01       |
| Aroclor-1016-2       | (2) | 4.96       | 4.96      | 4.86      | 5.06 | 0.00       |
| Aroclor-1016-3       | (3) | 5.08       | 5.08      | 4.98      | 5.18 | 0.00       |
| Aroclor-1016-4       | (4) | 5.12       | 5.12      | 5.02      | 5.22 | 0.00       |
| Aroclor-1016-5       | (5) | 5.33       | 5.34      | 5.24      | 5.44 | 0.01       |
| Aroclor-1260-1       | (1) | 6.55       | 6.56      | 6.46      | 6.66 | 0.01       |
| Aroclor-1260-2       | (2) | 6.71       | 6.71      | 6.61      | 6.81 | 0.01       |
| Aroclor-1260-3       | (3) | 6.92       | 6.92      | 6.82      | 7.02 | 0.01       |
| Aroclor-1260-4       | (4) | 7.18       | 7.18      | 7.08      | 7.28 | 0.00       |
| Aroclor-1260-5       | (5) | 7.41       | 7.42      | 7.32      | 7.52 | 0.01       |
| Tetrachloro-m-xylene |     | 3.80       | 3.81      | 3.71      | 3.91 | 0.01       |
| Decachlorobiphenyl   |     | 8.82       | 8.83      | 8.73      | 8.93 | 0.01       |