

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112896.D **Time Analyzed:** 08:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.751	4.655	4.855	526.960	500.000	5.4
Aroclor-1016-2	4.769	4.674	4.874	547.510	500.000	9.5
Aroclor-1016-3	4.827	4.731	4.931	528.710	500.000	5.7
Aroclor-1016-4	4.945	4.851	5.051	536.170	500.000	7.2
Aroclor-1016-5	5.202	5.108	5.308	543.210	500.000	8.6
Aroclor-1260-1	6.239	6.145	6.345	535.780	500.000	7.2
Aroclor-1260-2	6.428	6.333	6.533	521.900	500.000	4.4
Aroclor-1260-3	6.794	6.700	6.900	558.270	500.000	11.7
Aroclor-1260-4	7.054	6.959	7.159	519.310	500.000	3.9
Aroclor-1260-5	7.297	7.203	7.403	511.980	500.000	2.4
Decachlorobiphenyl	8.686	8.593	8.793	48.410	50.000	-3.2
Tetrachloro-m-xylene	3.666	3.569	3.769	56.200	50.000	12.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112896.D **Time Analyzed:** 08:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.732	4.638	4.838	490.590	500.000	-1.9
Aroclor-1016-2	4.750	4.656	4.856	495.730	500.000	-0.9
Aroclor-1016-3	4.925	4.831	5.031	481.470	500.000	-3.7
Aroclor-1016-4	4.967	4.873	5.073	485.500	500.000	-2.9
Aroclor-1016-5	5.179	5.085	5.285	492.900	500.000	-1.4
Aroclor-1260-1	6.208	6.113	6.313	500.910	500.000	0.2
Aroclor-1260-2	6.396	6.302	6.502	500.730	500.000	0.1
Aroclor-1260-3	6.546	6.453	6.653	521.660	500.000	4.3
Aroclor-1260-4	7.016	6.923	7.123	523.900	500.000	4.8
Aroclor-1260-5	7.258	7.164	7.364	542.550	500.000	8.5
Decachlorobiphenyl	8.633	8.540	8.740	51.130	50.000	2.3
Tetrachloro-m-xylene	3.658	3.562	3.762	52.790	50.000	5.6