

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112930.D **Time Analyzed:** 08:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.750	4.655	4.855	529.710	500.000	5.9
Aroclor-1016-2	4.769	4.674	4.874	529.250	500.000	5.9
Aroclor-1016-3	4.826	4.731	4.931	519.760	500.000	4.0
Aroclor-1016-4	4.945	4.851	5.051	529.550	500.000	5.9
Aroclor-1016-5	5.202	5.108	5.308	542.490	500.000	8.5
Aroclor-1260-1	6.239	6.145	6.345	531.990	500.000	6.4
Aroclor-1260-2	6.429	6.333	6.533	518.600	500.000	3.7
Aroclor-1260-3	6.794	6.700	6.900	551.750	500.000	10.4
Aroclor-1260-4	7.053	6.959	7.159	513.930	500.000	2.8
Aroclor-1260-5	7.297	7.203	7.403	504.540	500.000	0.9
Decachlorobiphenyl	8.686	8.593	8.793	48.920	50.000	-2.2
Tetrachloro-m-xylene	3.665	3.569	3.769	55.480	50.000	11.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/23/2025 07/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112930.D **Time Analyzed:** 08:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.732	4.638	4.838	485.220	500.000	-3.0
Aroclor-1016-2	4.750	4.656	4.856	489.550	500.000	-2.1
Aroclor-1016-3	4.925	4.831	5.031	489.200	500.000	-2.2
Aroclor-1016-4	4.967	4.873	5.073	484.210	500.000	-3.2
Aroclor-1016-5	5.179	5.085	5.285	491.820	500.000	-1.6
Aroclor-1260-1	6.207	6.113	6.313	491.660	500.000	-1.7
Aroclor-1260-2	6.395	6.302	6.502	494.220	500.000	-1.2
Aroclor-1260-3	6.547	6.453	6.653	514.790	500.000	3.0
Aroclor-1260-4	7.016	6.923	7.123	513.570	500.000	2.7
Aroclor-1260-5	7.258	7.164	7.364	528.080	500.000	5.6
Decachlorobiphenyl	8.632	8.540	8.740	51.160	50.000	2.3
Tetrachloro-m-xylene	3.658	3.562	3.762	52.000	50.000	4.0