

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074334.D **Time Analyzed:** 09:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.809	5.712	5.912	519.460	500.000	3.9
Aroclor-1016-2	5.830	5.733	5.933	523.810	500.000	4.8
Aroclor-1016-3	5.893	5.796	5.996	532.470	500.000	6.5
Aroclor-1016-4	5.990	5.893	6.093	534.840	500.000	7.0
Aroclor-1016-5	6.283	6.186	6.386	527.000	500.000	5.4
Aroclor-1260-1	7.400	7.303	7.503	504.090	500.000	0.8
Aroclor-1260-2	7.653	7.555	7.755	484.890	500.000	-3.0
Aroclor-1260-3	8.011	7.913	8.113	492.480	500.000	-1.5
Aroclor-1260-4	8.239	8.142	8.342	494.160	500.000	-1.2
Aroclor-1260-5	8.565	8.468	8.668	474.660	500.000	-5.1
Decachlorobiphenyl	10.436	10.338	10.538	48.910	50.000	-2.2
Tetrachloro-m-xylene	4.657	4.560	4.760	54.280	50.000	8.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074334.D **Time Analyzed:** 09:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.900	4.806	5.006	567.910	500.000	13.6
Aroclor-1016-2	4.958	4.863	5.063	577.260	500.000	15.5
Aroclor-1016-3	5.079	4.983	5.183	563.770	500.000	12.8
Aroclor-1016-4	5.119	5.023	5.223	570.360	500.000	14.1
Aroclor-1016-5	5.334	5.239	5.439	591.500	500.000	18.3
Aroclor-1260-1	6.550	6.455	6.655	545.640	500.000	9.1
Aroclor-1260-2	6.705	6.609	6.809	538.090	500.000	7.6
Aroclor-1260-3	6.915	6.819	7.019	564.730	500.000	12.9
Aroclor-1260-4	7.175	7.079	7.279	549.320	500.000	9.9
Aroclor-1260-5	7.414	7.318	7.518	552.870	500.000	10.6
Decachlorobiphenyl	8.819	8.726	8.926	52.780	50.000	5.6
Tetrachloro-m-xylene	3.801	3.705	3.905	59.680	50.000	19.4