

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL08 **Date Analyzed:** 08/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074368.D **Time Analyzed:** 23:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.810	5.712	5.912	532.220	500.000	6.4
Aroclor-1016-2	5.832	5.733	5.933	531.780	500.000	6.4
Aroclor-1016-3	5.895	5.796	5.996	543.380	500.000	8.7
Aroclor-1016-4	5.992	5.893	6.093	542.160	500.000	8.4
Aroclor-1016-5	6.284	6.186	6.386	532.190	500.000	6.4
Aroclor-1260-1	7.401	7.303	7.503	512.180	500.000	2.4
Aroclor-1260-2	7.654	7.555	7.755	494.700	500.000	-1.1
Aroclor-1260-3	8.012	7.913	8.113	501.400	500.000	0.3
Aroclor-1260-4	8.240	8.142	8.342	507.860	500.000	1.6
Aroclor-1260-5	8.566	8.468	8.668	486.610	500.000	-2.7
Decachlorobiphenyl	10.437	10.338	10.538	49.940	50.000	-0.1
Tetrachloro-m-xylene	4.659	4.560	4.760	55.440	50.000	10.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL08 **Date Analyzed:** 08/13/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074368.D **Time Analyzed:** 23:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.902	4.806	5.006	575.690	500.000	15.1
Aroclor-1016-2	4.960	4.863	5.063	595.280	500.000	19.1
Aroclor-1016-3	5.080	4.983	5.183	593.040	500.000	18.6
Aroclor-1016-4	5.121	5.023	5.223	572.950	500.000	14.6
Aroclor-1016-5	5.336	5.239	5.439	591.030	500.000	18.2
Aroclor-1260-1	6.552	6.455	6.655	570.950	500.000	14.2
Aroclor-1260-2	6.706	6.609	6.809	562.770	500.000	12.6
Aroclor-1260-3	6.916	6.819	7.019	577.230	500.000	15.4
Aroclor-1260-4	7.176	7.079	7.279	563.830	500.000	12.8
Aroclor-1260-5	7.414	7.318	7.518	581.180	500.000	16.2
Decachlorobiphenyl	8.819	8.726	8.926	54.620	50.000	9.2
Tetrachloro-m-xylene	3.803	3.705	3.905	59.950	50.000	19.9