

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL10 **Date Analyzed:** 08/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074387.D **Time Analyzed:** 14:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.810	5.712	5.912	456.760	500.000	-8.6
Aroclor-1016-2	5.832	5.733	5.933	454.890	500.000	-9.0
Aroclor-1016-3	5.894	5.796	5.996	460.980	500.000	-7.8
Aroclor-1016-4	5.992	5.893	6.093	463.390	500.000	-7.3
Aroclor-1016-5	6.284	6.186	6.386	445.880	500.000	-10.8
Aroclor-1260-1	7.402	7.303	7.503	429.860	500.000	-14.0
Aroclor-1260-2	7.654	7.555	7.755	410.640	500.000	-17.9
Aroclor-1260-3	8.012	7.913	8.113	417.030	500.000	-16.6
Aroclor-1260-4	8.241	8.142	8.342	421.890	500.000	-15.6
Aroclor-1260-5	8.567	8.468	8.668	405.160	500.000	-19.0
Decachlorobiphenyl	10.440	10.338	10.538	42.840	50.000	-14.3
Tetrachloro-m-xylene	4.658	4.560	4.760	49.610	50.000	-0.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/01/2025 08/01/2025

Client Sample No.: CCAL10 **Date Analyzed:** 08/14/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074387.D **Time Analyzed:** 14:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.901	4.806	5.006	524.770	500.000	5.0
Aroclor-1016-2	4.959	4.863	5.063	530.180	500.000	6.0
Aroclor-1016-3	5.079	4.983	5.183	518.870	500.000	3.8
Aroclor-1016-4	5.119	5.023	5.223	527.160	500.000	5.4
Aroclor-1016-5	5.334	5.239	5.439	523.280	500.000	4.7
Aroclor-1260-1	6.551	6.455	6.655	444.180	500.000	-11.2
Aroclor-1260-2	6.705	6.609	6.809	444.670	500.000	-11.1
Aroclor-1260-3	6.915	6.819	7.019	459.000	500.000	-8.2
Aroclor-1260-4	7.175	7.079	7.279	442.510	500.000	-11.5
Aroclor-1260-5	7.414	7.318	7.518	433.540	500.000	-13.3
Decachlorobiphenyl	8.819	8.726	8.926	42.850	50.000	-14.3
Tetrachloro-m-xylene	3.801	3.705	3.905	54.230	50.000	8.5