

Surrogate Summary

SDG No.: **Q2832**

Client: **Portal Partners Tri-Venture**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO112413.D	PIBLK-PO112413.D	Tetrachloro-m-xyl	1	20	19.1	95		70 (60)	130 (140)
		Decachlorobiphen	1	20	20.0	100		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.5	98		70 (60)	130 (140)
		Decachlorobiphen	2	20	20.4	102		70 (60)	130 (140)
I.BLK-PO112900.D	PIBLK-PO112900.D	Tetrachloro-m-xyl	1	20	20.7	104		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.2	96		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	20.3	101		70 (60)	130 (140)
		Decachlorobiphen	2	20	20.6	103		70 (60)	130 (140)
Q2832-07	TG-S04	Tetrachloro-m-xyl	1	20	21.4	107		30 (32)	150 (144)
		Decachlorobiphen	1	20	17.2	86		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	20.7	104		30 (32)	150 (144)
		Decachlorobiphen	2	20	20.6	103		30 (32)	150 (175)
Q2832-09	TG-S05	Tetrachloro-m-xyl	1	20	16.7	84		30 (32)	150 (144)
		Decachlorobiphen	1	20	15.9	79		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	16.3	81		30 (32)	150 (144)
		Decachlorobiphen	2	20	22.4	112		30 (32)	150 (175)
I.BLK-PO112915.D	PIBLK-PO112915.D	Tetrachloro-m-xyl	1	20	19.9	100		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.2	96		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.4	97		70 (60)	130 (140)
		Decachlorobiphen	2	20	21.3	106		70 (60)	130 (140)
I.BLK-PO112934.D	PIBLK-PO112934.D	Tetrachloro-m-xyl	1	20	20.6	103		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.7	98		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.9	99		70 (60)	130 (140)
		Decachlorobiphen	2	20	21.0	105		70 (60)	130 (140)
Q2832-07DL	TG-S04DL	Tetrachloro-m-xyl	1	20	18.6	93		30 (32)	150 (144)
		Decachlorobiphen	1	20	16.3	82		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	17.8	89		30 (32)	150 (144)
		Decachlorobiphen	2	20	19.3	97		30 (32)	150 (175)
I.BLK-PO112949.D	PIBLK-PO112949.D	Tetrachloro-m-xyl	1	20	20.7	104		70 (60)	130 (140)
		Decachlorobiphen	1	20	18.4	92		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	20.2	101		70 (60)	130 (140)
		Decachlorobiphen	2	20	20.7	104		70 (60)	130 (140)
I.BLK-PP074167.D	PIBLK-PP074167.D	Tetrachloro-m-xyl	1	20	19.1	95		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.7	98		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	17.3	86		70 (60)	130 (140)
		Decachlorobiphen	2	20	18.7	93		70 (60)	130 (140)
I.BLK-PP074337.D	PIBLK-PP074337.D	Tetrachloro-m-xyl	1	20	19.3	96		70 (60)	130 (140)
		Decachlorobiphen	1	20	18.5	92		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	16.4	82		70 (60)	130 (140)
		Decachlorobiphen	2	20	16.2	81		70 (60)	130 (140)
PB169227BS	PB169227BS	Tetrachloro-m-xyl	1	20	25.4	127		30 (32)	150 (144)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: **Q2832**

Client: **Portal Partners Tri-Venture**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
PB169227BS	PB169227BS	Decachlorobiphen	1	20	23.6	118		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	22.8	114		30 (32)	150 (144)
		Decachlorobiphen	2	20	21.8	109		30 (32)	150 (175)
Q2832-01	TG-S01	Tetrachloro-m-xyl	1	20	28.0	140		30 (32)	150 (144)
		Decachlorobiphen	1	20	26.4	132		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	26.5	133		30 (32)	150 (144)
		Decachlorobiphen	2	20	24.7	123		30 (32)	150 (175)
Q2832-03	TG-S02	Tetrachloro-m-xyl	1	20	27.2	136		30 (32)	150 (144)
		Decachlorobiphen	1	20	24.6	123		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	28.4	142		30 (32)	150 (144)
		Decachlorobiphen	2	20	23.4	117		30 (32)	150 (175)
Q2832-03MS	TG-S02MS	Tetrachloro-m-xyl	1	20	24.3	121		30 (32)	150 (144)
		Decachlorobiphen	1	20	23.4	117		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	22.3	112		30 (32)	150 (144)
		Decachlorobiphen	2	20	22.7	114		30 (32)	150 (175)
Q2832-03MSD	TG-S02MSD	Tetrachloro-m-xyl	1	20	24.2	121		30 (32)	150 (144)
		Decachlorobiphen	1	20	22.9	115		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	21.9	110		30 (32)	150 (144)
		Decachlorobiphen	2	20	21.9	110		30 (32)	150 (175)
Q2832-05	TG-S03	Tetrachloro-m-xyl	1	20	22.1	110		30 (32)	150 (144)
		Decachlorobiphen	1	20	21.9	110		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	28.3	142		30 (32)	150 (144)
		Decachlorobiphen	2	20	22.0	110		30 (32)	150 (175)
I.BLK-PP074351.D	PIBLK-PP074351.D	Tetrachloro-m-xyl	1	20	19.5	98		70 (60)	130 (140)
		Decachlorobiphen	1	20	18.2	91		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	16.7	83		70 (60)	130 (140)
		Decachlorobiphen	2	20	16.3	82		70 (60)	130 (140)
I.BLK-PP074365.D	PIBLK-PP074365.D	Tetrachloro-m-xyl	1	20	19.7	99		70 (60)	130 (140)
		Decachlorobiphen	1	20	18.7	94		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	17.0	85		70 (60)	130 (140)
		Decachlorobiphen	2	20	16.9	84		70 (60)	130 (140)
PB169227BL	PB169227BL	Tetrachloro-m-xyl	1	20	24.6	123		30 (32)	150 (144)
		Decachlorobiphen	1	20	23.4	117		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	22.3	112		30 (32)	150 (144)
		Decachlorobiphen	2	20	22.1	110		30 (32)	150 (175)
I.BLK-PP074371.D	PIBLK-PP074371.D	Tetrachloro-m-xyl	1	20	19.7	99		70 (60)	130 (140)
		Decachlorobiphen	1	20	18.8	94		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	16.6	83		70 (60)	130 (140)
		Decachlorobiphen	2	20	17.1	86		70 (60)	130 (140)
I.BLK-PP074376.D	PIBLK-PP074376.D	Tetrachloro-m-xyl	1	20	23.1	115		70 (60)	130 (140)
		Decachlorobiphen	1	20	21.1	106		70 (60)	130 (140)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: Q2832

Client: Portal Partners Tri-Venture

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP074376.D	PIBLK-PP074376.D	Tetrachloro-m-xyl	2	20	19.3	97		70 (60)	130 (140)
		Decachlorobiphen	2	20	19.5	97		70 (60)	130 (140)
Q2832-03DL	TG-S02DL	Tetrachloro-m-xyl	1	20	26.6	133		30 (32)	150 (144)
		Decachlorobiphen	1	20	23.3	117		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	24.1	120		30 (32)	150 (144)
		Decachlorobiphen	2	20	20.3	102		30 (32)	150 (175)
Q2832-05DL	TG-S03DL	Tetrachloro-m-xyl	1	20	28.4	142		30 (32)	150 (144)
		Decachlorobiphen	1	20	23.0	115		30 (32)	150 (175)
		Tetrachloro-m-xyl	2	20	26.3	132		30 (32)	150 (144)
		Decachlorobiphen	2	20	21.6	108		30 (32)	150 (175)
I.BLK-PP074390.D	PIBLK-PP074390.D	Tetrachloro-m-xyl	1	20	20.9	104		70 (60)	130 (140)
		Decachlorobiphen	1	20	19.4	97		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	18.7	93		70 (60)	130 (140)
		Decachlorobiphen	2	20	18.1	90		70 (60)	130 (140)