

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2832</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Portal Partners Tri-Venture</u>	Datafile :	<u>PP074340.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169227BS (Column 1)	AR1016	166.6	193	ug/kg	116				40 (71)	140 (120)	
	AR1260	166.6	173	ug/kg	104				40 (65)	140 (130)	
PB169227BS (Column 2)	AR1016	166.6	195	ug/kg	117				40 (71)	140 (120)	
	AR1260	166.6	189	ug/kg	113				40 (65)	140 (130)	