

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: PORT06

Lab Code: ACE

SDG NO.: Q2832

Continuing Calib Date: 08/15/2025

Initial Calibration Date(s): 08/07/2025 08/07/2025

Continuing Calib Time: 19:29

Initial Calibration Time(s): 12:02 13:38

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.29	7.30	7.20	7.40	0.01
2,4-D	8.42	8.42	8.32	8.52	0.00
2,4,5-TP(Silvex)	9.29	9.30	9.20	9.40	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: PORT06

Lab Code: ACE

SDG NO.: Q2832

Continuing Calib Date: 08/15/2025

Initial Calibration Date(s): 08/07/2025 08/07/2025

Continuing Calib Time: 19:29

Initial Calibration Time(s): 12:02 13:38

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.74	7.75	7.65	7.85	0.01
2,4-D	9.00	9.01	8.91	9.11	0.01
2,4,5-TP(Silvex)	9.90	9.90	9.80	10.00	0.00