

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: PORT06

Lab Code: ACE

SDG NO.: Q2832

Continuing Calib Date: 08/18/2025

Initial Calibration Date(s): 08/07/2025

08/07/2025

Continuing Calib Time: 09:58

Initial Calibration Time(s): 12:02

13:38

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|------------------|------------|-----------|------------------------|------|------------|
| 2,4-DCAA         | 7.29       | 7.30      | 7.20                   | 7.40 | 0.01       |
| 2,4-D            | 8.42       | 8.42      | 8.32                   | 8.52 | 0.00       |
| 2,4,5-TP(Silvex) | 9.30       | 9.30      | 9.20                   | 9.40 | 0.00       |

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance

**Contract:** PORT06

**Lab Code:** ACE

**SDG NO.:** Q2832

**Continuing Calib Date:** 08/18/2025

**Initial Calibration Date(s):** 08/07/2025 08/07/2025

**Continuing Calib Time:** 09:58

**Initial Calibration Time(s):** 12:02 13:38

**GC Column:** RTX-CLP2 **ID:** 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|------------------|------------|-----------|-----------|-------|------------|
|                  |            |           | FROM      | TO    |            |
| 2,4-DCAA         | 7.74       | 7.75      | 7.65      | 7.85  | 0.01       |
| 2,4-D            | 9.00       | 9.01      | 8.91      | 9.11  | 0.01       |
| 2,4,5-TP(Silvex) | 9.90       | 9.90      | 9.80      | 10.00 | 0.00       |