

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/15/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031461.D **Time Analyzed:** 19:29

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.293	9.200	9.400	725.120	712.500	1.8
2,4-D	8.417	8.324	8.524	711.490	705.000	0.9
2,4-DCAA	7.290	7.195	7.395	730.280	750.000	-2.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/15/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031461.D **Time Analyzed:** 19:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.897	9.803	10.003	667.310	712.500	-6.3
2,4-D	9.001	8.906	9.106	632.480	705.000	-10.3
2,4-DCAA	7.744	7.649	7.849	681.040	750.000	-9.2