

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/16/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031473.D **Time Analyzed:** 01:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.292	9.200	9.400	734.760	712.500	3.1
2,4-D	8.416	8.324	8.524	719.350	705.000	2.0
2,4-DCAA	7.289	7.195	7.395	748.270	750.000	-0.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 08/16/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031473.D **Time Analyzed:** 01:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.896	9.803	10.003	695.480	712.500	-2.4
2,4-D	9.000	8.906	9.106	646.840	705.000	-8.2
2,4-DCAA	7.744	7.649	7.849	703.470	750.000	-6.2