

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031491.D **Time Analyzed:** 18:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.292	9.200	9.400	689.180	712.500	-3.3
2,4-D	8.418	8.324	8.524	669.990	705.000	-5.0
2,4-DCAA	7.289	7.195	7.395	706.000	750.000	-5.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** PORT06
Lab Code: ACE **SDG NO.:** Q2832
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/07/2025 08/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031491.D **Time Analyzed:** 18:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.895	9.803	10.003	657.320	712.500	-7.7
2,4-D	9.001	8.906	9.106	602.080	705.000	-14.6
2,4-DCAA	7.744	7.649	7.849	660.480	750.000	-11.9