

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS080725

Review By	Abdul	Review On	8/8/2025 8:24:12 AM
Supervise By	mohammad	Supervise On	8/11/2025 1:55:16 AM
SubDirectory	PS080725	HP Acquire Method	HP Processing Method ps080725 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031377.D	07 Aug 2025 11:14		AR\AJ	Ok
2	I.BLK	I.BLK	PS031378.D	07 Aug 2025 11:38		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS031379.D	07 Aug 2025 12:02		AR\AJ	Ok
4	HSTDICC500	HSTDICC500	PS031380.D	07 Aug 2025 12:26		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS031381.D	07 Aug 2025 12:50		AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS031382.D	07 Aug 2025 13:14		AR\AJ	Ok
7	HSTDICC1500	HSTDICC1500	PS031383.D	07 Aug 2025 13:38		AR\AJ	Ok,M
8	HSTDICV750	ICVPS080725	PS031384.D	07 Aug 2025 14:02		AR\AJ	Ok
9	I.BLK	I.BLK	PS031385.D	07 Aug 2025 14:27	Looks like hexan	AR\AJ	Not Ok
10	HSTDCCC750	HSTDCCC750	PS031386.D	07 Aug 2025 14:51		AR\AJ	Ok
11	PB169160BL	PB169160BL	PS031387.D	07 Aug 2025 16:03	Opening IBLK fail	AR\AJ	Not Ok
12	PB169160BS	PB169160BS	PS031388.D	07 Aug 2025 16:27	Opening IBLK fail	AR\AJ	Not Ok
13	PB169160BSD	PB169160BSD	PS031389.D	07 Aug 2025 16:51	Opening IBLK fail	AR\AJ	Not Ok
14	Q2759-04	LIQUID-DRUM	PS031390.D	07 Aug 2025 17:15	Opening IBLK fail	AR\AJ	Not Ok
15	I.BLK	I.BLK	PS031391.D	07 Aug 2025 17:40	Opening IBLK fail	AR\AJ	Not Ok
16	HSTDCCC750	HSTDCCC750	PS031392.D	07 Aug 2025 18:04		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS081525

Review By	Abdul	Review On	8/18/2025 7:49:18 AM
Supervise By	mohammad	Supervise On	8/19/2025 1:27:14 AM
SubDirectory	PS081525	HP Acquire Method	HP Processing Method ps080725 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24802,PP24556,PP24557,PP24558,PP24559,PP24560,PP24803,PP24804
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24804
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031452.D	15 Aug 2025 10:08		AR\AJ	Ok
2	I.BLK	I.BLK	PS031453.D	15 Aug 2025 10:32		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS031454.D	15 Aug 2025 10:56		AR\AJ	Ok
4	PB169278BL	PB169278BL	PS031455.D	15 Aug 2025 17:04		AR\AJ	Ok
5	PB169278BS	PB169278BS	PS031456.D	15 Aug 2025 17:28		AR\AJ	Ok
6	PB169212TB	PB169212TB	PS031457.D	15 Aug 2025 17:52		AR\AJ	Ok
7	Q2732-03	WC-A7-01-C	PS031458.D	15 Aug 2025 18:16	surrogate fail	AR\AJ	ReRun
8	Q2832-02	TG-S01	PS031459.D	15 Aug 2025 18:40	surrogate fail, Surrogate fail confirm with MS/MSD	AR\AJ	Ok
9	I.BLK	I.BLK	PS031460.D	15 Aug 2025 19:05		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS031461.D	15 Aug 2025 19:29		AR\AJ	Ok
11	Q2832-02MS	TG-S01MS	PS031462.D	15 Aug 2025 20:41	Surrogate Fail , recovery fail for some compounds	AR\AJ	Ok,M
12	Q2832-02MSD	TG-S01MSD	PS031463.D	15 Aug 2025 21:05	Surrogate Fail , recovery fail for some compounds	AR\AJ	Ok,M
13	Q2832-04	TG-S02	PS031464.D	15 Aug 2025 21:29	surrogate fail	AR\AJ	ReRun
14	Q2832-06	TG-S03	PS031465.D	15 Aug 2025 21:54	surrogate fail	AR\AJ	ReRun
15	Q2832-08	TG-S04	PS031466.D	15 Aug 2025 22:18	surrogate fail	AR\AJ	ReRun
16	Q2832-10	TG-S05	PS031467.D	15 Aug 2025 22:42	surrogate fail	AR\AJ	ReRun
17	Q2836-03	WC-A2-15-C	PS031468.D	15 Aug 2025 23:06		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS081525

Review By	Abdul	Review On	8/18/2025 7:49:18 AM
Supervise By	mohammad	Supervise On	8/19/2025 1:27:14 AM
SubDirectory	PS081525	HP Acquire Method	HP Processing Method ps080725 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24802,PP24556,PP24557,PP24558,PP24559,PP24560,PP24803,PP24804		
CCC	PP24559		
Internal Standard/PEM	PP24804		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q2836-07	WC-A2-16-C	PS031469.D	15 Aug 2025 23:30	surrogate fail	AR\AJ	Not Ok
19	Q2836-11	WC-A2-17-C	PS031470.D	15 Aug 2025 23:54	surrogate fail	AR\AJ	Not Ok
20	Q2836-15	WC-A5-02-C	PS031471.D	16 Aug 2025 00:18	surrogate fail	AR\AJ	ReRun
21	I.BLK	I.BLK	PS031472.D	16 Aug 2025 00:43		AR\AJ	Ok
22	HSTDCCC750	HSTDCCC750	PS031473.D	16 Aug 2025 01:07		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS081825

Review By	Review On
Supervise By	Supervise On
SubDirectory	PS081825
HP Acquire Method	HP Processing Method
	ps080725 8151
STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24802,PP24556,PP24557,PP24558,PP24559,PP24560,PP24803,PP24804
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24804
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS031474.D	18 Aug 2025 08:36		AR\AJ	Ok
2	I.BLK	I.BLK	PS031475.D	18 Aug 2025 09:00		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS031476.D	18 Aug 2025 09:58		AR\AJ	Ok,NS
4	Q2732-03	WC-A7-01-C	PS031477.D	18 Aug 2025 10:23		AR\AJ	Ok,NS
5	Q2836-07	WC-A2-16-C	PS031478.D	18 Aug 2025 10:47		AR\AJ	Ok
6	Q2836-11	WC-A2-17-C	PS031479.D	18 Aug 2025 11:11		AR\AJ	Ok
7	Q2836-15	WC-A5-02-C	PS031480.D	18 Aug 2025 11:35		AR\AJ	Ok
8	I.BLK	I.BLK	PS031481.D	18 Aug 2025 11:59		AR\AJ	Ok
9	HSTDCCC750	HSTDCCC750	PS031482.D	18 Aug 2025 12:23		AR\AJ	Ok
10	Q2832-02RE	TG-S01RE	PS031483.D	18 Aug 2025 14:44	surrogate fail, not required	AR\AJ	Not Ok
11	Q2832-02MS	TG-S01MS	PS031484.D	18 Aug 2025 15:08	not required	AR\AJ	Not Ok
12	Q2832-02MSD	TG-S01MSD	PS031485.D	18 Aug 2025 15:32	not required	AR\AJ	Not Ok
13	Q2832-04RE	TG-S02RE	PS031486.D	18 Aug 2025 15:57	surrogate fail	AR\AJ	Confirms
14	Q2832-06RE	TG-S03RE	PS031487.D	18 Aug 2025 16:21	surrogate fail	AR\AJ	Confirms
15	Q2832-08RE	TG-S04RE	PS031488.D	18 Aug 2025 16:45	surrogate fail	AR\AJ	Confirms
16	Q2832-10RE	TG-S05RE	PS031489.D	18 Aug 2025 17:09	surrogate fail	AR\AJ	Confirms
17	I.BLK	I.BLK	PS031490.D	18 Aug 2025 17:44		AR\AJ	Ok
18	HSTDCCC750	HSTDCCC750	PS031491.D	18 Aug 2025 18:09		AR\AJ	Ok,NS

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS081825

Review By		Review On	
Supervise By		Supervise On	
SubDirectory	PS081825	HP Acquire Method	HP Processing Method ps080725 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24802,PP24556,PP24557,PP24558,PP24559,PP24560,PP24803,PP24804		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard	PP24804		
MS/MSD Standard			
LCS Standard			

M : Manual Integration