

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: PORT06

Lab Code: ACE

SDG NO.: Q2832

Continuing Calib Date: 08/16/2025

Initial Calibration Date(s): 07/31/2025

07/31/2025

Continuing Calib Time: 00:05

Initial Calibration Time(s): 11:47

12:41

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 9.07       | 9.07      | 8.97                   | 9.17 | 0.00       |
| Tetrachloro-m-xylene | 3.55       | 3.55      | 3.45                   | 3.65 | 0.00       |
| gamma-BHC (Lindane)  | 4.33       | 4.33      | 4.23                   | 4.43 | 0.00       |
| Heptachlor           | 4.93       | 4.93      | 4.83                   | 5.03 | 0.00       |
| Heptachlor epoxide   | 5.69       | 5.69      | 5.59                   | 5.79 | 0.00       |
| Endrin               | 6.57       | 6.57      | 6.47                   | 6.67 | 0.00       |
| Methoxychlor         | 7.49       | 7.49      | 7.39                   | 7.59 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: PORT06

Lab Code: ACE

SDG NO.: Q2832

Continuing Calib Date: 08/16/2025

Initial Calibration Date(s): 07/31/2025

07/31/2025

Continuing Calib Time: 00:05

Initial Calibration Time(s): 11:47

12:41

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 8.07       | 8.07      | 7.97                   | 8.17 | 0.00       |
| Tetrachloro-m-xylene | 2.88       | 2.88      | 2.78                   | 2.98 | 0.00       |
| gamma-BHC (Lindane)  | 3.73       | 3.73      | 3.63                   | 3.83 | 0.00       |
| Heptachlor           | 4.08       | 4.08      | 3.98                   | 4.18 | 0.00       |
| Heptachlor epoxide   | 4.87       | 4.87      | 4.77                   | 4.97 | 0.00       |
| Endrin               | 5.79       | 5.79      | 5.69                   | 5.89 | 0.00       |
| Methoxychlor         | 6.75       | 6.75      | 6.65                   | 6.85 | 0.00       |